


Data Path : Z:\svoasrv\HPCHEM1\BNA M\Data\BM041521\
 Data File : BM029466.D
 Acq On : 14 Apr 2021 14:31
 Operator : CG/JU
 Sample : DFTPP32
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 DFTPP32

Integration File: LSCINT.P

Method : Z:\SVOASRV\HPCHEM1\BNA_M\METHODS\SOM-EPA-BM040721MA.M
 Title : SVOA CALIBRATION
 Last Update : Wed Apr 14 17:30:14 2021



AutoFind: Scans 2485, 2486, 2487; Background Corrected with Scan 2480

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	62.1	107421	PASS
68	69	0.00	2	1.9	2026	PASS
69	198	0.00	100	62.7	108432	PASS
70	69	0.00	2	0.5	534	PASS
127	198	10	80	60.6	104877	PASS
197	198	0.00	2	0.2	357	PASS
198	198	100	100	100.0	172930	PASS
199	198	5	9	7.0	12186	PASS
275	198	10	60	26.5	45821	PASS
365	198	1	100	4.4	7562	PASS
441	198	0.01	100	10.0	17260	PASS
442	198	50	100	64.4	111416	PASS
443	442	15	24	19.5	21722	PASS

m/z	Abundance
36.90	403.0
38.00	1579.0
39.10	8838.0
40.10	766.0
41.10	318.0
44.00	1078.0
44.90	364.0
49.00	1087.0
50.00	28824.0
51.10	98008.0
52.00	5098.0
53.00	237.0
55.00	585.0
56.00	2836.0
57.00	6188.0
58.00	301.0
61.00	1162.0
62.00	1231.0
63.10	4117.0
64.00	745.0
65.00	1448.0
66.00	153.0
66.90	167.0
68.00	1652.0
69.00	102376.0
70.10	518.0
73.00	1587.0
74.00	12224.0
75.00	16592.0
76.10	6362.0
77.10	117520.0
78.10	8291.0
79.00	8694.0
80.00	5664.0
81.00	8988.0
82.00	2332.0
83.00	1812.0
84.00	372.0
85.00	1975.0
86.00	2148.0
87.00	960.0
88.00	505.0
89.00	281.0
91.00	1621.0
92.00	2019.0
93.00	15318.0
94.00	822.0
94.90	168.0
96.00	544.0
97.00	309.0
98.00	12541.0
99.00	9248.0
100.00	634.0
101.00	3906.0
102.00	280.0
102.90	1189.0
104.00	2667.0
105.00	2895.0
106.00	987.0
107.00	32984.0
108.00	5039.0
109.10	1060.0
110.00	51064.0
111.00	8815.0
112.00	1186.0
112.80	335.0
114.90	186.0
116.00	1552.0
117.00	40184.0
118.00	2329.0
118.90	310.0
119.90	275.0
120.90	163.0
122.00	2230.0
123.00	3598.0
124.00	1408.0
125.00	1637.0
127.00	103736.0
128.00	7667.0
129.00	42624.0
130.00	3498.0
131.00	862.0
132.00	366.0
133.10	235.0
133.90	1133.0
134.90	3354.0
136.00	1555.0
137.00	1783.0
137.90	358.0
139.00	193.0
140.00	714.0
141.00	6088.0

Instrument :
BNA_M
ClientSampleId :
DFTPP32

142.00	1609.0
143.00	1191.0
144.00	450.0
144.90	244.0
146.00	1250.0
147.00	3276.0
148.00	9577.0
149.00	1517.0
149.90	398.0
151.00	709.0
151.60	403.0
153.00	1799.0
154.00	2084.0
155.00	3652.0
156.00	4729.0
156.90	578.0
158.00	1313.0
159.00	988.0
160.00	2112.0
161.00	2998.0
162.00	699.0
163.00	183.0
164.00	280.0
165.00	2690.0
165.90	1717.0
167.00	15635.0
168.00	8428.0
169.00	1145.0
170.00	382.0
170.80	744.0
172.00	1322.0
173.00	1591.0
174.00	2695.0
175.10	4954.0
176.00	1072.0
177.00	2147.0
178.00	1084.0
179.00	9516.0
180.00	6504.0
181.00	2906.0
181.90	497.0
182.90	236.0
184.00	638.0
185.00	4544.0
186.00	31296.0
187.00	9411.0
188.00	1227.0
189.00	2353.0
190.00	355.0
190.90	1232.0
192.00	3318.0
193.00	3652.0
194.00	605.0
195.00	381.0
196.00	5814.0
198.00	184448.0
199.00	12308.0
200.00	1168.0
201.10	619.0
201.60	694.0
203.00	2099.0
204.10	9029.0
205.00	14066.0
206.10	58824.0
207.10	8601.0
208.00	2609.0
209.00	971.0
210.10	1312.0
210.90	2072.0
211.80	452.0
212.90	229.0
215.00	1054.0
216.00	1965.0
217.00	17384.0
218.00	2466.0
221.00	6262.0
222.00	1937.0
223.00	3707.0
224.00	34520.0
225.00	8550.0
226.10	1134.0
227.00	17088.0
228.00	2256.0
229.00	3336.0
229.80	346.0
231.00	1190.0
231.90	300.0
233.00	342.0
233.90	756.0
235.00	1413.0
236.00	982.0
237.00	1295.0
238.00	199.0
239.00	548.0
240.00	557.0
241.00	903.0

Instrument :
BNA_M
ClientSampleId :
DFTPP32

242.00	1615.0
243.00	1977.0
244.10	23856.0
245.10	3388.0
246.00	6210.0
246.90	1448.0
248.00	335.0
248.90	1026.0
249.80	190.0
251.00	382.0
251.90	425.0
253.00	691.0
255.00	125776.0
256.00	18312.0
257.00	1575.0
258.00	9261.0
259.00	1693.0
260.00	183.0
261.00	213.0
264.00	246.0
265.00	4182.0
266.00	658.0
270.00	278.0
271.00	304.0
272.00	342.0
273.00	3581.0
274.00	9971.0
275.00	52216.0
275.90	6839.0
277.00	5685.0
277.90	1342.0
279.00	220.0
280.90	263.0
281.80	255.0
283.10	589.0
283.90	675.0
285.00	954.0
288.90	211.0
290.00	199.0
291.90	235.0
293.10	1130.0
294.00	281.0
294.80	362.0
296.00	22360.0
297.00	3125.0
298.00	215.0
300.90	318.0
302.10	286.0
303.00	1943.0
303.90	625.0
307.90	355.0
308.80	217.0
309.90	210.0
313.10	242.0
314.00	1167.0
314.90	2323.0
315.90	918.0
317.00	258.0
320.90	541.0
322.00	159.0
323.00	5312.0
323.90	799.0
325.80	165.0
327.00	1261.0
328.00	460.0
332.00	424.0
333.00	495.0
333.90	3424.0
335.00	960.0
338.80	243.0
341.00	754.0
342.00	273.0
346.00	1505.0
346.90	294.0
352.00	1435.0
353.00	869.0
354.00	1839.0
355.00	336.0
359.00	177.0
365.00	9561.0
366.00	1292.0
370.80	226.0
372.00	2416.0
373.00	638.0
383.00	729.0
390.00	329.0
391.00	180.0
392.00	215.0
402.00	943.0
403.00	1573.0
404.00	630.0
421.00	990.0
421.90	1269.0
423.00	7820.0
424.00	2001.0
424.90	213.0

Instrument :
BNA_M
ClientSampleId :
DFTPP32

439.60	183.0
441.00	22464.0
442.00	149376.0
443.10	28720.0
444.00	2648.0

Instrument :
BNA_M
ClientSampleId :
DFTPP32

m/z	Abundance
36.00	531.0
38.00	2077.0
39.10	10919.0
40.00	765.0
41.00	340.0
44.00	730.0
45.00	319.0
49.10	1394.0
50.10	35824.0
51.10	129776.0
52.10	6942.0
55.00	868.0
56.00	3610.0
57.00	8229.0
58.00	350.0
59.00	220.0
61.00	1618.0
62.00	2049.0
63.00	4899.0
64.00	776.0
65.00	2081.0
66.10	202.0
67.00	217.0
68.10	2410.0
69.00	129064.0
70.00	540.0
73.00	1746.0
74.00	16122.0
75.00	21528.0
76.10	8477.0
77.10	147776.0
78.00	9615.0
79.00	9927.0
80.00	7166.0
81.00	10220.0
82.00	2627.0
83.00	2735.0
84.00	272.0
85.00	1993.0
86.00	3771.0
87.00	1353.0
87.90	424.0
89.00	304.0
91.00	2097.0
92.10	2766.0
93.00	18632.0
93.90	1431.0
95.00	344.0
96.10	727.0
97.00	344.0
98.00	15408.0
99.00	11492.0
100.00	937.0
101.00	5636.0
102.00	290.0
103.00	1772.0
104.00	3355.0
105.00	3426.0
106.00	1136.0
107.00	42744.0
108.00	5685.0
109.10	1123.0
110.00	62672.0
111.00	10172.0
112.00	1333.0
112.90	492.0
116.00	2315.0
117.00	49648.0
118.00	3089.0
118.90	323.0
120.00	455.0
120.90	194.0
122.00	2591.0
123.00	4513.0
124.00	1877.0
125.00	1555.0
127.00	127568.0
128.00	9655.0
129.00	51888.0
130.00	4593.0
131.00	1134.0
131.90	374.0
134.00	1508.0
135.00	3874.0
136.00	1939.0
137.00	2887.0
137.90	608.0
138.90	214.0
140.00	1058.0
141.00	6756.0
141.90	2122.0
143.00	1812.0

Instrument :
BNA_M
ClientSampleId :
DFTPP32

144.10	823.0
144.80	417.0
146.00	1279.0
147.00	4041.0
148.00	11630.0
149.00	2146.0
149.90	550.0
151.00	894.0
151.60	468.0
152.10	496.0
153.00	2590.0
154.10	1672.0
155.00	4536.0
156.00	5778.0
157.10	877.0
158.00	1230.0
159.00	912.0
159.90	2253.0
161.00	3184.0
161.90	693.0
163.10	309.0
163.90	409.0
165.00	2917.0
166.00	1994.0
167.00	18064.0
168.00	9144.0
169.00	1402.0
170.00	571.0
170.90	856.0
172.00	1298.0
173.00	1802.0
174.00	3195.0
175.10	5944.0
176.00	1276.0
177.00	2056.0
178.00	1163.0
179.00	11355.0
180.00	8064.0
181.00	3242.0
182.10	681.0
182.80	334.0
183.90	1185.0
185.00	5023.0
186.00	35848.0
187.00	10389.0
187.90	949.0
189.00	2749.0
190.10	497.0
191.00	1304.0
192.00	3242.0
193.00	4131.0
194.10	697.0
194.90	398.0
196.00	6402.0
198.00	207360.0
199.00	14785.0
200.00	1150.0
201.40	799.0
203.00	2078.0
204.00	11286.0
205.00	16984.0
206.10	64000.0
207.00	9553.0
208.00	2723.0
209.00	1230.0
210.10	1430.0
211.00	2823.0
212.10	389.0
212.90	307.0
214.90	911.0
216.00	2260.0
217.00	20992.0
218.00	2358.0
219.10	163.0
221.00	5542.0
221.80	2404.0
223.00	4152.0
224.10	36880.0
225.10	9488.0
226.10	1162.0
227.00	18568.0
228.00	2536.0
229.00	3742.0
229.90	482.0
231.10	1161.0
232.10	334.0
232.90	383.0
234.00	1252.0
235.10	1316.0
236.00	779.0
237.00	2073.0
238.10	278.0
239.00	838.0
240.00	617.0
241.00	943.0
242.00	2446.0

Instrument :
BNA_M
ClientSampleId :
DFTPP32

243.00	1934.0
244.10	25840.0
245.00	3332.0
246.00	6871.0
247.00	1235.0
249.00	1155.0
250.90	216.0
252.00	324.0
252.90	649.0
254.10	1517.0
255.00	133632.0
256.00	19392.0
257.00	1845.0
258.00	9861.0
259.10	1829.0
260.00	175.0
261.00	258.0
263.90	267.0
265.00	4123.0
265.90	552.0
270.00	245.0
271.00	307.0
272.00	454.0
273.00	4179.0
274.00	10519.0
275.00	54400.0
276.00	6937.0
277.00	5605.0
277.90	1259.0
279.00	307.0
281.00	325.0
281.90	152.0
283.00	726.0
284.00	506.0
285.00	1001.0
285.90	228.0
289.00	210.0
290.00	222.0
290.90	181.0
292.00	266.0
293.00	1485.0
294.10	316.0
295.10	524.0
296.00	20392.0
297.00	2848.0
298.00	204.0
300.80	245.0
301.10	255.0
301.90	448.0
303.00	1734.0
304.10	631.0
305.10	156.0
308.10	359.0
308.80	192.0
310.00	242.0
313.10	200.0
314.00	821.0
315.00	2311.0
316.10	1064.0
317.20	194.0
321.00	602.0
322.10	306.0
323.00	4703.0
323.90	804.0
326.90	1307.0
328.00	496.0
332.00	438.0
332.90	602.0
334.00	3326.0
334.90	855.0
335.90	152.0
340.90	517.0
341.90	209.0
345.90	1055.0
346.90	208.0
352.00	1221.0
353.00	1064.0
353.90	1354.0
355.00	325.0
365.00	8755.0
365.90	1095.0
370.00	223.0
370.90	336.0
372.00	2675.0
373.00	690.0
382.80	607.0
384.10	240.0
390.10	250.0
391.00	228.0
392.00	176.0
402.00	1088.0
403.00	1358.0
404.00	555.0
421.00	1048.0
422.10	949.0
423.00	7144.0

Instrument :
BNA_M
ClientSampleId :
DFTPP32

424.00	2377.0
425.00	474.0
440.20	209.0
441.00	20352.0
442.00	125480.0
443.00	24696.0
444.10	2214.0
445.00	161.0

Instrument :
BNA_M
ClientSampleId :
DFTPP32

m/z	Abundance
37.00	640.0
38.10	1374.0
39.10	8666.0
40.00	879.0
41.00	262.0
44.00	889.0
45.00	173.0
46.90	197.0
49.10	1082.0
50.10	26392.0
51.10	94480.0
52.10	4377.0
53.00	260.0
55.00	600.0
56.00	2821.0
57.00	6063.0
58.10	169.0
59.00	154.0
61.00	1093.0
62.00	1492.0
63.00	3859.0
64.00	576.0
65.00	1736.0
66.10	158.0
67.00	160.0
68.10	2016.0
69.00	93856.0
70.00	546.0
73.00	1278.0
74.00	11304.0
75.00	15379.0
76.10	5428.0
77.10	105784.0
78.10	7382.0
79.00	7253.0
80.00	5505.0
81.00	7292.0
82.00	1941.0
83.00	1527.0
84.00	456.0
85.00	1472.0
86.00	2706.0
87.00	966.0
88.00	449.0
89.10	153.0
91.00	1544.0
92.00	2275.0
93.00	13239.0
93.90	873.0
96.00	500.0
98.00	10071.0
99.00	8360.0
100.00	734.0
101.00	3985.0
102.00	214.0
103.00	1123.0
104.00	2269.0
105.00	2483.0
106.00	959.0
107.00	27744.0
108.00	3815.0
108.90	733.0
110.00	41520.0
111.00	7543.0
112.00	1242.0
112.90	269.0
114.80	211.0
116.00	1663.0
117.00	33544.0
118.00	2473.0
118.90	295.0
120.10	292.0
122.00	1619.0
122.90	2757.0
124.00	1561.0
124.90	1020.0
127.00	83328.0
128.00	6680.0
129.00	33520.0
130.00	3227.0
130.90	590.0
131.90	325.0
132.90	325.0
134.00	1074.0
135.00	2796.0
136.00	1395.0
137.00	1662.0
137.90	227.0
138.70	220.0
139.90	376.0
141.00	5346.0
141.90	1416.0

Instrument :
BNA_M
ClientSampleId :
DFTPP32

143.00	843.0
143.90	294.0
145.10	151.0
146.00	964.0
147.00	2662.0
148.00	8114.0
149.00	1758.0
150.10	550.0
151.10	752.0
152.00	373.0
153.00	1900.0
153.90	1066.0
155.00	2638.0
156.00	3175.0
156.90	537.0
158.00	873.0
159.00	711.0
160.00	1535.0
161.00	2454.0
162.00	524.0
163.00	223.0
163.80	217.0
164.90	1844.0
166.00	1216.0
167.00	11421.0
168.00	7495.0
168.90	955.0
169.80	256.0
171.00	340.0
172.00	1038.0
173.00	1121.0
174.00	1956.0
175.00	3078.0
176.00	623.0
177.00	1562.0
178.00	592.0
179.00	7128.0
180.00	5474.0
181.00	2101.0
182.00	286.0
184.00	514.0
185.00	3032.0
186.00	22016.0
187.00	6181.0
187.90	921.0
189.00	1735.0
190.00	378.0
190.80	634.0
192.00	2454.0
193.00	2455.0
194.00	670.0
195.00	305.0
196.00	3494.0
196.70	891.0
197.10	1073.0
198.00	126984.0
199.00	9465.0
200.00	749.0
201.20	418.0
201.60	371.0
203.00	1270.0
204.00	5836.0
205.10	10446.0
206.10	38320.0
207.10	6062.0
208.00	1906.0
209.00	941.0
210.10	772.0
211.00	1846.0
211.70	307.0
212.10	288.0
215.00	491.0
216.00	944.0
217.00	12467.0
218.00	1667.0
219.90	212.0
221.00	3987.0
222.00	1359.0
223.00	2351.0
224.00	21712.0
225.00	5956.0
226.00	846.0
227.00	11474.0
228.00	1779.0
229.00	2341.0
230.00	293.0
231.00	749.0
232.90	281.0
233.90	845.0
234.90	792.0
236.00	651.0
237.00	1164.0
238.00	185.0
239.00	494.0
240.00	481.0
241.00	507.0

Instrument :
BNA_M
ClientSampleId :
DFTPP32

242.00	1101.0
243.00	1187.0
244.00	15792.0
245.10	2352.0
246.00	4438.0
247.00	865.0
248.00	369.0
249.00	538.0
252.10	219.0
252.90	384.0
253.20	360.0
254.10	843.0
255.00	78488.0
256.00	11595.0
257.00	1322.0
258.00	5909.0
259.00	1254.0
259.90	200.0
265.00	2482.0
266.00	453.0
270.00	163.0
270.90	188.0
271.80	324.0
273.00	2533.0
274.00	6012.0
275.00	30848.0
276.00	3812.0
276.90	3800.0
277.90	794.0
278.90	286.0
281.00	183.0
281.80	177.0
283.00	366.0
284.00	225.0
285.00	508.0
289.90	154.0
292.90	563.0
294.10	169.0
294.80	244.0
296.00	11132.0
297.00	1511.0
302.00	202.0
303.10	1154.0
304.10	262.0
307.90	198.0
314.10	515.0
315.00	1208.0
316.00	664.0
317.00	233.0
321.00	322.0
323.00	2713.0
323.90	491.0
326.80	676.0
327.80	296.0
331.80	239.0
332.80	176.0
334.00	1488.0
335.00	460.0
340.90	289.0
346.00	559.0
346.90	275.0
352.10	747.0
353.00	661.0
354.10	603.0
364.90	4370.0
366.00	662.0
370.80	196.0
372.10	1065.0
373.00	221.0
383.10	411.0
401.90	523.0
403.00	551.0
403.80	193.0
404.10	223.0
420.90	498.0
422.00	558.0
423.00	3430.0
423.90	1732.0
425.00	335.0
441.00	8965.0
442.00	59392.0
443.00	11752.0
444.10	1220.0

Instrument :
BNA_M
ClientSampleId :
DFTPP32

Scan 2480 (17.474 min): BM029466.D
DFTPP32

m/z	Abundance
40.00	462.0
44.00	891.0
49.00	156.0
73.00	437.0
207.10	898.0
207.90	207.0
281.00	260.0

Instrument :
BNA_M
ClientSampleId :
DFTPP32