

Data Path : Z:\SVOASRV\HPCHEM1\BNA_M\DATA\BM041719\
 Data File : BM019937.D
 Acq On : 17 Apr 2019 14:22
 Operator : JU/SJ
 Sample : K2421-01
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 OIL-WATER-SEPERATOR-1

Quant Time: Apr 17 18:00:13 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_M\METHODS\8270-BM041519.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Apr 15 16:16:29 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.53	152	152691	20.00	ng	0.00
21) Naphthalene-d8	10.30	136	574069	20.00	ng	0.00
39) Acenaphthene-d10	14.19	164	289911	20.00	ng	0.00
64) Phenanthrene-d10	16.96	188	551535	20.00	ng	0.00
76) Chrysene-d12	21.17	240	504506	20.00	ng	0.00
87) Perylene-d12	23.37	264	662019	20.00	ng	-0.01
System Monitoring Compounds						
5) 2-Fluorophenol	5.14	112	956357	101.51	ng	0.00
7) Phenol-d6	6.73	99	1006913	71.01	ng	0.00
23) Nitrobenzene-d5	8.70	82	1021713	79.59	ng	0.00
42) 2,4,6-Tribromophenol	15.70	330	479290	102.70	ng	0.00
45) 2-Fluorobiphenyl	12.80	172	1937304	83.28	ng	0.00
79) Terphenyl-d14	19.60	244	2323248	81.22	ng	0.00
Target Compounds						
50) Dimethylphthalate	13.67	163	94272	3.687	ng	Qvalue 99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA_M\DATA\BM041719\
Data File : BM019937.D
Acq On : 17 Apr 2019 14:22
Operator : JU/SJ
Sample : K2421-01
Misc :
ALS Vial : 5 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
OIL-WATER-SEPERATOR-1

Quant Time: Apr 17 18:00:13 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_M\METHODS\8270-BM041519.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Apr 15 16:16:29 2019
Response via : Initial Calibration

