

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM041824\
 Data File : BM045367.D
 Acq On : 18 Apr 2024 22:45
 Operator : MA/JU
 Sample : P2100-04
 Misc :
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 C0AP2

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 04/19/2024
 Supervised By :mohammad ahmed 04/22/2024

Quant Time: Apr 18 23:17:19 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\SFAM-EPA-SIM-BM040424.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Apr 18 22:19:59 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.576	152	1165	0.400	ng/ul	-0.03
4) Naphthalene-d8	10.342	136	3432	0.400	ng/ul	-0.04
9) Acenaphthene-d10	14.223	164	1777m	0.400	ng/ul	-0.02
13) Phenanthrene-d10	16.984	188	3754m	0.400	ng/ul	-0.03
17) Chrysene-d12	21.217	240	3148m	0.400	ng/ul	0.00
23) Perylene-d12	23.400	264	3426m	0.400	ng/ul	-0.02
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.175	96	5029	3.286	ng/ul	0.00
6) 2-Methylnaphthalene-d10	11.964	152	1483	0.319	ng/ul	-0.03
18) Fluoranthene-d10	19.029	212	3315	0.417	ng/ul	-0.02
Target Compounds					Qvalue	
2) 1,4-Dioxane	3.209	88	1069	0.729	ng/ul#	82

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM041824\
Data File : BM045367.D
Acq On : 18 Apr 2024 22:45
Operator : MA/JU
Sample : P2100-04
Misc :
ALS Vial : 8 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
C0AP2

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 04/19/2024
Supervised By :mohammad ahmed 04/22/2024

Quant Time: Apr 18 23:17:19 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\SFAM-EPA-SIM-BM040424.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Apr 18 22:19:59 2024
Response via : Initial Calibration

