

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM041824\
 Data File : BM045390.D
 Acq On : 19 Apr 2024 13:11
 Operator : MA/JU
 Sample : SP6477
 Misc : SFAM SIM SURROGATE
 ALS Vial : 33 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 SP6477

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 04/22/2024
 Supervised By :mohammad ahmed 04/22/2024

Quant Time: Apr 19 14:33:55 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\SFAM-EPA-SIM-BM040424.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Apr 18 22:19:59 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.660	152	419	0.400	ng/ul	# 0.05
4) Naphthalene-d8	10.523	136	1535	0.400	ng/ul	# 0.14
9) Acenaphthene-d10	14.283	164	907m	0.400	ng/ul	0.04
13) Phenanthrene-d10	17.077	188	1675m	0.400	ng/ul	0.06
17) Chrysene-d12	21.248	240	1550m	0.400	ng/ul	0.03
23) Perylene-d12	23.438	264	1867m	0.400	ng/ul	0.01
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.179	96	1385	2.516	ng/ul	0.00
6) 2-Methylnaphthalene-d10	12.085	152	1778	0.855	ng/ul	0.09
18) Fluoranthene-d10	19.066	212	4127m	1.054	ng/ul	0.01

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM041824\
Data File : BM045390.D
Acq On : 19 Apr 2024 13:11
Operator : MA/JU
Sample : SP6477
Misc : SFAM SIM SURROGATE
ALS Vial : 33 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
SP6477

Quant Time: Apr 19 14:33:55 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\SFAM-EPA-SIM-BM040424.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Apr 18 22:19:59 2024
Response via : Initial Calibration

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 04/22/2024
Supervised By :mohammad ahmed 04/22/2024

