

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM042122\
 Data File : BM034731.D
 Acq On : 20 Apr 2022 19:35
 Operator : CG/JU
 Sample : SSTD01026
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 SSTD010026

Quant Time: Apr 21 03:33:46 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\SFAM-EPA-BM042122.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Thu Apr 21 03:29:13 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.951	152	229512	20.000	ng/u1	0.00
20) Naphthalene-d8	10.751	136	918890	20.000	ng/u1	0.00
38) Acenaphthene-d10	14.580	164	566009	20.000	ng/u1	0.00
64) Phenanthrene-d10	17.321	188	1053058	20.000	ng/u1	0.00
79) Chrysene-d12	21.492	240	894839	20.000	ng/u1	0.00
88) Perylene-d12	23.892	264	879780	20.000	ng/u1	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.387	96	21206	3.975	ng/uL	0.00
4) Pyridine-d5	3.799	84	126486	10.104	ng/u1	0.00
7) Phenol-d5	7.104	99	171374	10.198	ng/u1	0.00
9) Bis-(2-Chloroethyl)eth...	7.275	67	102346	10.829	ng/u1	0.00
11) 2-Chlorophenol-d4	7.481	132	144863	9.969	ng/u1	0.00
15) 4-Methylphenol-d8	8.651	113	135239	9.922	ng/u1	0.00
21) Nitrobenzene-d5	9.104	128	63603	10.177	ng/u1	0.00
24) 2-Nitrophenol-d4	9.834	143	62747	9.845	ng/u1	0.00
28) 2,4-Dichlorophenol-d3	10.369	165	139224	9.180	ng/u1	0.00
31) 4-Chloroaniline-d4	10.881	131	194387	9.806	ng/u1	0.00
46) Dimethylphthalate-d6	13.992	166	400033	9.430	ng/u1	0.00
49) Acenaphthylene-d8	14.275	160	519156	9.589	ng/u1	0.00
54) 4-Nitrophenol-d4	14.757	143	56704	8.662	ng/u1	0.00
60) Fluorene-d10	15.569	176	352080	9.567	ng/u1	0.00
65) 4,6-Dinitro-2-methylph...	15.674	200	42903	7.956	ng/u1	0.00
73) Anthracene-d10	17.415	188	501181	9.737	ng/u1	0.00
81) Pyrene-d10	19.704	212	500646	10.476	ng/u1	0.00
92) Benzo(a)pyrene-d12	23.739	264	432541	9.570	ng/u1	0.00
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.422	88	20963	3.865	ng/uL	94
5) Pyridine	3.816	79	137089	10.569	ng/u1	99
6) Benzaldehyde	7.081	77	118470	13.466	ng/u1	98
8) Phenol	7.134	94	170972	10.310	ng/u1	99
10) Bis(2-Chloroethyl)ether	7.369	93	134174	10.290	ng/u1	100
12) 2-Chlorophenol	7.510	128	148688	10.117	ng/u1	97
13) 2-Methylphenol	8.387	108	128404	10.060	ng/u1	99
14) 2,2'-oxybis(1-Chloropr...	8.487	45	201759	11.812	ng/u1	99
16) Acetophenone	8.769	105	219119	10.656	ng/u1	99
17) N-Nitroso-di-n-propyla...	8.757	70	104030	10.696	ng/u1	96
18) 4-Methylphenol	8.716	108	141047	10.343	ng/u1	95
19) Hexachloroethane	9.040	117	62022	10.173	ng/u1	96
22) Nitrobenzene	9.145	77	150682	10.652	ng/u1	100
23) Isophorone	9.681	82	277133	10.089	ng/u1	99
25) 2-Nitrophenol	9.863	139	70173	9.885	ng/u1	97
26) 2,4-Dimethylphenol	9.922	107	159702	10.042	ng/u1	98
27) Bis(2-Chloroethoxy)met...	10.163	93	178628	10.336	ng/u1	99
29) 2,4-Dichlorophenol	10.392	162	138919	9.563	ng/u1	99
30) Naphthalene	10.804	128	469601	9.938	ng/u1	99
32) 4-Chloroaniline	10.904	127	194557	9.946	ng/u1	99
33) Hexachlorobutadiene	11.104	225	96580	8.845	ng/u1	99
34) Caprolactam	11.669	113	30018	8.635	ng/u1	97
35) 4-Chloro-3-methylphenol	12.028	107	133376	9.836	ng/u1	99
36) 2-Methylnaphthalene	12.410	142	324006	9.849	ng/u1	98


Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM042122\
 Data File : BM034731.D
 Acq On : 20 Apr 2022 19:35
 Operator : CG/JU
 Sample : SSTD01026
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 SSTD010026

Quant Time: Apr 21 03:33:46 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\SFAM-EPA-BM042122.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Thu Apr 21 03:29:13 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
37) 1-Methylnaphthalene	12.628	142	332026	10.002	ng/ul	98
39) 1,2,4,5-Tetrachloroben...	12.781	216	170267	8.393	ng/ul	100
40) Hexachlorocyclopentadiene	12.769	237	105812	7.586	ng/ul	97
41) 2,4,6-Trichlorophenol	13.016	196	101021	8.458	ng/ul	96
42) 2,4,5-Trichlorophenol	13.080	196	106813	8.480	ng/ul	99
43) 1,1'-Biphenyl	13.422	154	438116	9.300	ng/ul	100
44) 2-Chloronaphthalene	13.457	162	339280	9.259	ng/ul	99
45) 2-Nitroaniline	13.651	65	68766	10.055	ng/ul	98
47) Dimethylphthalate	14.039	163	397184	9.729	ng/ul	99
48) 2,6-Dinitrotoluene	14.145	165	65666	9.720	ng/ul	88
50) Acenaphthylene	14.298	152	523985	9.826	ng/ul	99
51) 3-Nitroaniline	14.475	138	68646	11.848	ng/ul	98
52) Acenaphthene	14.645	153	348146	9.971	ng/ul	99
53) 2,4-Dinitrophenol	14.675	184	27018	7.150	ng/ul	97
55) 4-Nitrophenol	14.769	109	48793	9.783	ng/ul	98
56) Dibenzofuran	14.974	168	488661	9.673	ng/ul	98
57) 2,4-Dinitrotoluene	14.927	165	91282	9.606	ng/ul	95
58) 2,3,4,6-Tetrachlorophenol	15.198	232	81089	7.996	ng/ul	99
59) Diethylphthalate	15.404	149	388840	9.798	ng/ul	100
61) Fluorene	15.627	166	396773	9.712	ng/ul	98
62) 4-Chlorophenyl-phenyle...	15.621	204	198463	9.066	ng/ul	97
63) 4-Nitroaniline	15.627	138	65722	12.392	ng/ul	95
66) 4,6-Dinitro-2-methylph...	15.692	198	43743	7.959	ng/ul	97
67) N-Nitrosodiphenylamine	15.833	169	329236	10.431	ng/ul	99
68) 4-Bromophenyl-phenylether	16.516	248	111239	9.002	ng/ul	98
69) Hexachlorobenzene	16.627	284	126217	8.676	ng/ul	99
70) Atrazine	16.780	200	103718	8.530	ng/ul	99
71) Pentachlorophenol	16.968	266	63354	7.134	ng/ul	98
72) Phenanthrene	17.363	178	579285	10.026	ng/ul	99
74) Anthracene	17.451	178	581443	9.919	ng/ul	100
75) 1,2,3,4-Tetrachloroben...	13.380	216	172951	9.172	ng/ul	100
76) Pentachlorobenzene	14.898	250	158308	9.114	ng/ul	99
77) Carbazole	17.715	167	503321	9.822	ng/ul	99
78) Di-n-butylphthalate	18.298	149	565600	9.730	ng/ul	100
80) Fluoranthene	19.374	202	606369	11.004	ng/ul	99
82) Pyrene	19.733	202	622106	10.994	ng/ul	100
83) Butylbenzylphthalate	20.639	149	205704	11.110	ng/ul	98
84) 3,3'-Dichlorobenzidine	21.409	252	177325	9.567	ng/ul	99
85) Benzo(a)anthracene	21.480	228	557169	10.049	ng/ul	99
86) Bis(2-ethylhexyl)phtha...	21.427	149	323887	10.944	ng/ul	98
87) Chrysene	21.533	228	546385	10.135	ng/ul	100
89) Di-n-octyl phthalate	22.350	149	502908	10.501	ng/ul	100
90) Benzo(b)fluoranthene	23.162	252	542860	9.910	ng/ul	98
91) Benzo(k)fluoranthene	23.209	252	526365	10.242	ng/ul	100
93) Benzo(a)pyrene	23.786	252	528737	9.970	ng/ul#	96
94) Indeno(1,2,3-cd)pyrene	26.380	276	714135	11.203	ng/ul	98
95) Dibenzo(a,h)anthracene	26.397	278	618029	11.256	ng/ul	100
96) Benzo(g,h,i)perylene	27.144	276	612652	11.241	ng/ul	99





(#) = qualifier out of range (m) = manual integration (+) = signals summed



#2
1,4-Dioxane
Concen: 3.865 ng/uL
RT: 3.422 min Scan# 51
Delta R.T. 0.006 min
Lab File: BM034731.D
Acq: 20 Apr 2022 19:35




Instrument :
BNA_M
ClientSampleId :
SSTD010026


Tgt Ion: 88 Resp: 20963
Ion Ratio Lower Upper
88 100
43 31.1 27.6 41.4
58 69.7 59.5 89.3



#5
Pyridine
Concen: 10.569 ng/uL
RT: 3.816 min Scan# 124
Delta R.T. 0.000 min
Lab File: BM034731.D
Acq: 20 Apr 2022 19:35

Tgt Ion: 79 Resp: 137089
Ion Ratio Lower Upper
79 100
52 81.3 64.3 96.5
51 41.1 32.4 48.6









#6
Benzaldehyde
Concen: 13.466 ng/ul
RT: 7.081 min Scan# 679
Delta R.T. 0.000 min
Lab File: BM034731.D
Acq: 20 Apr 2022 19:35




Instrument :
BNA_M
ClientSampleId :
SSTD010026


Tgt Ion: 77 Resp: 118470
Ion Ratio Lower Upper
77 100
105 94.7 77.4 116.0
106 89.8 73.4 110.2



#8
Phenol
Concen: 10.310 ng/ul
RT: 7.134 min Scan# 688
Delta R.T. 0.000 min
Lab File: BM034731.D
Acq: 20 Apr 2022 19:35

Tgt Ion: 94 Resp: 170972
Ion Ratio Lower Upper
94 100
65 30.5 23.8 35.6
66 40.0 31.3 46.9









#10
 Bis(2-Chloroethyl)ether
 Concen: 10.290 ng/ul
 RT: 7.369 min Scan# 710
 Delta R.T. 0.000 min
 Lab File: BM034731.D
 Acq: 20 Apr 2022 19:35

Instrument :
 BNA_M
 ClientSampleId :
 SSTD010026

Tgt Ion: 93 Resp: 134174




Ion	Ratio	Lower	Upper
93	100		
63	78.5	62.5	93.7
95	32.5	25.8	38.8




#12
 2-Chlorophenol
 Concen: 10.117 ng/ul
 RT: 7.510 min Scan# 752
 Delta R.T. 0.000 min
 Lab File: BM034731.D
 Acq: 20 Apr 2022 19:35

Tgt Ion:128 Resp: 148688

Ion	Ratio	Lower	Upper
128	100		
64	43.5	37.1	55.7
130	32.7	26.2	39.4









#13
2-Methylphenol
Concen: 10.060 ng/ul
RT: 8.387 min Scan# 901
Delta R.T. 0.000 min
Lab File: BM034731.D
Acq: 20 Apr 2022 19:35




Instrument :
BNA_M
ClientSampleId :
SSTD010026


Tgt Ion:108 Resp: 128404
Ion Ratio Lower Upper
108 100
107 87.4 70.8 106.2



#14
2,2'-oxybis(1-Chloropropane)
Concen: 11.812 ng/ul
RT: 8.487 min Scan# 918
Delta R.T. 0.006 min
Lab File: BM034731.D
Acq: 20 Apr 2022 19:35

Tgt Ion: 45 Resp: 201759
Ion Ratio Lower Upper
45 100
77 14.0 10.6 15.8
79 10.3 8.1 12.1









#16
Acetophenone
Concen: 10.656 ng/ul
RT: 8.769 min Scan# 966
Delta R.T. 0.000 min
Lab File: BM034731.D
Acq: 20 Apr 2022 19:35




Instrument :
BNA_M
ClientSampleId :
SSTD010026


Tgt Ion:105 Resp: 219119
Ion Ratio Lower Upper
105 100
77 76.7 62.4 93.6
51 28.9 23.8 35.6



#17
N-Nitroso-di-n-propylamine
Concen: 10.696 ng/ul
RT: 8.757 min Scan# 964
Delta R.T. -0.006 min
Lab File: BM034731.D
Acq: 20 Apr 2022 19:35

Tgt Ion: 70 Resp: 104030
Ion Ratio Lower Upper
70 100
42 69.5 53.3 79.9
101 10.9 8.2 12.4
130 20.9 18.3 27.5









#18
4-Methylphenol
Concen: 10.343 ng/ul
RT: 8.716 min Scan# 91
Delta R.T. 0.000 min
Lab File: BM034731.D
Acq: 20 Apr 2022 19:35




Instrument :
BNA_M
ClientSampleId :
SSTD010026


Tgt Ion:108 Resp: 141047
Ion Ratio Lower Upper
108 100
107 107.0 89.7 134.5



#19
Hexachloroethane
Concen: 10.173 ng/ul
RT: 9.040 min Scan# 1012
Delta R.T. 0.000 min
Lab File: BM034731.D
Acq: 20 Apr 2022 19:35

Tgt Ion:117 Resp: 62022
Ion Ratio Lower Upper
117 100
201 99.2 81.3 121.9
199 61.5 53.4 80.0









#22
Nitrobenzene
Concen: 10.652 ng/ul
RT: 9.145 min Scan# 1030
Delta R.T. 0.000 min
Lab File: BM034731.D
Acq: 20 Apr 2022 19:35




Instrument :
BNA_M
ClientSampleId :
SSTD010026


Tgt Ion: 77 Resp: 150682
Ion Ratio Lower Upper
77 100
123 45.0 36.1 54.1
65 13.7 11.0 16.6



#23
Isophorone
Concen: 10.089 ng/ul
RT: 9.681 min Scan# 1121
Delta R.T. 0.000 min
Lab File: BM034731.D
Acq: 20 Apr 2022 19:35

Tgt Ion: 82 Resp: 277133
Ion Ratio Lower Upper
82 100
95 7.4 6.2 9.4
138 19.7 15.3 22.9









#25
2-Nitrophenol
Concen: 9.885 ng/ul
RT: 9.863 min Scan# 1152
Delta R.T. 0.000 min
Lab File: BM034731.D
Acq: 20 Apr 2022 19:35




Instrument :
BNA_M
ClientSampleId :
SSTD010026


Tgt Ion:139 Resp: 70173
Ion Ratio Lower Upper
139 100
65 50.1 38.9 58.3
109 32.1 24.1 36.1



#26
2,4-Dimethylphenol
Concen: 10.042 ng/ul
RT: 9.922 min Scan# 1162
Delta R.T. -0.006 min
Lab File: BM034731.D
Acq: 20 Apr 2022 19:35

Tgt Ion:107 Resp: 159702
Ion Ratio Lower Upper
107 100
121 49.5 40.3 60.5
122 85.9 66.6 100.0









#27
 Bis(2-Chloroethoxy)methane
 Concen: 10.336 ng/ul
 RT: 10.163 min Scan# 1203
 Delta R.T. 0.000 min
 Lab File: BM034731.D
 Acq: 20 Apr 2022 19:35




Instrument :
 BNA_M
 ClientSampleId :
 SSTD010026


Tgt Ion: 93 Resp: 178628
 Ion Ratio Lower Upper
 93 100
 95 31.4 25.8 38.6
 123 11.2 9.2 13.8



#29
 2,4-Dichlorophenol
 Concen: 9.563 ng/ul
 RT: 10.392 min Scan# 1242
 Delta R.T. -0.006 min
 Lab File: BM034731.D
 Acq: 20 Apr 2022 19:35

Tgt Ion: 162 Resp: 138919
 Ion Ratio Lower Upper
 162 100
 164 64.9 53.0 79.4
 98 33.5 27.4 41.0









#30
Naphthalene
Concen: 9.938 ng/ul
RT: 10.804 min Scan# 1312
Delta R.T. 0.000 min
Lab File: BM034731.D
Acq: 20 Apr 2022 19:35




Instrument :
BNA_M
ClientSampleId :
SSTD010026


Tgt Ion:128 Resp: 469601
Ion Ratio Lower Upper
128 100
129 10.8 9.1 13.7
127 13.2 10.6 16.0



#32
4-Chloroaniline
Concen: 9.946 ng/ul
RT: 10.904 min Scan# 1329
Delta R.T. 0.000 min
Lab File: BM034731.D
Acq: 20 Apr 2022 19:35

Tgt Ion:127 Resp: 194557
Ion Ratio Lower Upper
127 100
129 32.5 25.4 38.2









#33
Hexachlorobutadiene
Concen: 8.845 ng/u1
RT: 11.104 min Scan# 11
Delta R.T. 0.000 min
Lab File: BM034731.D
Acq: 20 Apr 2022 19:35




Instrument :
BNA_M
ClientSampleId :
SSTD010026


Tgt Ion:225 Resp: 96580
Ion Ratio Lower Upper
225 100
223 60.8 49.8 74.8
227 61.4 49.6 74.4



#34
Caprolactam
Concen: 8.635 ng/u1
RT: 11.669 min Scan# 1459
Delta R.T. 0.000 min
Lab File: BM034731.D
Acq: 20 Apr 2022 19:35

Tgt Ion:113 Resp: 30018
Ion Ratio Lower Upper
113 100
55 156.8 130.6 195.8
56 131.8 104.3 156.5









#35
 4-Chloro-3-methylphenol
 Concen: 9.836 ng/ul
 RT: 12.028 min Scan# 1520
 Delta R.T. 0.000 min
 Lab File: BM034731.D
 Acq: 20 Apr 2022 19:35




Instrument :
 BNA_M
 ClientSampleId :
 SST010026


Tgt Ion:107 Resp: 133376
 Ion Ratio Lower Upper
 107 100
 144 27.4 21.5 32.3
 142 84.0 66.8 100.2



#36
 2-Methylnaphthalene
 Concen: 9.849 ng/ul
 RT: 12.410 min Scan# 1585
 Delta R.T. 0.000 min
 Lab File: BM034731.D
 Acq: 20 Apr 2022 19:35

Tgt Ion:142 Resp: 324006
 Ion Ratio Lower Upper
 142 100
 141 88.5 72.2 108.2





#37

1-Methylnaphthalene

Concen: 10.002 ng/ul

RT: 12.628 min Scan# 1

Delta R.T. 0.000 min

Lab File: BM034731.D

Acq: 20 Apr 2022 19:35

Instrument :

BNA_M

ClientSampleId :

SSTD010026





Tgt Ion:142 Resp: 332026

Ion Ratio Lower Upper

142 100

141 92.9 73.0 109.4

116 4.0 3.2 4.8



#39

1,2,4,5-Tetrachlorobenzene

Concen: 8.393 ng/ul

RT: 12.781 min Scan# 1648

Delta R.T. 0.000 min

Lab File: BM034731.D

Acq: 20 Apr 2022 19:35

Tgt Ion:216 Resp: 170267




Ion Ratio Lower Upper


216 100

214 80.5 64.5 96.7

179 22.1 17.9 26.9

108 18.5 15.0 22.4









#40
Hexachlorocyclopentadiene
Concen: 7.586 ng/ul
RT: 12.769 min Scan# 1646
Delta R.T. 0.000 min
Lab File: BM034731.D
Acq: 20 Apr 2022 19:35




Instrument :
BNA_M
ClientSampleId :
SSTD010026


Tgt Ion:237 Resp: 105812
Ion Ratio Lower Upper
237 100
235 63.9 49.2 73.8
272 13.4 10.0 15.0



#41
2,4,6-Trichlorophenol
Concen: 8.458 ng/ul
RT: 13.016 min Scan# 1688
Delta R.T. 0.000 min
Lab File: BM034731.D
Acq: 20 Apr 2022 19:35

Tgt Ion:196 Resp: 101021
Ion Ratio Lower Upper
196 100
198 92.6 78.5 117.7
200 30.7 25.0 37.6









#42
2,4,5-Trichlorophenol
Concen: 8.480 ng/ul
RT: 13.080 min Scan# 1699
Delta R.T. 0.000 min
Lab File: BM034731.D
Acq: 20 Apr 2022 19:35




Instrument :
BNA_M
ClientSampleId :
SSTD010026


Tgt Ion:196 Resp: 106813
Ion Ratio Lower Upper
196 100
198 97.5 77.7 116.5
200 30.1 25.0 37.4



#43
1,1'-Biphenyl
Concen: 9.300 ng/ul
RT: 13.422 min Scan# 1757
Delta R.T. 0.000 min
Lab File: BM034731.D
Acq: 20 Apr 2022 19:35

Tgt Ion:154 Resp: 438116
Ion Ratio Lower Upper
154 100
153 40.0 32.2 48.2
76 11.1 8.6 12.8









#44
2-Chloronaphthalene
Concen: 9.259 ng/ul
RT: 13.457 min Scan# 1763
Delta R.T. 0.000 min
Lab File: BM034731.D
Acq: 20 Apr 2022 19:35




Instrument :
BNA_M
ClientSampleId :
SSTD010026


Tgt Ion:162 Resp: 339280
Ion Ratio Lower Upper
162 100
164 32.6 25.7 38.5
127 37.9 29.8 44.8



#45
2-Nitroaniline
Concen: 10.055 ng/ul
RT: 13.651 min Scan# 1796
Delta R.T. 0.000 min
Lab File: BM034731.D
Acq: 20 Apr 2022 19:35

Tgt Ion: 65 Resp: 68766
Ion Ratio Lower Upper
65 100
92 67.5 55.9 83.9
138 110.1 87.3 130.9









#47
Dimethylphthalate
Concen: 9.729 ng/ul
RT: 14.039 min Scan# 1862
Delta R.T. 0.000 min
Lab File: BM034731.D
Acq: 20 Apr 2022 19:35




Instrument :
BNA_M
ClientSampleId :
SSTD010026


Tgt Ion:163 Resp: 397184
Ion Ratio Lower Upper
163 100
194 3.6 3.4 5.0
164 10.4 8.2 12.4



#48
2,6-Dinitrotoluene
Concen: 9.720 ng/ul
RT: 14.145 min Scan# 1880
Delta R.T. -0.006 min
Lab File: BM034731.D
Acq: 20 Apr 2022 19:35

Tgt Ion:165 Resp: 65666
Ion Ratio Lower Upper
165 100
89 64.7 43.1 64.7
121 23.0 16.6 24.8









#50
Acenaphthylene
Concen: 9.826 ng/ul
RT: 14.298 min Scan# 1906
Delta R.T. -0.006 min
Lab File: BM034731.D
Acq: 20 Apr 2022 19:35




Instrument :
BNA_M
ClientSampleId :
SSTD010026


Tgt Ion:152 Resp: 523985
Ion Ratio Lower Upper
152 100
151 19.4 15.3 22.9
153 13.4 10.2 15.4



#51
3-Nitroaniline
Concen: 11.848 ng/ul
RT: 14.475 min Scan# 1936
Delta R.T. 0.000 min
Lab File: BM034731.D
Acq: 20 Apr 2022 19:35

Tgt Ion:138 Resp: 68646
Ion Ratio Lower Upper
138 100
108 13.6 10.2 15.4
92 108.5 85.4 128.0









#52
Acenaphthene
Concen: 9.971 ng/ul
RT: 14.645 min Scan# 1965
Delta R.T. 0.000 min
Lab File: BM034731.D
Acq: 20 Apr 2022 19:35




Instrument :
BNA_M
ClientSampleId :
SSTD010026


Tgt Ion:153 Resp: 348146
Ion Ratio Lower Upper
153 100
152 46.5 37.5 56.3
154 87.1 71.0 106.4



#53
2,4-Dinitrophenol
Concen: 7.150 ng/ul
RT: 14.675 min Scan# 1970
Delta R.T. 0.000 min
Lab File: BM034731.D
Acq: 20 Apr 2022 19:35

Tgt Ion:184 Resp: 27018
Ion Ratio Lower Upper
184 100
63 61.5 51.8 77.6
154 66.1 54.3 81.5









#55
4-Nitrophenol
Concen: 9.783 ng/ul
RT: 14.769 min Scan# 1987
Delta R.T. -0.006 min
Lab File: BM034731.D
Acq: 20 Apr 2022 19:35




Instrument :
BNA_M
ClientSampleId :
SSTD010026


Tgt Ion:109 Resp: 48793
Ion Ratio Lower Upper
109 100
139 131.0 104.6 156.8
65 129.9 100.6 150.8



#56
Dibenzofuran
Concen: 9.673 ng/ul
RT: 14.974 min Scan# 2021
Delta R.T. -0.006 min
Lab File: BM034731.D
Acq: 20 Apr 2022 19:35

Tgt Ion:168 Resp: 488661
Ion Ratio Lower Upper
168 100
139 38.8 30.0 45.0





#57

2,4-Dinitrotoluene

Concen: 9.606 ng/ul

RT: 14.927 min Scan# 20

Delta R.T. -0.006 min

Lab File: BM034731.D

Acq: 20 Apr 2022 19:35

Instrument :

BNA_M

ClientSampleId :

SSTD010026





Tgt Ion:165 Resp: 91282

Ion Ratio Lower Upper

165 100

63 46.4 34.2 51.4

182 2.8 2.2 3.2



#58

2,3,4,6-Tetrachlorophenol

Concen: 7.996 ng/ul

RT: 15.198 min Scan# 2059

Delta R.T. 0.000 min

Lab File: BM034731.D

Acq: 20 Apr 2022 19:35

Tgt Ion:232 Resp: 81089




Ion Ratio Lower Upper


232 100

131 48.3 39.6 59.4

130 2.6 2.1 3.1

166 34.8 27.7 41.5









#59
Diethylphthalate
Concen: 9.798 ng/ul
RT: 15.404 min Scan# 2094
Delta R.T. 0.000 min
Lab File: BM034731.D
Acq: 20 Apr 2022 19:35




Instrument :
BNA_M
ClientSampleId :
SSTD010026


Tgt Ion:149 Resp: 388840
Ion Ratio Lower Upper
149 100
177 21.3 17.2 25.8
150 12.1 9.8 14.6



#61
Fluorene
Concen: 9.712 ng/ul
RT: 15.627 min Scan# 2132
Delta R.T. 0.000 min
Lab File: BM034731.D
Acq: 20 Apr 2022 19:35

Tgt Ion:166 Resp: 396773
Ion Ratio Lower Upper
166 100
165 99.1 77.9 116.9
167 13.9 10.3 15.5









#62
4-Chlorophenyl-phenylether
Concen: 9.066 ng/ul
RT: 15.621 min Scan# 2131
Delta R.T. 0.000 min
Lab File: BM034731.D
Acq: 20 Apr 2022 19:35




Instrument :
BNA_M
ClientSampleId :
SSTD010026


Tgt Ion:204 Resp: 198463
Ion Ratio Lower Upper
204 100
206 31.1 26.2 39.4
141 59.9 49.5 74.3



#63
4-Nitroaniline
Concen: 12.392 ng/ul
RT: 15.627 min Scan# 2132
Delta R.T. 0.000 min
Lab File: BM034731.D
Acq: 20 Apr 2022 19:35

Tgt Ion:138 Resp: 65722
Ion Ratio Lower Upper
138 100
92 49.7 37.2 55.8
108 77.1 58.4 87.6









#66
 4,6-Dinitro-2-methylphenol
 Concen: 7.959 ng/ul
 RT: 15.692 min Scan# 2143
 Delta R.T. 0.000 min
 Lab File: BM034731.D
 Acq: 20 Apr 2022 19:35




Instrument :
 BNA_M
 ClientSampleId :
 SSTD010026


Tgt Ion:198 Resp: 43743
 Ion Ratio Lower Upper
 198 100
 182 5.7 4.6 6.8
 77 25.6 22.2 33.2



#67
 N-Nitrosodiphenylamine
 Concen: 10.431 ng/ul
 RT: 15.833 min Scan# 2167
 Delta R.T. 0.000 min
 Lab File: BM034731.D
 Acq: 20 Apr 2022 19:35

Tgt Ion:169 Resp: 329236
 Ion Ratio Lower Upper
 169 100
 168 67.6 54.8 82.2
 167 35.6 28.9 43.3









#68
4-Bromophenyl-phenylether
Concen: 9.002 ng/ul
RT: 16.516 min Scan# 211
Delta R.T. 0.000 min
Lab File: BM034731.D
Acq: 20 Apr 2022 19:35




Instrument :
BNA_M
ClientSampleId :
SSTD010026


Tgt Ion:248 Resp: 111239
Ion Ratio Lower Upper
248 100
250 98.9 77.9 116.9
141 77.6 61.0 91.4



#69
Hexachlorobenzene
Concen: 8.676 ng/ul
RT: 16.627 min Scan# 2302
Delta R.T. 0.000 min
Lab File: BM034731.D
Acq: 20 Apr 2022 19:35

Tgt Ion:284 Resp: 126217
Ion Ratio Lower Upper
284 100
142 40.5 31.2 46.8
249 32.5 25.9 38.9





#70

Atrazine

Concen: 8.530 ng/ul

RT: 16.780 min Scan# 21

Delta R.T. 0.000 min

Lab File: BM034731.D

Acq: 20 Apr 2022 19:35

Instrument :

BNA_M

ClientSampleId :

SSTD010026





Tgt Ion:200 Resp: 103718

Ion Ratio Lower Upper

200 100

173 27.5 21.3 31.9

215 47.2 37.1 55.7



#71

Pentachlorophenol

Concen: 7.134 ng/ul

RT: 16.968 min Scan# 2360

Delta R.T. 0.000 min

Lab File: BM034731.D

Acq: 20 Apr 2022 19:35




Tgt Ion:266 Resp: 63354


Ion Ratio Lower Upper

266 100

264 62.6 48.4 72.6

268 64.8 50.6 76.0





#72

Phenanthrene

Concen: 10.026 ng/u1

RT: 17.363 min Scan# 2427

Delta R.T. 0.000 min

Lab File: BM034731.D

Acq: 20 Apr 2022 19:35

Instrument :

BNA_M

ClientSampleId :

SSTD010026





Tgt Ion:178 Resp: 579285

Ion Ratio Lower Upper

178 100

179 15.7 12.2 18.4

176 19.2 15.0 22.6



#74

Anthracene

Concen: 9.919 ng/u1

RT: 17.451 min Scan# 2442

Delta R.T. 0.000 min

Lab File: BM034731.D

Acq: 20 Apr 2022 19:35




Tgt Ion:178 Resp: 581443


Ion Ratio Lower Upper

178 100

179 15.4 12.2 18.4

176 18.5 14.6 22.0









#75
1,2,3,4-Tetrachlorobenzene
Concen: 9.172 ng/uL
RT: 13.380 min Scan# 1750
Delta R.T. 0.000 min
Lab File: BM034731.D
Acq: 20 Apr 2022 19:35




Instrument :
BNA_M
ClientSampleId :
SSTD010026


Tgt Ion:216 Resp: 172951
Ion Ratio Lower Upper
216 100
214 78.4 62.6 94.0
218 45.7 37.1 55.7



#76
Pentachlorobenzene
Concen: 9.114 ng/uL
RT: 14.898 min Scan# 2008
Delta R.T. 0.000 min
Lab File: BM034731.D
Acq: 20 Apr 2022 19:35

Tgt Ion:250 Resp: 158308
Ion Ratio Lower Upper
250 100
248 63.2 49.9 74.9
252 64.4 51.5 77.3









#77
 Carbazole
 Concen: 9.822 ng/ul
 RT: 17.715 min Scan# 2487
 Delta R.T. 0.000 min
 Lab File: BM034731.D
 Acq: 20 Apr 2022 19:35




Instrument :
 BNA_M
 ClientSampleId :
 SSTD010026


Tgt Ion:167 Resp: 503321
 Ion Ratio Lower Upper
 167 100
 166 21.6 17.1 25.7
 139 12.7 10.5 15.7



#78
 Di-n-butylphthalate
 Concen: 9.730 ng/ul
 RT: 18.298 min Scan# 2586
 Delta R.T. 0.000 min
 Lab File: BM034731.D
 Acq: 20 Apr 2022 19:35

Tgt Ion:149 Resp: 565600
 Ion Ratio Lower Upper
 149 100
 150 8.9 7.1 10.7
 104 5.1 4.1 6.1









#80
 Fluoranthene
 Concen: 11.004 ng/u1
 RT: 19.374 min Scan# 21
 Delta R.T. 0.000 min
 Lab File: BM034731.D
 Acq: 20 Apr 2022 19:35




Instrument :
 BNA_M
 ClientSampleId :
 SSTD010026


Tgt Ion:202 Resp: 606369
 Ion Ratio Lower Upper
 202 100
 101 9.8 7.4 11.2
 100 7.2 5.8 8.6



#82
 Pyrene
 Concen: 10.994 ng/u1
 RT: 19.733 min Scan# 2830
 Delta R.T. 0.000 min
 Lab File: BM034731.D
 Acq: 20 Apr 2022 19:35

Tgt Ion:202 Resp: 622106
 Ion Ratio Lower Upper
 202 100
 101 11.7 9.4 14.0
 100 9.6 7.8 11.6









#83
Butylbenzylphthalate
Concen: 11.110 ng/ul
RT: 20.639 min Scan# 2984
Delta R.T. 0.000 min
Lab File: BM034731.D
Acq: 20 Apr 2022 19:35




Instrument :
BNA_M
ClientSampleId :
SSTD010026


Tgt Ion:149 Resp: 205704
Ion Ratio Lower Upper
149 100
91 66.1 54.3 81.5
206 20.4 15.8 23.6



#84
3,3'-Dichlorobenzidine
Concen: 9.567 ng/ul
RT: 21.409 min Scan# 3115
Delta R.T. 0.000 min
Lab File: BM034731.D
Acq: 20 Apr 2022 19:35

Tgt Ion:252 Resp: 177325
Ion Ratio Lower Upper
252 100
254 64.4 51.0 76.4
126 11.5 9.0 13.6





#85

Benzo(a)anthracene

Concen: 10.049 ng/ul

RT: 21.480 min Scan# 3127

Delta R.T. 0.000 min

Lab File: BM034731.D

Acq: 20 Apr 2022 19:35

Instrument :

BNA_M

ClientSampleId :

SSTD010026





Tgt Ion:228 Resp: 557169

Ion Ratio Lower Upper

228 100

229 18.7 15.4 23.0

226 27.1 21.7 32.5



#86

Bis(2-ethylhexyl)phthalate

Concen: 10.944 ng/ul

RT: 21.427 min Scan# 3118

Delta R.T. 0.000 min

Lab File: BM034731.D

Acq: 20 Apr 2022 19:35




Tgt Ion:149 Resp: 323887


Ion Ratio Lower Upper

149 100

167 27.9 21.4 32.0

279 4.5 3.4 5.2









#87
Chrysene
Concen: 10.135 ng/ul
RT: 21.533 min Scan# 3136
Delta R.T. 0.000 min
Lab File: BM034731.D
Acq: 20 Apr 2022 19:35




Instrument :
BNA_M
ClientSampleId :
SSTD010026


Tgt Ion:228 Resp: 546385
Ion Ratio Lower Upper
228 100
226 29.5 23.7 35.5
229 19.4 15.7 23.5



#89
Di-n-octyl phthalate
Concen: 10.501 ng/ul
RT: 22.350 min Scan# 3275
Delta R.T. 0.000 min
Lab File: BM034731.D
Acq: 20 Apr 2022 19:35

Tgt Ion:149 Resp: 502908









#90
Benzo(b)fluoranthene
Concen: 9.910 ng/ul
RT: 23.162 min Scan# 3414
Delta R.T. -0.006 min
Lab File: BM034731.D
Acq: 20 Apr 2022 19:35




Instrument :
BNA_M
ClientSampleId :
SSTD010026


Tgt Ion:252 Resp: 542860
Ion Ratio Lower Upper
252 100
253 22.7 17.4 26.0
125 10.0 7.6 11.4



#91
Benzo(k)fluoranthene
Concen: 10.242 ng/ul
RT: 23.209 min Scan# 3421
Delta R.T. -0.006 min
Lab File: BM034731.D
Acq: 20 Apr 2022 19:35

Tgt Ion:252 Resp: 526365
Ion Ratio Lower Upper
252 100
253 21.9 17.4 26.2
125 10.1 7.8 11.6









#93
Benzo(a)pyrene
Concen: 9.970 ng/ul
RT: 23.786 min Scan# 31
Delta R.T. -0.006 min
Lab File: BM034731.D
Acq: 20 Apr 2022 19:35




Instrument :
BNA_M
ClientSampleId :
SSTD010026


Tgt Ion:252 Resp: 528737
Ion Ratio Lower Upper
252 100
253 21.9 16.6 25.0
125 10.4 6.2 9.2#



#94
Indeno(1,2,3-cd)pyrene
Concen: 11.203 ng/ul
RT: 26.380 min Scan# 3960
Delta R.T. -0.006 min
Lab File: BM034731.D
Acq: 20 Apr 2022 19:35

Tgt Ion:276 Resp: 714135
Ion Ratio Lower Upper
276 100
138 17.7 15.3 22.9
277 24.4 19.6 29.4









#95
 Dibenzo(a,h)anthracene
 Concen: 11.256 ng/ul
 RT: 26.397 min Scan# 3963
 Delta R.T. -0.006 min
 Lab File: BM034731.D
 Acq: 20 Apr 2022 19:35

Instrument :
 BNA_M
 ClientSampleId :
 SSTD010026

Tgt Ion:278 Resp: 618029
 Ion Ratio Lower Upper
 278 100
 139 12.1 9.7 14.5
 279 23.6 18.8 28.2



#96
 Benzo(g,h,i)perylene
 Concen: 11.241 ng/ul
 RT: 27.144 min Scan# 4090
 Delta R.T. 0.000 min
 Lab File: BM034731.D
 Acq: 20 Apr 2022 19:35

Tgt Ion:276 Resp: 612652
 Ion Ratio Lower Upper
 276 100
 138 15.5 13.0 19.6
 277 23.5 18.6 28.0

