

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM042122\
 Data File : BM034734.D
 Acq On : 20 Apr 2022 21:23
 Operator : CG/JU
 Sample : SSTD08029
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 SSTD080029

Quant Time: Apr 21 03:34:16 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\SFAM-EPA-BM042122.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Thu Apr 21 03:29:13 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.951	152	232218	20.000	ng/u1	0.00
20) Naphthalene-d8	10.757	136	820999	20.000	ng/u1	# 0.00
38) Acenaphthene-d10	14.580	164	522328	20.000	ng/u1	0.00
64) Phenanthrene-d10	17.321	188	1010903	20.000	ng/u1	# 0.00
79) Chrysene-d12	21.497	240	1143062	20.000	ng/u1	# 0.00
88) Perylene-d12	23.897	264	1249153	20.000	ng/u1	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.381	96	150764	27.929	ng/uL	0.00
4) Pyridine-d5	3.793	84	990182	78.176	ng/u1	0.00
7) Phenol-d5	7.110	99	1317722	77.497	ng/u1	0.00
9) Bis-(2-Chloroethyl)eth...	7.281	67	681907	71.313	ng/u1	0.00
11) 2-Chlorophenol-d4	7.481	132	1175998	79.983	ng/u1	0.00
15) 4-Methylphenol-d8	8.663	113	1083407	78.558	ng/u1	0.00
21) Nitrobenzene-d5	9.110	128	548206	98.172	ng/u1	0.00
24) 2-Nitrophenol-d4	9.834	143	623529	109.498	ng/u1	0.00
28) 2,4-Dichlorophenol-d3	10.375	165	1206224	89.020	ng/u1	0.00
31) 4-Chloroaniline-d4	10.886	131	1422071	80.290	ng/u1	0.00
46) Dimethylphthalate-d6	13.998	166	3227136	82.437	ng/u1	0.00
49) Acenaphthylene-d8	14.274	160	4061041	81.284	ng/u1	0.00
54) 4-Nitrophenol-d4	14.774	143	542693	89.832	ng/u1	0.02
60) Fluorene-d10	15.574	176	2673935	78.732	ng/u1	0.00
65) 4,6-Dinitro-2-methylph...	15.686	200	566220	109.373	ng/u1	0.00
73) Anthracene-d10	17.421	188	4079112	82.556	ng/u1	0.00
81) Pyrene-d10	19.709	212	4836515	79.229	ng/u1	0.00
92) Benzo(a)pyrene-d12	23.750	264	5295480	82.516	ng/u1	0.01
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.416	88	148852	27.124	ng/uL	91
5) Pyridine	3.816	79	1016876	77.481	ng/u1	88
6) Benzaldehyde	7.087	77	449806	50.531	ng/u1	92
8) Phenol	7.140	94	1288563	76.795	ng/u1	98
10) Bis(2-Chloroethyl)ether	7.375	93	977957	74.129	ng/u1	93
12) 2-Chlorophenol	7.516	128	1170693	78.726	ng/u1	94
13) 2-Methylphenol	8.392	108	985029	76.278	ng/u1	95
14) 2,2'-oxybis(1-Chloropr...	8.492	45	1095833	63.410	ng/u1#	91
16) Acetophenone	8.775	105	1620383	77.881	ng/u1	94
17) N-Nitroso-di-n-propyla...	8.781	70	758868	77.118	ng/u1	90
18) 4-Methylphenol	8.734	108	1075091	77.920	ng/u1	98
19) Hexachloroethane	9.039	117	476915	77.316	ng/u1#	76
22) Nitrobenzene	9.151	77	1094695	86.614	ng/u1	93
23) Isophorone	9.692	82	2017361	82.196	ng/u1	97
25) 2-Nitrophenol	9.869	139	639325	100.802	ng/u1	94
26) 2,4-Dimethylphenol	9.928	107	1180455	83.079	ng/u1	97
27) Bis(2-Chloroethoxy)met...	10.169	93	1202529	77.877	ng/u1	99
29) 2,4-Dichlorophenol	10.404	162	1144157	88.154	ng/u1	97
30) Naphthalene	10.810	128	3380223	80.064	ng/u1	99
32) 4-Chloroaniline	10.910	127	1397373	79.956	ng/u1	99
33) Hexachlorobutadiene	11.104	225	902465	92.505	ng/u1	99
34) Caprolactam	11.698	113	140630	45.275	ng/u1	78
35) 4-Chloro-3-methylphenol	12.039	107	1030970	85.098	ng/u1	98
36) 2-Methylnaphthalene	12.416	142	2394458	81.467	ng/u1	99


Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM042122\
 Data File : BM034734.D
 Acq On : 20 Apr 2022 21:23
 Operator : CG/JU
 Sample : SSTD08029
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 SSTD080029

Quant Time: Apr 21 03:34:16 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\SFAM-EPA-BM042122.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Thu Apr 21 03:29:13 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
37) 1-Methylnaphthalene	12.633	142	2381559	80.295	ng/ul	98
39) 1,2,4,5-Tetrachloroben...	12.786	216	1569388	83.826	ng/ul#	95
40) Hexachlorocyclopentadiene	12.769	237	1130240	87.808	ng/ul	99
41) 2,4,6-Trichlorophenol	13.016	196	956078	86.743	ng/ul	99
42) 2,4,5-Trichlorophenol	13.092	196	1016415	87.447	ng/ul	100
43) 1,1'-Biphenyl	13.422	154	3227003	74.228	ng/ul	99
44) 2-Chloronaphthalene	13.463	162	2557641	75.638	ng/ul	98
45) 2-Nitroaniline	13.657	65	556279	88.146	ng/ul	87
47) Dimethylphthalate	14.045	163	3081274	81.790	ng/ul	100
48) 2,6-Dinitrotoluene	14.157	165	650498	104.335	ng/ul	91
50) Acenaphthylene	14.304	152	4001386	81.312	ng/ul	98
51) 3-Nitroaniline	14.480	138	509813	95.350	ng/ul	92
52) Acenaphthene	14.645	153	2580798	80.094	ng/ul	99
53) 2,4-Dinitrophenol	14.680	184	386923	110.955	ng/ul#	83
55) 4-Nitrophenol	14.786	109	395480	85.925	ng/ul	95
56) Dibenzofuran	14.980	168	3744303	80.319	ng/ul	97
57) 2,4-Dinitrotoluene	14.939	165	899668	102.598	ng/ul#	86
58) 2,3,4,6-Tetrachlorophenol	15.204	232	885323	94.595	ng/ul#	84
59) Diethylphthalate	15.410	149	2928182	79.952	ng/ul	96
61) Fluorene	15.633	166	2959230	78.490	ng/ul	100
62) 4-Chlorophenyl-phenyle...	15.627	204	1660598	82.201	ng/ul	90
63) 4-Nitroaniline	15.645	138	434335	88.744	ng/ul	98
66) 4,6-Dinitro-2-methylph...	15.704	198	557004	105.567	ng/ul#	83
67) N-Nitrosodiphenylamine	15.833	169	2461308	81.232	ng/ul	99
68) 4-Bromophenyl-phenylether	16.515	248	1022140	86.168	ng/ul#	92
69) Hexachlorobenzene	16.633	284	1179260	84.437	ng/ul#	87
70) Atrazine	16.792	200	994980	85.244	ng/ul	96
71) Pentachlorophenol	16.974	266	770967	90.432	ng/ul	99
72) Phenanthrene	17.368	178	4441017	80.070	ng/ul	100
74) Anthracene	17.457	178	4624151	82.174	ng/ul	99
75) 1,2,3,4-Tetrachloroben...	13.386	216	1588601	87.761	ng/ul	99
76) Pentachlorobenzene	14.904	250	1486155	89.126	ng/ul	99
77) Carbazole	17.721	167	4028116	81.883	ng/ul	98
78) Di-n-butylphthalate	18.298	149	4777329	85.615	ng/ul	99
80) Fluoranthene	19.374	202	5481575	77.872	ng/ul	97
82) Pyrene	19.739	202	5583283	77.240	ng/ul	95
83) Butylbenzylphthalate	20.639	149	2113771	89.373	ng/ul#	91
84) 3,3'-Dichlorobenzidine	21.415	252	1931824	81.592	ng/ul#	98
85) Benzo(a)anthracene	21.486	228	5603078	79.112	ng/ul	98
86) Bis(2-ethylhexyl)phtha...	21.427	149	3231173	85.470	ng/ul#	95
87) Chrysene	21.539	228	5560754	80.747	ng/ul	99
89) Di-n-octyl phthalate	22.356	149	5709765	83.967	ng/ul	100
90) Benzo(b)fluoranthene	23.174	252	6360913	81.780	ng/ul	98
91) Benzo(k)fluoranthene	23.221	252	5956357	81.627	ng/ul	98
93) Benzo(a)pyrene	23.803	252	6260277	83.142	ng/ul	98
94) Indeno(1,2,3-cd)pyrene	26.403	276	7374839	81.485	ng/ul	99
95) Dibenzo(a,h)anthracene	26.427	278	6377133	81.802	ng/ul	99
96) Benzo(g,h,i)perylene	27.174	276	6189498	79.985	ng/ul	99





(#) = qualifier out of range (m) = manual integration (+) = signals summed



#2
1,4-Dioxane
Concen: 27.124 ng/uL
RT: 3.416 min Scan# 51
Delta R.T. -0.000 min
Lab File: BM034734.D
Acq: 20 Apr 2022 21:23




Instrument :
BNA_M
ClientSampleId :
SSTD080029


Tgt Ion: 88 Resp: 148852
Ion Ratio Lower Upper
88 100
43 27.7 27.6 41.4
58 67.7 59.5 89.3



#5
Pyridine
Concen: 77.481 ng/uL
RT: 3.816 min Scan# 124
Delta R.T. -0.000 min
Lab File: BM034734.D
Acq: 20 Apr 2022 21:23

Tgt Ion: 79 Resp: 1016876
Ion Ratio Lower Upper
79 100
52 68.8 64.3 96.5
51 34.0 32.4 48.6









#6
Benzaldehyde
Concen: 50.531 ng/ul
RT: 7.087 min Scan# 679
Delta R.T. 0.006 min
Lab File: BM034734.D
Acq: 20 Apr 2022 21:23




Instrument :
BNA_M
ClientSampleId :
SSTD080029


Tgt Ion: 77 Resp: 449806
Ion Ratio Lower Upper
77 100
105 104.6 77.4 116.0
106 100.0 73.4 110.2



#8
Phenol
Concen: 76.795 ng/ul
RT: 7.140 min Scan# 689
Delta R.T. 0.006 min
Lab File: BM034734.D
Acq: 20 Apr 2022 21:23

Tgt Ion: 94 Resp: 1288563
Ion Ratio Lower Upper
94 100
65 27.6 23.8 35.6
66 38.4 31.3 46.9









#10
Bis(2-Chloroethyl)ether
Concen: 74.129 ng/ul
RT: 7.375 min Scan# 71
Delta R.T. 0.006 min
Lab File: BM034734.D
Acq: 20 Apr 2022 21:23




Instrument :
BNA_M
ClientSampleId :
SSTD080029


Tgt Ion: 93 Resp: 977957
Ion Ratio Lower Upper
93 100
63 70.0 62.5 93.7
95 33.0 25.8 38.8



#12
2-Chlorophenol
Concen: 78.726 ng/ul
RT: 7.516 min Scan# 753
Delta R.T. 0.006 min
Lab File: BM034734.D
Acq: 20 Apr 2022 21:23

Tgt Ion:128 Resp: 1170693
Ion Ratio Lower Upper
128 100
64 39.7 37.1 55.7
130 33.1 26.2 39.4





#13

2-Methylphenol

Concen: 76.278 ng/ul

RT: 8.392 min Scan# 901

Delta R.T. 0.006 min

Lab File: BM034734.D

Acq: 20 Apr 2022 21:23

Instrument :

BNA_M

ClientSampleId :





SSTD080029

Tgt Ion:108 Resp: 985029

Ion Ratio Lower Upper

108 100

107 93.0 70.8 106.2



#14

2,2'-oxybis(1-Chloropropane)

Concen: 63.410 ng/ul

RT: 8.492 min Scan# 919

Delta R.T. 0.012 min

Lab File: BM034734.D

Acq: 20 Apr 2022 21:23




Tgt Ion: 45 Resp: 1095833


Ion Ratio Lower Upper

45 100

77 17.5 10.6 15.8#

79 13.0 8.1 12.1#









#16
Acetophenone
Concen: 77.881 ng/ul
RT: 8.775 min Scan# 966
Delta R.T. 0.006 min
Lab File: BM034734.D
Acq: 20 Apr 2022 21:23




Instrument :
BNA_M
ClientSampleId :
SSTD080029


Tgt Ion:105 Resp: 1620383
Ion Ratio Lower Upper
105 100
77 73.9 62.4 93.6
51 24.3 23.8 35.6



#17
N-Nitroso-di-n-propylamine
Concen: 77.118 ng/ul
RT: 8.781 min Scan# 968
Delta R.T. 0.018 min
Lab File: BM034734.D
Acq: 20 Apr 2022 21:23

Tgt Ion: 70 Resp: 758868
Ion Ratio Lower Upper
70 100
42 56.3 53.3 79.9
101 10.9 8.2 12.4
130 26.0 18.3 27.5









#18
4-Methylphenol
Concen: 77.920 ng/ul
RT: 8.734 min Scan# 90
Delta R.T. 0.018 min
Lab File: BM034734.D
Acq: 20 Apr 2022 21:23




Instrument :
BNA_M
ClientSampleId :
SSTD080029


Tgt Ion:108 Resp: 1075091
Ion Ratio Lower Upper
108 100
107 110.1 89.7 134.5



#19
Hexachloroethane
Concen: 77.316 ng/ul
RT: 9.039 min Scan# 1012
Delta R.T. -0.000 min
Lab File: BM034734.D
Acq: 20 Apr 2022 21:23

Tgt Ion:117 Resp: 476915
Ion Ratio Lower Upper
117 100
201 130.1 81.3 121.9#
199 80.9 53.4 80.0#









#22
Nitrobenzene
Concen: 86.614 ng/ul
RT: 9.151 min Scan# 1031
Delta R.T. 0.006 min
Lab File: BM034734.D
Acq: 20 Apr 2022 21:23




Instrument :
BNA_M
ClientSampleId :
SSTD080029


Tgt Ion: 77 Resp: 1094695
Ion Ratio Lower Upper
77 100
123 50.6 36.1 54.1
65 13.5 11.0 16.6



#23
Isophorone
Concen: 82.196 ng/ul
RT: 9.692 min Scan# 1123
Delta R.T. 0.012 min
Lab File: BM034734.D
Acq: 20 Apr 2022 21:23

Tgt Ion: 82 Resp: 2017361
Ion Ratio Lower Upper
82 100
95 8.1 6.2 9.4
138 21.2 15.3 22.9









#25
2-Nitrophenol
Concen: 100.802 ng/ul
RT: 9.869 min Scan# 1152
Delta R.T. 0.006 min
Lab File: BM034734.D
Acq: 20 Apr 2022 21:23




Instrument :
BNA_M
ClientSampleId :
SSTD080029


Tgt Ion:139 Resp: 639325
Ion Ratio Lower Upper
139 100
65 42.8 38.9 58.3
109 29.0 24.1 36.1



#26
2,4-Dimethylphenol
Concen: 83.079 ng/ul
RT: 9.928 min Scan# 1163
Delta R.T. -0.000 min
Lab File: BM034734.D
Acq: 20 Apr 2022 21:23

Tgt Ion:107 Resp: 1180455
Ion Ratio Lower Upper
107 100
121 52.4 40.3 60.5
122 85.4 66.6 100.0





#27

Bis(2-Chloroethoxy)methane

Concen: 77.877 ng/ul

RT: 10.169 min Scan# 1203

Delta R.T. 0.006 min

Lab File: BM034734.D

Acq: 20 Apr 2022 21:23

Instrument :





BNA_M

ClientSampleId :

SSTD080029

Tgt Ion: 93 Resp: 1202529

Ion	Ratio	Lower	Upper
93	100		
95	32.7	25.8	38.6
123	12.4	9.2	13.8



#29

2,4-Dichlorophenol

Concen: 88.154 ng/ul

RT: 10.404 min Scan# 1244




Delta R.T. 0.006 min


Lab File: BM034734.D

Acq: 20 Apr 2022 21:23

Tgt Ion:162 Resp: 1144157

Ion	Ratio	Lower	Upper
162	100		
164	65.5	53.0	79.4
98	30.3	27.4	41.0









#30
Naphthalene
Concen: 80.064 ng/ul
RT: 10.810 min Scan# 11
Delta R.T. 0.006 min
Lab File: BM034734.D
Acq: 20 Apr 2022 21:23




Instrument :
BNA_M
ClientSampleId :
SSTD080029


Tgt Ion:128 Resp: 3380223
Ion Ratio Lower Upper
128 100
129 10.7 9.1 13.7
127 13.0 10.6 16.0



#32
4-Chloroaniline
Concen: 79.956 ng/ul
RT: 10.910 min Scan# 1330
Delta R.T. 0.006 min
Lab File: BM034734.D
Acq: 20 Apr 2022 21:23

Tgt Ion:127 Resp: 1397373
Ion Ratio Lower Upper
127 100
129 32.3 25.4 38.2









#33
Hexachlorobutadiene
Concen: 92.505 ng/ul
RT: 11.104 min Scan# 11
Delta R.T. -0.000 min
Lab File: BM034734.D
Acq: 20 Apr 2022 21:23




Instrument :
BNA_M
ClientSampleId :
SSTD080029


Tgt Ion:225 Resp: 902465
Ion Ratio Lower Upper
225 100
223 62.4 49.8 74.8
227 63.4 49.6 74.4



#34
Caprolactam
Concen: 45.275 ng/ul
RT: 11.698 min Scan# 1464
Delta R.T. 0.029 min
Lab File: BM034734.D
Acq: 20 Apr 2022 21:23

Tgt Ion:113 Resp: 140630
Ion Ratio Lower Upper
113 100
55 133.6 130.6 195.8
56 104.8 104.3 156.5









#35
4-Chloro-3-methylphenol
Concen: 85.098 ng/ul
RT: 12.039 min Scan# 1522
Delta R.T. 0.012 min
Lab File: BM034734.D
Acq: 20 Apr 2022 21:23




Instrument :
BNA_M
ClientSampleId :
SSTD080029


Tgt Ion:107 Resp: 1030970
Ion Ratio Lower Upper
107 100
144 27.6 21.5 32.3
142 85.6 66.8 100.2



#36
2-Methylnaphthalene
Concen: 81.467 ng/ul
RT: 12.416 min Scan# 1586
Delta R.T. 0.006 min
Lab File: BM034734.D
Acq: 20 Apr 2022 21:23

Tgt Ion:142 Resp: 2394458
Ion Ratio Lower Upper
142 100
141 89.6 72.2 108.2





#37

1-Methylnaphthalene

Concen: 80.295 ng/ul

RT: 12.633 min Scan# 1

Delta R.T. 0.006 min

Lab File: BM034734.D

Acq: 20 Apr 2022 21:23

Instrument :

BNA_M

ClientSampleId :

SSTD080029





Tgt Ion:142 Resp: 2381559

Ion Ratio Lower Upper

142 100

141 93.6 73.0 109.4

116 3.5 3.2 4.8



#39

1,2,4,5-Tetrachlorobenzene

Concen: 83.826 ng/ul

RT: 12.786 min Scan# 1649

Delta R.T. 0.006 min

Lab File: BM034734.D

Acq: 20 Apr 2022 21:23

Tgt Ion:216 Resp: 1569388




Ion Ratio Lower Upper


216 100

214 77.9 64.5 96.7

179 19.6 17.9 26.9

108 12.2 15.0 22.4#









#40
Hexachlorocyclopentadiene
Concen: 87.808 ng/ul
RT: 12.769 min Scan# 1
Delta R.T. -0.000 min
Lab File: BM034734.D
Acq: 20 Apr 2022 21:23




Instrument :
BNA_M
ClientSampleId :
SSTD080029


Tgt Ion:237 Resp: 1130240
Ion Ratio Lower Upper
237 100
235 62.4 49.2 73.8
272 12.9 10.0 15.0



#41
2,4,6-Trichlorophenol
Concen: 86.743 ng/ul
RT: 13.016 min Scan# 1688
Delta R.T. -0.000 min
Lab File: BM034734.D
Acq: 20 Apr 2022 21:23

Tgt Ion:196 Resp: 956078
Ion Ratio Lower Upper
196 100
198 96.8 78.5 117.7
200 31.1 25.0 37.6









#42
2,4,5-Trichlorophenol
Concen: 87.447 ng/ul
RT: 13.092 min Scan# 1701
Delta R.T. 0.012 min
Lab File: BM034734.D
Acq: 20 Apr 2022 21:23




Instrument :
BNA_M
ClientSampleId :
SSTD080029


Tgt Ion:196 Resp: 1016415
Ion Ratio Lower Upper
196 100
198 96.9 77.7 116.5
200 31.4 25.0 37.4



#43
1,1'-Biphenyl
Concen: 74.228 ng/ul
RT: 13.422 min Scan# 1757
Delta R.T. -0.000 min
Lab File: BM034734.D
Acq: 20 Apr 2022 21:23

Tgt Ion:154 Resp: 3227003
Ion Ratio Lower Upper
154 100
153 40.8 32.2 48.2
76 9.5 8.6 12.8





#44

2-Chloronaphthalene

Concen: 75.638 ng/ul

RT: 13.463 min Scan# 1

Delta R.T. 0.006 min

Lab File: BM034734.D

Acq: 20 Apr 2022 21:23

Instrument :

BNA_M

ClientSampleId :

SSTD080029





Tgt Ion:162 Resp: 2557641

Ion Ratio Lower Upper

162 100

164 33.1 25.7 38.5

127 36.2 29.8 44.8



#45

2-Nitroaniline

Concen: 88.146 ng/ul

RT: 13.657 min Scan# 1797

Delta R.T. 0.006 min

Lab File: BM034734.D

Acq: 20 Apr 2022 21:23




Tgt Ion: 65 Resp: 556279


Ion Ratio Lower Upper

65 100

92 73.3 55.9 83.9

138 128.4 87.3 130.9





#47

Dimethylphthalate

Concen: 81.790 ng/ul

RT: 14.045 min Scan# 1863

Delta R.T. 0.006 min

Lab File: BM034734.D

Acq: 20 Apr 2022 21:23

Instrument :

BNA_M

ClientSampleId :

SSTD080029





Tgt Ion:163 Resp: 3081274

Ion Ratio Lower Upper

163 100

194 4.4 3.4 5.0

164 10.3 8.2 12.4



#48

2,6-Dinitrotoluene

Concen: 104.335 ng/ul

RT: 14.157 min Scan# 1882

Delta R.T. 0.006 min

Lab File: BM034734.D

Acq: 20 Apr 2022 21:23




Tgt Ion:165 Resp: 650498


Ion Ratio Lower Upper

165 100

89 46.8 43.1 64.7

121 18.4 16.6 24.8









#50
Acenaphthylene
Concen: 81.312 ng/ul
RT: 14.304 min Scan# 1907
Delta R.T. -0.000 min
Lab File: BM034734.D
Acq: 20 Apr 2022 21:23




Instrument :
BNA_M
ClientSampleId :
SSTD080029


Tgt Ion:152 Resp: 4001386
Ion Ratio Lower Upper
152 100
151 19.8 15.3 22.9
153 13.5 10.2 15.4



#51
3-Nitroaniline
Concen: 95.350 ng/ul
RT: 14.480 min Scan# 1937
Delta R.T. 0.006 min
Lab File: BM034734.D
Acq: 20 Apr 2022 21:23

Tgt Ion:138 Resp: 509813
Ion Ratio Lower Upper
138 100
108 10.8 10.2 15.4
92 98.3 85.4 128.0









#52
Acenaphthene
Concen: 80.094 ng/ul
RT: 14.645 min Scan# 1965
Delta R.T. -0.000 min
Lab File: BM034734.D
Acq: 20 Apr 2022 21:23




Instrument :
BNA_M
ClientSampleId :
SSTD080029


Tgt Ion:153 Resp: 2580798
Ion Ratio Lower Upper
153 100
152 46.3 37.5 56.3
154 88.1 71.0 106.4



#53
2,4-Dinitrophenol
Concen: 110.955 ng/ul
RT: 14.680 min Scan# 1971
Delta R.T. 0.006 min
Lab File: BM034734.D
Acq: 20 Apr 2022 21:23

Tgt Ion:184 Resp: 386923
Ion Ratio Lower Upper
184 100
63 49.0 51.8 77.6#
154 57.0 54.3 81.5









#55
4-Nitrophenol
Concen: 85.925 ng/ul
RT: 14.786 min Scan# 1987
Delta R.T. 0.012 min
Lab File: BM034734.D
Acq: 20 Apr 2022 21:23




Instrument :
BNA_M
ClientSampleId :
SSTD080029


Tgt Ion:109 Resp: 395480
Ion Ratio Lower Upper
109 100
139 136.4 104.6 156.8
65 119.3 100.6 150.8



#56
Dibenzofuran
Concen: 80.319 ng/ul
RT: 14.980 min Scan# 2022
Delta R.T. -0.000 min
Lab File: BM034734.D
Acq: 20 Apr 2022 21:23

Tgt Ion:168 Resp: 3744303
Ion Ratio Lower Upper
168 100
139 35.8 30.0 45.0









#57
2,4-Dinitrotoluene
Concen: 102.598 ng/ul
RT: 14.939 min Scan# 20
Delta R.T. 0.006 min
Lab File: BM034734.D
Acq: 20 Apr 2022 21:23




Instrument :
BNA_M
ClientSampleId :
SSTD080029


Tgt Ion:165 Resp: 899668
Ion Ratio Lower Upper
165 100
63 33.3 34.2 51.4#
182 3.1 2.2 3.2



#58
2,3,4,6-Tetrachlorophenol
Concen: 94.595 ng/ul
RT: 15.204 min Scan# 2060
Delta R.T. 0.006 min
Lab File: BM034734.D
Acq: 20 Apr 2022 21:23

Tgt Ion:232 Resp: 885323
Ion Ratio Lower Upper
232 100
131 35.7 39.6 59.4#
130 1.8 2.1 3.1#
166 28.0 27.7 41.5









#59
Diethylphthalate
Concen: 79.952 ng/ul
RT: 15.410 min Scan# 2095
Delta R.T. 0.006 min
Lab File: BM034734.D
Acq: 20 Apr 2022 21:23




Instrument :
BNA_M
ClientSampleId :
SSTD080029


Tgt Ion:149 Resp: 2928182
Ion Ratio Lower Upper
149 100
177 23.8 17.2 25.8
150 12.7 9.8 14.6



#61
Fluorene
Concen: 78.490 ng/ul
RT: 15.633 min Scan# 2133
Delta R.T. 0.006 min
Lab File: BM034734.D
Acq: 20 Apr 2022 21:23

Tgt Ion:166 Resp: 2959230
Ion Ratio Lower Upper
166 100
165 97.6 77.9 116.9
167 13.5 10.3 15.5









#62
4-Chlorophenyl-phenylether
Concen: 82.201 ng/ul
RT: 15.627 min Scan# 2131
Delta R.T. 0.006 min
Lab File: BM034734.D
Acq: 20 Apr 2022 21:23




Instrument :
BNA_M
ClientSampleId :
SSTD080029


Tgt Ion:204 Resp: 1660598
Ion Ratio Lower Upper
204 100
206 33.0 26.2 39.4
141 50.3 49.5 74.3



#63
4-Nitroaniline
Concen: 88.744 ng/ul
RT: 15.645 min Scan# 2135
Delta R.T. 0.018 min
Lab File: BM034734.D
Acq: 20 Apr 2022 21:23

Tgt Ion:138 Resp: 434335
Ion Ratio Lower Upper
138 100
92 46.3 37.2 55.8
108 75.5 58.4 87.6









#66
4,6-Dinitro-2-methylphenol
Concen: 105.567 ng/ul
RT: 15.704 min Scan# 2145
Delta R.T. 0.012 min
Lab File: BM034734.D
Acq: 20 Apr 2022 21:23




Instrument :
BNA_M
ClientSampleId :
SSTD080029


Tgt Ion:198 Resp: 557004
Ion Ratio Lower Upper
198 100
182 3.2 4.6 6.8#
77 18.0 22.2 33.2#



#67
N-Nitrosodiphenylamine
Concen: 81.232 ng/ul
RT: 15.833 min Scan# 2167
Delta R.T. -0.000 min
Lab File: BM034734.D
Acq: 20 Apr 2022 21:23

Tgt Ion:169 Resp: 2461308
Ion Ratio Lower Upper
169 100
168 67.5 54.8 82.2
167 35.9 28.9 43.3









#68
4-Bromophenyl-phenylether
Concen: 86.168 ng/ul
RT: 16.515 min Scan# 21
Delta R.T. -0.000 min
Lab File: BM034734.D
Acq: 20 Apr 2022 21:23




Instrument :
BNA_M
ClientSampleId :
SSTD080029


Tgt Ion:248 Resp: 1022140
Ion Ratio Lower Upper
248 100
250 97.0 77.9 116.9
141 60.4 61.0 91.4#



#69
Hexachlorobenzene
Concen: 84.437 ng/ul
RT: 16.633 min Scan# 2303
Delta R.T. 0.006 min
Lab File: BM034734.D
Acq: 20 Apr 2022 21:23

Tgt Ion:284 Resp: 1179260
Ion Ratio Lower Upper
284 100
142 25.6 31.2 46.8#
249 31.2 25.9 38.9





#70

Atrazine

Concen: 85.244 ng/ul

RT: 16.792 min Scan# 21

Delta R.T. 0.012 min

Lab File: BM034734.D

Acq: 20 Apr 2022 21:23

Instrument :

BNA_M

ClientSampleId :

SSTD080029





Tgt Ion:200 Resp: 994980

Ion Ratio Lower Upper

200 100

173 24.2 21.3 31.9

215 49.0 37.1 55.7



#71

Pentachlorophenol

Concen: 90.432 ng/ul

RT: 16.974 min Scan# 2361

Delta R.T. 0.006 min

Lab File: BM034734.D

Acq: 20 Apr 2022 21:23




Tgt Ion:266 Resp: 770967


Ion Ratio Lower Upper

266 100

264 62.0 48.4 72.6

268 63.8 50.6 76.0









#72
Phenanthrene
Concen: 80.070 ng/ul
RT: 17.368 min Scan# 2427
Delta R.T. 0.006 min
Lab File: BM034734.D
Acq: 20 Apr 2022 21:23




Instrument :
BNA_M
ClientSampleId :
SSTD080029


Tgt Ion:178 Resp: 4441017
Ion Ratio Lower Upper
178 100
179 15.5 12.2 18.4
176 18.7 15.0 22.6



#74
Anthracene
Concen: 82.174 ng/ul
RT: 17.457 min Scan# 2443
Delta R.T. 0.006 min
Lab File: BM034734.D
Acq: 20 Apr 2022 21:23

Tgt Ion:178 Resp: 4624151
Ion Ratio Lower Upper
178 100
179 15.6 12.2 18.4
176 18.5 14.6 22.0





#75

1,2,3,4-Tetrachlorobenzene

Concen: 87.761 ng/uL

RT: 13.386 min Scan# 1

Delta R.T. 0.006 min

Lab File: BM034734.D

Acq: 20 Apr 2022 21:23

Instrument :

BNA_M

ClientSampleId :

SSTD080029





Tgt Ion:216 Resp: 1588601

Ion Ratio Lower Upper

216 100

214 78.4 62.6 94.0

218 48.1 37.1 55.7



#76

Pentachlorobenzene

Concen: 89.126 ng/uL

RT: 14.904 min Scan# 2009

Delta R.T. 0.006 min

Lab File: BM034734.D

Acq: 20 Apr 2022 21:23




Tgt Ion:250 Resp: 1486155


Ion Ratio Lower Upper

250 100

248 63.6 49.9 74.9

252 64.2 51.5 77.3









#77
 Carbazole
 Concen: 81.883 ng/ul
 RT: 17.721 min Scan# 2488
 Delta R.T. 0.006 min
 Lab File: BM034734.D
 Acq: 20 Apr 2022 21:23




Instrument :
 BNA_M
 ClientSampleId :
 SSTD080029


Tgt Ion:167 Resp: 4028116
 Ion Ratio Lower Upper
 167 100
 166 22.0 17.1 25.7
 139 12.1 10.5 15.7



#78
 Di-n-butylphthalate
 Concen: 85.615 ng/ul
 RT: 18.298 min Scan# 2586
 Delta R.T. -0.000 min
 Lab File: BM034734.D
 Acq: 20 Apr 2022 21:23

Tgt Ion:149 Resp: 4777329
 Ion Ratio Lower Upper
 149 100
 150 9.4 7.1 10.7
 104 4.8 4.1 6.1









#80
Fluoranthene
Concen: 77.872 ng/u1
RT: 19.374 min Scan# 21
Delta R.T. -0.000 min
Lab File: BM034734.D
Acq: 20 Apr 2022 21:23




Instrument :
BNA_M
ClientSampleId :
SSTD080029


Tgt Ion:202 Resp: 5481575
Ion Ratio Lower Upper
202 100
101 8.0 7.4 11.2
100 6.5 5.8 8.6



#82
Pyrene
Concen: 77.240 ng/u1
RT: 19.739 min Scan# 2831
Delta R.T. 0.006 min
Lab File: BM034734.D
Acq: 20 Apr 2022 21:23

Tgt Ion:202 Resp: 5583283
Ion Ratio Lower Upper
202 100
101 9.5 9.4 14.0
100 8.0 7.8 11.6









#83
Butylbenzylphthalate
Concen: 89.373 ng/ul
RT: 20.639 min Scan# 2984
Delta R.T. -0.000 min
Lab File: BM034734.D
Acq: 20 Apr 2022 21:23




Instrument :
BNA_M
ClientSampleId :
SSTD080029


Tgt Ion:149 Resp: 2113771
Ion Ratio Lower Upper
149 100
91 60.7 54.3 81.5
206 24.1 15.8 23.6#



#84
3,3'-Dichlorobenzidine
Concen: 81.592 ng/ul
RT: 21.415 min Scan# 3116
Delta R.T. 0.006 min
Lab File: BM034734.D
Acq: 20 Apr 2022 21:23

Tgt Ion:252 Resp: 1931824
Ion Ratio Lower Upper
252 100
254 64.5 51.0 76.4
126 7.8 9.0 13.6#





#85

Benzo(a)anthracene

Concen: 79.112 ng/ul

RT: 21.486 min Scan# 3127

Delta R.T. 0.006 min

Lab File: BM034734.D

Acq: 20 Apr 2022 21:23

Instrument :

BNA_M

ClientSampleId :

SSTD080029





Tgt Ion:228 Resp: 5603078

Ion Ratio Lower Upper

228 100

229 20.2 15.4 23.0

226 27.7 21.7 32.5



#86

Bis(2-ethylhexyl)phthalate

Concen: 85.470 ng/ul

RT: 21.427 min Scan# 3118

Delta R.T. -0.000 min

Lab File: BM034734.D

Acq: 20 Apr 2022 21:23




Tgt Ion:149 Resp: 3231173


Ion Ratio Lower Upper

149 100

167 29.4 21.4 32.0

279 5.8 3.4 5.2#





#87

Chrysene

Concen: 80.747 ng/ul

RT: 21.539 min Scan# 31

Delta R.T. 0.006 min

Lab File: BM034734.D

Acq: 20 Apr 2022 21:23

Instrument :

BNA_M

ClientSampleId :

SSTD080029





Tgt Ion:228 Resp: 5560754

Ion Ratio Lower Upper

228 100

226 30.4 23.7 35.5

229 19.9 15.7 23.5



#89

Di-n-octyl phthalate

Concen: 83.967 ng/ul




RT: 22.356 min Scan# 3276


Delta R.T. 0.006 min

Lab File: BM034734.D

Acq: 20 Apr 2022 21:23

Tgt Ion:149 Resp: 5709765









#90
Benzo(b)fluoranthene
Concen: 81.780 ng/ul
RT: 23.174 min Scan# 3415
Delta R.T. 0.006 min
Lab File: BM034734.D
Acq: 20 Apr 2022 21:23




Instrument :
BNA_M
ClientSampleId :
SSTD080029


Tgt Ion:252 Resp: 6360913
Ion Ratio Lower Upper
252 100
253 22.1 17.4 26.0
125 7.7 7.6 11.4



#91
Benzo(k)fluoranthene
Concen: 81.627 ng/ul
RT: 23.221 min Scan# 3423
Delta R.T. 0.006 min
Lab File: BM034734.D
Acq: 20 Apr 2022 21:23

Tgt Ion:252 Resp: 5956357
Ion Ratio Lower Upper
252 100
253 22.2 17.4 26.2
125 7.9 7.8 11.6





#93

Benzo(a)pyrene

Concen: 83.142 ng/ul

RT: 23.803 min Scan# 3

Delta R.T. 0.012 min

Lab File: BM034734.D

Acq: 20 Apr 2022 21:23

Instrument :

BNA_M

ClientSampleId :

SSTD080029





Tgt Ion:252 Resp: 6260277

Ion Ratio Lower Upper

252 100

253 22.0 16.6 25.0

125 8.6 6.2 9.2



#94

Indeno(1,2,3-cd)pyrene

Concen: 81.485 ng/ul

RT: 26.403 min Scan# 3964

Delta R.T. 0.018 min

Lab File: BM034734.D

Acq: 20 Apr 2022 21:23




Tgt Ion:276 Resp: 7374839


Ion Ratio Lower Upper

276 100

138 19.5 15.3 22.9

277 25.2 19.6 29.4









#95
Dibenzo(a,h)anthracene
Concen: 81.802 ng/ul
RT: 26.427 min Scan# 3968
Delta R.T. 0.023 min
Lab File: BM034734.D
Acq: 20 Apr 2022 21:23

Instrument :
BNA_M
ClientSampleId :
SSTD080029

Tgt Ion:278 Resp: 6377133
Ion Ratio Lower Upper
278 100
139 12.5 9.7 14.5
279 24.1 18.8 28.2



#96
Benzo(g,h,i)perylene
Concen: 79.985 ng/ul
RT: 27.174 min Scan# 4095
Delta R.T. 0.029 min
Lab File: BM034734.D
Acq: 20 Apr 2022 21:23

Tgt Ion:276 Resp: 6189498
Ion Ratio Lower Upper
276 100
138 16.5 13.0 19.6
277 23.6 18.6 28.0

