

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM042122\
 Data File : BM034743.D
 Acq On : 21 Apr 2022 12:42
 Operator : CG/JU
 Sample : N2473-01
 Misc :
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 C0J06

Quant Time: Apr 22 01:11:07 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\SFAM-EPA-BM042122.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Thu Apr 21 05:03:15 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.945	152	206179	20.000	ng/u1	0.00
20) Naphthalene-d8	10.751	136	835421	20.000	ng/u1	0.00
38) Acenaphthene-d10	14.580	164	515451	20.000	ng/u1	0.00
64) Phenanthrene-d10	17.315	188	993280	20.000	ng/u1	0.00
79) Chrysene-d12	21.492	240	882430	20.000	ng/u1	0.00
88) Perylene-d12	23.885	264	895388	20.000	ng/u1	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.381	96	22963	4.894	ng/uL	0.00
4) Pyridine-d5	3.799	84	93948	8.066	ng/u1	0.00
7) Phenol-d5	7.104	99	93023	5.974	ng/u1	0.00
9) Bis-(2-Chloroethyl)eth...	7.275	67	263905	30.219	ng/u1	0.00
11) 2-Chlorophenol-d4	7.475	132	283982	21.547	ng/u1	0.00
15) 4-Methylphenol-d8	8.645	113	174728	14.023	ng/u1	-0.01
21) Nitrobenzene-d5	9.104	128	169108	27.710	ng/u1	0.00
24) 2-Nitrophenol-d4	9.828	143	162801	25.694	ng/u1	0.00
28) 2,4-Dichlorophenol-d3	10.369	165	293426	21.984	ng/u1	0.00
31) 4-Chloroaniline-d4	10.880	131	438544	24.300	ng/u1	0.00
46) Dimethylphthalate-d6	13.986	166	1089666	29.039	ng/u1	0.00
49) Acenaphthylene-d8	14.268	160	1268141	26.278	ng/u1	0.00
54) 4-Nitrophenol-d4	14.751	143	30424	4.830	ng/u1	0.00
60) Fluorene-d10	15.568	176	965860	29.746	ng/u1	0.00
65) 4,6-Dinitro-2-methylph...	15.674	200	129372	22.702	ng/u1	0.00
73) Anthracene-d10	17.415	188	1522784	31.484	ng/u1	0.00
81) Pyrene-d10	19.703	212	1622458	33.783	ng/u1	0.00
92) Benzo(a)pyrene-d12	23.738	264	1517253	33.504	ng/u1	0.00

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM042122\
Data File : BM034743.D
Acq On : 21 Apr 2022 12:42
Operator : CG/JU
Sample : N2473-01
Misc :
ALS Vial : 7 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
C0J06

Quant Time: Apr 22 01:11:07 2022
Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\SFAM-EPA-BM042122.M
Quant Title : SVOA CALIBRATION
QLast Update : Thu Apr 21 05:03:15 2022
Response via : Initial Calibration

