


Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM042222\
 Data File : BM034764.D
 Acq On : 22 Apr 2022 08:17
 Operator : CG/JU
 Sample : DFTPP050
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 DFTPP050

Integration File: rteint.p

Method : Z:\SVOASRV\HPCHEM1\BNA_M\METHODS\SFAM-EPA-SIM-BM042222.M
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 Last Update : Fri Apr 22 16:54:34 2022



AutoFind: Scans 2495, 2496, 2497; Background Corrected with Scan 2489

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	40.6	164042	PASS
68	69	0.00	2	1.7	2768	PASS
69	198	0.00	100	40.4	163539	PASS
70	69	0.00	2	0.5	874	PASS
127	198	10	80	49.7	201067	PASS
197	198	0.00	2	0.4	1653	PASS
198	198	100	100	100.0	404480	PASS
199	198	5	9	6.9	27968	PASS
275	198	10	60	27.8	112397	PASS
365	198	1	100	3.8	15365	PASS
441	198	0.01	100	12.7	51325	PASS
442	198	50	100	83.6	337963	PASS
443	442	15	24	19.3	65307	PASS

m/z	Abundance
35.90	215.0
40.00	922.0
41.00	268.0
43.10	331.0
44.00	1710.0
48.90	601.0
50.90	267.0
55.10	159.0
73.10	1133.0
75.10	215.0
76.90	164.0
77.90	151.0
81.00	159.0
81.90	170.0
83.90	546.0
86.00	281.0
95.90	541.0
97.20	202.0
103.00	156.0
114.90	182.0
133.00	536.0
134.90	241.0
146.90	351.0
149.00	158.0
162.90	150.0
176.90	213.0
178.80	159.0
191.00	708.0
192.00	178.0
192.90	547.0
207.00	6015.0
208.00	1360.0
209.00	779.0
220.90	204.0
281.00	1282.0
282.00	256.0
283.10	183.0

Instrument :
BNA_M
ClientSampleId :
DFTPP050

m/z	Abundance
36.10	177.0
37.10	678.0
38.00	2212.0
39.00	10946.0
40.00	1436.0
41.00	423.0
43.00	290.0
44.00	1578.0
45.00	279.0
46.90	289.0
48.10	157.0
49.00	1677.0
50.00	36112.0
51.00	131136.0
52.00	7294.0
53.00	349.0
55.00	865.0
56.00	4035.0
57.00	8890.0
58.00	262.0
59.10	270.0
61.00	1837.0
62.00	1723.0
63.00	5413.0
64.10	687.0
65.00	2244.0
66.90	233.0
68.00	2005.0
69.00	127224.0
70.00	703.0
73.00	4428.0
74.00	13844.0
75.00	21136.0
76.10	6525.0
77.10	137344.0
78.10	9509.0
79.00	9801.0
80.00	7646.0
81.00	11105.0
82.00	2514.0
83.00	2776.0
83.90	695.0
85.00	1837.0
86.00	3270.0
87.00	1395.0
88.10	410.0
89.00	290.0
91.00	1984.0
92.00	2959.0
93.00	18640.0
94.00	1154.0
94.90	255.0
96.10	1129.0
96.90	328.0
98.00	13924.0
99.00	12889.0
99.90	845.0
101.00	5785.0
102.00	425.0
102.90	2233.0
104.00	4138.0
105.00	3827.0
106.10	1357.0
107.00	44624.0
108.00	6398.0
109.00	1064.0
110.00	75240.0
111.00	11655.0
112.10	1463.0
113.00	623.0
114.10	204.0
115.00	271.0
116.00	2461.0
117.00	43200.0
118.00	3140.0
118.90	549.0
120.00	655.0
121.00	282.0
122.00	2801.0
123.00	4788.0
124.00	2254.0
125.00	2040.0
127.00	145536.0
128.00	10992.0
129.00	58528.0
130.00	4910.0
130.90	1028.0
131.90	618.0
132.90	835.0
134.00	2060.0
135.00	4951.0
136.00	2430.0

Instrument :
BNA_M
ClientSampleId :
DFTPP050

137.00	3290.0
138.10	636.0
138.90	324.0
140.00	685.0
141.00	8135.0
142.00	2131.0
142.90	1563.0
144.10	518.0
145.00	487.0
146.00	1654.0
147.00	6072.0
148.00	10773.0
149.00	2600.0
150.00	970.0
151.00	1011.0
151.80	576.0
153.00	2729.0
154.00	1622.0
155.00	4371.0
156.00	5982.0
157.00	1362.0
158.00	1445.0
159.00	1022.0
160.00	2542.0
161.00	3992.0
161.90	1165.0
162.90	479.0
163.80	324.0
165.00	3288.0
166.00	2615.0
167.00	16201.0
168.00	11430.0
169.00	1559.0
170.00	506.0
171.10	835.0
172.00	1678.0
173.00	1783.0
174.00	3396.0
175.10	6120.0
175.90	1820.0
177.00	3260.0
178.00	1336.0
179.00	11159.0
180.00	9015.0
181.00	3950.0
182.10	557.0
182.70	214.0
183.20	224.0
184.00	995.0
185.00	5761.0
186.00	42248.0
187.00	11777.0
188.00	1667.0
189.00	2864.0
189.90	550.0
191.00	1945.0
192.00	3898.0
193.00	5001.0
194.00	912.0
194.90	333.0
196.00	7445.0
196.80	1354.0
197.90	264256.0
199.00	18736.0
200.00	1613.0
201.40	1172.0
203.00	2534.0
204.00	10765.0
205.00	19032.0
206.00	75056.0
207.00	17648.0
208.00	4397.0
209.00	2031.0
210.10	1098.0
211.00	3197.0
212.00	498.0
215.00	883.0
216.00	1482.0
217.00	20304.0
218.00	2450.0
219.00	383.0
221.00	10320.0
221.90	2782.0
223.00	4065.0
224.00	43016.0
225.00	10650.0
226.00	1484.0
227.00	18912.0
228.00	2960.0
229.00	3783.0
230.10	416.0
231.00	1175.0
232.90	305.0
233.90	1276.0
235.00	1169.0
236.00	1003.0

Instrument :
BNA_M
ClientSampleId :
DFTPP050

237.00	1910.0
238.00	396.0
239.00	809.0
240.00	772.0
241.00	930.0
242.00	1971.0
243.00	2535.0
244.00	28528.0
245.00	4845.0
246.00	6523.0
246.90	1637.0
247.90	435.0
248.90	1250.0
249.80	156.0
251.00	269.0
252.00	394.0
252.90	654.0
254.10	828.0
255.00	165632.0
256.00	24872.0
257.00	2170.0
258.00	10671.0
258.90	1365.0
260.00	324.0
261.10	156.0
263.90	398.0
265.00	4060.0
266.00	684.0
267.00	554.0
267.90	244.0
269.90	174.0
270.90	321.0
272.00	468.0
273.00	4612.0
274.00	11843.0
275.00	68680.0
276.00	9958.0
277.00	6594.0
278.00	1296.0
279.10	375.0
281.00	1880.0
282.00	861.0
283.00	1167.0
284.00	594.0
285.00	1137.0
286.00	306.0
288.90	303.0
289.80	195.0
290.90	268.0
292.00	232.0
293.00	1270.0
293.90	644.0
295.10	565.0
296.00	20448.0
297.00	3446.0
297.90	216.0
301.00	185.0
301.90	371.0
303.00	2496.0
304.10	647.0
308.00	448.0
309.00	184.0
311.20	153.0
312.80	239.0
314.10	732.0
315.00	2457.0
316.00	1231.0
317.00	286.0
321.00	465.0
322.00	211.0
323.10	5812.0
324.00	1363.0
325.00	272.0
325.90	221.0
327.00	1377.0
328.00	886.0
332.00	499.0
332.90	483.0
334.00	3679.0
335.00	1039.0
341.00	1135.0
341.90	347.0
343.00	232.0
346.00	1382.0
347.00	322.0
352.00	1537.0
353.00	1138.0
354.00	1652.0
355.10	1419.0
356.10	477.0
357.10	282.0
365.00	8478.0
365.90	1149.0
369.90	207.0
370.90	437.0
372.00	3162.0

Instrument :
BNA_M
ClientSampleId :
DFTPP050

373.10	510.0
383.00	759.0
384.10	297.0
390.00	306.0
390.90	345.0
401.00	338.0
402.00	1154.0
403.00	1570.0
404.00	671.0
405.00	280.0
415.00	162.0
421.00	1091.0
422.10	1258.0
423.00	8976.0
424.00	3002.0
425.10	645.0
429.10	219.0
441.10	23360.0
442.10	162048.0
443.10	31312.0
444.10	3621.0

Instrument :
BNA_M
ClientSampleId :
DFTPP050

m/z	Abundance
36.00	412.0
37.00	1011.0
38.00	2466.0
39.00	16512.0
40.00	1335.0
41.00	547.0
43.00	369.0
44.00	2281.0
45.00	549.0
47.90	213.0
49.10	1927.0
50.00	52176.0
51.10	196032.0
52.00	9693.0
53.00	539.0
55.10	1310.0
56.00	6010.0
57.00	12027.0
57.90	658.0
59.00	265.0
59.90	249.0
61.00	2727.0
62.00	3258.0
63.00	7793.0
64.00	989.0
65.10	3621.0
65.90	373.0
67.00	269.0
68.00	3571.0
69.00	194688.0
70.00	781.0
71.00	161.0
71.90	168.0
73.10	4384.0
74.00	21880.0
75.00	32312.0
76.10	10847.0
77.10	217472.0
78.10	14706.0
79.00	15300.0
80.00	12748.0
81.00	16165.0
82.00	4450.0
83.00	3212.0
83.90	784.0
85.00	2654.0
86.00	4838.0
87.00	2115.0
88.00	840.0
89.00	613.0
91.00	3526.0
92.00	4280.0
93.00	29056.0
94.00	1902.0
95.00	461.0
96.00	1867.0
96.90	607.0
98.00	22288.0
99.00	18064.0
100.00	1571.0
101.00	9299.0
102.00	452.0
103.00	3344.0
104.00	6382.0
105.00	5996.0
106.10	1687.0
107.00	69624.0
108.00	9883.0
109.10	2044.0
110.00	120816.0
111.00	19720.0
112.00	2628.0
113.00	1183.0
113.80	214.0
114.70	406.0
116.00	3974.0
117.00	69344.0
118.00	4788.0
118.90	643.0
120.00	984.0
121.00	423.0
122.00	5150.0
123.00	7822.0
124.00	3354.0
125.00	3194.0
126.10	942.0
127.00	240192.0
128.00	19088.0
129.00	95136.0
130.00	9037.0
131.00	1713.0
132.00	1094.0

Instrument :
BNA_M
ClientSampleId :
DFTPP050

133.00	1206.0
134.00	2753.0
135.00	8731.0
136.00	3260.0
137.00	4929.0
138.10	853.0
139.00	604.0
140.00	1218.0
141.00	13334.0
142.00	4639.0
143.00	3585.0
143.90	715.0
145.00	640.0
145.90	2339.0
147.00	8054.0
148.00	17216.0
149.00	3742.0
149.90	845.0
151.00	1786.0
151.90	1004.0
153.00	4354.0
154.00	3043.0
155.00	7297.0
156.00	10738.0
157.00	2294.0
158.00	2390.0
159.00	1771.0
160.00	4338.0
161.00	6150.0
162.00	1797.0
163.00	713.0
164.00	831.0
165.00	5064.0
166.00	4132.0
167.00	27400.0
168.00	16824.0
169.00	2248.0
170.00	893.0
171.00	1149.0
172.00	2648.0
173.00	3262.0
174.00	5725.0
175.00	10895.0
176.00	2675.0
177.00	5174.0
177.90	1834.0
179.00	18856.0
180.00	15039.0
181.00	6026.0
182.00	1235.0
183.10	636.0
183.90	2026.0
185.00	9750.0
186.00	73048.0
187.00	21760.0
188.00	2603.0
189.00	5104.0
189.90	929.0
191.00	2930.0
192.00	7356.0
193.00	7418.0
194.00	1777.0
194.90	704.0
196.00	13591.0
197.00	2833.0
198.00	475392.0
199.00	33248.0
199.90	2994.0
201.30	1683.0
203.00	4321.0
204.00	19104.0
205.00	34336.0
206.00	137920.0
207.00	24832.0
208.00	6353.0
209.00	2292.0
210.00	2269.0
210.90	5491.0
211.90	1234.0
213.10	441.0
215.00	1816.0
216.10	2749.0
217.00	37520.0
218.00	4497.0
219.00	506.0
221.00	18088.0
221.90	4595.0
223.00	8468.0
224.00	73464.0
225.00	19880.0
226.00	2129.0
227.00	34976.0
228.00	4479.0
229.00	7018.0
229.90	934.0
231.00	2655.0

Instrument :
BNA_M
ClientSampleId :
DFTPP050

232.00	337.0
233.10	555.0
234.00	2439.0
235.00	2736.0
236.00	1734.0
237.00	3601.0
238.00	663.0
238.90	1493.0
240.00	1315.0
240.90	2050.0
242.00	3809.0
243.00	3666.0
244.00	55704.0
245.00	7636.0
246.00	12330.0
247.00	2677.0
248.00	673.0
249.00	1956.0
249.90	459.0
251.00	528.0
252.00	649.0
253.00	1334.0
254.10	2059.0
255.00	302592.0
256.00	45128.0
257.00	3781.0
258.00	19088.0
259.00	3272.0
260.00	578.0
261.00	431.0
262.80	167.0
264.00	632.0
265.00	8038.0
265.90	1145.0
267.00	510.0
268.00	184.0
269.90	327.0
271.00	595.0
272.00	1103.0
273.00	8527.0
274.00	23776.0
275.00	130848.0
276.00	17576.0
277.00	11969.0
278.00	1973.0
279.00	443.0
281.10	1765.0
282.10	900.0
283.00	1534.0
283.90	1019.0
284.90	2056.0
286.00	399.0
289.00	566.0
290.00	378.0
291.10	300.0
292.00	683.0
293.00	2800.0
294.00	887.0
295.00	889.0
296.00	43032.0
297.00	5667.0
298.00	315.0
301.00	501.0
302.10	618.0
303.00	4918.0
304.10	1035.0
305.00	253.0
308.00	476.0
308.90	403.0
310.00	577.0
310.90	201.0
311.90	252.0
312.90	418.0
314.10	2159.0
315.00	4860.0
316.00	2622.0
317.00	476.0
321.00	1313.0
322.00	568.0
323.10	10930.0
324.00	2104.0
324.90	273.0
326.00	277.0
327.00	2608.0
328.00	1328.0
328.90	276.0
332.00	888.0
332.90	1134.0
334.00	8330.0
335.00	1865.0
336.00	191.0
339.00	246.0
340.10	260.0
341.00	1904.0
342.00	387.0
343.00	177.0

Instrument :
BNA_M
ClientSampleId :
DFTPP050

346.00	2387.0
347.00	726.0
350.00	170.0
351.00	185.0
352.00	3502.0
353.10	2513.0
354.00	3195.0
355.10	1479.0
356.10	202.0
356.90	215.0
359.10	238.0
364.00	190.0
365.00	17176.0
366.00	2551.0
370.00	288.0
370.90	756.0
372.00	6022.0
373.10	1498.0
373.80	221.0
383.00	1580.0
384.00	671.0
384.80	171.0
390.10	781.0
391.00	289.0
391.90	479.0
401.00	677.0
402.00	2240.0
403.00	3486.0
404.00	1517.0
404.90	383.0
415.10	218.0
421.00	3048.0
422.00	2547.0
423.10	18848.0
424.00	5985.0
425.00	649.0
428.90	157.0
441.10	57176.0
442.10	370048.0
443.10	70856.0
444.10	7369.0
445.00	413.0

Instrument :
BNA_M
ClientSampleId :
DFTPP050

m/z	Abundance
36.00	264.0
37.00	905.0
38.00	2586.0
39.00	14126.0
40.00	1580.0
41.00	504.0
42.90	233.0
44.00	2196.0
44.90	527.0
47.00	165.0
48.00	236.0
49.00	1597.0
50.10	44456.0
51.10	165760.0
52.00	7919.0
53.00	572.0
55.00	1114.0
56.00	4816.0
57.00	10964.0
58.00	674.0
59.00	217.0
60.00	237.0
61.00	2156.0
62.00	2071.0
63.00	6706.0
64.00	1123.0
65.00	3341.0
66.00	358.0
66.90	285.0
68.00	2727.0
69.00	168704.0
70.00	1138.0
71.10	231.0
73.00	3136.0
74.00	17912.0
75.00	27496.0
76.10	9132.0
77.10	185728.0
78.10	11947.0
79.00	13848.0
80.00	9872.0
81.00	13936.0
82.00	3824.0
83.00	2899.0
83.90	737.0
85.00	2430.0
85.90	4696.0
87.00	2120.0
88.00	674.0
88.80	305.0
90.00	195.0
91.00	3108.0
92.00	3834.0
93.00	24992.0
94.00	1737.0
95.10	387.0
96.00	1871.0
96.90	550.0
98.00	20456.0
99.00	14282.0
100.00	1186.0
101.00	7856.0
101.90	567.0
103.00	2915.0
104.00	5139.0
105.00	5168.0
106.00	1700.0
107.00	63080.0
108.00	9641.0
109.10	1708.0
110.00	106368.0
111.00	15892.0
112.00	2575.0
113.00	882.0
114.90	420.0
116.00	3674.0
117.00	61736.0
118.00	4842.0
119.00	548.0
120.00	653.0
121.10	418.0
122.00	4583.0
123.00	6133.0
124.00	2897.0
125.00	3030.0
126.10	905.0
127.00	217472.0
128.00	16640.0
129.00	88024.0
130.00	7456.0
131.00	1571.0
132.10	588.0

Instrument :
BNA_M
ClientSampleId :
DFTPP050

133.10	942.0
134.00	2880.0
135.00	7677.0
136.00	3542.0
137.00	4266.0
138.00	972.0
138.90	335.0
140.00	1438.0
141.00	12328.0
142.00	3736.0
142.90	2424.0
144.00	577.0
145.00	678.0
146.00	2316.0
147.00	7291.0
148.00	15299.0
149.00	3539.0
150.00	839.0
151.00	1328.0
151.90	1045.0
153.00	4254.0
153.90	3362.0
155.00	6956.0
156.00	9785.0
157.00	1557.0
158.00	2126.0
159.00	2189.0
160.00	4032.0
161.00	5479.0
162.00	1539.0
163.00	675.0
163.90	776.0
165.00	4772.0
166.10	3673.0
167.00	27304.0
168.00	15115.0
169.00	2000.0
169.90	1171.0
171.00	1365.0
172.00	2356.0
173.00	2733.0
174.00	6005.0
175.00	10095.0
176.00	2730.0
177.00	5047.0
178.00	2056.0
179.00	20976.0
180.00	13154.0
181.00	6028.0
182.00	956.0
182.90	589.0
184.00	2006.0
185.00	9465.0
186.00	69400.0
187.00	20800.0
188.00	2452.0
189.00	4775.0
190.00	1071.0
191.00	2831.0
192.00	6310.0
193.00	7876.0
194.00	1546.0
194.90	652.0
196.00	13789.0
197.00	2127.0
198.00	473792.0
199.00	31920.0
200.00	2477.0
201.40	1748.0
203.00	3644.0
204.00	20568.0
205.00	32712.0
206.00	136384.0
207.00	23768.0
208.00	6342.0
209.10	2665.0
210.00	2174.0
211.00	4998.0
211.80	1293.0
212.90	407.0
213.90	193.0
215.00	1385.0
216.00	3380.0
217.00	38464.0
218.00	4865.0
219.00	523.0
221.00	17344.0
222.00	4437.0
223.00	8271.0
224.00	77192.0
225.00	20112.0
226.00	2243.0
227.00	35208.0
228.00	5179.0
229.00	7064.0
230.00	925.0

Instrument :
BNA_M
ClientSampleId :
DFTPP050

231.00	3047.0
231.90	702.0
233.00	866.0
234.00	2590.0
235.00	2397.0
236.00	1820.0
237.00	3135.0
238.00	467.0
239.00	1460.0
240.00	1248.0
241.00	2128.0
242.00	3512.0
243.00	4567.0
244.00	59664.0
245.00	7338.0
246.00	12504.0
247.00	2732.0
248.00	587.0
249.00	2159.0
249.90	319.0
250.90	631.0
251.90	549.0
253.00	1361.0
254.10	2090.0
255.00	318976.0
256.00	46008.0
257.00	3729.0
258.00	19544.0
259.00	3091.0
260.00	568.0
261.10	442.0
262.90	156.0
264.00	522.0
265.00	8699.0
265.90	1291.0
267.00	413.0
269.10	181.0
270.00	493.0
270.90	783.0
272.00	732.0
273.00	9377.0
274.00	24344.0
275.00	137664.0
276.00	18584.0
277.00	12541.0
278.00	2593.0
279.00	518.0
281.00	1302.0
282.00	625.0
283.00	1688.0
284.10	1076.0
285.00	2617.0
286.00	350.0
289.00	579.0
290.10	385.0
290.90	347.0
292.00	550.0
292.90	2989.0
293.90	921.0
295.00	1032.0
296.00	46472.0
297.00	6141.0
297.90	368.0
300.90	530.0
302.10	650.0
303.00	4771.0
304.00	1113.0
305.10	265.0
307.90	618.0
309.00	532.0
309.90	531.0
311.00	173.0
312.00	185.0
312.80	428.0
314.00	2011.0
315.00	5108.0
316.10	2962.0
316.90	567.0
321.00	1275.0
322.00	782.0
323.00	13824.0
324.00	2472.0
324.90	413.0
326.00	415.0
327.00	3081.0
328.00	1257.0
328.90	234.0
332.00	1191.0
333.00	1159.0
334.00	9098.0
335.00	2105.0
335.90	409.0
339.00	392.0
340.00	179.0
341.10	1508.0
341.90	553.0

Instrument :
BNA_M
ClientSampleId :
DFTPP050

346.00	2891.0
347.10	661.0
350.90	410.0
352.10	3851.0
353.10	2877.0
354.10	4394.0
355.00	1327.0
355.90	203.0
357.10	222.0
358.90	232.0
365.00	20440.0
366.10	2669.0
366.90	274.0
370.00	430.0
371.00	841.0
372.00	6950.0
373.00	1529.0
374.10	194.0
383.00	1863.0
384.00	641.0
385.00	255.0
390.10	862.0
391.00	713.0
392.00	339.0
401.00	640.0
402.00	2838.0
403.00	4148.0
404.00	1445.0
405.00	368.0
415.00	237.0
421.00	3188.0
422.00	3499.0
423.00	23008.0
424.10	6844.0
425.10	988.0
440.20	152.0
441.10	73440.0
442.10	481792.0
443.10	93752.0
444.00	8788.0
445.00	473.0

Instrument :

BNA_M

ClientSampleId :

DFTPP050