

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM042223\
 Data File : BM039606.D
 Acq On : 23 Apr 2023 10:21
 Operator : CG/JU
 Sample : SSTDCCC020
 Misc :
 ALS Vial : 34 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 SSTD020120

Quant Time: Apr 24 01:02:35 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\SFAM-EPA-BM041823.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Sun Apr 23 00:18:16 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.813	152	238306	20.000	ng/u1	0.00
20) Naphthalene-d8	10.619	136	1055464	20.000	ng/u1	0.00
38) Acenaphthene-d10	14.472	164	614109	20.000	ng/u1	0.00
64) Phenanthrene-d10	17.224	188	1150438	20.000	ng/u1	0.00
79) Chrysene-d12	21.418	240	850168	20.000	ng/u1	0.00
88) Perylene-d12	23.783	264	869329	20.000	ng/u1	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.208	96	52552	8.465	ng/uL	0.00
4) Pyridine-d5	3.631	84	358858	20.571	ng/u1	0.00
7) Phenol-d5	6.996	99	453481	20.791	ng/u1	0.00
9) Bis-(2-Chloroethyl)eth...	7.148	67	308963	21.432	ng/u1	0.00
11) 2-Chlorophenol-d4	7.348	132	363681	21.256	ng/u1	0.00
15) 4-Methylphenol-d8	8.542	113	356695	20.302	ng/u1	0.00
21) Nitrobenzene-d5	8.984	128	184162	21.688	ng/u1	0.00
24) 2-Nitrophenol-d4	9.707	143	207805	21.794	ng/u1	0.00
28) 2,4-Dichlorophenol-d3	10.254	165	352377	21.478	ng/u1	0.00
31) 4-Chloroaniline-d4	10.772	131	485012	19.793	ng/u1	0.00
46) Dimethylphthalate-d6	13.883	166	1003641	20.264	ng/u1	0.00
49) Acenaphthylene-d8	14.166	160	1209316	21.735	ng/u1	0.00
54) 4-Nitrophenol-d4	14.719	143	100012	13.343	ng/u1	-0.01
60) Fluorene-d10	15.466	176	838641	20.668	ng/u1	0.00
65) 4,6-Dinitro-2-methylph...	15.607	200	152558	19.955	ng/u1	0.00
73) Anthracene-d10	17.324	188	1191983	21.544	ng/u1	0.00
81) Pyrene-d10	19.624	212	1225874	21.699	ng/u1	0.00
92) Benzo(a)pyrene-d12	23.624	264	990283	21.099	ng/u1	0.00
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.243	88	58961	8.678	ng/uL	96
5) Pyridine	3.649	79	390754	21.244	ng/u1	98
6) Benzaldehyde	6.954	77	266330	28.581	ng/u1	95
8) Phenol	7.025	94	478551	21.074	ng/u1	99
10) Bis(2-Chloroethyl)ether	7.243	93	401676	21.215	ng/u1	99
12) 2-Chlorophenol	7.378	128	371406	21.573	ng/u1	97
13) 2-Methylphenol	8.272	108	363114	20.587	ng/u1	99
14) 2,2'-oxybis(1-Chloropr...	8.348	45	589429	22.414	ng/u1	99
16) Acetophenone	8.648	105	616283	21.355	ng/u1	99
17) N-Nitroso-di-n-propyla...	8.631	70	338263	21.178	ng/u1	96
18) 4-Methylphenol	8.607	108	390290	20.495	ng/u1	99
19) Hexachloroethane	8.884	117	166319	21.270	ng/u1	98
22) Nitrobenzene	9.025	77	474575	22.635	ng/u1	96
23) Isophorone	9.548	82	943631	21.106	ng/u1	99
25) 2-Nitrophenol	9.736	139	213313	21.441	ng/u1	93
26) 2,4-Dimethylphenol	9.807	107	443725	21.826	ng/u1	99
27) Bis(2-Chloroethoxy)met...	10.036	93	546501	20.787	ng/u1	100
29) 2,4-Dichlorophenol	10.284	162	345582	21.383	ng/u1	99
30) Naphthalene	10.672	128	1255394	21.497	ng/u1	99
32) 4-Chloroaniline	10.795	127	480496	19.901	ng/u1	100
33) Hexachlorobutadiene	10.942	225	226279	21.249	ng/u1	99
34) Caprolactam	11.578	113	109548	18.202	ng/u1	97
35) 4-Chloro-3-methylphenol	11.936	107	385006	20.404	ng/u1	99
36) 2-Methylnaphthalene	12.283	142	809789	20.685	ng/u1	99

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM042223\
 Data File : BM039606.D
 Acq On : 23 Apr 2023 10:21
 Operator : CG/JU
 Sample : SSTDCCC020
 Misc :
 ALS Vial : 34 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 SSTD020120

Quant Time: Apr 24 01:02:35 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\SFAM-EPA-BM041823.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Sun Apr 23 00:18:16 2023
 Response via : Initial Calibration


Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
37) 1-Methylnaphthalene	12.507	142	809132	20.567	ng/ul	97
39) 1,2,4,5-Tetrachloroben...	12.654	216	419960	22.783	ng/ul	99
40) Hexachlorocyclopentadiene	12.625	237	232230	21.145	ng/ul	99
41) 2,4,6-Trichlorophenol	12.907	196	265279	21.542	ng/ul	99
42) 2,4,5-Trichlorophenol	12.995	196	270760	21.155	ng/ul	99
43) 1,1'-Biphenyl	13.301	154	1060382	22.343	ng/ul	100
44) 2-Chloronaphthalene	13.348	162	822648	22.097	ng/ul	99
45) 2-Nitroaniline	13.566	65	237261	22.084	ng/ul	96
47) Dimethylphthalate	13.930	163	999963	20.392	ng/ul	100
48) 2,6-Dinitrotoluene	14.060	165	198539	20.019	ng/ul	95
50) Acenaphthylene	14.195	152	1313187	22.064	ng/ul	100
51) 3-Nitroaniline	14.401	138	175210	18.849	ng/ul	98
52) Acenaphthene	14.536	153	892838	21.789	ng/ul	99
53) 2,4-Dinitrophenol	14.619	184	110662	20.962	ng/ul	96
55) 4-Nitrophenol	14.736	109	76283	13.398	ng/ul	91
56) Dibenzofuran	14.871	168	1196229	21.478	ng/ul	100
57) 2,4-Dinitrotoluene	14.854	165	276629	20.074	ng/ul	91
58) 2,3,4,6-Tetrachlorophenol	15.107	232	227733	19.878	ng/ul	98
59) Diethylphthalate	15.295	149	994120	19.557	ng/ul	99
61) Fluorene	15.524	166	984322	21.364	ng/ul	100
62) 4-Chlorophenyl-phenyle...	15.518	204	475099	20.805	ng/ul	95
63) 4-Nitroaniline	15.566	138	131612	18.463	ng/ul	98
66) 4,6-Dinitro-2-methylph...	15.618	198	145980	19.861	ng/ul#	87
67) N-Nitrosodiphenylamine	15.736	169	784491	22.887	ng/ul	99
68) 4-Bromophenyl-phenylether	16.413	248	270345	21.757	ng/ul	99
69) Hexachlorobenzene	16.524	284	310872	21.778	ng/ul	96
70) Atrazine	16.689	200	258909	20.357	ng/ul	100
71) Pentachlorophenol	16.883	266	170471	20.429	ng/ul	98
72) Phenanthrene	17.265	178	1389824	21.760	ng/ul	99
74) Anthracene	17.360	178	1402566	21.890	ng/ul	100
75) 1,2,3,4-Tetrachloroben...	13.266	216	425394	24.828	ng/uL	98
76) Pentachlorobenzene	14.789	250	404087	23.961	ng/uL	99
77) Carbazole	17.642	167	1201135	21.096	ng/ul	99
78) Di-n-butylphthalate	18.195	149	1505999	18.701	ng/ul	100
80) Fluoranthene	19.289	202	1446459	21.844	ng/ul	98
82) Pyrene	19.654	202	1518047	22.145	ng/ul	97
83) Butylbenzylphthalate	20.553	149	586691	18.132	ng/ul	99
84) 3,3'-Dichlorobenzidine	21.342	252	399076	21.168	ng/ul	100
85) Benzo(a)anthracene	21.400	228	1262374	20.529	ng/ul	99
86) Bis(2-ethylhexyl)phtha...	21.324	149	814007	16.916	ng/ul	100
87) Chrysene	21.453	228	1240773	21.154	ng/ul	100
89) Di-n-octyl phthalate	22.236	149	1356966	15.536	ng/ul	100
90) Benzo(b)fluoranthene	23.059	252	1197669	19.580	ng/ul	99
91) Benzo(k)fluoranthene	23.106	252	1232105	20.511	ng/ul	99
93) Benzo(a)pyrene	23.677	252	1077073	20.890	ng/ul	99
94) Indeno(1,2,3-cd)pyrene	26.230	276	1374504	24.237	ng/ul	99
95) Dibenzo(a,h)anthracene	26.235	278	1147373	24.401	ng/ul	99
96) Benzo(g,h,i)perylene	26.977	276	1121126	24.816	ng/ul	100


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM042223\
 Data File : BM039606.D
 Acq On : 23 Apr 2023 10:21
 Operator : CG/JU
 Sample : SSTDCCC020
 Misc :
 ALS Vial : 34 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 SSTD020120

Quant Time: Apr 24 01:02:35 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\SFAM-EPA-BM041823.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Sun Apr 23 00:18:16 2023
 Response via : Initial Calibration









#2
1,4-Dioxane
Concen: 8.678 ng/uL
RT: 3.243 min Scan# 31
Delta R.T. -0.000 min
Lab File: BM039606.D
Acq: 23 Apr 2023 10:21




Instrument :
BNA_M
ClientSampleId :
SSTD020120


Tgt Ion: 88 Resp: 58961
Ion Ratio Lower Upper
88 100
43 30.1 24.6 36.8
58 75.0 56.7 85.1



#5
Pyridine
Concen: 21.244 ng/uL
RT: 3.649 min Scan# 104
Delta R.T. -0.006 min
Lab File: BM039606.D
Acq: 23 Apr 2023 10:21

Tgt Ion: 79 Resp: 390754
Ion Ratio Lower Upper
79 100
52 76.3 61.9 92.9
51 38.4 29.7 44.5









#6
Benzaldehyde
Concen: 28.581 ng/ul
RT: 6.954 min Scan# 666
Delta R.T. -0.006 min
Lab File: BM039606.D
Acq: 23 Apr 2023 10:21




Instrument :
BNA_M
ClientSampleId :
SSTD020120


Tgt Ion: 77 Resp: 266330
Ion Ratio Lower Upper
77 100
105 91.5 76.3 114.5
106 89.1 75.4 113.0



#8
Phenol
Concen: 21.074 ng/ul
RT: 7.025 min Scan# 678
Delta R.T. -0.006 min
Lab File: BM039606.D
Acq: 23 Apr 2023 10:21

Tgt Ion: 94 Resp: 478551
Ion Ratio Lower Upper
94 100
65 28.7 22.2 33.4
66 37.4 29.6 44.4









#10
Bis(2-Chloroethyl)ether
Concen: 21.215 ng/ul
RT: 7.243 min Scan# 715
Delta R.T. -0.000 min
Lab File: BM039606.D
Acq: 23 Apr 2023 10:21




Instrument :
BNA_M
ClientSampleId :
SSTD020120


Tgt Ion: 93 Resp: 401676
Ion Ratio Lower Upper
93 100
63 78.2 61.9 92.9
95 32.5 25.8 38.8



#12
2-Chlorophenol
Concen: 21.573 ng/ul
RT: 7.378 min Scan# 738
Delta R.T. -0.006 min
Lab File: BM039606.D
Acq: 23 Apr 2023 10:21

Tgt Ion:128 Resp: 371406
Ion Ratio Lower Upper
128 100
64 51.1 38.8 58.2
130 32.1 25.4 38.0





#13

2-Methylphenol

Concen: 20.587 ng/ul

RT: 8.272 min Scan# 891

Delta R.T. -0.006 min

Lab File: BM039606.D

Acq: 23 Apr 2023 10:21

Instrument :

BNA_M

ClientSampleId :





SSTD020120

Tgt Ion:108 Resp: 363114

Ion Ratio Lower Upper

108 100

107 90.7 71.7 107.5



#14

2,2'-oxybis(1-Chloropropane)

Concen: 22.414 ng/ul

RT: 8.348 min Scan# 903

Delta R.T. -0.006 min

Lab File: BM039606.D

Acq: 23 Apr 2023 10:21




Tgt Ion: 45 Resp: 589429


Ion Ratio Lower Upper

45 100

77 15.7 12.3 18.5

79 11.6 9.8 14.6









#16
Acetophenone
Concen: 21.355 ng/ul
RT: 8.648 min Scan# 91
Delta R.T. -0.000 min
Lab File: BM039606.D
Acq: 23 Apr 2023 10:21




Instrument :
BNA_M
ClientSampleId :
SSTD020120


Tgt Ion:105 Resp: 616283
Ion Ratio Lower Upper
105 100
77 81.3 64.1 96.1
51 28.4 22.2 33.4



#17
N-Nitroso-di-n-propylamine
Concen: 21.178 ng/ul
RT: 8.631 min Scan# 951
Delta R.T. -0.000 min
Lab File: BM039606.D
Acq: 23 Apr 2023 10:21

Tgt Ion: 70 Resp: 338263
Ion Ratio Lower Upper
70 100
42 57.0 43.0 64.4
101 8.9 7.4 11.2
130 19.2 16.2 24.2









#18
4-Methylphenol
Concen: 20.495 ng/ul
RT: 8.607 min Scan# 948
Delta R.T. -0.006 min
Lab File: BM039606.D
Acq: 23 Apr 2023 10:21




Instrument :
BNA_M
ClientSampleId :
SSTD020120


Tgt Ion:108 Resp: 390290
Ion Ratio Lower Upper
108 100
107 111.4 88.6 133.0



#19
Hexachloroethane
Concen: 21.270 ng/ul
RT: 8.884 min Scan# 994
Delta R.T. -0.006 min
Lab File: BM039606.D
Acq: 23 Apr 2023 10:21

Tgt Ion:117 Resp: 166319
Ion Ratio Lower Upper
117 100
201 95.0 78.3 117.5
199 59.1 47.4 71.0









#22
Nitrobenzene
Concen: 22.635 ng/ul
RT: 9.025 min Scan# 1018
Delta R.T. -0.006 min
Lab File: BM039606.D
Acq: 23 Apr 2023 10:21




Instrument :
BNA_M
ClientSampleId :
SSTD020120


Tgt Ion: 77 Resp: 474575
Ion Ratio Lower Upper
77 100
123 42.0 36.1 54.1
65 13.7 10.2 15.2



#23
Isophorone
Concen: 21.106 ng/ul
RT: 9.548 min Scan# 1107
Delta R.T. -0.006 min
Lab File: BM039606.D
Acq: 23 Apr 2023 10:21

Tgt Ion: 82 Resp: 943631
Ion Ratio Lower Upper
82 100
95 6.6 5.4 8.2
138 15.9 13.2 19.8









#25
2-Nitrophenol
Concen: 21.441 ng/ul
RT: 9.736 min Scan# 1139
Delta R.T. -0.006 min
Lab File: BM039606.D
Acq: 23 Apr 2023 10:21




Instrument :
BNA_M
ClientSampleId :
SSTD020120


Tgt Ion:139 Resp: 213313
Ion Ratio Lower Upper
139 100
65 57.8 40.7 61.1
109 31.4 23.9 35.9



#26
2,4-Dimethylphenol
Concen: 21.826 ng/ul
RT: 9.807 min Scan# 1151
Delta R.T. -0.006 min
Lab File: BM039606.D
Acq: 23 Apr 2023 10:21

Tgt Ion:107 Resp: 443725
Ion Ratio Lower Upper
107 100
121 49.0 39.1 58.7
122 83.9 68.1 102.1









#27
 Bis(2-Chloroethoxy)methane
 Concen: 20.787 ng/ul
 RT: 10.036 min Scan# 1190
 Delta R.T. -0.006 min
 Lab File: BM039606.D
 Acq: 23 Apr 2023 10:21




Instrument :
 BNA_M
 ClientSampleId :
 SSTD020120


Tgt Ion: 93 Resp: 546501
 Ion Ratio Lower Upper
 93 100
 95 32.3 25.8 38.6
 123 10.9 8.8 13.2



#29
 2,4-Dichlorophenol
 Concen: 21.383 ng/ul
 RT: 10.284 min Scan# 1232
 Delta R.T. -0.006 min
 Lab File: BM039606.D
 Acq: 23 Apr 2023 10:21

Tgt Ion: 162 Resp: 345582
 Ion Ratio Lower Upper
 162 100
 164 64.9 52.4 78.6
 98 37.1 29.4 44.0









#30
Naphthalene
Concen: 21.497 ng/ul
RT: 10.672 min Scan# 11
Delta R.T. -0.000 min
Lab File: BM039606.D
Acq: 23 Apr 2023 10:21




Instrument :
BNA_M
ClientSampleId :
SSTD020120


Tgt Ion:128 Resp: 1255394
Ion Ratio Lower Upper
128 100
129 10.6 8.9 13.3
127 13.0 10.5 15.7



#32
4-Chloroaniline
Concen: 19.901 ng/ul
RT: 10.795 min Scan# 1319
Delta R.T. -0.006 min
Lab File: BM039606.D
Acq: 23 Apr 2023 10:21

Tgt Ion:127 Resp: 480496
Ion Ratio Lower Upper
127 100
129 32.3 25.8 38.8





#33

Hexachlorobutadiene

Concen: 21.249 ng/ul

RT: 10.942 min Scan# 11

Delta R.T. -0.006 min

Lab File: BM039606.D

Acq: 23 Apr 2023 10:21

Instrument :

BNA_M

ClientSampleId :

SSTD020120





Tgt Ion:225 Resp: 226279

Ion Ratio Lower Upper

225 100

223 64.2 50.1 75.1

227 62.9 50.4 75.6



#34

Caprolactam

Concen: 18.202 ng/ul

RT: 11.578 min Scan# 1452

Delta R.T. -0.006 min

Lab File: BM039606.D

Acq: 23 Apr 2023 10:21




Tgt Ion:113 Resp: 109548


Ion Ratio Lower Upper

113 100

55 176.3 145.5 218.3

56 132.2 103.9 155.9









#35
4-Chloro-3-methylphenol
Concen: 20.404 ng/ul
RT: 11.936 min Scan# 11
Delta R.T. -0.006 min
Lab File: BM039606.D
Acq: 23 Apr 2023 10:21




Instrument :
BNA_M
ClientSampleId :
SSTD020120


Tgt Ion:107 Resp: 385006
Ion Ratio Lower Upper
107 100
144 25.6 21.0 31.4
142 78.2 62.9 94.3



#36
2-Methylnaphthalene
Concen: 20.685 ng/ul
RT: 12.283 min Scan# 1572
Delta R.T. -0.006 min
Lab File: BM039606.D
Acq: 23 Apr 2023 10:21

Tgt Ion:142 Resp: 809789
Ion Ratio Lower Upper
142 100
141 87.6 71.1 106.7





#37

1-Methylnaphthalene

Concen: 20.567 ng/ul

RT: 12.507 min Scan# 1

Delta R.T. -0.006 min

Lab File: BM039606.D

Acq: 23 Apr 2023 10:21

Instrument :

BNA_M

ClientSampleId :

SSTD020120





Tgt Ion:142 Resp: 809132

Ion Ratio Lower Upper

142 100

141 93.0 72.0 108.0

116 3.9 3.2 4.8



#39

1,2,4,5-Tetrachlorobenzene

Concen: 22.783 ng/ul

RT: 12.654 min Scan# 1635

Delta R.T. -0.006 min

Lab File: BM039606.D

Acq: 23 Apr 2023 10:21

Tgt Ion:216 Resp: 419960




Ion Ratio Lower Upper


216 100

214 78.0 62.2 93.2

179 21.1 16.5 24.7

108 19.2 14.0 21.0









#40
Hexachlorocyclopentadiene
Concen: 21.145 ng/ul
RT: 12.625 min Scan# 1630
Delta R.T. -0.006 min
Lab File: BM039606.D
Acq: 23 Apr 2023 10:21




Instrument :
BNA_M
ClientSampleId :
SSTD020120


Tgt Ion:237 Resp: 232230
Ion Ratio Lower Upper
237 100
235 63.1 50.9 76.3
272 12.6 10.3 15.5



#41
2,4,6-Trichlorophenol
Concen: 21.542 ng/ul
RT: 12.907 min Scan# 1678
Delta R.T. -0.006 min
Lab File: BM039606.D
Acq: 23 Apr 2023 10:21

Tgt Ion:196 Resp: 265279
Ion Ratio Lower Upper
196 100
198 95.8 76.7 115.1
200 31.8 23.8 35.6





#42

2,4,5-Trichlorophenol

Concen: 21.155 ng/ul

RT: 12.995 min Scan# 1

Delta R.T. -0.006 min

Lab File: BM039606.D

Acq: 23 Apr 2023 10:21

Instrument :

BNA_M

ClientSampleId :

SSTD020120





Tgt Ion:196 Resp: 270760

Ion Ratio Lower Upper

196 100

198 93.9 75.7 113.5

200 29.6 24.6 36.8



#43

1,1'-Biphenyl

Concen: 22.343 ng/ul

RT: 13.301 min Scan# 1745

Delta R.T. -0.006 min

Lab File: BM039606.D

Acq: 23 Apr 2023 10:21




Tgt Ion:154 Resp: 1060382


Ion Ratio Lower Upper

154 100

153 40.1 31.9 47.9

76 14.7 11.5 17.3





#44

2-Chloronaphthalene

Concen: 22.097 ng/ul

RT: 13.348 min Scan# 11

Delta R.T. -0.000 min

Lab File: BM039606.D

Acq: 23 Apr 2023 10:21

Instrument :

BNA_M

ClientSampleId :

SSTD020120





Tgt Ion:162 Resp: 822648

Ion Ratio Lower Upper

162 100

164 31.9 26.1 39.1

127 39.5 31.4 47.2



#45

2-Nitroaniline

Concen: 22.084 ng/ul

RT: 13.566 min Scan# 1790

Delta R.T. -0.006 min

Lab File: BM039606.D

Acq: 23 Apr 2023 10:21




Tgt Ion: 65 Resp: 237261


Ion Ratio Lower Upper

65 100

92 68.6 53.4 80.2

138 98.1 83.2 124.8









#47
Dimethylphthalate
Concen: 20.392 ng/ul
RT: 13.930 min Scan# 1852
Delta R.T. -0.006 min
Lab File: BM039606.D
Acq: 23 Apr 2023 10:21




Instrument :
BNA_M
ClientSampleId :
SSTD020120


Tgt Ion:163 Resp: 999963
Ion Ratio Lower Upper
163 100
194 3.9 3.1 4.7
164 10.3 8.2 12.2



#48
2,6-Dinitrotoluene
Concen: 20.019 ng/ul
RT: 14.060 min Scan# 1874
Delta R.T. -0.006 min
Lab File: BM039606.D
Acq: 23 Apr 2023 10:21

Tgt Ion:165 Resp: 198539
Ion Ratio Lower Upper
165 100
89 67.8 50.6 76.0
121 23.1 17.3 25.9





#50

Acenaphthylene

Concen: 22.064 ng/ul

RT: 14.195 min Scan# 1897

Delta R.T. -0.000 min

Lab File: BM039606.D

Acq: 23 Apr 2023 10:21

Instrument :

BNA_M

ClientSampleId :

SSTD020120





Tgt Ion:152 Resp: 1313187

Ion Ratio Lower Upper

152 100

151 19.9 15.8 23.8

153 13.0 10.4 15.6



#51

3-Nitroaniline

Concen: 18.849 ng/ul

RT: 14.401 min Scan# 1932

Delta R.T. -0.000 min

Lab File: BM039606.D

Acq: 23 Apr 2023 10:21




Tgt Ion:138 Resp: 175210


Ion Ratio Lower Upper

138 100

108 9.9 8.8 13.2

92 121.0 95.5 143.3









#52
Acenaphthene
Concen: 21.789 ng/ul
RT: 14.536 min Scan# 1955
Delta R.T. -0.000 min
Lab File: BM039606.D
Acq: 23 Apr 2023 10:21




Instrument :
BNA_M
ClientSampleId :
SSTD020120


Tgt Ion:153 Resp: 892838
Ion Ratio Lower Upper
153 100
152 48.0 38.2 57.4
154 87.0 71.0 106.4



#53
2,4-Dinitrophenol
Concen: 20.962 ng/ul
RT: 14.619 min Scan# 1969
Delta R.T. -0.006 min
Lab File: BM039606.D
Acq: 23 Apr 2023 10:21

Tgt Ion:184 Resp: 110662
Ion Ratio Lower Upper
184 100
63 80.6 60.0 90.0
154 64.4 51.8 77.6









#55
4-Nitrophenol
Concen: 13.398 ng/ul
RT: 14.736 min Scan# 1990
Delta R.T. -0.006 min
Lab File: BM039606.D
Acq: 23 Apr 2023 10:21




Instrument :
BNA_M
ClientSampleId :
SSTD020120


Tgt Ion:109 Resp: 76283
Ion Ratio Lower Upper
109 100
139 125.9 106.4 159.6
65 146.1 104.8 157.2



#56
Dibenzofuran
Concen: 21.478 ng/ul
RT: 14.871 min Scan# 2012
Delta R.T. -0.006 min
Lab File: BM039606.D
Acq: 23 Apr 2023 10:21

Tgt Ion:168 Resp: 1196229
Ion Ratio Lower Upper
168 100
139 38.7 30.7 46.1





#57

2,4-Dinitrotoluene

Concen: 20.074 ng/ul

RT: 14.854 min Scan# 2010

Delta R.T. -0.006 min

Lab File: BM039606.D

Acq: 23 Apr 2023 10:21

Instrument :

BNA_M

ClientSampleId :

SSTD020120





Tgt Ion:165 Resp: 276629

Ion Ratio Lower Upper

165 100

63 55.5 39.4 59.2

182 2.5 2.0 3.0



#58

2,3,4,6-Tetrachlorophenol

Concen: 19.878 ng/ul

RT: 15.107 min Scan# 2052

Delta R.T. -0.006 min

Lab File: BM039606.D

Acq: 23 Apr 2023 10:21

Tgt Ion:232 Resp: 227733




Ion Ratio Lower Upper


232 100

131 49.3 38.0 57.0

130 2.8 2.0 3.0

166 31.9 24.8 37.2









#59
Diethylphthalate
Concen: 19.557 ng/ul
RT: 15.295 min Scan# 2084
Delta R.T. -0.006 min
Lab File: BM039606.D
Acq: 23 Apr 2023 10:21




Instrument :
BNA_M
ClientSampleId :
SSTD020120


Tgt Ion:149 Resp: 994120
Ion Ratio Lower Upper
149 100
177 21.0 17.3 25.9
150 12.4 9.8 14.6



#61
Fluorene
Concen: 21.364 ng/ul
RT: 15.524 min Scan# 2123
Delta R.T. -0.000 min
Lab File: BM039606.D
Acq: 23 Apr 2023 10:21

Tgt Ion:166 Resp: 984322
Ion Ratio Lower Upper
166 100
165 97.3 78.2 117.2
167 13.2 10.7 16.1









#62
4-Chlorophenyl-phenylether
Concen: 20.805 ng/ul
RT: 15.518 min Scan# 2122
Delta R.T. -0.000 min
Lab File: BM039606.D
Acq: 23 Apr 2023 10:21




Instrument :
BNA_M
ClientSampleId :
SSTD020120


Tgt Ion:204 Resp: 475099
Ion Ratio Lower Upper
204 100
206 31.4 26.8 40.2
141 56.2 48.5 72.7



#63
4-Nitroaniline
Concen: 18.463 ng/ul
RT: 15.566 min Scan# 2130
Delta R.T. -0.000 min
Lab File: BM039606.D
Acq: 23 Apr 2023 10:21

Tgt Ion:138 Resp: 131612
Ion Ratio Lower Upper
138 100
92 56.4 44.1 66.1
108 82.8 64.6 97.0









#66
4,6-Dinitro-2-methylphenol
Concen: 19.861 ng/ul
RT: 15.618 min Scan# 2140
Delta R.T. -0.006 min
Lab File: BM039606.D
Acq: 23 Apr 2023 10:21




Instrument :
BNA_M
ClientSampleId :
SSTD020120


Tgt Ion:198 Resp: 145980
Ion Ratio Lower Upper
198 100
182 5.1 3.1 4.7#
77 33.0 20.6 31.0#



#67
N-Nitrosodiphenylamine
Concen: 22.887 ng/ul
RT: 15.736 min Scan# 2159
Delta R.T. -0.006 min
Lab File: BM039606.D
Acq: 23 Apr 2023 10:21

Tgt Ion:169 Resp: 784491
Ion Ratio Lower Upper
169 100
168 67.1 53.2 79.8
167 36.5 28.9 43.3









#68
4-Bromophenyl-phenylether
Concen: 21.757 ng/ul
RT: 16.413 min Scan# 21
Delta R.T. -0.006 min
Lab File: BM039606.D
Acq: 23 Apr 2023 10:21




Instrument :
BNA_M
ClientSampleId :
SSTD020120


Tgt Ion:248 Resp: 270345
Ion Ratio Lower Upper
248 100
250 96.4 77.7 116.5
141 77.3 61.0 91.4



#69
Hexachlorobenzene
Concen: 21.778 ng/ul
RT: 16.524 min Scan# 2293
Delta R.T. -0.006 min
Lab File: BM039606.D
Acq: 23 Apr 2023 10:21

Tgt Ion:284 Resp: 310872
Ion Ratio Lower Upper
284 100
142 39.2 28.9 43.3
249 31.6 23.9 35.9





#70

Atrazine

Concen: 20.357 ng/ul

RT: 16.689 min Scan# 21

Delta R.T. -0.006 min

Lab File: BM039606.D

Acq: 23 Apr 2023 10:21

Instrument :

BNA_M

ClientSampleId :

SSTD020120





Tgt Ion:200 Resp: 258909

Ion Ratio Lower Upper

200 100

173 27.8 22.2 33.2

215 48.6 38.9 58.3



#71

Pentachlorophenol

Concen: 20.429 ng/ul

RT: 16.883 min Scan# 2354

Delta R.T. -0.006 min

Lab File: BM039606.D

Acq: 23 Apr 2023 10:21




Tgt Ion:266 Resp: 170471


Ion Ratio Lower Upper

266 100

264 64.7 50.2 75.4

268 65.5 51.9 77.9









#72
Phenanthrene
Concen: 21.760 ng/ul
RT: 17.265 min Scan# 2419
Delta R.T. -0.006 min
Lab File: BM039606.D
Acq: 23 Apr 2023 10:21




Instrument :
BNA_M
ClientSampleId :
SSTD020120


Tgt Ion:178 Resp: 1389824
Ion Ratio Lower Upper
178 100
179 15.2 12.2 18.4
176 19.3 15.8 23.8



#74
Anthracene
Concen: 21.890 ng/ul
RT: 17.360 min Scan# 2435
Delta R.T. -0.000 min
Lab File: BM039606.D
Acq: 23 Apr 2023 10:21

Tgt Ion:178 Resp: 1402566
Ion Ratio Lower Upper
178 100
179 15.4 12.1 18.1
176 18.9 15.0 22.6





#75

1,2,3,4-Tetrachlorobenzene

Concen: 24.828 ng/uL

RT: 13.266 min Scan# 1

Delta R.T. -0.006 min

Lab File: BM039606.D

Acq: 23 Apr 2023 10:21

Instrument :

BNA_M

ClientSampleId :

SSTD020120





Tgt Ion:216 Resp: 425394

Ion Ratio Lower Upper

216 100

214 77.2 63.0 94.4

218 48.5 38.1 57.1



#76

Pentachlorobenzene

Concen: 23.961 ng/uL

RT: 14.789 min Scan# 1998

Delta R.T. -0.006 min

Lab File: BM039606.D

Acq: 23 Apr 2023 10:21




Tgt Ion:250 Resp: 404087


Ion Ratio Lower Upper

250 100

248 62.9 50.4 75.6

252 64.5 50.2 75.2





#77

Carbazole

Concen: 21.096 ng/ul

RT: 17.642 min Scan# 2483

Delta R.T. -0.000 min

Lab File: BM039606.D

Acq: 23 Apr 2023 10:21

Instrument :

BNA_M

ClientSampleId :

SSTD020120





Tgt Ion:167 Resp: 1201135

Ion Ratio Lower Upper

167 100

166 21.0 17.0 25.4

139 12.7 10.6 15.8



#78

Di-n-butylphthalate

Concen: 18.701 ng/ul

RT: 18.195 min Scan# 2577

Delta R.T. -0.000 min

Lab File: BM039606.D

Acq: 23 Apr 2023 10:21




Tgt Ion:149 Resp: 1505999


Ion Ratio Lower Upper

149 100

150 8.9 7.0 10.6

104 5.3 4.1 6.1









#80
Fluoranthene
Concen: 21.844 ng/ul
RT: 19.289 min Scan# 21
Delta R.T. -0.000 min
Lab File: BM039606.D
Acq: 23 Apr 2023 10:21




Instrument :
BNA_M
ClientSampleId :
SSTD020120


Tgt Ion:202 Resp: 1446459
Ion Ratio Lower Upper
202 100
101 12.2 10.4 15.6
100 9.3 7.7 11.5



#82
Pyrene
Concen: 22.145 ng/ul
RT: 19.654 min Scan# 2825
Delta R.T. -0.000 min
Lab File: BM039606.D
Acq: 23 Apr 2023 10:21

Tgt Ion:202 Resp: 1518047
Ion Ratio Lower Upper
202 100
101 13.5 12.1 18.1
100 11.3 9.6 14.4









#83
 Butylbenzylphthalate
 Concen: 18.132 ng/ul
 RT: 20.553 min Scan# 2978
 Delta R.T. -0.000 min
 Lab File: BM039606.D
 Acq: 23 Apr 2023 10:21




Instrument :
 BNA_M
 ClientSampleId :
 SSTD020120


Tgt Ion:149 Resp: 586691
 Ion Ratio Lower Upper
 149 100
 91 75.0 59.4 89.0
 206 20.2 15.8 23.8



#84
 3,3'-Dichlorobenzidine
 Concen: 21.168 ng/ul
 RT: 21.342 min Scan# 3112
 Delta R.T. -0.000 min
 Lab File: BM039606.D
 Acq: 23 Apr 2023 10:21

Tgt Ion:252 Resp: 399076
 Ion Ratio Lower Upper
 252 100
 254 63.6 50.9 76.3
 126 12.3 10.4 15.6





#85

Benzo(a)anthracene

Concen: 20.529 ng/ul

RT: 21.400 min Scan# 3123

Delta R.T. -0.006 min

Lab File: BM039606.D

Acq: 23 Apr 2023 10:21

Instrument :

BNA_M

ClientSampleId :

SSTD020120





Tgt Ion:228 Resp: 1262374

Ion Ratio Lower Upper

228 100

229 19.7 15.4 23.2

226 27.1 21.6 32.4



#86

Bis(2-ethylhexyl)phthalate

Concen: 16.916 ng/ul

RT: 21.324 min Scan# 3109

Delta R.T. -0.000 min

Lab File: BM039606.D

Acq: 23 Apr 2023 10:21




Tgt Ion:149 Resp: 814007


Ion Ratio Lower Upper

149 100

167 27.4 21.8 32.6

279 3.9 3.1 4.7





#87

Chrysene

Concen: 21.154 ng/ul

RT: 21.453 min Scan# 31

Delta R.T. -0.006 min

Lab File: BM039606.D

Acq: 23 Apr 2023 10:21

Instrument :

BNA_M

ClientSampleId :

SSTD020120





Tgt Ion:228 Resp: 1240773

Ion Ratio Lower Upper

228 100

226 30.1 23.9 35.9

229 19.9 15.8 23.6



#89

Di-n-octyl phthalate

Concen: 15.536 ng/ul




RT: 22.236 min Scan# 3264


Delta R.T. -0.006 min

Lab File: BM039606.D

Acq: 23 Apr 2023 10:21

Tgt Ion:149 Resp: 1356966









#90
Benzo(b)fluoranthene
Concen: 19.580 ng/ul
RT: 23.059 min Scan# 3404
Delta R.T. -0.006 min
Lab File: BM039606.D
Acq: 23 Apr 2023 10:21




Instrument :
BNA_M
ClientSampleId :
SSTD020120


Tgt Ion:252 Resp: 1197669
Ion Ratio Lower Upper
252 100
253 22.2 17.4 26.2
125 11.7 8.7 13.1



#91
Benzo(k)fluoranthene
Concen: 20.511 ng/ul
RT: 23.106 min Scan# 3412
Delta R.T. -0.006 min
Lab File: BM039606.D
Acq: 23 Apr 2023 10:21

Tgt Ion:252 Resp: 1232105
Ion Ratio Lower Upper
252 100
253 21.7 17.6 26.4
125 11.3 8.6 12.8









#93
Benzo(a)pyrene
Concen: 20.890 ng/ul
RT: 23.677 min Scan# 31
Delta R.T. -0.006 min
Lab File: BM039606.D
Acq: 23 Apr 2023 10:21




Instrument :
BNA_M
ClientSampleId :
SSTD020120


Tgt Ion:252 Resp: 1077073
Ion Ratio Lower Upper
252 100
253 21.5 17.4 26.0
125 12.2 9.4 14.2



#94
Indeno(1,2,3-cd)pyrene
Concen: 24.237 ng/ul
RT: 26.230 min Scan# 3943
Delta R.T. -0.006 min
Lab File: BM039606.D
Acq: 23 Apr 2023 10:21

Tgt Ion:276 Resp: 1374504
Ion Ratio Lower Upper
276 100
138 26.8 21.2 31.8
277 24.1 20.1 30.1









#95
Dibenzo(a,h)anthracene
Concen: 24.401 ng/ul
RT: 26.235 min Scan# 3944
Delta R.T. -0.012 min
Lab File: BM039606.D
Acq: 23 Apr 2023 10:21

Instrument :
BNA_M
ClientSampleId :
SSTD020120

Tgt Ion:278 Resp: 1147373
Ion Ratio Lower Upper
278 100
139 17.6 13.5 20.3
279 23.7 18.9 28.3



#96
Benzo(g,h,i)perylene
Concen: 24.816 ng/ul
RT: 26.977 min Scan# 4070
Delta R.T. -0.012 min
Lab File: BM039606.D
Acq: 23 Apr 2023 10:21

Tgt Ion:276 Resp: 1121126
Ion Ratio Lower Upper
276 100
138 23.7 18.7 28.1
277 23.4 18.7 28.1

