


Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM042424\
Data File : BM045524.D
Acq On : 24 Apr 2024 11:00
Operator : MA/JU
Sample : DFTPP018
Misc :
ALS Vial : 1 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
DFTPP018

Integration File: rteint.p

Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\SFAM-EPA-SIM-BM042024.M
Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
Last Update : Wed Apr 24 05:17:39 2024



AutoFind: Scans 2434, 2435, 2436; Background Corrected with Scan 2428

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	67.1	20621	PASS
68	69	0.00	2	1.1	220	PASS
69	198	0.00	100	62.8	19293	PASS
70	69	0.00	2	0.3	56	PASS
127	198	10	80	67.4	20702	PASS
197	198	0.00	2	0.6	191	PASS
198	198	100	100	100.0	30731	PASS
199	198	5	9	7.5	2290	PASS
275	198	10	60	26.8	8239	PASS
365	198	1	100	5.9	1813	PASS
441	198	0.01	100	14.5	4464	PASS
442	198	50	100	90.1	27685	PASS
443	442	15	24	18.3	5061	PASS

m/z	Abundance
40.10	222.0
40.90	156.0
43.10	167.0
43.80	372.0
73.00	377.0
132.90	180.0
207.00	634.0
280.90	240.0

Instrument :

BNA_M

ClientSampleId :

DFTPP018

m/z	Abundance
37.00	168.0
37.90	507.0
39.00	1983.0
39.90	239.0
40.90	191.0
43.00	246.0
44.00	524.0
48.80	155.0
50.00	8140.0
51.00	26792.0
52.00	1464.0
54.80	346.0
56.00	800.0
57.00	1393.0
61.10	286.0
62.00	534.0
63.00	1006.0
64.90	388.0
67.90	409.0
68.90	23832.0
69.90	169.0
72.90	613.0
74.00	3673.0
75.00	4836.0
76.00	2154.0
77.00	36824.0
78.00	2857.0
78.90	1607.0
80.00	1739.0
81.00	1894.0
81.90	527.0
83.00	458.0
84.90	417.0
85.90	843.0
87.00	248.0
90.80	350.0
91.90	500.0
93.00	3910.0
93.80	280.0
95.90	216.0
97.90	3445.0
98.90	2209.0
99.90	285.0
100.90	1222.0
103.00	408.0
103.90	825.0
104.80	504.0
105.80	187.0
106.10	172.0
107.00	8663.0
108.00	1117.0
109.00	348.0
109.90	11831.0
111.00	2340.0
115.80	525.0
116.90	11973.0
117.90	775.0
119.80	191.0
122.00	695.0
123.00	774.0
123.80	493.0
124.80	312.0
127.00	23672.0
127.90	2045.0
128.90	10527.0
130.00	794.0
130.90	197.0
133.90	345.0
134.90	809.0
135.90	366.0
136.80	496.0
140.80	1082.0
142.00	292.0
142.90	257.0
145.80	353.0
146.90	907.0
147.90	2794.0
148.90	597.0
152.90	312.0
153.90	506.0
154.90	894.0
156.00	1072.0
157.00	209.0
157.90	174.0
159.90	470.0
160.90	570.0
164.90	673.0
166.00	416.0
166.90	4309.0
168.00	2779.0
168.90	303.0
170.90	202.0

Instrument :
BNA_M
ClientSampleId :
DFTPP018

171.80	248.0
172.90	442.0
173.90	481.0
174.90	912.0
176.00	305.0
176.80	570.0
178.90	2201.0
179.90	1619.0
180.80	577.0
184.00	206.0
185.00	1081.0
186.00	6745.0
187.00	1576.0
187.80	297.0
188.90	386.0
190.90	291.0
191.90	1115.0
192.90	918.0
193.70	164.0
196.00	994.0
196.90	300.0
197.90	34176.0
198.90	2552.0
199.90	293.0
201.40	272.0
202.90	345.0
204.00	1946.0
205.00	2689.0
206.00	11354.0
207.00	1947.0
208.00	650.0
209.00	231.0
209.90	275.0
210.80	433.0
215.80	459.0
216.90	3390.0
217.90	463.0
221.00	1594.0
221.90	484.0
222.90	713.0
224.00	6563.0
224.90	1492.0
225.90	260.0
226.90	3347.0
227.90	387.0
229.00	739.0
230.90	207.0
234.70	170.0
235.80	333.0
236.90	219.0
239.00	164.0
240.80	196.0
241.90	508.0
242.90	274.0
244.00	4787.0
244.90	571.0
246.00	1050.0
246.90	164.0
248.80	268.0
253.00	153.0
254.00	164.0
255.00	24424.0
255.90	3844.0
256.80	244.0
257.90	1311.0
258.90	190.0
264.90	624.0
272.90	932.0
273.90	1658.0
275.00	8702.0
275.90	1307.0
276.90	1243.0
277.90	198.0
280.90	166.0
282.80	213.0
284.90	180.0
293.00	298.0
295.90	3727.0
296.90	460.0
302.90	473.0
314.80	538.0
316.00	164.0
322.90	1055.0
323.90	226.0
327.00	158.0
333.00	177.0
333.90	690.0
345.90	259.0
352.00	383.0
353.90	257.0
364.90	1873.0
365.80	266.0
372.00	487.0
401.80	167.0
420.70	158.0
421.80	259.0

Instrument :
BNA_M
ClientSampleId :
DFTPP018

423.00	1351.0
423.90	502.0
441.00	3644.0
442.00	22672.0
443.00	4531.0
444.00	606.0

Instrument :
BNA_M
ClientSampleId :
DFTPP018

m/z	Abundance
37.10	201.0
38.00	358.0
39.00	1869.0
39.90	194.0
40.80	197.0
43.00	201.0
43.90	381.0
44.90	157.0
48.90	252.0
50.00	7513.0
51.00	24096.0
52.00	1170.0
55.00	220.0
55.80	687.0
57.00	1138.0
60.90	217.0
61.90	381.0
63.00	1006.0
64.90	297.0
68.10	361.0
69.00	22816.0
72.90	806.0
74.00	4054.0
74.90	4233.0
76.10	1888.0
77.00	35880.0
78.00	2140.0
79.00	1953.0
79.90	1557.0
81.00	1982.0
82.00	496.0
83.00	307.0
84.90	436.0
86.00	811.0
86.90	319.0
87.80	164.0
90.90	402.0
91.90	422.0
93.00	3946.0
93.90	240.0
96.00	168.0
98.00	2959.0
99.00	2035.0
99.90	214.0
100.90	1057.0
102.80	185.0
104.00	607.0
104.80	765.0
106.10	200.0
107.00	8732.0
107.90	1200.0
108.90	278.0
110.00	10612.0
111.00	1831.0
111.90	293.0
115.90	447.0
116.90	11960.0
118.00	746.0
122.00	466.0
122.90	861.0
123.80	301.0
124.90	357.0
127.00	24672.0
127.90	1845.0
128.90	10323.0
129.90	659.0
130.80	236.0
133.90	442.0
135.00	975.0
135.90	329.0
136.90	508.0
140.00	152.0
140.90	1865.0
142.10	503.0
142.90	278.0
145.80	401.0
146.90	882.0
147.90	2794.0
148.80	360.0
153.00	467.0
153.80	313.0
154.90	871.0
155.90	1298.0
158.00	309.0
160.00	425.0
161.10	753.0
161.90	246.0
164.80	633.0
165.90	432.0
166.90	4210.0
167.90	2795.0
169.00	346.0

Instrument :
BNA_M
ClientSampleId :
DFTPP018

171.80	230.0
173.00	248.0
173.90	681.0
175.00	1029.0
176.00	174.0
177.00	554.0
178.90	1964.0
179.90	1603.0
181.00	820.0
181.80	169.0
184.90	864.0
185.90	6050.0
186.90	2068.0
187.90	297.0
188.90	531.0
190.90	347.0
191.80	769.0
192.90	1000.0
193.90	273.0
195.90	1449.0
197.00	273.0
197.90	36568.0
198.90	2747.0
199.90	300.0
201.10	218.0
201.50	155.0
202.90	436.0
203.90	2197.0
204.90	2966.0
206.00	12355.0
206.90	2331.0
207.80	545.0
208.80	191.0
209.70	206.0
210.70	483.0
215.80	320.0
216.90	4189.0
217.80	585.0
221.00	1616.0
221.80	615.0
222.80	837.0
224.00	7392.0
225.00	1914.0
226.90	3553.0
227.90	600.0
228.90	949.0
230.80	277.0
233.90	301.0
234.80	228.0
235.80	253.0
236.90	241.0
238.90	233.0
240.90	211.0
241.80	636.0
242.90	423.0
244.00	5241.0
244.90	609.0
245.90	1581.0
246.80	202.0
252.80	205.0
253.90	181.0
254.90	28656.0
255.90	4189.0
257.00	416.0
257.90	2120.0
258.90	437.0
265.00	835.0
265.90	188.0
272.90	684.0
273.90	2258.0
274.90	9720.0
275.90	1280.0
277.00	1129.0
277.80	282.0
284.90	219.0
292.90	433.0
295.90	4630.0
297.00	571.0
303.00	402.0
313.90	227.0
314.90	789.0
315.80	196.0
323.00	1382.0
324.10	314.0
327.00	251.0
333.00	204.0
333.90	956.0
335.00	196.0
340.90	199.0
341.20	199.0
345.80	343.0
351.90	431.0
353.00	237.0
354.00	299.0
364.90	2239.0
365.90	203.0

Instrument :
BNA_M
ClientSampleId :
DFTPP018

371.90	700.0
372.80	171.0
402.00	276.0
403.00	375.0
421.00	291.0
421.90	235.0
423.00	1615.0
423.90	506.0
424.90	166.0
441.00	5149.0
442.00	33824.0
443.00	5876.0
443.90	262.0

Instrument :
BNA_M
ClientSampleId :
DFTPP018

m/z	Abundance
39.10	935.0
40.90	224.0
42.90	214.0
44.00	415.0
50.00	3177.0
51.00	10974.0
52.00	648.0
55.00	204.0
56.00	246.0
57.00	647.0
61.00	183.0
63.00	521.0
68.00	298.0
69.00	11230.0
73.00	359.0
74.00	1676.0
75.00	2461.0
75.90	847.0
77.00	16768.0
78.00	1269.0
78.90	1092.0
79.90	759.0
80.90	1019.0
81.90	485.0
83.00	184.0
85.10	174.0
86.10	339.0
90.90	159.0
92.10	168.0
93.00	2154.0
93.90	229.0
97.90	1522.0
98.90	1155.0
100.80	570.0
103.80	306.0
104.80	338.0
105.90	154.0
107.00	4277.0
108.00	764.0
109.90	5576.0
111.00	1029.0
112.00	203.0
116.00	255.0
116.90	6307.0
117.90	553.0
121.80	205.0
122.90	427.0
123.70	194.0
124.90	204.0
127.00	13761.0
128.00	868.0
128.90	5604.0
129.90	537.0
131.00	161.0
133.90	194.0
135.00	448.0
137.00	249.0
141.00	750.0
146.90	511.0
147.90	1433.0
148.80	349.0
153.80	181.0
155.00	438.0
155.90	709.0
160.00	257.0
160.90	381.0
164.80	285.0
165.80	202.0
166.90	2010.0
167.90	1775.0
169.00	272.0
171.90	152.0
173.00	275.0
175.00	632.0
176.80	362.0
177.80	160.0
178.90	1120.0
179.90	938.0
180.80	235.0
185.00	527.0
185.90	3566.0
186.90	965.0
188.80	313.0
190.90	184.0
191.90	505.0
192.90	418.0
195.80	571.0
197.90	21448.0
198.80	1571.0
203.00	246.0
203.90	1346.0
205.00	1584.0

Instrument :
BNA_M
ClientSampleId :
DFTPP018

206.00	7463.0
207.00	1297.0
208.00	355.0
208.90	233.0
210.90	255.0
216.00	168.0
216.90	2417.0
217.90	362.0
221.00	719.0
221.70	308.0
222.80	404.0
224.00	4549.0
224.90	990.0
225.90	182.0
226.90	2146.0
227.90	302.0
228.90	351.0
233.90	229.0
234.80	186.0
236.00	204.0
236.90	243.0
241.90	432.0
242.90	275.0
243.90	3033.0
244.90	485.0
246.00	645.0
246.80	150.0
248.90	242.0
254.90	16552.0
255.90	2195.0
257.90	1170.0
258.90	241.0
264.90	779.0
272.90	547.0
273.90	1323.0
274.90	6296.0
275.90	865.0
276.90	784.0
280.80	178.0
282.80	151.0
284.90	150.0
293.00	190.0
295.90	3399.0
297.00	284.0
302.80	360.0
314.80	277.0
322.80	717.0
323.90	176.0
332.80	214.0
333.90	603.0
346.00	259.0
351.90	403.0
353.90	271.0
364.90	1328.0
365.90	269.0
372.00	469.0
401.90	305.0
402.90	268.0
403.90	177.0
420.90	388.0
423.00	1555.0
423.90	520.0
441.00	4598.0
442.00	26560.0
443.00	4775.0
443.90	603.0

Instrument :
BNA_M
ClientSampleId :
DFTPP018