

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM042522\
 Data File : BM034796.D
 Acq On : 25 Apr 2022 13:23
 Operator : CG/JU
 Sample : PB144304BS
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 SLCS304

Quant Time: Apr 26 03:03:50 2022
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_M\METHODS\SFAM-EPA-SIM-BM042222.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Apr 22 16:54:34 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.943	152	6626	0.400	ng/ul	0.00
4) Naphthalene-d8	10.743	136	16673	0.400	ng/ul #	0.00
9) Acenaphthene-d10	14.570	164	8245	0.400	ng/ul	0.00
13) Phenanthrene-d10	17.313	188	14974	0.400	ng/ul #	0.00
17) Chrysene-d12	21.491	240	10260	0.400	ng/ul #	0.00
23) Perylene-d12	23.888	264	10552	0.400	ng/ul	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.382	96	5026	0.686	ng/ul	0.00
6) 2-Methylnaphthalene-d10	12.332	152	8750	0.373	ng/ul	0.00
18) Fluoranthene-d10	19.336	212	13462	0.416	ng/ul	0.00
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.415	88	12770	1.831	ng/ul#	81
5) Naphthalene	10.792	128	17535	0.368	ng/ul#	88
7) 2-Methylnaphthalene	12.404	142	10895	0.344	ng/ul	100
8) 1-Methylnaphthalene	12.624	142	10641	0.342	ng/ul	100
10) Acenaphthylene	14.293	152	14936	0.415	ng/ul#	90
11) Acenaphthene	14.635	153	11159	0.366	ng/ul	95
12) Fluorene	15.621	166	12076	0.352	ng/ul#	98
14) Pentachlorophenol	16.963	266	3323	0.748	ng/ul	97
15) Phenanthrene	17.355	178	18545	0.375	ng/ul	95
16) Anthracene	17.444	178	16345	0.371	ng/ul#	92
19) Fluoranthene	19.368	202	19811	0.402	ng/ul	97
20) Pyrene	19.731	202	19717	0.399	ng/ul	100
21) Benzo(a)anthracene	21.476	228	16931	0.424	ng/ul	98
22) Chrysene	21.529	228	17648	0.415	ng/ul	98
24) Benzo(b)fluoranthene	23.157	252	19046	0.395	ng/ul	91
25) Benzo(k)fluoranthene	23.207	252	17828	0.370	ng/ul#	89
26) Benzo(a)pyrene	23.780	252	17440	0.444	ng/ul#	85
27) Indeno(1,2,3-cd)pyrene	26.370	276	22172	0.402	ng/ul#	81
28) Dibenzo(a,h)anthracene	26.390	278	17760	0.395	ng/ul#	91
29) Benzo(g,h,i)perylene	27.128	276	19325	0.406	ng/ul	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM042522\
Data File : BM034796.D
Acq On : 25 Apr 2022 13:23
Operator : CG/JU
Sample : PB144304BS
Misc :
ALS Vial : 5 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
SLCS304

Quant Time: Apr 26 03:03:50 2022
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_M\METHODS\SFAM-EPA-SIM-BM042222.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Apr 22 16:54:34 2022
Response via : Initial Calibration

