

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM042522\
 Data File : BM034805.D
 Acq On : 25 Apr 2022 18:51
 Operator : CG/JU
 Sample : N2475-09
 Misc :
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 A0BQ4

Quant Time: Apr 26 03:06:05 2022
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_M\METHODS\SFAM-EPA-SIM-BM042222.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Apr 22 16:54:34 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.942	152	6128	0.400	ng/ul	0.00
4) Naphthalene-d8	10.743	136	15387	0.400	ng/ul	# 0.00
9) Acenaphthene-d10	14.571	164	8032	0.400	ng/ul	0.00
13) Phenanthrene-d10	17.310	188	13448	0.400	ng/ul	# 0.00
17) Chrysene-d12	21.487	240	9129	0.400	ng/ul	# 0.00
23) Perylene-d12	23.881	264	10290	0.400	ng/ul	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.386	96	27562	4.067	ng/ul	0.00
6) 2-Methylnaphthalene-d10	12.332	152	7849	0.363	ng/ul	0.00
18) Fluoranthene-d10	19.336	212	13561	0.471	ng/ul	0.00

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM042522\
Data File : BM034805.D
Acq On : 25 Apr 2022 18:51
Operator : CG/JU
Sample : N2475-09
Misc :
ALS Vial : 13 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
A0BQ4

Quant Time: Apr 26 03:06:05 2022
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_M\METHODS\SFAM-EPA-SIM-BM042222.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Apr 22 16:54:34 2022
Response via : Initial Calibration

