

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM042522\
 Data File : BM034826.D
 Acq On : 26 Apr 2022 07:56
 Operator : CG/JU
 Sample : N2373-09
 Misc :
 ALS Vial : 35 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 BFFE2

Manual Integrations
 APPROVED

Reviewed By :Jagrut Upadhyay 04/26/2022
 Supervised By :Yogesh Patel 04/29/2022

Quant Time: Apr 26 08:33:32 2022
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_M\METHODS\SFAM-EPA-SIM-BM042222.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Apr 22 16:54:34 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.942	152	6711	0.400	ng/ul	0.00
4) Naphthalene-d8	10.743	136	17510	0.400	ng/ul	# 0.00
9) Acenaphthene-d10	14.570	164	9849	0.400	ng/ul	0.00
13) Phenanthrene-d10	17.309	188	19965	0.400	ng/ul	# 0.00
17) Chrysene-d12	21.488	240	16229	0.400	ng/ul	# 0.00
23) Perylene-d12	23.882	264	18684	0.400	ng/ul	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.382	96	10670	1.438	ng/ul	0.00
6) 2-Methylnaphthalene-d10	12.327	152	3341	0.136	ng/ul	0.00
18) Fluoranthene-d10	19.336	212	11524	0.225	ng/ul	0.00
Target Compounds						
					Qvalue	
5) Naphthalene	10.792	128	7947	0.159	ng/ul#	90
7) 2-Methylnaphthalene	12.404	142	6113	0.184	ng/ul	99
8) 1-Methylnaphthalene	12.618	142	5928	0.181	ng/ul	100
10) Acenaphthylene	14.288	152	66867	1.553	ng/ul#	90
11) Acenaphthene	14.630	153	6600	0.181	ng/ul	98
12) Fluorene	15.616	166	16202	0.395	ng/ul	99
14) Pentachlorophenol	16.963	266	1073	0.181	ng/ul	97
15) Phenanthrene	17.351	178	421147	6.381	ng/ul	93
16) Anthracene	17.444	178	78834	1.341	ng/ul#	91
19) Fluoranthene	19.363	202	979691	12.575	ng/ul	97
20) Pyrene	19.726	202	1047110	13.382	ng/ul	100
21) Benzo(a)anthracene	21.470	228	437536	6.930	ng/ul	98
22) Chrysene	21.523	228	643825	9.564	ng/ul	99
24) Benzo(b)fluoranthene	23.157	252	949996	11.136	ng/ul	86
25) Benzo(k)fluoranthene	23.198	252	288088m	3.380	ng/ul	
26) Benzo(a)pyrene	23.777	252	578244	8.306	ng/ul#	76
27) Indeno(1,2,3-cd)pyrene	26.360	276	555216	5.681	ng/ul#	94
28) Dibenzo(a,h)anthracene	26.376	278	125390m	1.577	ng/ul	
29) Benzo(g,h,i)perylene	27.128	276	517773	6.137	ng/ul	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM042522\
Data File : BM034826.D
Acq On : 26 Apr 2022 07:56
Operator : CG/JU
Sample : N2373-09
Misc :
ALS Vial : 35 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
BFFE2

Manual Integrations
APPROVED

Reviewed By :Jagrut Upadhyay 04/26/2022
Supervised By :Yogesh Patel 04/29/2022

Quant Time: Apr 26 08:33:32 2022
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_M\METHODS\SFAM-EPA-SIM-BM042222.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Apr 22 16:54:34 2022
Response via : Initial Calibration

