


Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM050120\
Data File : BM025963.D
Acq On : 01 May 2020 10:35
Operator : MA/SJ
Sample : DFTPP01
Misc :
ALS Vial : 1 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
DFTPP01

Integration File: rteint.p

Method : Z:\SVOASRV\HPCHEM1\BNA_M\METHODS\SOM-EPA-SIM-BM041520.M
Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
Last Update : Fri May 01 12:24:43 2020



AutoFind: Scans 2503, 2504, 2505; Background Corrected with Scan 2496

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	50.8	66226	PASS
68	69	0.00	2	1.8	1113	PASS
69	198	0.00	100	46.2	60229	PASS
70	69	0.00	2	0.6	388	PASS
127	198	10	80	48.5	63224	PASS
197	198	0.00	2	0.7	867	PASS
198	198	100	100	100.0	130368	PASS
199	198	5	9	6.9	8944	PASS
275	198	10	60	27.7	36133	PASS
365	198	1	100	4.2	5420	PASS
441	198	0.01	100	12.2	15957	PASS
442	198	50	100	81.7	106453	PASS
443	442	15	24	19.7	20970	PASS

m/z	Abundance
50.10	20608.0
51.10	78976.0
52.10	4238.0
55.00	448.0
56.00	2365.0
57.00	5526.0
58.10	216.0
61.00	1084.0
62.00	1383.0
63.00	2503.0
64.10	393.0
65.10	1243.0
68.00	1245.0
69.00	69592.0
69.90	366.0
73.00	967.0
74.00	7509.0
75.00	11319.0
76.10	3619.0
77.10	75248.0
78.10	4804.0
79.00	6127.0
80.00	4761.0
81.00	6010.0
82.00	1503.0
83.00	1166.0
84.00	229.0
85.00	892.0
86.00	2029.0
87.10	820.0
87.80	313.0
89.10	172.0
91.00	1477.0
92.10	1690.0
93.00	11255.0
94.00	715.0
95.00	182.0
96.00	605.0
97.20	166.0
98.00	8290.0
99.00	7149.0
100.10	568.0
101.00	3100.0
103.00	983.0
104.00	1797.0
105.00	1650.0
105.90	533.0
107.00	23528.0
108.00	3459.0
109.00	804.0
110.00	37312.0
111.00	6445.0
112.10	849.0
113.10	268.0
116.10	1600.0
117.00	23000.0
118.00	2006.0
119.00	168.0
120.00	201.0
122.00	1738.0
123.00	2061.0
123.90	1003.0
125.00	983.0
126.20	356.0
127.00	69096.0
128.00	5088.0
129.00	29440.0
130.00	2843.0
130.90	455.0
132.70	170.0
134.00	931.0
135.00	2687.0
136.00	915.0
137.00	1562.0
137.90	206.0
139.00	215.0
140.00	408.0
141.00	4197.0
142.00	1313.0
143.00	768.0
144.10	367.0
144.90	312.0
146.10	452.0
147.00	2049.0
148.00	5300.0
149.10	973.0
150.10	306.0
151.10	535.0
151.60	476.0
153.00	1548.0
154.00	1134.0
155.00	2250.0

Instrument :
BNA_M
ClientSampleId :
DFTPP01

156.00	3123.0
157.00	584.0
158.00	623.0
158.90	578.0
160.00	1372.0
161.00	2160.0
162.00	567.0
163.00	219.0
163.90	242.0
165.10	1507.0
166.10	1037.0
167.00	8468.0
168.00	5464.0
169.00	760.0
170.00	323.0
171.00	389.0
171.90	716.0
172.90	1075.0
174.00	1706.0
175.00	2823.0
176.10	827.0
177.00	1533.0
178.00	548.0
179.00	5911.0
180.00	4760.0
181.00	1912.0
182.20	284.0
182.80	151.0
183.20	196.0
183.90	478.0
185.00	2583.0
186.00	19664.0
187.00	5854.0
188.00	960.0
189.00	1517.0
190.00	366.0
190.90	739.0
192.00	1774.0
193.00	2168.0
193.90	395.0
194.90	235.0
195.90	4194.0
197.10	929.0
198.00	135360.0
199.00	9345.0
200.10	804.0
201.40	609.0
202.00	260.0
203.10	980.0
204.00	5787.0
205.00	9491.0
206.10	38424.0
207.00	5605.0
208.00	1812.0
208.90	468.0
210.10	786.0
211.00	1361.0
212.00	285.0
215.00	503.0
215.90	829.0
217.00	11311.0
218.00	1451.0
221.10	4298.0
221.90	1235.0
223.00	2479.0
224.00	21432.0
225.00	5588.0
226.00	1013.0
227.00	11644.0
228.00	1607.0
229.00	1996.0
230.00	323.0
231.00	781.0
233.00	165.0
233.90	556.0
234.90	867.0
235.90	390.0
236.90	1310.0
239.00	411.0
239.80	492.0
240.80	550.0
242.00	1070.0
243.00	973.0
244.00	16188.0
245.00	2029.0
246.00	3744.0
246.90	966.0
248.90	570.0
251.20	156.0
252.10	207.0
253.10	335.0
254.10	825.0
255.00	84544.0
256.00	12756.0
257.00	1056.0
258.00	5696.0

Instrument :
BNA_M
ClientSampleId :
DFTPP01

259.10	980.0
265.00	2131.0
265.90	278.0
271.00	224.0
273.00	2235.0
274.10	5983.0
275.00	35744.0
276.00	4836.0
277.00	3557.0
278.10	709.0
278.90	157.0
280.90	241.0
282.90	199.0
283.10	198.0
284.00	258.0
285.10	620.0
292.00	161.0
293.00	818.0
293.90	338.0
294.90	263.0
296.00	11174.0
297.00	1502.0
298.10	193.0
301.00	226.0
301.90	206.0
303.10	1504.0
304.00	316.0
308.00	202.0
309.90	154.0
314.10	323.0
314.90	1408.0
316.10	580.0
320.90	414.0
321.90	213.0
323.00	2740.0
324.00	644.0
327.00	563.0
328.00	368.0
331.90	225.0
333.10	326.0
334.10	2124.0
335.20	421.0
341.10	365.0
342.20	157.0
346.00	952.0
346.90	195.0
352.10	813.0
353.00	739.0
354.00	966.0
355.10	176.0
365.00	4499.0
365.90	587.0
370.90	306.0
372.00	1199.0
373.00	410.0
383.00	297.0
383.90	153.0
389.90	309.0
401.90	560.0
403.00	816.0
404.00	188.0
404.80	163.0
421.00	773.0
422.10	796.0
423.00	4342.0
424.00	1889.0
425.10	250.0
439.80	220.0
441.00	13265.0
442.10	88184.0
443.00	16448.0
444.00	1699.0

Instrument :
BNA_M
ClientSampleId :
DFTPP01

Scan 2496 (17.369 min): BM025963.D
DFTPP01

m/z	Abundance
207.00	258.0
281.00	177.0

Instrument :
BNA_M
ClientSampleId :
DFTPP01

m/z	Abundance
50.10	21624.0
51.10	78576.0
52.10	4132.0
53.10	214.0
55.00	438.0
56.00	2554.0
57.00	5332.0
57.90	404.0
61.00	1075.0
62.00	1010.0
63.00	2743.0
64.00	287.0
65.00	1636.0
68.00	1427.0
69.00	72008.0
70.00	527.0
73.00	951.0
74.00	7967.0
75.10	11744.0
76.10	4001.0
77.10	75336.0
78.10	5244.0
79.10	6149.0
80.00	4620.0
81.10	5712.0
82.00	1221.0
83.10	1191.0
84.10	235.0
85.10	826.0
86.00	1678.0
87.00	920.0
87.90	222.0
90.90	1280.0
92.00	1442.0
93.00	10671.0
94.10	557.0
95.00	252.0
96.00	433.0
97.00	227.0
98.00	9146.0
99.00	6587.0
100.00	638.0
101.00	2949.0
102.10	208.0
103.00	1073.0
104.00	2118.0
105.00	1799.0
106.10	585.0
107.00	23488.0
108.00	3268.0
109.10	747.0
110.00	37480.0
111.00	5942.0
112.00	824.0
113.20	285.0
116.00	1417.0
117.00	23472.0
118.00	1761.0
119.10	167.0
120.00	319.0
122.10	1333.0
123.00	2305.0
124.00	1178.0
125.00	986.0
126.20	469.0
127.00	77168.0
128.00	5877.0
129.00	30832.0
130.00	2335.0
131.10	488.0
132.00	258.0
132.80	156.0
133.90	854.0
135.00	2512.0
136.00	1162.0
137.10	1731.0
138.00	192.0
139.00	213.0
140.00	396.0
141.00	3712.0
142.10	1010.0
143.00	942.0
144.00	291.0
145.00	215.0
146.00	1040.0
147.00	2177.0
148.00	5907.0
149.00	1656.0
149.90	405.0
151.00	608.0
151.70	412.0
152.90	1234.0

Instrument :
BNA_M
ClientSampleId :
DFTPP01

154.00	1092.0
155.00	2874.0
156.00	3155.0
157.10	722.0
157.90	1051.0
158.90	489.0
159.90	1493.0
161.00	1780.0
162.00	652.0
162.90	288.0
164.00	261.0
164.90	1680.0
166.10	1299.0
167.00	9190.0
168.00	6396.0
169.00	487.0
169.90	349.0
170.90	272.0
171.90	877.0
173.00	908.0
174.00	2035.0
175.00	3617.0
175.90	813.0
177.00	1676.0
178.00	478.0
179.00	6957.0
180.00	4659.0
180.90	2042.0
182.10	249.0
182.70	188.0
183.90	603.0
185.00	3688.0
186.00	23496.0
187.00	7055.0
188.00	735.0
189.00	1964.0
189.90	342.0
191.10	970.0
192.00	1893.0
193.00	2537.0
194.10	495.0
196.00	4106.0
197.10	942.0
198.00	157312.0
199.00	10591.0
199.90	869.0
201.30	443.0
201.70	578.0
203.00	1448.0
204.10	6315.0
205.00	10772.0
206.00	44944.0
207.00	5701.0
208.00	2102.0
209.00	660.0
210.10	766.0
211.00	1697.0
211.80	382.0
213.20	159.0
214.90	542.0
216.00	1338.0
217.00	14054.0
218.00	1604.0
219.00	227.0
221.00	5496.0
221.90	2173.0
223.00	3189.0
224.00	24768.0
225.00	6695.0
226.10	694.0
227.00	12260.0
227.90	1626.0
229.00	1965.0
230.00	281.0
231.00	1108.0
232.10	169.0
233.00	276.0
234.00	634.0
235.00	898.0
236.00	507.0
236.90	1078.0
238.10	209.0
238.90	436.0
239.90	463.0
240.90	758.0
242.00	1266.0
243.10	1541.0
244.00	18368.0
245.10	2233.0
246.00	4213.0
246.90	1028.0
249.00	780.0
250.90	307.0
252.90	511.0
254.10	1026.0
255.00	100824.0

Instrument :
BNA_M
ClientSampleId :
DFTPP01

256.00	15036.0
257.00	1243.0
258.00	7279.0
259.10	1170.0
259.90	223.0
265.00	2741.0
265.90	424.0
270.90	264.0
272.20	400.0
273.00	2844.0
274.00	8444.0
275.00	43096.0
276.00	5829.0
277.00	4332.0
278.10	786.0
278.90	159.0
281.10	156.0
283.00	471.0
284.00	313.0
285.00	857.0
286.00	155.0
289.00	270.0
292.00	307.0
293.00	824.0
293.90	395.0
295.00	410.0
296.00	14397.0
297.00	1844.0
298.10	166.0
301.00	192.0
303.10	1408.0
304.00	602.0
310.00	249.0
314.00	674.0
315.00	2036.0
315.90	762.0
321.10	472.0
323.10	4171.0
323.90	835.0
327.10	841.0
327.90	271.0
332.00	290.0
332.90	372.0
334.10	2734.0
335.00	681.0
341.00	676.0
341.90	150.0
345.90	930.0
346.90	190.0
352.10	1307.0
353.10	1093.0
354.10	1644.0
355.10	254.0
365.00	6831.0
365.90	1204.0
370.90	320.0
372.00	2297.0
373.00	598.0
383.00	545.0
384.00	172.0
389.90	223.0
391.00	176.0
402.10	621.0
403.00	1086.0
404.10	344.0
421.00	1043.0
422.00	1009.0
423.00	6544.0
424.00	2301.0
424.90	254.0
439.90	185.0
441.10	19472.0
442.10	126584.0
443.10	25064.0
444.00	2267.0
444.90	160.0

Instrument :
BNA_M
ClientSampleId :
DFTPP01

m/z	Abundance
50.10	11429.0
51.10	41128.0
52.10	1944.0
55.00	235.0
56.00	1211.0
57.00	2793.0
61.00	446.0
61.90	471.0
63.00	1402.0
64.00	223.0
65.10	613.0
68.00	668.0
69.00	39088.0
70.00	272.0
73.10	680.0
74.00	4462.0
75.00	6541.0
76.00	2254.0
77.10	41280.0
78.10	3135.0
79.00	3083.0
80.00	2309.0
81.00	3420.0
82.00	604.0
82.90	814.0
85.10	452.0
86.00	1247.0
87.10	430.0
87.90	190.0
91.10	740.0
91.90	882.0
93.10	5648.0
94.00	390.0
94.90	154.0
96.00	321.0
98.00	4558.0
99.00	3893.0
100.00	222.0
101.00	1736.0
102.90	580.0
104.00	1493.0
105.00	999.0
105.90	301.0
107.00	13201.0
108.00	2046.0
109.00	354.0
110.00	22808.0
111.00	3666.0
111.90	225.0
116.10	866.0
117.00	13983.0
118.00	979.0
119.90	205.0
122.00	943.0
123.00	1322.0
124.00	572.0
125.00	633.0
126.20	286.0
127.00	43408.0
128.00	3342.0
129.00	17704.0
130.00	1271.0
131.00	273.0
133.00	155.0
134.10	405.0
135.00	1259.0
136.00	610.0
137.10	1077.0
138.00	153.0
138.80	195.0
140.00	354.0
141.00	2572.0
142.00	663.0
143.00	521.0
145.90	493.0
147.00	1349.0
148.00	3364.0
148.80	621.0
151.00	304.0
151.60	241.0
153.10	815.0
154.00	557.0
154.90	1134.0
156.00	1824.0
157.10	225.0
157.90	354.0
159.00	371.0
160.00	816.0
161.00	1262.0
162.00	380.0
165.00	1050.0
166.00	779.0

Instrument :
BNA_M
ClientSampleId :
DFTPP01

167.00	5593.0
168.00	3579.0
169.10	546.0
169.90	176.0
170.90	236.0
171.90	658.0
173.00	624.0
174.00	1159.0
175.00	1953.0
176.10	800.0
176.90	885.0
178.00	403.0
179.00	4313.0
180.00	3065.0
181.00	1220.0
181.80	220.0
184.00	299.0
185.10	1612.0
186.00	15145.0
187.00	4624.0
188.00	344.0
189.00	1161.0
189.90	215.0
190.90	661.0
191.90	1440.0
192.90	1353.0
194.10	322.0
195.00	213.0
196.00	2189.0
197.10	730.0
198.00	98432.0
199.00	6897.0
200.00	700.0
201.50	431.0
203.00	889.0
204.10	4471.0
205.00	6866.0
206.10	27720.0
207.10	3964.0
208.00	1035.0
209.10	317.0
210.00	449.0
211.00	1339.0
211.90	191.0
214.90	415.0
216.00	660.0
217.00	8568.0
217.90	1152.0
219.80	165.0
221.00	3345.0
221.90	1201.0
223.00	2412.0
224.00	17200.0
225.00	4342.0
226.00	353.0
227.00	8630.0
227.90	1049.0
229.00	1231.0
230.00	190.0
231.00	531.0
233.10	185.0
234.00	513.0
234.90	563.0
235.90	299.0
237.00	813.0
238.90	305.0
239.90	316.0
240.90	389.0
242.00	966.0
243.00	759.0
244.10	12207.0
245.00	1239.0
246.00	3013.0
247.00	543.0
248.90	359.0
252.00	201.0
252.90	465.0
254.20	425.0
255.00	69368.0
256.00	9421.0
257.00	1058.0
258.10	4494.0
259.00	764.0
265.00	1855.0
266.10	198.0
271.00	254.0
273.00	1855.0
274.00	5175.0
275.00	29560.0
276.00	3754.0
277.00	2771.0
277.90	471.0
281.90	161.0
283.00	406.0
283.80	319.0
285.00	417.0

Instrument :
BNA_M
ClientSampleId :
DFTPP01

293.00	631.0
294.00	215.0
294.90	219.0
296.00	9968.0
296.90	1317.0
302.00	177.0
303.00	1137.0
304.00	455.0
308.00	173.0
314.00	508.0
315.00	1087.0
316.00	659.0
321.00	371.0
322.20	207.0
323.00	2560.0
324.00	354.0
327.00	565.0
328.00	321.0
331.80	193.0
333.00	298.0
334.00	2078.0
335.00	457.0
340.90	488.0
346.00	548.0
352.00	799.0
353.00	654.0
353.90	1025.0
355.00	216.0
365.00	4931.0
365.90	824.0
371.00	229.0
372.10	1575.0
373.10	355.0
383.00	439.0
402.10	631.0
403.00	1006.0
404.00	470.0
421.00	826.0
422.00	916.0
423.10	5665.0
424.10	1993.0
424.80	179.0
439.90	252.0
441.00	15136.0
442.10	104592.0
443.00	21400.0
444.00	2084.0

Instrument :
BNA_M
ClientSampleId :
DFTPP01