

Data Path : Z:\SVOASRV\HPCHEM1\BNA M\DATA\BM050121\
 Data File : BM029757.D
 Acq On : 01 May 2021 21:50
 Operator : CG/JU
 Sample : PB135733BS
 Misc :
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 SLCS733

Manual Integrations
 APPROVED

mohammad
 5/3/2021 12:24:43 PM

Quant Time: May 02 01:18:15 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_M\METHODS\SFAM-EPA-BM050121.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Sat May 01 16:23:05 2021
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.58	152	51821	20.00	ng/ul	0.00
20) Naphthalene-d8	10.35	136	188362	20.00	ng/ul	0.00
38) Acenaphthene-d10	14.22	164	143446	20.00	ng/ul	0.00
64) Phenanthrene-d10	16.97	188	320851	20.00	ng/ul	0.00
79) Chrysene-d12	21.15	240	382247	20.00	ng/ul	0.00
88) Perylene-d12	23.32	264	402750	20.00	ng/ul	0.00

System Monitoring Compounds

3) 1,4-Dioxane-d8	3.08	96	6318	5.96	ng/uL	0.00
4) Pyridine-d5	3.49	84	72357	31.52	ng/ul	0.00
7) Phenol-d5	6.76	99	111279	31.89	ng/ul	0.00
9) Bis-(2-Chloroethyl)ether-d	6.92	67	57544	31.49	ng/ul	0.00
11) 2-Chlorophenol-d4	7.11	132	102951	32.62	ng/ul	0.00
15) 4-Methylphenol-d8	8.29	113	96408	32.26	ng/ul	0.00
21) Nitrobenzene-d5	8.73	128	47146	34.37	ng/ul	0.00
24) 2-Nitrophenol-d4	9.45	143	61486	35.41	ng/ul	0.00
28) 2,4-Dichlorophenol-d3	9.99	165	130221	35.89	ng/ul	0.00
31) 4-Chloroaniline-d4	10.50	131	120263	29.39	ng/ul	0.00
46) Dimethylphthalate-d6	13.64	166	394163	34.69	ng/ul	0.00
49) Acenaphthylene-d8	13.92	160	481069	34.34	ng/ul	0.00
54) 4-Nitrophenol-d4	14.45	143	53987	35.92	ng/ul	0.00
60) Fluorene-d10	15.22	176	342917	34.89	ng/ul	0.00
65) 4,6-Dinitro-2-methylphenol	15.36	200	81641	35.00	ng/ul	0.00
73) Anthracene-d10	17.07	188	564597	34.91	ng/ul	0.00
81) Pyrene-d10	19.36	212	709156	36.33	ng/ul	0.00
92) Benzo(a)pyrene-d12	23.19	264	830782	35.95	ng/ul	0.00

Target Compounds

					Ovalue
2) 1,4-Dioxane	3.11	88	12672	11.359	ng/uL 98
5) Pyridine	3.51	79	76626	31.364	ng/ul 95
6) Benzaldehyde	6.73	77	52537	26.253	ng/ul 95
8) Phenol	6.79	94	112074	32.406	ng/ul 96
10) Bis(2-Chloroethyl)ether	7.01	93	85805	32.419	ng/ul 96
12) 2-Chlorophenol	7.15	128	105444	33.578	ng/ul 96
13) 2-Methylphenol	8.03	108	89861	32.370	ng/ul 95
14) 2,2'-oxybis(1-Chloropropan	8.11	45	79494	30.597	ng/ul 96
16) Acetophenone	8.40	105	147469	32.012	ng/ul 99
17) N-Nitroso-di-n-propylamine	8.39	70	70543	31.774	ng/ul 99
18) 4-Methylphenol	8.36	108	99454	32.882	ng/ul 96
19) Hexachloroethane	8.64	117	44197	31.126	ng/ul 97
22) Nitrobenzene	8.78	77	104532	34.500	ng/ul 96
23) Isophorone	9.30	82	198861	34.023	ng/ul 99
25) 2-Nitrophenol	9.48	139	62914	34.913	ng/ul 100
26) 2,4-Dimethylphenol	9.55	107	110914	33.810	ng/ul 98
27) Bis(2-Chloroethoxy)methane	9.78	93	117404	34.506	ng/ul 98
29) 2,4-Dichlorophenol	10.01	162	123190	36.354	ng/ul 96
30) Naphthalene	10.40	128	337731	33.734	ng/ul 99
32) 4-Chloroaniline	10.52	127	116462	28.279	ng/ul 99
33) Hexachlorobutadiene	10.69	225	101236	34.565	ng/ul 99
34) Caprolactam	11.30	113	32540m	34.778	ng/ul

Data Path : Z:\SVOASRV\HPCHEM1\BNA M\DATA\BM050121\
 Data File : BM029757.D
 Acq On : 01 May 2021 21:50
 Operator : CG/JU
 Sample : PB135733BS
 Misc :
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 SLCS733

Manual Integrations
 APPROVED

mohammad
 5/3/2021 12:24:43 PM

Quant Time: May 02 01:18:15 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_M\METHODS\SFAM-EPA-BM050121.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Sat May 01 16:23:05 2021
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
35) 4-Chloro-3-methylphenol	11.66	107	103901	35.999	ng/ul	96
36) 2-Methylnaphthalene	12.02	142	261149	33.775	ng/ul	97
37) 1-Methylnaphthalene	12.25	142	254197	33.775	ng/ul	98
39) 1,2,4,5-Tetrachlorobenzene	12.40	216	188996	33.808	ng/ul	100
40) Hexachlorocyclopentadiene	12.37	237	78028	28.434	ng/ul	91
41) 2,4,6-Trichlorophenol	12.65	196	116196	36.014	ng/ul	93
42) 2,4,5-Trichlorophenol	12.73	196	121407	35.884	ng/ul	99
43) 1,1'-Biphenyl	13.05	154	363762	33.912	ng/ul	99
44) 2-Chloronaphthalene	13.09	162	291093	33.686	ng/ul	100
45) 2-Nitroaniline	13.31	65	58067	34.799	ng/ul	91
47) Dimethylphthalate	13.69	163	385362	34.967	ng/ul	99
48) 2,6-Dinitrotoluene	13.81	165	81739	36.797	ng/ul	95
50) Acenaphthylene	13.95	152	428761	32.874	ng/ul	99
51) 3-Nitroaniline	14.14	138	58312	29.488	ng/ul	94
52) Acenaphthene	14.29	153	310789	34.167	ng/ul	99
53) 2,4-Dinitrophenol	14.36	184	46674	35.244	ng/ul	93
55) 4-Nitrophenol	14.46	109	43109	40.553	ng/ul	88
56) Dibenzofuran	14.62	168	465758	34.459	ng/ul	99
57) 2,4-Dinitrotoluene	14.60	165	118168	37.941	ng/ul	99
58) 2,3,4,6-Tetrachlorophenol	14.86	232	119037	37.259	ng/ul	96
59) Diethylphthalate	15.06	149	367913	35.168	ng/ul	99
61) Fluorene	15.27	166	388371	34.445	ng/ul	97
62) 4-Chlorophenyl-phenylether	15.27	204	226099	35.901	ng/ul	96
63) 4-Nitroaniline	15.31	138	67997m	32.324	ng/ul	
66) 4,6-Dinitro-2-methylphenol	15.37	198	78624	34.983	ng/ul#	95
67) N-Nitrosodiphenylamine	15.49	169	342742	34.593	ng/ul	99
68) 4-Bromophenyl-phenylether	16.17	248	139564	35.460	ng/ul	95
69) Hexachlorobenzene	16.28	284	165252	36.356	ng/ul	95
70) Atrazine	16.45	200	145466	35.632	ng/ul	98
71) Pentachlorophenol	16.63	266	94981	39.711	ng/ul	97
72) Phenanthrene	17.01	178	654573	35.571	ng/ul	99
74) Anthracene	17.10	178	668200	35.335	ng/ul	99
75) 1,2,3,4-Tetrachlorobenzene	13.02	216	192223	32.215	ng/uL	97
76) Pentachlorobenzene	14.55	250	188555	33.742	ng/uL	98
77) Carbazole	17.37	167	562272	34.070	ng/ul	99
78) Di-n-butylphthalate	17.95	149	638447	36.423	ng/ul	100
80) Fluoranthene	19.03	202	841652	36.659	ng/ul	99
82) Pyrene	19.39	202	880701	36.049	ng/ul	99
83) Butylbenzylphthalate	20.30	149	296468	34.972	ng/ul	96
84) 3,3'-Dichlorobenzidine	21.08	252	292384	29.801	ng/ul	97
85) Benzo(a)anthracene	21.14	228	891952	35.660	ng/ul	100
86) Bis(2-ethylhexyl)phthalate	21.08	149	453756	34.881	ng/ul	99
87) Chrysene	21.19	228	888320	35.539	ng/ul	99
89) Di-n-octyl phthalate	21.94	149	797017	34.483	ng/ul	100
90) Benzo(b)fluoranthene	22.67	252	988457	37.659	ng/ul	100
91) Benzo(k)fluoranthene	22.72	252	916572	35.450	ng/ul	99
93) Benzo(a)pyrene	23.23	252	862408	36.316	ng/ul	99
94) Indeno(1,2,3-cd)pyrene	25.50	276	1152063	37.045	ng/ul	100
95) Dibenzo(a,h)anthracene	25.50	278	960478	37.003	ng/ul	99
96) Benzo(g,h,i)perylene	26.16	276	944874	36.993	ng/ul	99

Data Path : Z:\SVOASRV\HPCHEM1\BNA M\DATA\BM050121\
Data File : BM029757.D
Acq On : 01 May 2021 21:50
Operator : CG/JU
Sample : PB135733BS
Misc :
ALS Vial : 7 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
SLCS733


Manual Integrations
APPROVED

mohammad
5/3/2021 12:24:43 PM

Quant Time: May 02 01:18:15 2021
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_M\METHODS\SFAM-EPA-BM050121.M
Quant Title : SVOA CALIBRATION
QLast Update : Sat May 01 16:23:05 2021
Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
--------------------	------	------	----------	------	-------	----------

(#) = qualifier out of range (m) = manual integration (+) = signals summed



#2

1,4-Dioxane

Concen: 11.359 ng/uL

RT: 3.11 min Scan# 38

Delta R.T. 0.00 min

Lab File: BM029757.D

Acq: 01 May 2021 21:50

Instrument :

BNA_M





ClientSampleId :

SLCS733

Tot Ion:	88	Resp:	12672
Ion	Ratio	Lower	Upper
88	100		
43	29.0	22.6	33.8
58	55.3	45.5	68.3

Manual Integrations
APPROVED

mohammad
5/3/2021 12:24:43 PM



#5

Pyridine

Concen: 31.364 ng/uL




RT: 3.51 min Scan# 106


Delta R.T. -0.01 min

Lab File: BM029757.D

Acq: 01 May 2021 21:50

Tgt Ion:	79	Resp:	76626
Ion	Ratio	Lower	Upper
79	100		
52	60.7	51.4	77.2
51	31.2	27.4	41.0





#6

Benzaldehyde

Concen: 26.253 ng/ul

RT: 6.73 min Scan# 653

Delta R.T. -0.01 min

Lab File: BM029757.D

Acq: 01 May 2021 21:50

Instrument :

BNA_M





ClientSampleId :

SLCS733

Tot Ion:	77	Resp:	52537
Ion Ratio	Lower	Upper	
77	100		
105	104.3	81.5	122.3
106	95.4	81.8	122.8

Manual Integrations
APPROVED

mohammad
5/3/2021 12:24:43 PM



#8

Phenol

Concen: 32.406 ng/ul




RT: 6.79 min Scan# 663


Delta R.T. -0.01 min

Lab File: BM029757.D

Acq: 01 May 2021 21:50

Tgt Ion:	94	Resp:	112074
Ion Ratio	Lower	Upper	
94	100		
65	28.1	21.4	32.2
66	39.0	29.0	43.4









#10
Bis(2-Chloroethyl)ether
Concen: 32.419 ng/ul
RT: 7.01 min Scan# 701
Delta R.T. -0.01 min
Lab File: BM029757.D
Acq: 01 May 2021 21:50

Instrument :
BNA_M
ClientSampleId :
SLCS733

Tot Ion:	93	Resp:	85805
Ion Ratio	Lower	Upper	
93	100		
63	68.1	51.0	76.6
95	32.4	25.6	38.4




Manual Integrations
APPROVED


mohammad
5/3/2021 12:24:43 PM



#12
2-Chlorophenol
Concen: 33.578 ng/ul
RT: 7.15 min Scan# 724
Delta R.T. 0.00 min
Lab File: BM029757.D
Acq: 01 May 2021 21:50

Tgt Ion:	128	Resp:	105444
Ion Ratio	Lower	Upper	
128	100		
64	35.2	30.0	45.0
130	32.1	27.1	40.7





#13

2-Methylphenol

Concen: 32.370 ng/ul

RT: 8.03 min Scan# 874

Delta R.T. -0.01 min

Lab File: BM029757.D

Acq: 01 May 2021 21:50

Instrument :

BNA_M

ClientSampleId :

SLCS733

Tot Ion:108 Resp: 89861

Ion Ratio Lower Upper

108 100


107 94.2 71.4 107.2

Manual Integrations

APPROVED

mohammad

5/3/2021 12:24:43 PM



Abundance Ion 108.00 (107.70 to 108.70): E

Ion 107.00 (106.70 to 107.70): E

60000

50000

40000

30000

20000



10000

0

Time-->

7.95 8.00 8.05 8.10

8.03



#14

2,2'-oxybis(1-chloropropane)

Concen: 30.597 ng/ul

RT: 8.11 min Scan# 888

Delta R.T. 0.00 min

Lab File: BM029757.D

Acq: 01 May 2021 21:50


Tgt Ion: 45 Resp: 79494

Ion Ratio Lower Upper

45 100

77 23.8 16.5 24.7

79 17.2 13.5 20.3



Abundance Ion 45.00 (44.70 to 45.70): BM0

Ion 77.00 (76.70 to 77.70): BM0

Ion 79.00 (78.70 to 79.70): BM0

40000

30000

20000


10000


0

Time-->

8.05 8.10 8.15

8.11









#16
Acetophenone
Concen: 32.012 ng/ul
RT: 8.40 min Scan# 937
Delta R.T. -0.01 min
Lab File: BM029757.D
Acq: 01 May 2021 21:50

Instrument :
BNA_M
ClientSampleId :
SLCS733

Tot Ion	Ion	Ratio	Resp	Lower	Upper
105	105	100	147469		
77	77	75.9	60.1	90.1	
51	51	21.8	17.0	25.4	




Manual Integrations
APPROVED


mohammad
5/3/2021 12:24:43 PM



#17
N-Nitroso-di-n-propylamine
Concen: 31.774 ng/ul
RT: 8.39 min Scan# 935
Delta R.T. 0.00 min
Lab File: BM029757.D
Acq: 01 May 2021 21:50

Tgt Ion	Ion	Ratio	Resp	Lower	Upper
70	70	100	70543		
42	42	49.1	40.1	60.1	
101	101	10.9	8.7	13.1	
130	130	25.6	19.8	29.6	





#18

4-Methylphenol

Concen: 32.882 ng/ul

RT: 8.36 min Scan# 930

Delta R.T. 0.00 min

Lab File: BM029757.D

Acq: 01 May 2021 21:50

Instrument :

BNA_M

ClientSampleId :

SLCS733

Tot Ion:108 Resp: 99454

Ion Ratio Lower Upper

108 100


107 113.0 94.2 141.4

Manual Integrations

APPROVED




mohammad

5/3/2021 12:24:43 PM



Abundance Ion 108.00 (107.70 to 108.70): E

Ion 107.00 (106.70 to 107.70): E



#19

Hexachloroethane

Concen: 31.126 ng/ul

RT: 8.64 min Scan# 978

Delta R.T. 0.00 min

Lab File: BM029757.D

Acq: 01 May 2021 21:50


Tgt Ion:117 Resp: 44197

Ion Ratio Lower Upper

117 100

201 150.0 117.4 176.2



199 94.7 73.4 110.0




Abundance Ion 117.00 (116.70 to 117.70): E

Ion 201.00 (200.70 to 201.70): E

Ion 199.00 (198.70 to 199.70): E





#22

Nitrobenzene

Concen: 34.500 ng/ul

RT: 8.78 min Scan# 1001

Delta R.T. 0.00 min

Lab File: BM029757.D

Acq: 01 May 2021 21:50

Instrument :

BNA_M





ClientSampleId :

SLCS733

Tot Ion:	77	Resp:	104532
Ion Ratio	Lower	Upper	
77	100		
123	48.7	36.6	54.8
65	12.5	10.2	15.2

Manual Integrations
APPROVED

mohammad
5/3/2021 12:24:43 PM



#23

Isophorone

Concen: 34.023 ng/ul




RT: 9.30 min Scan# 1090


Delta R.T. 0.00 min

Lab File: BM029757.D

Acq: 01 May 2021 21:50

Tgt Ion:	82	Resp:	198861
Ion Ratio	Lower	Upper	
82	100		
95	7.8	6.0	9.0
138	21.4	16.5	24.7









#25
2-Nitrophenol
Concen: 34.913 ng/ul
RT: 9.48 min Scan# 1121
Delta R.T. 0.00 min
Lab File: BM029757.D
Acq: 01 May 2021 21:50

Instrument :
BNA_M
ClientSampleId :
SLCS733

Tot Ion	Ratio	Lower	Upper
139	100		
65	42.9	34.6	52.0
109	28.3	22.6	34.0




Manual Integrations
APPROVED


mohammad
5/3/2021 12:24:43 PM



#26
2,4-Dimethylphenol
Concen: 33.810 ng/ul
RT: 9.55 min Scan# 1132
Delta R.T. -0.01 min
Lab File: BM029757.D
Acq: 01 May 2021 21:50

Tgt Ion	Ratio	Lower	Upper
107	100		
121	51.0	41.7	62.5
122	86.1	70.3	105.5





#27

Bis(2-Chloroethoxy)methane

Concen: 34.506 ng/ul

RT: 9.78 min Scan# 1172

Delta R.T. 0.00 min

Lab File: BM029757.D

Acq: 01 May 2021 21:50

Instrument :

BNA_M





ClientSampleId :

SLCS733

Tot Ion:	93	Resp:	117404
Ion	Ratio	Lower	Upper
93	100		
95	32.7	27.0	40.6
123	11.8	8.9	13.3

Manual Integrations
APPROVED

mohammad
5/3/2021 12:24:43 PM



#29

2,4-Dichlorophenol

Concen: 36.354 ng/ul




RT: 10.01 min Scan# 1211


Delta R.T. -0.01 min

Lab File: BM029757.D

Acq: 01 May 2021 21:50

Tgt Ion:	162	Resp:	123190
Ion	Ratio	Lower	Upper
162	100		
164	67.8	56.9	85.3
98	28.7	21.9	32.9





#30

Naphthalene

Concen: 33.734 ng/ul

RT: 10.40 min Scan# 1278

Delta R.T. 0.00 min

Lab File: BM029757.D

Acq: 01 May 2021 21:50

Instrument :

BNA_M





ClientSampled :

SLCS733

Tot Ion	Ion	Ratio	Resp	Lower	Upper
128	100		337731		
129	10.9	8.6	12.8		
127	13.6	11.3	16.9		

Manual Integrations
APPROVED

mohammad
5/3/2021 12:24:43 PM



#32

4-Chloroaniline

Concen: 28.279 ng/ul




RT: 10.52 min Scan# 1298


Delta R.T. -0.01 min

Lab File: BM029757.D

Acq: 01 May 2021 21:50

Tgt Ion	Ion	Ratio	Resp	Lower	Upper
127	100		116462		
129	30.8	25.3	37.9		









#33
Hexachlorobutadiene
Concen: 34.565 ng/ul
RT: 10.69 min Scan# 1326
Delta R.T. 0.00 min
Lab File: BM029757.D
Acq: 01 May 2021 21:50

Instrument :
BNA_M
ClientSampleId :
SLCS733

Tot Ion	Ratio	Resp	Lower	Upper
225	100	101236		
223	62.6	50.0	75.0	
227	64.3	52.6	79.0	




Manual Integrations
APPROVED


mohammad
5/3/2021 12:24:43 PM



#34
Caprolactam
Concen: 34.778 ng/ul m
RT: 11.30 min Scan# 1431
Delta R.T. -0.01 min
Lab File: BM029757.D
Acq: 01 May 2021 21:50

Tgt Ion	Ratio	Resp	Lower	Upper
113	100	32540		
55	108.3	90.8	136.2	
56	91.9	72.3	108.5	









#35
4-Chloro-3-methylphenol
Concen: 35.999 ng/ul
RT: 11.66 min Scan# 1491
Delta R.T. -0.01 min
Lab File: BM029757.D
Acq: 01 May 2021 21:50

Instrument :
BNA_M
ClientSampleId :
SLCS733

Tot Ion	Ratio	Lower	Upper
107	100		
144	27.6	22.9	34.3
142	88.5	74.1	111.1




Manual Integrations
APPROVED


mohammad
5/3/2021 12:24:43 PM



#36
2-Methylnaphthalene
Concen: 33.775 ng/ul
RT: 12.02 min Scan# 1553
Delta R.T. -0.01 min
Lab File: BM029757.D
Acq: 01 May 2021 21:50

Tgt Ion	Ratio	Lower	Upper
142	100		
141	88.7	69.0	103.6





#37

1-Methylnaphthalene

Concen: 33.775 ng/ul

RT: 12.25 min Scan# 1591

Delta R.T. -0.01 min

Lab File: BM029757.D

Acq: 01 May 2021 21:50

Instrument :

BNA_M


ClientSampleId :

SLCS733




Tot Ion:	142	Resp:	254197
Ion Ratio	Lower	Upper	
142	100		
141	92.5	75.8	113.6
116	3.5	2.7	4.1

Manual Integrations
APPROVED

mohammad
5/3/2021 12:24:43 PM



Abundance Ion 142.00 (141.70 to 142.70): E
Ion 141.00 (140.70 to 141.70): E
Ion 116.00 (115.70 to 116.70): E



#39

1,2,4,5-Tetrachlorobenzene

Concen: 33.808 ng/ul

RT: 12.40 min Scan# 1617




Delta R.T. -0.01 min


Lab File: BM029757.D

Acq: 01 May 2021 21:50

Tgt Ion:	216	Resp:	188996
Ion Ratio	Lower	Upper	
216	100		
214	76.3	61.1	91.7
179	18.2	14.8	22.2
108	10.2	8.6	13.0

Abundance Ion 216.00 (215.70 to 216.70): E
Ion 214.00 (213.70 to 214.70): E
Ion 179.00 (178.70 to 179.70): E
Ion 108.00 (107.70 to 108.70): E









#40
Hexachlorocyclopentadiene
Concen: 28.434 ng/ul
RT: 12.37 min Scan# 1613
Delta R.T. -0.01 min
Lab File: BM029757.D
Acq: 01 May 2021 21:50

Instrument :
BNA_M
ClientSampleId :
SLCS733

Tot Ion	Ratio	Resp Lower	Upper
237	100		
235	64.2	45.6	68.4
272	12.9	8.9	13.3




Manual Integrations
APPROVED


mohammad
5/3/2021 12:24:43 PM



#41
2,4,6-Trichlorophenol
Concen: 36.014 ng/ul
RT: 12.65 min Scan# 1660
Delta R.T. -0.01 min
Lab File: BM029757.D
Acq: 01 May 2021 21:50

Tgt Ion	Ratio	Resp Lower	Upper
196	100		
198	99.6	72.7	109.1
200	30.8	24.7	37.1









#42
 2,4,5-Trichlorophenol
 Concen: 35.884 ng/ul
 RT: 12.73 min Scan# 1673
 Delta R.T. -0.01 min
 Lab File: BM029757.D
 Acq: 01 May 2021 21:50

Instrument :
 BNA_M
 ClientSampled :
 SLCS733

Tot Ion:	196	Resp:	121407
Ion Ratio	Lower	Upper	
196	100		
198	98.4	77.5	116.3
200	31.7	25.6	38.4




Manual Integrations
 APPROVED


mohammad
 5/3/2021 12:24:43 PM



#43
 1,1'-Biphenyl
 Concen: 33.912 ng/ul
 RT: 13.05 min Scan# 1728
 Delta R.T. -0.01 min
 Lab File: BM029757.D
 Acq: 01 May 2021 21:50

Tgt Ion:	154	Resp:	363762
Ion Ratio	Lower	Upper	
154	100		
153	40.7	31.9	47.9
76	9.0	7.0	10.4









#44
2-Chloronaphthalene
Concen: 33.686 ng/ul
RT: 13.09 min Scan# 1735
Delta R.T. 0.00 min
Lab File: BM029757.D
Acq: 01 May 2021 21:50

Instrument :
BNA_M
ClientSampleId :
SLCS733

Tot Ion	Ratio	Resp	Lower	Upper
162	100	291093		
164	33.7		27.0	40.4
127	32.4		26.3	39.5




Manual Integrations
APPROVED


mohammad
5/3/2021 12:24:43 PM



#45
2-Nitroaniline
Concen: 34.799 ng/ul
RT: 13.31 min Scan# 1772
Delta R.T. 0.00 min
Lab File: BM029757.D
Acq: 01 May 2021 21:50

Tgt Ion	Ratio	Resp	Lower	Upper
65	100	58067		
92	78.8		55.9	83.9
138	136.3		102.2	153.2






#47
Dimethylphthalate
Concen: 34.967 ng/ul
RT: 13.69 min Scan# 1837
Delta R.T. 0.00 min
Lab File: BM029757.D
Acq: 01 May 2021 21:50

Instrument :
BNA_M
ClientSampleId :
SLCS733


Tot Ion:	163	Resp:	385362
Ion Ratio	Lower	Upper	
163	100		
194	4.8	3.7	5.5
164	9.6	8.0	12.0

Manual Integrations
APPROVED



mohammad
5/3/2021 12:24:43 PM



Abundance Ion 163.00 (162.70 to 163.70): E
Ion 194.00 (193.70 to 194.70): E
Ion 164.00 (163.70 to 164.70): E




Time--> 13.60 13.70 13.80




#48
2,6-Dinitrotoluene
Concen: 36.797 ng/ul
RT: 13.81 min Scan# 1857
Delta R.T. 0.00 min
Lab File: BM029757.D
Acq: 01 May 2021 21:50


Tgt Ion:	165	Resp:	81739
Ion Ratio	Lower	Upper	
165	100		
89	45.6	40.2	60.2
121	17.7	14.6	22.0




Abundance Ion 165.00 (164.70 to 165.70): E
Ion 89.00 (88.70 to 89.70): BM
Ion 121.00 (120.70 to 121.70): E



Time--> 13.75 13.80 13.85 13.90





#50

Acenaphthylene

Concen: 32.874 ng/ul

RT: 13.95 min Scan# 1880

Delta R.T. 0.00 min

Lab File: BM029757.D

Acq: 01 May 2021 21:50

Instrument :

BNA_M





ClientSampleId :

SLCS733

Tot Ion:	152	Resp:	428761
Ion Ratio	Lower	Upper	
152	100		
151	20.0	15.9	23.9
153	12.9	10.7	16.1

Manual Integrations
APPROVED

mohammad
5/3/2021 12:24:43 PM



#51

3-Nitroaniline

Concen: 29.488 ng/ul




RT: 14.14 min Scan# 1913


Delta R.T. -0.01 min

Lab File: BM029757.D

Acq: 01 May 2021 21:50

Tgt Ion:	138	Resp:	58312
Ion Ratio	Lower	Upper	
138	100		
108	9.4	7.5	11.3
92	102.5	87.4	131.0





#52

Acenaphthene

Concen: 34.167 ng/ul

RT: 14.29 min Scan# 1938

Delta R.T. -0.01 min

Lab File: BM029757.D

Acq: 01 May 2021 21:50

Instrument :

BNA_M





ClientSampled :

SLCS733

Tot Ion:	153	Resp:	310789
Ion Ratio	Lower	Upper	
153	100		
152	46.3	37.8	56.8
154	88.6	71.4	107.2

Manual Integrations
APPROVED

mohammad
5/3/2021 12:24:43 PM



#53

2,4-Dinitrophenol

Concen: 35.244 ng/ul




RT: 14.36 min Scan# 1950


Delta R.T. -0.01 min

Lab File: BM029757.D

Acq: 01 May 2021 21:50

Tgt Ion:	184	Resp:	46674
Ion Ratio	Lower	Upper	
184	100		
63	46.6	42.2	63.4
154	58.1	43.0	64.4









#55
4-Nitrophenol
Concen: 40.553 ng/ul
RT: 14.46 min Scan# 1968
Delta R.T. -0.01 min
Lab File: BM029757.D
Acq: 01 May 2021 21:50

Instrument :
BNA_M
ClientSampleId :
SLCS733

Tot Ion	Ratio	Lower	Upper
109	100		
139	137.5	96.5	144.7
65	120.8	89.9	134.9




Manual Integrations
APPROVED


mohammad
5/3/2021 12:24:43 PM



#56
Dibenzofuran
Concen: 34.459 ng/ul
RT: 14.62 min Scan# 1995
Delta R.T. -0.01 min
Lab File: BM029757.D
Acq: 01 May 2021 21:50

Tgt Ion	Ratio	Lower	Upper
168	100		
139	31.9	26.0	39.0









#57
 2,4-Dinitrotoluene
 Concen: 37.941 ng/ul
 RT: 14.60 min Scan# 1992
 Delta R.T. -0.01 min
 Lab File: BM029757.D
 Acq: 01 May 2021 21:50

Instrument :
 BNA_M
 ClientSampled :
 SLCS733

Tot Ion:	165	Resp:	118168
Ion Ratio	Lower	Upper	
165	100		
63	33.4	27.4	41.0
182	3.6	2.6	4.0




Manual Integrations
 APPROVED


mohammad
 5/3/2021 12:24:43 PM



#58
 2,3,4,6-Tetrachlorophenol
 Concen: 37.259 ng/ul
 RT: 14.86 min Scan# 2035
 Delta R.T. -0.01 min
 Lab File: BM029757.D
 Acq: 01 May 2021 21:50

Tgt Ion:	232	Resp:	119037
Ion Ratio	Lower	Upper	
232	100		
131	29.7	26.2	39.2
130	1.6	1.4	2.2
166	25.4	21.3	31.9









#59
Diethylphthalate
Concen: 35.168 ng/ul
RT: 15.06 min Scan# 2069
Delta R.T. -0.01 min
Lab File: BM029757.D
Acq: 01 May 2021 21:50

Instrument :
BNA_M
ClientSampleId :
SLCS733

Tot Ion	Ratio	Lower	Upper
149	100		
177	25.7	19.8	29.8
150	12.3	9.8	14.6




Manual Integrations
APPROVED


mohammad
5/3/2021 12:24:43 PM



#61
Fluorene
Concen: 34.445 ng/ul
RT: 15.27 min Scan# 2106
Delta R.T. -0.01 min
Lab File: BM029757.D
Acq: 01 May 2021 21:50

Tgt Ion	Ratio	Lower	Upper
166	100		
165	99.5	77.3	115.9
167	14.0	10.4	15.6









#62
4-Chlorophenyl-phenylether
Concen: 35.901 ng/ul
RT: 15.27 min Scan# 2106
Delta R.T. 0.00 min
Lab File: BM029757.D
Acq: 01 May 2021 21:50

Instrument :
BNA_M
ClientSampled :
SLCS733

Tot Ion	Ratio	Resp	Lower	Upper
204	100	226099		
206	32.6		26.6	40.0
141	43.1		37.2	55.8




Manual Integrations
APPROVED


mohammad
5/3/2021 12:24:43 PM



#63
4-Nitroaniline
Concen: 32.324 ng/ul m
RT: 15.31 min Scan# 2112
Delta R.T. -0.01 min
Lab File: BM029757.D
Acq: 01 May 2021 21:50

Tgt Ion	Ratio	Resp	Lower	Upper
138	100	67997		
92	43.4		34.0	51.0
108	71.8		56.8	85.2









#66
 4,6-Dinitro-2-methylphenol
 Concen: 34.983 ng/ul
 RT: 15.37 min Scan# 2122
 Delta R.T. -0.01 min
 Lab File: BM029757.D
 Acq: 01 May 2021 21:50

Instrument :
 BNA_M
 ClientSampleId :
 SLCS733

Tot Ion	Ratio	Lower	Upper
198	100		
182	3.8	4.6	6.8#
77	18.1	13.0	19.4




Manual Integrations
 APPROVED


mohammad
 5/3/2021 12:24:43 PM



#67
 N-Nitrosodiphenylamine
 Concen: 34.593 ng/ul
 RT: 15.49 min Scan# 2143
 Delta R.T. 0.00 min
 Lab File: BM029757.D
 Acq: 01 May 2021 21:50

Tgt Ion	Ratio	Lower	Upper
169	100		
168	65.2	53.1	79.7
167	34.3	28.1	42.1






#68
4-Bromophenyl-phenylether
Concen: 35.460 ng/ul
RT: 16.17 min Scan# 2258
Delta R.T. 0.00 min
Lab File: BM029757.D
Acq: 01 May 2021 21:50

Instrument :
BNA_M
ClientSampled :
SLCS733

Tot Ion	Ratio	Lower	Upper
248	100		
250	98.9	77.3	115.9
141	49.0	44.3	66.5

Manual Integrations
APPROVED

mohammad
5/3/2021 12:24:43 PM






Abundance

Ion 248.00 (247.70 to 248.70): E


Ion 250.00 (249.70 to 250.70): E

Ion 141.00 (140.70 to 141.70): E



#69
Hexachlorobenzene
Concen: 36.356 ng/ul
RT: 16.28 min Scan# 2277
Delta R.T. 0.00 min
Lab File: BM029757.D
Acq: 01 May 2021 21:50

Tgt Ion	Ratio	Lower	Upper
284	100		
142	23.6	21.1	31.7
249	29.6	25.8	38.8






Abundance

Ion 284.00 (283.70 to 284.70): E

Ion 142.00 (141.70 to 142.70): E

Ion 249.00 (248.70 to 249.70): E





#70

Atrazine

Concen: 35.632 ng/ul

RT: 16.45 min Scan# 2306

Delta R.T. 0.00 min

Lab File: BM029757.D

Acq: 01 May 2021 21:50

Instrument :

BNA_M





ClientSampleId :

SLCS733

Tot Ion:	200	Resp:	145466
Ion Ratio	Lower	Upper	
200	100		
173	21.6	18.5	27.7
215	50.4	39.6	59.4

Manual Integrations
APPROVED

mohammad
5/3/2021 12:24:43 PM



#71

Pentachlorophenol

Concen: 39.711 ng/ul




RT: 16.63 min Scan# 2336


Delta R.T. -0.01 min

Lab File: BM029757.D

Acq: 01 May 2021 21:50

Tgt Ion:	266	Resp:	94981
Ion Ratio	Lower	Upper	
266	100		
264	65.0	51.1	76.7
268	62.9	52.8	79.2





#72

Phenanthrene

Concen: 35.571 ng/ul

RT: 17.01 min Scan# 2401

Delta R.T. -0.01 min

Lab File: BM029757.D

Acq: 01 May 2021 21:50

Instrument :

BNA_M





ClientSampleId :

SLCS733

Tot Ion:178	Resp:	654573
Ion Ratio	Lower	Upper
178	100	
179	15.6	12.3 18.5
176	18.4	15.1 22.7

Manual Integrations
APPROVED

mohammad
5/3/2021 12:24:43 PM



#74

Anthracene

Concen: 35.335 ng/ul




RT: 17.10 min Scan# 2417


Delta R.T. 0.00 min

Lab File: BM029757.D

Acq: 01 May 2021 21:50

Tgt Ion:178	Resp:	668200
Ion Ratio	Lower	Upper
178	100	
179	15.1	12.3 18.5
176	18.0	14.0 21.0





#75

1,2,3,4-Tetrachlorobenzene

Concen: 32.215 ng/uL

RT: 13.02 min Scan# 1722

Delta R.T. 0.00 min

Lab File: BM029757.D

Acq: 01 May 2021 21:50

Instrument :

BNA_M





ClientSampleId :

SLCS733

Tot Ion	Ion	Ratio	Resp	Lower	Upper
216	100		192223		
214	77.4	63.8	95.6		
218	47.2	39.1	58.7		

Manual Integrations
APPROVED

mohammad
5/3/2021 12:24:43 PM



#76

Pentachlorobenzene

Concen: 33.742 ng/uL




RT: 14.55 min Scan# 1982


Delta R.T. -0.01 min

Lab File: BM029757.D

Acq: 01 May 2021 21:50

Tgt Ion	Ion	Ratio	Resp	Lower	Upper
250	100		188555		
248	63.1	48.5	72.7		
252	64.1	52.0	78.0		





#77

Carbazole

Concen: 34.070 ng/ul

RT: 17.37 min Scan# 2463

Delta R.T. -0.01 min

Lab File: BM029757.D

Acq: 01 May 2021 21:50

Instrument :

BNA_M





ClientSampleId :

SLCS733

Tot Ion:	167	Resp:	562272
Ion Ratio	Lower	Upper	
167	100		
166	21.2	17.3	25.9
139	10.8	8.7	13.1

Manual Integrations
APPROVED

mohammad
5/3/2021 12:24:43 PM



#78

Di-n-butylphthalate

Concen: 36.423 ng/ul




RT: 17.95 min Scan# 2560


Delta R.T. 0.00 min

Lab File: BM029757.D

Acq: 01 May 2021 21:50

Tgt Ion:	149	Resp:	638447
Ion Ratio	Lower	Upper	
149	100		
150	9.3	7.4	11.0
104	4.3	3.7	5.5









#80
Fluoranthene
Concen: 36.659 ng/ul
RT: 19.03 min Scan# 2744
Delta R.T. -0.01 min
Lab File: BM029757.D
Acq: 01 May 2021 21:50

Instrument :
BNA_M
ClientSampleId :
SLCS733

Tot Ion	Ratio	Resp Lower	Upper
202	100		
101	6.7	5.6	8.4
100	5.1	4.5	6.7




Manual Integrations
APPROVED


mohammad
5/3/2021 12:24:43 PM



#82
Pyrene
Concen: 36.049 ng/ul
RT: 19.39 min Scan# 2806
Delta R.T. -0.01 min
Lab File: BM029757.D
Acq: 01 May 2021 21:50

Tgt Ion	Ratio	Resp Lower	Upper
202	100		
101	7.8	6.6	9.8
100	6.6	5.5	8.3









#83
 Butylbenzylphthalate
 Concen: 34.972 ng/ul
 RT: 20.30 min Scan# 2961
 Delta R.T. 0.00 min
 Lab File: BM029757.D
 Acq: 01 May 2021 21:50

Instrument :
 BNA_M
 ClientSampleId :
 SLCS733

Tot Ion:	149	Resp:	296468
Ion Ratio	Lower	Upper	
149	100		
91	58.7	49.4	74.0
206	26.7	19.9	29.9




Manual Integrations
 APPROVED


mohammad
 5/3/2021 12:24:43 PM



#84
 3,3'-Dichlorobenzidine
 Concen: 29.801 ng/ul
 RT: 21.08 min Scan# 3093
 Delta R.T. 0.00 min
 Lab File: BM029757.D
 Acq: 01 May 2021 21:50

Tgt Ion:	252	Resp:	292384
Ion Ratio	Lower	Upper	
252	100		
254	66.4	51.0	76.6
126	6.8	5.7	8.5









#85
 Benzo(a)anthracene
 Concen: 35.660 ng/ul
 RT: 21.14 min Scan# 3103
 Delta R.T. 0.00 min
 Lab File: BM029757.D
 Acq: 01 May 2021 21:50

Instrument :
 BNA_M
 ClientSampleId :
 SLCS733

Tot Ion:	228	Resp:	891952
Ion Ratio	Lower	Upper	
228	100		
229	19.4	15.6	23.4
226	26.9	21.4	32.0




Manual Integrations
 APPROVED


mohammad
 5/3/2021 12:24:43 PM



#86
 Bis(2-ethylhexyl)phthalate
 Concen: 34.881 ng/ul
 RT: 21.08 min Scan# 3093
 Delta R.T. 0.00 min
 Lab File: BM029757.D
 Acq: 01 May 2021 21:50

Tgt Ion:	149	Resp:	453756
Ion Ratio	Lower	Upper	
149	100		
167	29.6	23.3	34.9
279	6.1	4.6	6.8





#87

Chrysene

Concen: 35.539 ng/ul

RT: 21.19 min Scan# 3112

Delta R.T. 0.00 min

Lab File: BM029757.D

Acq: 01 May 2021 21:50

Instrument :

BNA_M





ClientSampleId :

SLCS733

Tot Ion:	228	Resp:	888320
Ion Ratio	Lower	Upper	
228	100		
226	29.8	23.8	35.8
229	18.8	15.6	23.4

Manual Integrations
APPROVED

mohammad
5/3/2021 12:24:43 PM



#89

Di-n-octyl phthalate

Concen: 34.483 ng/ul




RT: 21.94 min Scan# 3239


Delta R.T. 0.00 min

Lab File: BM029757.D

Acq: 01 May 2021 21:50

Tgt Ion:149 Resp: 797017





#90

Benzo(b)fluoranthene

Concen: 37.659 ng/ul

RT: 22.67 min Scan# 3364

Delta R.T. 0.00 min

Lab File: BM029757.D

Acq: 01 May 2021 21:50

Instrument :

BNA_M





ClientSampleId :

SLCS733

Tot Ion	Ratio	Lower	Upper
252	100		
253	21.6	17.3	25.9
125	6.5	5.2	7.8

Manual Integrations
APPROVED

mohammad
5/3/2021 12:24:43 PM



#91

Benzo(k)fluoranthene

Concen: 35.450 ng/ul




RT: 22.72 min Scan# 3371


Delta R.T. -0.01 min

Lab File: BM029757.D

Acq: 01 May 2021 21:50

Tgt Ion	Ratio	Lower	Upper
252	100		
253	21.9	17.0	25.6
125	6.3	5.0	7.4





#93

Benzo(a)pyrene

Concen: 36.316 ng/ul

RT: 23.23 min Scan# 3458

Delta R.T. 0.00 min

Lab File: BM029757.D

Acq: 01 May 2021 21:50

Instrument :

BNA_M


ClientSampleId :

SLCS733




Tot Ion:	252	Resp:	862408
Ion Ratio	Lower	Upper	
252	100		
253	21.8	17.6	26.4
125	7.0	5.9	8.9

Manual Integrations
APPROVED

mohammad
5/3/2021 12:24:43 PM



Abundance Ion 252.00 (251.70 to 252.70): E
Ion 253.00 (252.70 to 253.70): E
Ion 125.00 (124.70 to 125.70): E



#94

Indeno(1,2,3-cd)pyrene

Concen: 37.045 ng/ul

RT: 25.50 min Scan# 3844




Delta R.T. 0.01 min


Lab File: BM029757.D

Acq: 01 May 2021 21:50

Tgt Ion:	276	Resp:	1152063
Ion Ratio	Lower	Upper	
276	100		
138	17.2	13.7	20.5
277	25.4	20.4	30.6

Abundance Ion 276.00 (275.70 to 276.70): E
Ion 138.00 (137.70 to 138.70): E
Ion 277.00 (276.70 to 277.70): E









#95
Dibenzo(a,h)anthracene
Concen: 37.003 ng/ul
RT: 25.50 min Scan# 3845
Delta R.T. 0.00 min
Lab File: BM029757.D
Acq: 01 May 2021 21:50

Instrument :
BNA_M
ClientSampled :
SLCS733

Tot Ion:	278	Resp:	960478
Ion Ratio	Lower	Upper	
278	100		
139	10.5	8.9	13.3
279	23.9	18.7	28.1

Manual Integrations
APPROVED

mohammad
5/3/2021 12:24:43 PM



#96
Benzo(g,h,i)perylene
Concen: 36.993 ng/ul
RT: 26.16 min Scan# 3956
Delta R.T. 0.01 min
Lab File: BM029757.D
Acq: 01 May 2021 21:50

Tgt Ion:	276	Resp:	944874
Ion Ratio	Lower	Upper	
276	100		
138	14.7	12.4	18.6
277	23.8	19.1	28.7

