

Data Path : Z:\SVOASRV\HPCHEM1\BNA M\DATA\BM050121\
 Data File : BM029759.D
 Acq On : 01 May 2021 23:03
 Operator : CG/JU
 Sample : PB135930BS
 Misc :
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleID :
 SLCS930

Manual Integrations
 APPROVED

mohammad
 5/3/2021 12:24:47 PM

Quant Time: May 02 01:26:17 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_M\METHODS\SFAM-EPA-BM050121.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Sat May 01 16:23:05 2021
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.57	152	53300	20.00	ng/ul	0.00
20) Naphthalene-d8	10.35	136	197858	20.00	ng/ul	0.00
38) Acenaphthene-d10	14.22	164	150958	20.00	ng/ul	0.00
64) Phenanthrene-d10	16.97	188	337358	20.00	ng/ul	0.00
79) Chrysene-d12	21.16	240	402033	20.00	ng/ul	0.00
88) Perylene-d12	23.32	264	425631	20.00	ng/ul	0.00

System Monitoring Compounds

3) 1,4-Dioxane-d8	3.08	96	6494	5.96	ng/uL	0.00
4) Pyridine-d5	3.49	84	72799	30.84	ng/ul	-0.01
7) Phenol-d5	6.76	99	115365	32.14	ng/ul	0.00
9) Bis-(2-Chloroethyl)ether-d	6.92	67	59945	31.89	ng/ul	0.00
11) 2-Chlorophenol-d4	7.11	132	110833	34.14	ng/ul	0.00
15) 4-Methylphenol-d8	8.29	113	101128	32.90	ng/ul	0.00
21) Nitrobenzene-d5	8.73	128	49542	34.38	ng/ul	0.00
24) 2-Nitrophenol-d4	9.45	143	64630	35.43	ng/ul	0.00
28) 2,4-Dichlorophenol-d3	9.99	165	137951	36.19	ng/ul	0.00
31) 4-Chloroaniline-d4	10.50	131	126953	29.54	ng/ul	0.00
46) Dimethylphthalate-d6	13.65	166	419921	35.11	ng/ul	0.00
49) Acenaphthylene-d8	13.92	160	507662	34.43	ng/ul	0.00
54) 4-Nitrophenol-d4	14.45	143	56181	35.52	ng/ul	0.00
60) Fluorene-d10	15.22	176	364856	35.27	ng/ul	0.00
65) 4,6-Dinitro-2-methylphenol	15.36	200	84660	34.51	ng/ul	0.00
73) Anthracene-d10	17.07	188	598285	35.18	ng/ul	0.00
81) Pyrene-d10	19.36	212	739838	36.04	ng/ul	0.00
92) Benzo(a)pyrene-d12	23.19	264	860679	35.24	ng/ul	0.00

Target Compounds

					Ovalue
2) 1,4-Dioxane	3.11	88	13210	11.513	ng/uL 98
5) Pyridine	3.50	79	79412	31.603	ng/ul 99
6) Benzaldehyde	6.73	77	55488	26.958	ng/ul 97
8) Phenol	6.79	94	117953	33.159	ng/ul 95
10) Bis(2-Chloroethyl)ether	7.01	93	89651	32.932	ng/ul 95
12) 2-Chlorophenol	7.15	128	112319	34.774	ng/ul 97
13) 2-Methylphenol	8.03	108	94196	32.990	ng/ul 98
14) 2,2'-oxybis(1-Chloropropan	8.10	45	83898	31.396	ng/ul 97
16) Acetophenone	8.40	105	150637	31.792	ng/ul 99
17) N-Nitroso-di-n-propylamine	8.39	70	73894	32.360	ng/ul 99
18) 4-Methylphenol	8.36	108	103299	33.205	ng/ul 98
19) Hexachloroethane	8.64	117	47349	32.421	ng/ul 97
22) Nitrobenzene	8.77	77	110871	34.836	ng/ul 95
23) Isophorone	9.30	82	212420	34.598	ng/ul 99
25) 2-Nitrophenol	9.48	139	66627	35.199	ng/ul 99
26) 2,4-Dimethylphenol	9.55	107	122406	35.522	ng/ul 96
27) Bis(2-Chloroethoxy)methane	9.78	93	124902	34.948	ng/ul 99
29) 2,4-Dichlorophenol	10.01	162	131881	37.050	ng/ul 95
30) Naphthalene	10.40	128	354128	33.674	ng/ul 98
32) 4-Chloroaniline	10.52	127	121769	28.148	ng/ul 98
33) Hexachlorobutadiene	10.69	225	108655	35.317	ng/ul 99
34) Caprolactam	11.30	113	33718m	34.308	ng/ul

Data Path : Z:\SVOASRV\HPCHEM1\BNA M\DATA\BM050121\
 Data File : BM029759.D
 Acq On : 01 May 2021 23:03
 Operator : CG/JU
 Sample : PB135930BS
 Misc :
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 SLCS930

Manual Integrations
 APPROVED

mohammad
 5/3/2021 12:24:47 PM

Quant Time: May 02 01:26:17 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_M\METHODS\SFAM-EPA-BM050121.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Sat May 01 16:23:05 2021
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
35) 4-Chloro-3-methylphenol	11.66	107	108750	35.871	ng/ul	98
36) 2-Methylnaphthalene	12.02	142	274896	33.847	ng/ul	97
37) 1-Methylnaphthalene	12.25	142	267001	33.773	ng/ul	98
39) 1,2,4,5-Tetrachlorobenzene	12.40	216	202399	34.404	ng/ul	98
40) Hexachlorocyclopentadiene	12.37	237	88879	30.776	ng/ul	94
41) 2,4,6-Trichlorophenol	12.65	196	122400	36.049	ng/ul	92
42) 2,4,5-Trichlorophenol	12.72	196	125968	35.379	ng/ul	97
43) 1,1'-Biphenyl	13.05	154	389410	34.496	ng/ul	98
44) 2-Chloronaphthalene	13.09	162	314592	34.594	ng/ul	98
45) 2-Nitroaniline	13.31	65	62253	35.451	ng/ul	97
47) Dimethylphthalate	13.69	163	410544	35.398	ng/ul	100
48) 2,6-Dinitrotoluene	13.81	165	86137	36.847	ng/ul	92
50) Acenaphthylene	13.95	152	451813	32.918	ng/ul	99
51) 3-Nitroaniline	14.14	138	58662	28.189	ng/ul	97
52) Acenaphthene	14.29	153	326551	34.114	ng/ul	99
53) 2,4-Dinitrophenol	14.36	184	51446	36.914	ng/ul	93
55) 4-Nitrophenol	14.46	109	46755	41.794	ng/ul	86
56) Dibenzofuran	14.63	168	500860	35.212	ng/ul	98
57) 2,4-Dinitrotoluene	14.60	165	124890	38.104	ng/ul	99
58) 2,3,4,6-Tetrachlorophenol	14.86	232	128257	38.147	ng/ul	95
59) Diethylphthalate	15.06	149	394314	35.816	ng/ul	99
61) Fluorene	15.27	166	418574	35.277	ng/ul	99
62) 4-Chlorophenyl-phenylether	15.27	204	242495	36.588	ng/ul	97
63) 4-Nitroaniline	15.31	138	68195m	30.805	ng/ul	
66) 4,6-Dinitro-2-methylphenol	15.37	198	84437	35.731	ng/ul#	96
67) N-Nitrosodiphenylamine	15.49	169	361295	34.681	ng/ul	99
68) 4-Bromophenyl-phenylether	16.17	248	154525	37.341	ng/ul	96
69) Hexachlorobenzene	16.28	284	175300	36.679	ng/ul	98
70) Atrazine	16.45	200	155189	36.154	ng/ul	100
71) Pentachlorophenol	16.63	266	100357	39.906	ng/ul	98
72) Phenanthrene	17.01	178	688885	35.604	ng/ul	99
74) Anthracene	17.10	178	710915	35.755	ng/ul	99
75) 1,2,3,4-Tetrachlorobenzene	13.02	216	203906	32.501	ng/uL	98
76) Pentachlorobenzene	14.54	250	200524	34.128	ng/uL	98
77) Carbazole	17.38	167	582043	33.543	ng/ul	99
78) Di-n-butylphthalate	17.94	149	683313	37.075	ng/ul	100
80) Fluoranthene	19.03	202	882445	36.544	ng/ul	98
82) Pyrene	19.39	202	921574	35.866	ng/ul	99
83) Butylbenzylphthalate	20.30	149	316563	35.505	ng/ul	96
84) 3,3'-Dichlorobenzidine	21.08	252	306478	29.701	ng/ul	98
85) Benzo(a)anthracene	21.14	228	939048	35.695	ng/ul	99
86) Bis(2-ethylhexyl)phthalate	21.08	149	488916	35.735	ng/ul	100
87) Chrysene	21.19	228	933352	35.503	ng/ul	99
89) Di-n-octyl phthalate	21.94	149	837709	34.295	ng/ul	100
90) Benzo(b)fluoranthene	22.67	252	991394	35.740	ng/ul	99
91) Benzo(k)fluoranthene	22.72	252	988053	36.161	ng/ul	99
93) Benzo(a)pyrene	23.23	252	891991	35.542	ng/ul	99
94) Indeno(1,2,3-cd)pyrene	25.50	276	1175445	35.765	ng/ul	99
95) Dibenzo(a,h)anthracene	25.50	278	980687	35.751	ng/ul	99
96) Benzo(g,h,i)perylene	26.16	276	969194	35.906	ng/ul	98

Data Path : Z:\SVOASRV\HPCHEM1\BNA M\DATA\BM050121\
Data File : BM029759.D
Acq On : 01 May 2021 23:03
Operator : CG/JU
Sample : PB135930BS
Misc :
ALS Vial : 9 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
SLCS930


Manual Integrations
APPROVED

mohammad
5/3/2021 12:24:47 PM

Quant Time: May 02 01:26:17 2021
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_M\METHODS\SFAM-EPA-BM050121.M
Quant Title : SVOA CALIBRATION
QLast Update : Sat May 01 16:23:05 2021
Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
--------------------	------	------	----------	------	-------	----------

(#) = qualifier out of range (m) = manual integration (+) = signals summed



#2

1,4-Dioxane

Concen: 11.513 ng/uL

RT: 3.11 min Scan# 38

Delta R.T. 0.00 min

Lab File: BM029759.D

Acq: 01 May 2021 23:03

Instrument :

BNA_M





ClientSampled :

SLCS930

Tot Ion:	88	Resp:	13210
Ion	Ratio	Lower	Upper
88	100		
43	28.4	22.6	33.8
58	59.3	45.5	68.3

Manual Integrations
APPROVED

mohammad
5/3/2021 12:24:47 PM



#5

Pyridine

Concen: 31.603 ng/uL




RT: 3.50 min Scan# 105


Delta R.T. -0.02 min

Lab File: BM029759.D

Acq: 01 May 2021 23:03

Tgt Ion:	79	Resp:	79412
Ion	Ratio	Lower	Upper
79	100		
52	64.6	51.4	77.2
51	32.4	27.4	41.0





#6

Benzaldehyde

Concen: 26.958 ng/ul

RT: 6.73 min Scan# 653

Delta R.T. -0.01 min

Lab File: BM029759.D

Acq: 01 May 2021 23:03

Instrument :

BNA_M





ClientSampled :

SLCS930

Tot Ion:	77	Resp:	55488
Ion	Ratio	Lower	Upper
77	100		
105	100.8	81.5	122.3
106	97.1	81.8	122.8

Manual Integrations
APPROVED

mohammad
5/3/2021 12:24:47 PM



#8

Phenol

Concen: 33.159 ng/ul




RT: 6.79 min Scan# 663


Delta R.T. -0.01 min

Lab File: BM029759.D

Acq: 01 May 2021 23:03

Tgt Ion:	94	Resp:	117953
Ion	Ratio	Lower	Upper
94	100		
65	28.9	21.4	32.2
66	39.6	29.0	43.4









#10
Bis(2-Chloroethyl)ether
Concen: 32.932 ng/ul
RT: 7.01 min Scan# 701
Delta R.T. -0.01 min
Lab File: BM029759.D
Acq: 01 May 2021 23:03

Instrument :
BNA_M
ClientSampleId :
SLCS930

Tot Ion:	93	Resp:	89651
Ion	Ratio	Lower	Upper
93	100		
63	67.8	51.0	76.6
95	34.8	25.6	38.4




Manual Integrations
APPROVED


mohammad
5/3/2021 12:24:47 PM



#12
2-Chlorophenol
Concen: 34.774 ng/ul
RT: 7.15 min Scan# 724
Delta R.T. -0.00 min
Lab File: BM029759.D
Acq: 01 May 2021 23:03

Tgt Ion:	128	Resp:	112319
Ion	Ratio	Lower	Upper
128	100		
64	35.2	30.0	45.0
130	32.3	27.1	40.7





#13

2-Methylphenol

Concen: 32.990 ng/ul

RT: 8.03 min Scan# 874

Delta R.T. -0.01 min

Lab File: BM029759.D

Acq: 01 May 2021 23:03

Instrument :

BNA_M

ClientSampleId :

SLCS930

Tot Ion: 108 Resp: 94196

Ion Ratio Lower Upper

108 100


107 91.6 71.4 107.2

Manual Integrations

APPROVED

mohammad

5/3/2021 12:24:47 PM



Abundance Ion 108.00 (107.70 to 108.70): E

Ion 107.00 (106.70 to 107.70): E


60000

40000


20000

0

Time-->



Time-->



#14

2,2'-oxybis(1-Chloropropane)

Concen: 31.396 ng/ul

RT: 8.10 min Scan# 887

Delta R.T. -0.01 min

Lab File: BM029759.D

Acq: 01 May 2021 23:03


Tgt Ion: 45 Resp: 83898

Ion Ratio Lower Upper

45 100

77 23.3 16.5 24.7

79 16.9 13.5 20.3



Abundance Ion 45.00 (44.70 to 45.70): BM0

Ion 77.00 (76.70 to 77.70): BM0

Ion 79.00 (78.70 to 79.70): BM0

40000


30000


20000

10000

0

Time-->





#16

Acetophenone

Concen: 31.792 ng/ul

RT: 8.40 min Scan# 937

Delta R.T. -0.01 min

Lab File: BM029759.D

Acq: 01 May 2021 23:03

Instrument :

BNA_M


ClientSampleId :

SLCS930

Tot Ion:	105	Resp:	150637
Ion Ratio	Lower	Upper	
105	100		
77	75.2	60.1	90.1
51	22.5	17.0	25.4

Manual Integrations
APPROVED




mohammad
5/3/2021 12:24:47 PM



Abundance Ion 105.00 (104.70 to 105.70): E

Ion 77.00 (76.70 to 77.70): BM0

Ion 51.00 (50.70 to 51.70): BM0



#17

N-Nitroso-di-n-propylamine

Concen: 32.360 ng/ul

RT: 8.39 min Scan# 935

Delta R.T. -0.00 min

Lab File: BM029759.D

Acq: 01 May 2021 23:03




Tgt Ion:	70	Resp:	73894
Ion Ratio	Lower	Upper	
70	100		
42	50.5	40.1	60.1
101	11.2	8.7	13.1
130	23.4	19.8	29.6


Abundance Ion 70.00 (69.70 to 70.70): BM0

Ion 42.00 (41.70 to 42.70): BM0

Ion 101.00 (100.70 to 101.70): E

Ion 130.00 (129.70 to 130.70): E





#18

4-Methylphenol

Concen: 33.205 ng/ul

RT: 8.36 min Scan# 930

Delta R.T. -0.00 min

Lab File: BM029759.D

Acq: 01 May 2021 23:03

Instrument :

BNA_M

ClientSampleId :

SLCS930

Tot Ion:108 Resp: 103299

Ion Ratio Lower Upper

108 100


107 115.6 94.2 141.4

Manual Integrations

APPROVED




mohammad

5/3/2021 12:24:47 PM



Abundance Ion 108.00 (107.70 to 108.70): E

Ion 107.00 (106.70 to 107.70): E



#19

Hexachloroethane

Concen: 32.421 ng/ul

RT: 8.64 min Scan# 978

Delta R.T. -0.00 min

Lab File: BM029759.D

Acq: 01 May 2021 23:03


Tgt Ion:117 Resp: 47349

Ion Ratio Lower Upper

117 100

201 152.1 117.4 176.2



199 90.4 73.4 110.0




Abundance Ion 117.00 (116.70 to 117.70): E

Ion 201.00 (200.70 to 201.70): E

Ion 199.00 (198.70 to 199.70): E









#22
Nitrobenzene
Concen: 34.836 ng/ul
RT: 8.77 min Scan# 1001
Delta R.T. 0.00 min
Lab File: BM029759.D
Acq: 01 May 2021 23:03

Instrument :
BNA_M
ClientSampled :
SLCS930

Tot Ion	Ratio	Resp	Lower	Upper
77	100	110871		
123	50.0	36.6	54.8	
65	13.1	10.2	15.2	




Manual Integrations
APPROVED


mohammad
5/3/2021 12:24:47 PM



#23
Isophorone
Concen: 34.598 ng/ul
RT: 9.30 min Scan# 1090
Delta R.T. -0.00 min
Lab File: BM029759.D
Acq: 01 May 2021 23:03

Tgt Ion	Ratio	Resp	Lower	Upper
82	100	212420		
95	7.6	6.0	9.0	
138	21.2	16.5	24.7	









#25
2-Nitrophenol
Concen: 35.199 ng/ul
RT: 9.48 min Scan# 1121
Delta R.T. -0.00 min
Lab File: BM029759.D
Acq: 01 May 2021 23:03

Instrument :
BNA_M
ClientSampled :
SLCS930

Tot Ion	Ratio	Resp	Lower	Upper
139	100	66627		
65	42.7		34.6	52.0
109	28.1		22.6	34.0




Manual Integrations
APPROVED


mohammad
5/3/2021 12:24:47 PM



#26
2,4-Dimethylphenol
Concen: 35.522 ng/ul
RT: 9.55 min Scan# 1132
Delta R.T. -0.01 min
Lab File: BM029759.D
Acq: 01 May 2021 23:03

Tgt Ion	Ratio	Resp	Lower	Upper
107	100	122406		
121	50.7		41.7	62.5
122	83.6		70.3	105.5





#27

Bis(2-Chloroethoxy)methane

Concen: 34.948 ng/ul

RT: 9.78 min Scan# 1172

Delta R.T. 0.00 min

Lab File: BM029759.D

Acq: 01 May 2021 23:03

Instrument :

BNA_M





ClientSampleId :

SLCS930

Tot Ion:	93	Resp:	124902
Ion	Ratio	Lower	Upper
93	100		
95	33.2	27.0	40.6
123	12.2	8.9	13.3

Manual Integrations
APPROVED

mohammad
5/3/2021 12:24:47 PM



#29

2,4-Dichlorophenol

Concen: 37.050 ng/ul




RT: 10.01 min Scan# 1211


Delta R.T. -0.01 min

Lab File: BM029759.D

Acq: 01 May 2021 23:03

Tgt Ion:	162	Resp:	131881
Ion	Ratio	Lower	Upper
162	100		
164	66.0	56.9	85.3
98	26.6	21.9	32.9









#30
Naphthalene
Concen: 33.674 ng/ul
RT: 10.40 min Scan# 1278
Delta R.T. -0.00 min
Lab File: BM029759.D
Acq: 01 May 2021 23:03

Instrument :
BNA_M
ClientSampled :
SLCS930

Tot Ion	Ratio	Lower	Upper
128	100		
129	11.0	8.6	12.8
127	13.1	11.3	16.9




Manual Integrations
APPROVED


mohammad
5/3/2021 12:24:47 PM



#32
4-Chloroaniline
Concen: 28.148 ng/ul
RT: 10.52 min Scan# 1298
Delta R.T. -0.01 min
Lab File: BM029759.D
Acq: 01 May 2021 23:03

Tgt Ion	Ratio	Lower	Upper
127	100		
129	32.5	25.3	37.9









#33
Hexachlorobutadiene
Concen: 35.317 ng/ul
RT: 10.69 min Scan# 1326
Delta R.T. -0.00 min
Lab File: BM029759.D
Acq: 01 May 2021 23:03

Instrument :
BNA_M
ClientSampleId :
SLCS930

Tot Ion	Ratio	Resp	Lower	Upper
225	100	108655		
223	62.5		50.0	75.0
227	63.9		52.6	79.0




Manual Integrations
APPROVED


mohammad
5/3/2021 12:24:47 PM



#34
Caprolactam
Concen: 34.308 ng/ul m
RT: 11.30 min Scan# 1431
Delta R.T. -0.01 min
Lab File: BM029759.D
Acq: 01 May 2021 23:03

Tgt Ion	Ratio	Resp	Lower	Upper
113	100	33718		
55	107.2		90.8	136.2
56	95.6		72.3	108.5









#35
 4-Chloro-3-methylphenol
 Concen: 35.871 ng/ul
 RT: 11.66 min Scan# 1491
 Delta R.T. -0.01 min
 Lab File: BM029759.D
 Acq: 01 May 2021 23:03

Instrument :
 BNA_M
 ClientSampleId :
 SLCS930

Tot Ion	Ratio	Resp	Lower	Upper
107	100			
144	30.7	22.9	34.3	
142	94.1	74.1	111.1	




Manual Integrations
 APPROVED


mohammad
 5/3/2021 12:24:47 PM



#36
 2-Methylnaphthalene
 Concen: 33.847 ng/ul
 RT: 12.02 min Scan# 1553
 Delta R.T. -0.01 min
 Lab File: BM029759.D
 Acq: 01 May 2021 23:03

Tgt Ion	Ratio	Resp	Lower	Upper
142	100			
141	88.6	69.0	103.6	





#37

1-Methylnaphthalene

Concen: 33.773 ng/ul

RT: 12.25 min Scan# 1591

Delta R.T. -0.01 min

Lab File: BM029759.D

Acq: 01 May 2021 23:03

Instrument :

BNA_M


ClientSampleId :

SLCS930




Tot Ion:	142	Resp:	267001
Ion Ratio	Lower	Upper	
142	100		
141	93.1	75.8	113.6
116	3.6	2.7	4.1

Manual Integrations
APPROVED

mohammad
5/3/2021 12:24:47 PM



Abundance Ion 142.00 (141.70 to 142.70): E
Ion 141.00 (140.70 to 141.70): E
Ion 116.00 (115.70 to 116.70): E



#39

1,2,4,5-Tetrachlorobenzene

Concen: 34.404 ng/ul

RT: 12.40 min Scan# 1617




Delta R.T. -0.01 min


Lab File: BM029759.D

Acq: 01 May 2021 23:03

Tgt Ion:	216	Resp:	202399
Ion Ratio	Lower	Upper	
216	100		
214	78.5	61.1	91.7
179	18.7	14.8	22.2
108	11.1	8.6	13.0

Abundance Ion 216.00 (215.70 to 216.70): E
Ion 214.00 (213.70 to 214.70): E
Ion 179.00 (178.70 to 179.70): E
Ion 108.00 (107.70 to 108.70): E









#40
Hexachlorocyclopentadiene
Concen: 30.776 ng/ul
RT: 12.37 min Scan# 1613
Delta R.T. -0.01 min
Lab File: BM029759.D
Acq: 01 May 2021 23:03

Instrument :
BNA_M
ClientSampleId :
SLCS930

Tot Ion	Ratio	Lower	Upper
237	100		
235	62.4	45.6	68.4
272	11.8	8.9	13.3




Manual Integrations
APPROVED


mohammad
5/3/2021 12:24:47 PM



#41
2,4,6-Trichlorophenol
Concen: 36.049 ng/ul
RT: 12.65 min Scan# 1660
Delta R.T. -0.01 min
Lab File: BM029759.D
Acq: 01 May 2021 23:03

Tgt Ion	Ratio	Lower	Upper
196	100		
198	100.6	72.7	109.1
200	31.6	24.7	37.1









#42
2,4,5-Trichlorophenol
Concen: 35.379 ng/ul
RT: 12.72 min Scan# 1672
Delta R.T. -0.01 min
Lab File: BM029759.D
Acq: 01 May 2021 23:03

Instrument :
BNA_M
ClientSampled :
SLCS930

Tot Ion	Ratio	Resp	Lower	Upper
196	100	125968		
198	94.9	77.5	116.3	
200	29.4	25.6	38.4	




Manual Integrations
APPROVED


mohammad
5/3/2021 12:24:47 PM



#43
1,1'-Biphenyl
Concen: 34.496 ng/ul
RT: 13.05 min Scan# 1728
Delta R.T. -0.01 min
Lab File: BM029759.D
Acq: 01 May 2021 23:03

Tgt Ion	Ratio	Resp	Lower	Upper
154	100	389410		
153	41.0	31.9	47.9	
76	9.0	7.0	10.4	









#44
2-Chloronaphthalene
Concen: 34.594 ng/ul
RT: 13.09 min Scan# 1735
Delta R.T. 0.00 min
Lab File: BM029759.D
Acq: 01 May 2021 23:03

Instrument :
BNA_M
ClientSampled :
SLCS930

Tot Ion	Ratio	Lower	Upper
162	100		
164	32.8	27.0	40.4
127	31.0	26.3	39.5




Manual Integrations
APPROVED


mohammad
5/3/2021 12:24:47 PM



#45
2-Nitroaniline
Concen: 35.451 ng/ul
RT: 13.31 min Scan# 1772
Delta R.T. -0.00 min
Lab File: BM029759.D
Acq: 01 May 2021 23:03

Tgt Ion	Ratio	Lower	Upper
65	100		
92	75.1	55.9	83.9
138	128.8	102.2	153.2





#47

Dimethylphthalate

Concen: 35.398 ng/ul

RT: 13.69 min Scan# 1837

Delta R.T. -0.00 min

Lab File: BM029759.D

Acq: 01 May 2021 23:03

Instrument :

BNA_M


ClientSampleId :

SLCS930


Tot Ion:	163	Resp:	410544
Ion Ratio	Lower	Upper	
163	100		
194	4.8	3.7	5.5
164	10.1	8.0	12.0

Manual Integrations
APPROVED



mohammad
5/3/2021 12:24:47 PM



Abundance Ion 163.00 (162.70 to 163.70): E
400000
Ion 194.00 (193.70 to 194.70): E
Ion 164.00 (163.70 to 164.70): E



Time--> 13.60 13.70 13.80



#48

2,6-Dinitrotoluene

Concen: 36.847 ng/ul


RT: 13.81 min Scan# 1857

Delta R.T. 0.00 min


Lab File: BM029759.D

Acq: 01 May 2021 23:03


Tgt Ion:	165	Resp:	86137
Ion Ratio	Lower	Upper	
165	100		
89	43.9	40.2	60.2
121	15.6	14.6	22.0




Abundance Ion 165.00 (164.70 to 165.70): E
80000
Ion 89.00 (88.70 to 89.70): BM0
Ion 121.00 (120.70 to 121.70): E



Time--> 13.75 13.80 13.85





#50

Acenaphthylene

Concen: 32.918 ng/ul

RT: 13.95 min Scan# 1880

Delta R.T. 0.00 min

Lab File: BM029759.D

Acq: 01 May 2021 23:03

Instrument :

BNA_M





ClientSampleId :

SLCS930

Tot Ion:	152	Resp:	451813
Ion Ratio	Lower	Upper	
152	100		
151	19.7	15.9	23.9
153	12.8	10.7	16.1

Manual Integrations
APPROVED

mohammad
5/3/2021 12:24:47 PM



#51

3-Nitroaniline

Concen: 28.189 ng/ul




RT: 14.14 min Scan# 1913


Delta R.T. -0.01 min

Lab File: BM029759.D

Acq: 01 May 2021 23:03

Tgt Ion:	138	Resp:	58662
Ion Ratio	Lower	Upper	
138	100		
108	8.8	7.5	11.3
92	105.5	87.4	131.0









#52
Acenaphthene
Concen: 34.114 ng/ul
RT: 14.29 min Scan# 1938
Delta R.T. -0.01 min
Lab File: BM029759.D
Acq: 01 May 2021 23:03

Instrument :
BNA_M
ClientSampleId :
SLCS930

Tot Ion	Ratio	Lower	Upper
153	100		
152	45.4	37.8	56.8
154	89.3	71.4	107.2




Manual Integrations
APPROVED


mohammad
5/3/2021 12:24:47 PM



#53
2,4-Dinitrophenol
Concen: 36.914 ng/ul
RT: 14.36 min Scan# 1950
Delta R.T. -0.01 min
Lab File: BM029759.D
Acq: 01 May 2021 23:03

Tgt Ion	Ratio	Lower	Upper
184	100		
63	45.4	42.2	63.4
154	55.8	43.0	64.4









#55
4-Nitrophenol
Concen: 41.794 ng/ul
RT: 14.46 min Scan# 1968
Delta R.T. -0.01 min
Lab File: BM029759.D
Acq: 01 May 2021 23:03

Instrument :
BNA_M
ClientSampled :
SLCS930

Tot Ion	Ratio	Resp	Lower	Upper
109	100			
139	142.3	96.5	144.7	
65	120.7	89.9	134.9	




Manual Integrations
APPROVED


mohammad
5/3/2021 12:24:47 PM



#56
Dibenzofuran
Concen: 35.212 ng/ul
RT: 14.63 min Scan# 1996
Delta R.T. -0.00 min
Lab File: BM029759.D
Acq: 01 May 2021 23:03

Tgt Ion	Ratio	Resp	Lower	Upper
168	100			
139	31.3	26.0	39.0	





#57

2,4-Dinitrotoluene

Concen: 38.104 ng/ul

RT: 14.60 min Scan# 1992

Delta R.T. -0.01 min

Lab File: BM029759.D

Acq: 01 May 2021 23:03

Instrument :

BNA_M





ClientSampled :

SLCS930

Tot Ion:	165	Resp:	124890
Ion Ratio	Lower	Upper	
165	100		
63	33.6	27.4	41.0
182	3.3	2.6	4.0

Manual Integrations
APPROVED

mohammad
5/3/2021 12:24:47 PM



#58

2,3,4,6-Tetrachlorophenol

Concen: 38.147 ng/ul




RT: 14.86 min Scan# 2035


Delta R.T. -0.01 min

Lab File: BM029759.D

Acq: 01 May 2021 23:03

Tgt Ion:	232	Resp:	128257
Ion Ratio	Lower	Upper	
232	100		
131	29.1	26.2	39.2
130	1.6	1.4	2.2
166	24.4	21.3	31.9





#59

Diethylphthalate

Concen: 35.816 ng/ul

RT: 15.06 min Scan# 2070

Delta R.T. -0.00 min

Lab File: BM029759.D

Acq: 01 May 2021 23:03

Instrument :

BNA_M


ClientSampleId :

SLCS930




Tot Ion:	149	Resp:	394314
Ion Ratio	Lower	Upper	
149	100		
177	25.6	19.8	29.8
150	12.2	9.8	14.6

Manual Integrations
APPROVED

mohammad
5/3/2021 12:24:47 PM



Abundance Ion 149.00 (148.70 to 149.70): E
Ion 177.00 (176.70 to 177.70): E
Ion 150.00 (149.70 to 150.70): E



#61

Fluorene

Concen: 35.277 ng/ul


RT: 15.27 min Scan# 2106

Delta R.T. -0.01 min



Lab File: BM029759.D


Acq: 01 May 2021 23:03

Tgt Ion:	166	Resp:	418574
Ion Ratio	Lower	Upper	
166	100		
165	97.4	77.3	115.9
167	13.7	10.4	15.6



Abundance Ion 166.00 (165.70 to 166.70): E
Ion 165.00 (164.70 to 165.70): E
Ion 167.00 (166.70 to 167.70): E









#62
4-Chlorophenyl-phenylether
Concen: 36.588 ng/ul
RT: 15.27 min Scan# 2106
Delta R.T. -0.00 min
Lab File: BM029759.D
Acq: 01 May 2021 23:03

Instrument :
BNA_M
ClientSampled :
SLCS930

Tot Ion	Ratio	Resp	Lower	Upper
204	100	242495		
206	32.8	26.6	40.0	
141	43.6	37.2	55.8	




Manual Integrations
APPROVED


mohammad
5/3/2021 12:24:47 PM



#63
4-Nitroaniline
Concen: 30.805 ng/ul m
RT: 15.31 min Scan# 2112
Delta R.T. -0.01 min
Lab File: BM029759.D
Acq: 01 May 2021 23:03

Tgt Ion	Ratio	Resp	Lower	Upper
138	100	68195		
92	49.9	34.0	51.0	
108	78.6	56.8	85.2	









#66
4,6-Dinitro-2-methylphenol
Concen: 35.731 ng/ul
RT: 15.37 min Scan# 2122
Delta R.T. -0.01 min
Lab File: BM029759.D
Acq: 01 May 2021 23:03

Instrument :
BNA_M
ClientSampleId :
SLCS930

Tot Ion	Ratio	Resp	Lower	Upper
198	100	84437		
182	4.3		4.6	6.8#
77	18.0		13.0	19.4




Manual Integrations
APPROVED


mohammad
5/3/2021 12:24:47 PM



#67
N-Nitrosodiphenylamine
Concen: 34.681 ng/ul
RT: 15.49 min Scan# 2143
Delta R.T. -0.00 min
Lab File: BM029759.D
Acq: 01 May 2021 23:03

Tgt Ion	Ratio	Resp	Lower	Upper
169	100	361295		
168	65.2		53.1	79.7
167	35.2		28.1	42.1






#68
4-Bromophenyl-phenylether
Concen: 37.341 ng/ul
RT: 16.17 min Scan# 2258
Delta R.T. -0.00 min
Lab File: BM029759.D
Acq: 01 May 2021 23:03

Instrument :
BNA_M
ClientSampleId :
SLCS930

Tot Ion	Ratio	Lower	Upper
248	100		
250	97.5	77.3	115.9
141	48.8	44.3	66.5

Manual Integrations
APPROVED

mohammad
5/3/2021 12:24:47 PM






Abundance

Ion 248.00 (247.70 to 248.70): E

Ion 250.00 (249.70 to 250.70): E

Ion 141.00 (140.70 to 141.70): E



#69
Hexachlorobenzene
Concen: 36.679 ng/ul
RT: 16.28 min Scan# 2277
Delta R.T. -0.00 min
Lab File: BM029759.D
Acq: 01 May 2021 23:03




Tgt Ion	Ratio	Lower	Upper
284	100		
142	24.9	21.1	31.7
249	31.1	25.8	38.8


Abundance

Ion 284.00 (283.70 to 284.70): E

Ion 142.00 (141.70 to 142.70): E

Ion 249.00 (248.70 to 249.70): E









#70
Atrazine
Concen: 36.154 ng/ul
RT: 16.45 min Scan# 2306
Delta R.T. -0.00 min
Lab File: BM029759.D
Acq: 01 May 2021 23:03

Instrument :
BNA_M
ClientSampled :
SLCS930

Tot Ion	Ratio	Resp	Lower	Upper
200	100	155189		
173	23.2	18.5	27.7	
215	50.0	39.6	59.4	




Manual Integrations
APPROVED


mohammad
5/3/2021 12:24:47 PM



#71
Pentachlorophenol
Concen: 39.906 ng/ul
RT: 16.63 min Scan# 2336
Delta R.T. -0.01 min
Lab File: BM029759.D
Acq: 01 May 2021 23:03

Tgt Ion	Ratio	Resp	Lower	Upper
266	100	100357		
264	63.2	51.1	76.7	
268	63.4	52.8	79.2	





#72

Phenanthrene

Concen: 35.604 ng/ul

RT: 17.01 min Scan# 2401

Delta R.T. -0.01 min

Lab File: BM029759.D

Acq: 01 May 2021 23:03

Instrument :

BNA_M





ClientSampleId :

SLCS930

Tot Ion:178	Resp:	688885
Ion Ratio	Lower	Upper
178	100	
179	15.2	12.3 18.5
176	18.5	15.1 22.7

Manual Integrations
APPROVED

mohammad
5/3/2021 12:24:47 PM



#74

Anthracene

Concen: 35.755 ng/ul




RT: 17.10 min Scan# 2417


Delta R.T. -0.00 min

Lab File: BM029759.D

Acq: 01 May 2021 23:03

Tgt Ion:178	Resp:	710915
Ion Ratio	Lower	Upper
178	100	
179	15.6	12.3 18.5
176	18.1	14.0 21.0









#75
 1,2,3,4-Tetrachlorobenzene
 Concen: 32.501 ng/uL
 RT: 13.02 min Scan# 1722
 Delta R.T. 0.00 min
 Lab File: BM029759.D
 Acq: 01 May 2021 23:03

Instrument :
 BNA_M
 ClientSampleId :
 SLCS930

Tot Ion	Ratio	Resp Lower	Upper
216	100		
214	78.4	63.8	95.6
218	47.3	39.1	58.7




Manual Integrations
 APPROVED


mohammad
 5/3/2021 12:24:47 PM



#76
 Pentachlorobenzene
 Concen: 34.128 ng/uL
 RT: 14.54 min Scan# 1982
 Delta R.T. -0.01 min
 Lab File: BM029759.D
 Acq: 01 May 2021 23:03

Tgt Ion	Ratio	Resp Lower	Upper
250	100		
248	62.4	48.5	72.7
252	64.2	52.0	78.0









#77
 Carbazole
 Concen: 33.543 ng/ul
 RT: 17.38 min Scan# 2464
 Delta R.T. -0.00 min
 Lab File: BM029759.D
 Acq: 01 May 2021 23:03

Instrument :
 BNA_M
 ClientSampled :
 SLCS930

Tot Ion:	167	Resp:	582043
Ion Ratio	Lower	Upper	
167	100		
166	21.4	17.3	25.9
139	10.5	8.7	13.1




Manual Integrations
 APPROVED


mohammad
 5/3/2021 12:24:47 PM



#78
 Di-n-butylphthalate
 Concen: 37.075 ng/ul
 RT: 17.94 min Scan# 2560
 Delta R.T. -0.00 min
 Lab File: BM029759.D
 Acq: 01 May 2021 23:03

Tgt Ion:	149	Resp:	683313
Ion Ratio	Lower	Upper	
149	100		
150	9.2	7.4	11.0
104	4.5	3.7	5.5









#80
 Fluoranthene
 Concen: 36.544 ng/ul
 RT: 19.03 min Scan# 2745
 Delta R.T. -0.00 min
 Lab File: BM029759.D
 Acq: 01 May 2021 23:03

Instrument :
 BNA_M
 ClientSampleId :
 SLCS930

Tot Ion	Ion	Ratio	Resp	Lower	Upper
202	100		882445		
101	6.1	5.6	8.4		
100	5.0	4.5	6.7		




Manual Integrations
 APPROVED


mohammad
 5/3/2021 12:24:47 PM



#82
 Pyrene
 Concen: 35.866 ng/ul
 RT: 19.39 min Scan# 2806
 Delta R.T. -0.01 min
 Lab File: BM029759.D
 Acq: 01 May 2021 23:03

Tgt Ion	Ion	Ratio	Resp	Lower	Upper
202	100		921574		
101	7.8	6.6	9.8		
100	6.8	5.5	8.3		









#83
Butylbenzylphthalate
Concen: 35.505 ng/ul
RT: 20.30 min Scan# 2961
Delta R.T. -0.00 min
Lab File: BM029759.D
Acq: 01 May 2021 23:03

Instrument :
BNA_M
ClientSampleId :
SLCS930

Tot Ion	Ion	Ratio	Resp	Lower	Upper
149	100				
91	58.6		49.4	74.0	
206	26.5		19.9	29.9	




Manual Integrations
APPROVED


mohammad
5/3/2021 12:24:47 PM



#84
3,3'-Dichlorobenzidine
Concen: 29.701 ng/ul
RT: 21.08 min Scan# 3093
Delta R.T. -0.00 min
Lab File: BM029759.D
Acq: 01 May 2021 23:03

Tgt Ion	Ion	Ratio	Resp	Lower	Upper
252	100				
254	65.4		51.0	76.6	
126	6.6		5.7	8.5	









#85
Benzo(a)anthracene
Concen: 35.695 ng/ul
RT: 21.14 min Scan# 3103
Delta R.T. -0.00 min
Lab File: BM029759.D
Acq: 01 May 2021 23:03

Instrument :
BNA_M
ClientSampleId :
SLCS930

Tot Ion	Ratio	Lower	Upper
228	100		
229	19.7	15.6	23.4
226	27.4	21.4	32.0




Manual Integrations
APPROVED


mohammad
5/3/2021 12:24:47 PM



#86
Bis(2-ethylhexyl)phthalate
Concen: 35.735 ng/ul
RT: 21.08 min Scan# 3093
Delta R.T. -0.00 min
Lab File: BM029759.D
Acq: 01 May 2021 23:03

Tgt Ion	Ratio	Lower	Upper
149	100		
167	29.2	23.3	34.9
279	5.9	4.6	6.8









#87
Chrysene
Concen: 35.503 ng/ul
RT: 21.19 min Scan# 3112
Delta R.T. -0.00 min
Lab File: BM029759.D
Acq: 01 May 2021 23:03

Instrument :
BNA_M
ClientSampled :
SLCS930

Tot Ion	Ion	Ratio	Resp	Lower	Upper
228	100				
226	29.4	23.8	35.8		
229	19.7	15.6	23.4		




Manual Integrations
APPROVED


mohammad
5/3/2021 12:24:47 PM



#89
Di-n-octyl phthalate
Concen: 34.295 ng/ul
RT: 21.94 min Scan# 3239
Delta R.T. 0.00 min
Lab File: BM029759.D
Acq: 01 May 2021 23:03

Tgt Ion:149 Resp: 837709





#90

Benzo(b)fluoranthene

Concen: 35.740 ng/ul

RT: 22.67 min Scan# 3364

Delta R.T. -0.00 min

Lab File: BM029759.D

Acq: 01 May 2021 23:03

Instrument :

BNA_M





ClientSampleId :

SLCS930

Tot Ion:	252	Resp:	991394
Ion Ratio	Lower	Upper	
252	100		
253	22.0	17.3	25.9
125	6.4	5.2	7.8

Manual Integrations
APPROVED

mohammad
5/3/2021 12:24:47 PM



#91

Benzo(k)fluoranthene

Concen: 36.161 ng/ul




RT: 22.72 min Scan# 3372


Delta R.T. -0.00 min

Lab File: BM029759.D

Acq: 01 May 2021 23:03

Tgt Ion:	252	Resp:	988053
Ion Ratio	Lower	Upper	
252	100		
253	21.8	17.0	25.6
125	6.1	5.0	7.4









#93
Benzo(a)pyrene
Concen: 35.542 ng/ul
RT: 23.23 min Scan# 3459
Delta R.T. 0.01 min
Lab File: BM029759.D
Acq: 01 May 2021 23:03

Instrument :
BNA_M
ClientSampleId :
SLCS930

Tot Ion	Ratio	Lower	Upper
252	100		
253	21.5	17.6	26.4
125	6.7	5.9	8.9




Manual Integrations
APPROVED


mohammad
5/3/2021 12:24:47 PM



#94
Indeno(1,2,3-cd)pyrene
Concen: 35.765 ng/ul
RT: 25.50 min Scan# 3844
Delta R.T. 0.01 min
Lab File: BM029759.D
Acq: 01 May 2021 23:03

Tgt Ion	Ratio	Lower	Upper
276	100		
138	17.0	13.7	20.5
277	24.9	20.4	30.6





#95

Dibenzo(a,h)anthracene

Concen: 35.751 ng/ul

RT: 25.50 min Scan# 3845

Delta R.T. -0.00 min

Lab File: BM029759.D

Acq: 01 May 2021 23:03

Instrument :

BNA_M





ClientSampled :

SLCS930

Tot Ion:	278	Resp:	980687
Ion Ratio	Lower	Upper	
278	100		
139	11.0	8.9	13.3
279	23.9	18.7	28.1

Manual Integrations
APPROVED

mohammad
5/3/2021 12:24:47 PM



#96

Benzo(g,h,i)perylene

Concen: 35.906 ng/ul

RT: 26.16 min Scan# 3957

Delta R.T. 0.01 min

Lab File: BM029759.D

Acq: 01 May 2021 23:03

Tgt Ion:	276	Resp:	969194
Ion Ratio	Lower	Upper	
276	100		
138	13.9	12.4	18.6
277	23.1	19.1	28.7

