

Data Path : Z:\SVOASRV\HPCHEM1\BNA M\DATA\BM050121\
 Data File : BM029760.D
 Acq On : 01 May 2021 23:39
 Operator : CG/JU
 Sample : PB135931BL
 Misc :
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 SBLK931

Quant Time: May 02 01:31:51 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_M\METHODS\SFAM-EPA-BM050121.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Sun May 02 01:03:03 2021
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.57	152	56821	20.00	ng/ul	0.00
20) Naphthalene-d8	10.35	136	206715	20.00	ng/ul	0.00
38) Acenaphthene-d10	14.22	164	157486	20.00	ng/ul	0.00
64) Phenanthrene-d10	16.97	188	353421	20.00	ng/ul	0.00
79) Chrysene-d12	21.16	240	407410	20.00	ng/ul	0.00
88) Perylene-d12	23.32	264	458422	20.00	ng/ul	0.00

System Monitoring Compounds

3) 1,4-Dioxane-d8	3.08	96	6389	5.50	ng/uL	0.00
4) Pyridine-d5	3.49	84	70645	28.07	ng/ul	0.00
7) Phenol-d5	6.76	99	106189	27.75	ng/ul	0.00
9) Bis-(2-Chloroethyl)ether-d	6.92	67	57728	28.81	ng/ul	0.00
11) 2-Chlorophenol-d4	7.11	132	104326	30.15	ng/ul	0.00
15) 4-Methylphenol-d8	8.29	113	91190	27.83	ng/ul	0.00
21) Nitrobenzene-d5	8.73	128	45306	30.10	ng/ul	0.00
24) 2-Nitrophenol-d4	9.45	143	58482	30.69	ng/ul	0.00
28) 2,4-Dichlorophenol-d3	9.99	165	113410	28.48	ng/ul	0.00
31) 4-Chloroaniline-d4	10.50	131	126972	28.28	ng/ul	0.00
46) Dimethylphthalate-d6	13.64	166	395850	31.73	ng/ul	0.00
49) Acenaphthylene-d8	13.92	160	467628	30.40	ng/ul	0.00
54) 4-Nitrophenol-d4	14.46	143	35565	21.55	ng/ul	0.00
60) Fluorene-d10	15.22	176	344660	31.94	ng/ul	0.00
65) 4,6-Dinitro-2-methylphenol	15.36	200	61450	23.91	ng/ul	0.00
73) Anthracene-d10	17.07	188	564320	31.68	ng/ul	0.00
81) Pyrene-d10	19.36	212	705680	33.92	ng/ul	0.00
92) Benzo(a)pyrene-d12	23.19	264	826862	31.44	ng/ul	0.00

Target Compounds Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA M\DATA\BM050121\
Data File : BM029760.D
Acq On : 01 May 2021 23:39
Operator : CG/JU
Sample : PB135931BL
Misc :
ALS Vial : 10 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
SBLK931

Quant Time: May 02 01:31:51 2021
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_M\METHODS\SFAM-EPA-BM050121.M
Quant Title : SVOA CALIBRATION
QLast Update : Sun May 02 01:03:03 2021
Response via : Initial Calibration

