

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM050420\
 Data File : BM025973.D
 Acq On : 04 May 2020 15:16
 Operator : MA/SJ
 Sample : SSTDCCC0.4EC
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 SSTD0.486

Quant Time: May 04 15:47:33 2020
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_M\METHODS\SOM-EPA-SIM-BM041520.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon May 04 14:41:11 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.56	152	1486	0.40	ng/ul	0.00
2) Naphthalene-d8	10.32	136	5103	0.40	ng/ul	0.00
6) Acenaphthene-d10	14.19	164	2933	0.40	ng/ul	0.00
10) Phenanthrene-d10	16.94	188	6824	0.40	ng/ul	0.00
16) Chrysene-d12	21.13	240	6324	0.40	ng/ul	0.00
20) Perylene-d12	23.28	264	6179	0.40	ng/ul	0.00

System Monitoring Compounds						
4) 2-Methylnaphthalene-d10	11.92	152	3231	0.38	ng/ul	0.00
14) Fluoranthene-d10	18.97	212	7898	0.40	ng/ul	0.00

Target Compounds						Qvalue
3) Naphthalene	10.37	128	6318	0.39	ng/ul	99
5) 2-Methylnaphthalene	11.99	142	4116	0.39	ng/ul	98
7) Acenaphthylene	13.91	152	5201	0.38	ng/ul	99
8) Acenaphthene	14.25	153	4541	0.39	ng/ul	99
9) Fluorene	15.25	166	5312	0.39	ng/ul	100
11) Pentachlorophenol	16.61	266	630	0.37	ng/ul	96
12) Phenanthrene	16.98	178	9053	0.39	ng/ul	100
13) Anthracene	17.08	178	7655	0.39	ng/ul	99
15) Fluoranthene	19.00	202	10686	0.39	ng/ul	99
17) Pyrene	19.37	202	10730	0.35	ng/ul	99
18) Benzo(a)anthracene	21.12	228	9643	0.40	ng/ul	98
19) Chrysene	21.17	228	11325	0.38	ng/ul	99
21) Benzo(b)fluoranthene	22.64	252	10195	0.36	ng/ul	98
22) Benzo(k)fluoranthene	22.69	252	11836	0.38	ng/ul	97
23) Benzo(a)pyrene	23.19	252	9140	0.39	ng/ul	97
24) Indeno(1,2,3-cd)pyrene	25.40	276	11636	0.45	ng/ul#	97
25) Dibenzo(a,h)anthracene	25.41	278	9613	0.48	ng/ul	92
26) Benzo(g,h,i)perylene	26.04	276	9875	0.41	ng/ul	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM050420\
Data File : BM025973.D
Acq On : 04 May 2020 15:16
Operator : MA/SJ
Sample : SSTDCCC0.4EC
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
SSTD0.486

Quant Time: May 04 15:47:33 2020
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_M\METHODS\SOM-EPA-SIM-BM041520.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon May 04 14:41:11 2020
Response via : Initial Calibration

