


Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM050420\
 Data File : BM025969.D
 Acq On : 04 May 2020 10:22
 Operator : MA/SJ
 Sample : DFTPP02
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 DFTPP02

Integration File: rteint.p

Method : Z:\SVOASRV\HPCHEM1\BNA_M\METHODS\SOM-EPA-SIM-BM041520.M
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 Last Update : Mon May 04 14:41:11 2020



AutoFind: Scans 2502, 2503, 2504; Background Corrected with Scan 2496

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	51.0	69408	PASS
68	69	0.00	2	1.9	1204	PASS
69	198	0.00	100	47.1	64188	PASS
70	69	0.00	2	0.7	431	PASS
127	198	10	80	49.0	66709	PASS
197	198	0.00	2	0.8	1075	PASS
198	198	100	100	100.0	136138	PASS
199	198	5	9	7.0	9547	PASS
275	198	10	60	26.7	36306	PASS
365	198	1	100	3.4	4565	PASS
441	198	0.01	100	10.8	14744	PASS
442	198	50	100	73.7	100298	PASS
443	442	15	24	19.8	19854	PASS

m/z	Abundance
55.00	351.0
57.00	411.0
66.90	158.0
69.20	262.0
71.20	267.0
73.10	532.0
80.90	168.0
83.10	186.0
84.00	206.0
207.00	506.0
281.00	163.0

Instrument :
BNA_M
ClientSampleId :
DFTPP02

m/z	Abundance
50.10	15430.0
51.10	57376.0
52.10	3160.0
53.10	187.0
54.90	713.0
56.00	1640.0
57.00	4439.0
57.90	288.0
59.90	217.0
61.00	606.0
62.00	1023.0
63.00	2427.0
64.00	379.0
65.00	930.0
66.00	185.0
67.00	233.0
68.10	1079.0
69.00	51616.0
70.00	352.0
71.00	228.0
73.10	810.0
74.10	6046.0
75.00	8448.0
76.10	2666.0
77.10	54256.0
78.10	3829.0
79.00	3889.0
80.00	3013.0
81.00	4247.0
82.00	1293.0
83.00	1198.0
84.00	187.0
85.00	731.0
86.00	1277.0
87.00	495.0
87.90	213.0
91.00	1217.0
92.10	1102.0
93.00	7264.0
94.10	379.0
95.00	169.0
95.90	226.0
97.10	422.0
98.00	4989.0
99.10	4524.0
100.00	585.0
101.00	1856.0
103.00	628.0
104.00	1354.0
105.00	1116.0
106.00	452.0
107.00	16132.0
108.00	2825.0
109.10	636.0
110.00	27232.0
111.00	4631.0
112.00	510.0
113.20	194.0
116.00	1065.0
117.00	13922.0
118.00	1162.0
119.80	172.0
122.00	985.0
123.00	1677.0
123.90	711.0
125.00	716.0
126.20	442.0
127.00	51128.0
128.00	3717.0
129.00	21104.0
130.00	1902.0
130.90	391.0
132.00	279.0
132.90	335.0
133.90	664.0
135.00	1921.0
136.00	819.0
137.10	1005.0
139.90	431.0
141.00	2819.0
142.00	1041.0
142.90	551.0
143.90	201.0
146.00	745.0
147.00	1669.0
148.00	3376.0
149.10	730.0
149.90	286.0
151.20	291.0
151.90	240.0
153.00	778.0
154.00	552.0

Instrument :
BNA_M
ClientSampleId :
DFTPP02

155.00	1258.0
156.00	2225.0
156.90	333.0
158.00	611.0
159.00	383.0
160.00	1017.0
161.00	1330.0
162.00	323.0
162.90	175.0
163.90	218.0
165.00	1364.0
165.90	903.0
167.00	5812.0
168.00	3553.0
169.10	856.0
170.10	172.0
170.90	233.0
171.90	592.0
173.00	750.0
174.00	1422.0
175.00	2275.0
175.90	604.0
177.00	1334.0
178.10	318.0
179.00	4397.0
180.00	2820.0
181.00	1251.0
182.10	155.0
183.20	172.0
183.90	361.0
185.00	1564.0
186.00	14082.0
187.00	4269.0
188.00	460.0
188.90	763.0
189.90	222.0
191.00	503.0
191.90	1199.0
193.00	1406.0
193.90	363.0
194.90	164.0
196.00	2126.0
197.00	798.0
198.00	92320.0
199.00	7174.0
200.00	632.0
201.70	428.0
202.90	871.0
204.00	3602.0
205.00	6430.0
206.10	26808.0
207.00	4025.0
207.90	789.0
209.10	507.0
210.10	447.0
211.10	1200.0
211.80	219.0
215.00	260.0
216.00	707.0
217.00	7253.0
218.00	1183.0
219.90	169.0
221.00	3479.0
222.00	1053.0
222.90	1388.0
224.00	13650.0
225.00	3707.0
226.00	429.0
227.00	6038.0
228.00	901.0
229.00	1485.0
230.00	184.0
231.00	482.0
233.10	215.0
234.00	377.0
234.90	388.0
235.90	416.0
237.00	991.0
239.00	371.0
240.90	490.0
242.00	710.0
242.90	516.0
244.10	10134.0
245.10	1554.0
245.90	2612.0
246.80	517.0
249.00	442.0
253.10	219.0
254.10	452.0
255.00	54704.0
256.00	8220.0
257.00	1070.0
258.00	3511.0
259.00	561.0
265.00	1559.0
266.10	228.0

Instrument :
BNA_M
ClientSampleId :
DFTPP02

271.10	154.0
271.90	181.0
273.00	1451.0
274.00	4189.0
275.00	23968.0
275.90	3049.0
277.00	2176.0
277.90	411.0
280.90	274.0
282.90	267.0
284.10	166.0
285.00	345.0
290.00	170.0
292.90	495.0
294.00	215.0
295.10	180.0
296.00	6456.0
296.90	808.0
303.00	867.0
304.10	210.0
313.90	450.0
315.00	673.0
316.00	304.0
321.10	191.0
323.00	1771.0
324.00	523.0
326.90	420.0
328.00	194.0
332.00	179.0
334.00	1228.0
335.10	237.0
341.00	345.0
345.90	532.0
352.00	683.0
353.10	395.0
354.00	765.0
355.10	151.0
365.00	2835.0
366.00	357.0
370.90	182.0
372.00	825.0
373.00	501.0
383.10	411.0
390.00	157.0
402.10	414.0
402.90	464.0
403.10	463.0
404.20	175.0
420.90	338.0
422.10	444.0
423.10	2997.0
424.00	1268.0
424.90	187.0
441.00	7551.0
442.00	48944.0
443.10	9771.0
444.10	1175.0

Instrument :
BNA_M
ClientSampleId :
DFTPP02

m/z	Abundance
50.10	23320.0
51.10	84688.0
52.10	4689.0
55.10	1188.0
56.00	2826.0
57.10	6093.0
57.90	241.0
60.10	180.0
61.00	897.0
62.00	1279.0
63.10	3530.0
64.00	447.0
65.10	1617.0
67.00	351.0
68.10	1369.0
69.00	79128.0
70.00	630.0
71.10	270.0
72.00	172.0
73.00	883.0
74.00	7034.0
75.10	12228.0
76.10	4443.0
77.10	82288.0
78.10	5502.0
79.00	6245.0
80.00	4690.0
81.00	6467.0
82.00	1656.0
83.00	1304.0
84.00	406.0
85.00	1220.0
86.00	2105.0
86.90	913.0
88.00	505.0
89.10	226.0
91.00	1541.0
92.00	1762.0
93.00	10813.0
93.90	670.0
95.00	345.0
96.10	435.0
97.20	505.0
98.00	8192.0
99.00	6459.0
100.10	682.0
101.00	3327.0
101.90	221.0
103.00	1217.0
104.00	1898.0
105.00	1991.0
106.00	854.0
107.00	25920.0
108.00	3861.0
109.10	735.0
110.00	42848.0
111.00	7364.0
112.00	1165.0
113.00	448.0
114.80	248.0
115.20	170.0
116.00	1603.0
117.00	23520.0
118.00	1896.0
118.90	255.0
120.10	292.0
120.90	262.0
122.10	1559.0
123.10	2660.0
124.00	1305.0
125.00	980.0
126.00	413.0
127.00	80168.0
128.00	6072.0
129.00	31640.0
130.00	3092.0
130.90	505.0
132.00	215.0
133.00	217.0
134.00	1200.0
135.00	2802.0
136.00	1208.0
137.00	1710.0
138.00	440.0
140.10	391.0
141.00	4403.0
142.10	1472.0
143.00	1090.0
144.00	267.0
145.90	692.0
146.90	2610.0
148.00	5537.0

Instrument :
BNA_M
ClientSampleId :
DFTPP02

149.10	1054.0
149.90	260.0
150.90	676.0
151.80	231.0
152.00	230.0
153.00	1729.0
154.00	1415.0
155.10	2511.0
156.10	3858.0
156.90	684.0
157.90	694.0
159.00	629.0
160.00	1145.0
161.00	1846.0
162.00	473.0
164.90	1703.0
166.00	1251.0
167.00	8922.0
168.00	5192.0
169.00	1066.0
170.00	429.0
170.90	498.0
171.90	896.0
173.00	771.0
174.00	1841.0
175.00	3547.0
175.90	684.0
177.00	1651.0
178.10	651.0
179.00	6220.0
180.00	4694.0
181.00	2418.0
182.10	349.0
183.00	238.0
183.90	510.0
185.00	3807.0
186.00	24120.0
187.00	7372.0
188.00	739.0
189.00	1552.0
190.00	352.0
191.10	798.0
192.00	2050.0
193.10	2529.0
194.00	570.0
194.90	287.0
196.10	4232.0
197.10	1198.0
198.00	163392.0
199.00	10942.0
199.90	982.0
201.30	616.0
203.00	1079.0
204.00	6471.0
205.00	10962.0
206.00	44152.0
207.00	6093.0
208.00	1766.0
209.00	694.0
210.00	754.0
211.00	1821.0
214.90	770.0
216.00	1022.0
217.00	12150.0
217.90	1289.0
219.00	174.0
221.10	6216.0
221.80	1460.0
223.00	2675.0
224.00	24744.0
225.00	6639.0
226.00	890.0
227.00	11613.0
228.00	1691.0
229.00	2382.0
229.90	183.0
230.90	755.0
233.10	213.0
234.00	938.0
234.90	717.0
235.80	362.0
237.00	1402.0
237.90	161.0
239.00	410.0
239.80	289.0
241.00	496.0
241.90	1149.0
243.10	1654.0
244.10	19312.0
245.10	3100.0
246.00	4489.0
246.90	991.0
247.80	172.0
249.00	650.0
251.80	197.0
253.00	554.0

Instrument :
BNA_M
ClientSampleId :
DFTPP02

254.10	519.0
255.00	93960.0
256.00	15027.0
257.00	1056.0
258.00	6679.0
259.00	881.0
260.00	159.0
264.00	168.0
265.00	2381.0
266.00	502.0
271.00	428.0
272.00	261.0
273.00	2372.0
274.00	7588.0
275.00	40984.0
276.00	5477.0
277.10	3700.0
278.00	1014.0
281.00	197.0
283.10	580.0
284.00	374.0
285.00	884.0
291.90	186.0
293.00	1140.0
293.80	256.0
295.00	327.0
296.00	12928.0
297.00	1946.0
302.00	214.0
303.00	1608.0
303.90	545.0
309.00	173.0
313.90	561.0
315.00	1184.0
316.10	836.0
317.10	198.0
320.90	354.0
321.80	186.0
323.10	3848.0
324.10	674.0
325.00	156.0
326.90	558.0
328.00	452.0
332.00	368.0
333.10	360.0
334.00	2275.0
335.00	784.0
340.90	536.0
342.10	150.0
346.00	635.0
352.10	1109.0
353.00	785.0
354.00	1360.0
355.10	279.0
365.00	5182.0
366.00	1023.0
370.10	208.0
371.10	268.0
372.00	2015.0
373.00	491.0
383.00	507.0
384.10	179.0
389.90	330.0
400.80	175.0
401.90	813.0
403.10	1042.0
404.10	480.0
421.10	919.0
422.00	970.0
423.00	6001.0
424.00	2427.0
425.20	351.0
441.10	16299.0
442.00	113008.0
443.00	22584.0
444.10	2158.0

Instrument :
BNA_M
ClientSampleId :
DFTPP02

m/z	Abundance
50.10	17360.0
51.10	66160.0
52.10	3430.0
53.00	226.0
55.10	699.0
56.00	1621.0
57.10	4880.0
58.10	254.0
58.90	156.0
59.90	198.0
61.00	917.0
62.00	1109.0
63.00	2527.0
64.00	320.0
65.00	1248.0
67.10	201.0
68.10	1164.0
69.00	62608.0
70.00	312.0
71.10	390.0
72.00	156.0
73.10	1272.0
74.00	6268.0
75.10	9638.0
76.10	3273.0
77.10	65168.0
78.10	4488.0
79.00	4574.0
80.00	3918.0
81.00	5368.0
82.00	1141.0
83.00	1078.0
83.90	274.0
85.00	1123.0
86.00	1713.0
87.00	710.0
87.80	336.0
89.00	156.0
91.00	1191.0
92.00	1353.0
93.00	8230.0
94.00	612.0
95.20	339.0
96.00	372.0
97.00	237.0
98.00	6520.0
99.00	5292.0
99.90	523.0
101.00	2451.0
102.90	984.0
104.00	1797.0
105.00	1717.0
106.10	802.0
107.00	20376.0
108.00	3199.0
109.00	783.0
110.00	35664.0
111.00	5467.0
112.00	817.0
113.20	291.0
114.90	182.0
115.90	1390.0
117.00	19352.0
118.00	1473.0
119.00	225.0
120.00	514.0
122.00	1357.0
123.00	2650.0
124.00	967.0
125.10	941.0
126.10	441.0
127.00	68832.0
128.10	5335.0
129.00	28416.0
130.00	2756.0
131.00	369.0
131.90	203.0
132.90	340.0
134.00	719.0
134.90	2542.0
136.00	980.0
137.00	1623.0
137.80	266.0
138.10	275.0
139.10	177.0
140.10	271.0
141.00	3768.0
142.10	1171.0
143.00	968.0
144.10	267.0
145.00	260.0
146.00	511.0

Instrument :
BNA_M
ClientSampleId :
DFTPP02

147.00	2353.0
148.00	4391.0
149.10	914.0
149.90	166.0
151.20	517.0
152.00	260.0
152.90	1132.0
153.90	1095.0
154.90	1823.0
156.10	2809.0
156.90	648.0
158.10	762.0
158.90	573.0
159.90	1204.0
161.00	2360.0
161.90	476.0
163.00	159.0
164.00	165.0
165.00	1254.0
166.00	1018.0
167.00	7895.0
168.00	4620.0
169.00	844.0
170.00	330.0
170.90	396.0
171.90	1040.0
172.90	1048.0
174.00	1625.0
175.00	2890.0
176.10	635.0
177.00	1424.0
177.90	472.0
179.00	5494.0
180.00	4406.0
181.00	1679.0
182.00	458.0
183.90	626.0
185.00	2850.0
186.00	20632.0
187.00	5908.0
188.00	888.0
189.00	2107.0
190.00	236.0
191.10	665.0
192.00	1939.0
193.00	2041.0
194.00	410.0
194.90	284.0
196.00	3837.0
197.00	1231.0
198.00	152704.0
199.00	10527.0
199.90	635.0
200.80	272.0
201.40	803.0
202.90	1440.0
204.00	6796.0
205.00	9545.0
206.00	42096.0
207.10	6109.0
208.00	1700.0
209.00	533.0
210.00	915.0
211.10	1836.0
211.90	442.0
212.90	229.0
214.90	733.0
216.00	1120.0
216.90	11052.0
218.00	1577.0
219.00	169.0
221.10	6475.0
221.90	1445.0
223.00	2660.0
224.00	24272.0
225.10	6167.0
226.10	757.0
227.00	11593.0
228.00	1593.0
229.00	2347.0
230.00	322.0
231.10	787.0
233.00	164.0
234.00	582.0
235.00	748.0
236.00	365.0
237.00	881.0
237.90	230.0
238.90	573.0
240.00	247.0
241.00	717.0
242.00	1078.0
243.00	1207.0
244.10	18192.0
245.00	2313.0
246.00	3915.0

Instrument :
BNA_M
ClientSampleId :
DFTPP02

246.90	625.0
248.10	175.0
249.00	538.0
250.90	266.0
252.20	192.0
253.00	338.0
254.20	759.0
255.00	97824.0
256.00	14575.0
257.00	1058.0
258.00	5948.0
259.00	1153.0
259.90	273.0
263.90	162.0
264.20	183.0
265.00	2527.0
266.00	417.0
269.80	157.0
270.80	291.0
271.90	219.0
273.00	2672.0
274.00	7609.0
275.00	43968.0
276.00	5401.0
277.00	3925.0
277.90	762.0
281.90	157.0
282.90	464.0
284.10	263.0
284.90	849.0
290.00	168.0
292.10	187.0
293.00	1044.0
294.00	384.0
295.10	333.0
296.00	14640.0
297.00	1649.0
301.10	191.0
302.00	308.0
303.00	1668.0
303.90	367.0
308.10	254.0
314.00	725.0
315.00	1900.0
315.90	877.0
321.10	422.0
323.00	3650.0
324.00	614.0
327.10	785.0
328.10	296.0
332.10	378.0
333.00	479.0
334.00	2610.0
334.90	780.0
341.00	622.0
346.00	762.0
352.00	1439.0
353.00	893.0
354.00	1320.0
355.00	365.0
365.00	5679.0
365.90	738.0
369.90	163.0
371.00	364.0
372.00	2042.0
373.10	547.0
382.90	648.0
384.00	227.0
390.00	260.0
391.10	164.0
402.00	888.0
403.10	1387.0
403.90	524.0
421.00	1141.0
422.10	1073.0
423.10	7848.0
424.00	2334.0
425.00	388.0
439.80	158.0
441.00	20384.0
442.10	138944.0
443.10	27208.0
444.00	2395.0
444.90	154.0

Instrument :
BNA_M
ClientSampleId :
DFTPP02