


Data Path : Z:\SVOASRV\HPCHEM1\BNA M\DATA\BM051119\
 Data File : BM020267.D
 Acq On : 11 May 2019 11:17
 Operator : JU/SJ
 Sample : DFTPP
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 DFTPP

Integration File: rteint.p

Method : Z:\SVOASRV\HPCHEM1\BNA_M\METHODS\8270-BM043019.M
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 Last Update : Wed May 01 03:58:37 2019



AutoFind: Scans 2480, 2481, 2482; Background Corrected with Scan 2475

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	28.2	11234	PASS
68	69	0.00	2	1.7	342	PASS
69	198	0.00	100	50.7	20192	PASS
70	69	0.00	2	0.5	110	PASS
127	198	10	80	45.6	18186	PASS
197	198	0.00	2	0.0	0	PASS
198	198	100	100	100.0	39848	PASS
199	198	5	9	7.2	2876	PASS
275	198	10	60	29.4	11716	PASS
365	198	1	100	3.7	1456	PASS
441	198	0.01	100	12.7	5064	PASS
442	198	50	100	84.0	33456	PASS
443	442	15	24	18.9	6333	PASS