

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM051222\
 Data File : BM035026.D
 Acq On : 12 May 2022 20:40
 Operator : CG/JU
 Sample : SP5898
 Misc : SFAM SIM SURROGATE
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 SP5898

Quant Time: May 13 04:20:27 2022
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_M\METHODS\SFAM-EPA-SIM-BM051222.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu May 12 18:15:39 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.921	152	3583	0.400	ng/ul	0.00
4) Naphthalene-d8	10.721	136	10492	0.400	ng/ul	# 0.00
9) Acenaphthene-d10	14.552	164	6561	0.400	ng/ul	0.00
13) Phenanthrene-d10	17.296	188	14247	0.400	ng/ul	# 0.00
17) Chrysene-d12	21.473	240	11260	0.400	ng/ul	# 0.00
23) Perylene-d12	23.859	264	10441	0.400	ng/ul	# 0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.356	96	7219	1.682	ng/ul	0.00
6) 2-Methylnaphthalene-d10	12.310	152	13572	0.811	ng/ul	0.00
18) Fluoranthene-d10	19.322	212	31503	0.816	ng/ul	0.00

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM051222\
Data File : BM035026.D
Acq On : 12 May 2022 20:40
Operator : CG/JU
Sample : SP5898
Misc : SFAM SIM SURROGATE
ALS Vial : 12 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
SP5898

Quant Time: May 13 04:20:27 2022
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_M\METHODS\SFAM-EPA-SIM-BM051222.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu May 12 18:15:39 2022
Response via : Initial Calibration

