

Data Path : Z:\SVOASRV\HPCHEM1\BNA M\DATA\BM051319\
 Data File : BM020326.D
 Acq On : 13 May 2019 21:09
 Operator : JU/SJ
 Sample : K2769-03
 Misc :
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 P026-SS017-0612-01

Quant Time: May 14 03:50:55 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA M\METHODS\8270-BM043019.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed May 01 03:58:37 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	0.00	152	0	0.00	ng	-7.79
21) Naphthalene-d8	0.00	136	0	0.00	ng	-10.59
39) Acenaphthene-d10	0.00	164	0	0.00	ng	-14.43
64) Phenanthrene-d10	0.00	188	0	0.00	ng	-17.18
76) Chrysene-d12	21.36	240	412	20.00	ng	0.00
87) Perylene-d12	0.00	264	0	0.00	ng	-23.66
System Monitoring Compounds						
5) 2-Fluorophenol	0.00	112	0	0.00	ng	
7) Phenol-d6	0.00	99	0	0.00	ng	
23) Nitrobenzene-d5	0.00	82	0	0.00	ng	
42) 2,4,6-Tribromophenol	0.00	330	0	0.00	ng	
45) 2-Fluorobiphenyl	0.00	172	0	0.00	ng	
79) Terphenyl-d14	19.80	244	1205	51.01	ng	-0.01
Target Compounds						
78) Pyrene	19.59	202	113	4.085	ng	# 58
80) Butylbenzylphthalate	20.51	149	109	10.836	ng	# 19
81) Benzo(a)anthracene	21.34	228	448	15.505	ng	# 52
83) Chrysene	21.39	228	812	28.977	ng	# 50
84) Bis(2-ethylhexyl)phthalate	21.25	149	1371	87.834	ng	# 53
85) Di-n-octyl phthalate	22.13	149	116	4.471	ng	# 1

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA M\DATA\BM051319\
Data File : BM020326.D
Acq On : 13 May 2019 21:09
Operator : JU/SJ
Sample : K2769-03
Misc :
ALS Vial : 17 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
P026-SS017-0612-01

Quant Time: May 14 03:50:55 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA M\METHODS\8270-BM043019.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed May 01 03:58:37 2019
Response via : Initial Calibration

