


Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM051719\
 Data File : BM020354.D
 Acq On : 17 May 2019 08:33
 Operator : JU/SJ
 Sample : DFTPP68
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 DFTPP68

Integration File: LSCINT.P

Method : Z:\SVOASRV\HPCHEM1\BNA_M\METHODS\SOM-EPA-BM051619MA.M
 Title : SVOA CALIBRATION
 Last Update : Sat May 18 05:13:49 2019



AutoFind: Scans 2477, 2478, 2479; Background Corrected with Scan 2471

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	25.0	65192	PASS
68	69	0.00	2	1.8	2151	PASS
69	198	0.00	100	45.6	118850	PASS
70	69	0.00	2	0.5	566	PASS
127	198	10	80	40.9	106522	PASS
197	198	0.00	2	0.7	1766	PASS
198	198	100	100	100.0	260586	PASS
199	198	5	9	6.6	17307	PASS
275	198	10	60	31.8	82776	PASS
365	198	1	100	4.4	11530	PASS
441	198	0.01	100	14.7	38200	PASS
442	198	50	100	99.3	258752	PASS
443	442	15	24	19.0	49256	PASS

Scan 2471 (17.521 min): BM020354.D
DFTPP68

m/z	Abundance
39.90	302.0
44.00	2383.0
207.00	548.0

Instrument :
BNA_M
ClientSampleId :
DFTPP68

m/z	Abundance
37.00	455.0
38.10	1407.0
39.10	6047.0
40.00	625.0
44.00	2868.0
49.10	673.0
50.10	22808.0
51.10	82192.0
52.10	4354.0
55.00	583.0
56.00	2981.0
57.00	6326.0
58.00	447.0
61.00	1557.0
62.00	1782.0
63.00	4801.0
64.00	762.0
65.00	2370.0
68.00	2626.0
69.00	145792.0
70.00	719.0
73.00	1515.0
74.00	16031.0
75.00	23384.0
76.10	7252.0
77.10	143680.0
78.00	9843.0
79.00	11499.0
80.00	7986.0
81.00	10884.0
81.90	2873.0
82.90	2097.0
84.00	482.0
85.00	1912.0
86.00	3283.0
87.00	1245.0
88.00	542.0
91.00	2737.0
92.00	3012.0
93.00	18376.0
94.00	948.0
96.10	558.0
98.00	15162.0
99.00	10053.0
99.90	920.0
101.00	6419.0
101.90	309.0
103.10	2028.0
104.00	3861.0
105.00	3665.0
106.00	972.0
107.00	44240.0
108.00	5768.0
109.00	1056.0
110.00	68664.0
111.00	10900.0
112.00	1642.0
113.00	424.0
116.00	2204.0
117.00	40576.0
118.00	3135.0
118.90	414.0
121.90	2731.0
123.00	4616.0
124.00	2197.0
125.00	1847.0
126.20	906.0
127.00	126520.0
128.00	9691.0
129.00	53544.0
130.00	4111.0
131.00	903.0
132.20	358.0
134.00	1487.0
135.00	4258.0
136.00	1442.0
136.90	2040.0
138.00	565.0
138.90	355.0
140.00	740.0
141.00	7713.0
142.00	2164.0
143.00	1501.0
143.90	575.0
144.80	368.0
146.00	1450.0
146.90	4276.0
148.00	9362.0
149.00	2145.0
149.90	542.0
151.00	829.0

Instrument :
BNA_M
ClientSampleId :
DFTPP68

151.90	442.0
153.00	2572.0
154.00	1988.0
155.00	4545.0
156.00	6122.0
156.90	1310.0
158.00	1682.0
158.90	852.0
159.90	2313.0
161.00	3622.0
161.90	837.0
163.90	744.0
165.00	2911.0
166.00	2212.0
167.00	15150.0
168.00	8149.0
169.00	1286.0
170.10	645.0
171.10	671.0
171.90	1230.0
173.00	1813.0
174.00	2936.0
175.00	6404.0
176.00	1144.0
176.90	3161.0
177.90	1237.0
178.90	12208.0
180.00	7683.0
181.00	3504.0
181.80	453.0
182.80	314.0
183.90	1086.0
185.00	5950.0
186.00	41392.0
187.00	12894.0
187.90	1423.0
189.00	3173.0
190.00	532.0
191.00	1182.0
192.00	4204.0
193.00	4487.0
194.00	958.0
194.90	510.0
196.00	8679.0
197.10	2057.0
197.90	289024.0
198.90	19336.0
199.90	1845.0
201.30	1049.0
202.90	2567.0
204.00	12431.0
205.00	20080.0
206.00	82080.0
207.00	12330.0
207.90	2890.0
209.00	990.0
210.30	1448.0
211.00	3316.0
211.90	731.0
215.00	889.0
216.00	1904.0
216.90	23704.0
218.00	3125.0
219.00	346.0
221.00	12004.0
222.00	3170.0
223.00	5297.0
224.00	46624.0
225.00	12016.0
226.00	1438.0
227.00	22056.0
228.00	2614.0
229.00	4145.0
230.00	503.0
231.00	1670.0
232.90	350.0
233.90	1395.0
234.90	1833.0
235.90	1013.0
237.00	1986.0
239.00	916.0
239.90	518.0
241.00	1158.0
241.90	2958.0
243.00	2713.0
244.00	34592.0
245.00	4816.0
246.00	6854.0
246.90	1755.0
248.90	1230.0
249.80	407.0
251.00	445.0
251.90	340.0
252.90	604.0
254.10	1277.0

Instrument :
BNA_M
ClientSampleId :
DFTPP68

255.00	192640.0
256.00	28264.0
257.00	2671.0
258.00	11602.0
258.90	1660.0
260.90	329.0
265.00	4926.0
265.90	894.0
270.90	745.0
271.90	551.0
273.00	5213.0
274.00	13250.0
275.00	85136.0
276.00	10948.0
277.00	7747.0
278.00	1016.0
282.90	806.0
283.90	689.0
284.90	1057.0
291.80	378.0
293.00	1752.0
293.90	441.0
295.00	455.0
296.00	24520.0
296.90	3492.0
297.90	353.0
300.90	341.0
301.90	460.0
303.00	2720.0
303.90	790.0
307.90	317.0
309.90	311.0
314.00	1173.0
314.90	3129.0
316.00	1624.0
317.00	334.0
321.10	580.0
322.10	391.0
323.10	7419.0
324.00	1508.0
326.90	1458.0
328.10	669.0
331.90	500.0
332.90	658.0
334.00	4404.0
335.00	1076.0
340.90	1046.0
346.00	1383.0
346.90	364.0
352.00	2449.0
353.10	1462.0
354.00	2404.0
355.10	405.0
365.00	10835.0
365.90	1971.0
371.00	524.0
372.10	3529.0
373.00	654.0
382.90	989.0
384.00	342.0
390.00	398.0
402.10	1197.0
403.00	1993.0
404.00	778.0
421.00	1371.0
422.00	1869.0
423.00	11115.0
424.10	2985.0
425.00	345.0
441.00	32640.0
442.00	218240.0
443.00	42056.0
444.00	4129.0
445.10	394.0

Instrument :
BNA_M
ClientSampleId :
DFTPP68

m/z	Abundance
37.10	536.0
38.10	1243.0
39.10	6219.0
40.00	541.0
44.00	3045.0
49.10	698.0
50.10	19888.0
51.00	76192.0
52.00	3594.0
54.90	513.0
56.00	3272.0
57.00	6075.0
57.90	356.0
60.90	1593.0
62.00	2053.0
63.00	4493.0
64.10	623.0
65.00	2129.0
68.10	2502.0
69.00	139392.0
70.10	676.0
73.00	1655.0
74.00	14997.0
75.00	22624.0
76.10	7021.0
77.00	138496.0
78.00	9110.0
79.00	10544.0
80.00	7898.0
81.00	11047.0
82.00	2637.0
83.00	2420.0
85.00	1713.0
86.00	3168.0
86.90	1363.0
87.90	527.0
91.00	2595.0
91.90	2586.0
93.00	18824.0
94.00	1327.0
96.00	818.0
98.00	14197.0
99.00	9335.0
100.00	923.0
101.00	5950.0
102.10	305.0
103.00	2258.0
104.00	3587.0
105.00	3922.0
106.10	1020.0
107.00	42304.0
108.10	5633.0
109.10	1407.0
110.00	69320.0
111.00	10899.0
112.00	1093.0
113.00	557.0
115.90	2250.0
117.00	40008.0
118.00	2803.0
118.90	312.0
120.00	465.0
122.00	2898.0
123.00	4322.0
124.00	2057.0
125.00	1618.0
127.00	127592.0
128.00	10179.0
129.00	54352.0
130.00	4942.0
131.00	1112.0
132.00	418.0
133.10	307.0
133.90	1443.0
135.00	4336.0
135.90	2053.0
137.00	2138.0
137.90	379.0
138.90	410.0
140.00	750.0
141.00	8123.0
142.00	2717.0
143.00	1806.0
146.00	1674.0
147.00	4593.0
148.00	9275.0
149.00	1727.0
150.00	564.0
151.10	673.0
151.90	551.0
152.90	2617.0

Instrument :
BNA_M
ClientSampleId :
DFTPP68

153.90	2064.0
155.00	4708.0
156.00	6594.0
157.00	1313.0
157.90	1547.0
159.00	1157.0
160.00	2936.0
161.00	3696.0
161.90	1268.0
164.00	418.0
164.90	3607.0
166.00	2670.0
167.00	16584.0
168.00	7772.0
169.00	1653.0
169.90	645.0
171.00	821.0
172.00	1581.0
173.00	2398.0
173.90	3460.0
175.00	6919.0
176.10	1532.0
177.00	3143.0
177.90	1058.0
179.00	12619.0
180.00	8636.0
181.00	4087.0
182.10	758.0
183.00	361.0
184.00	1087.0
185.00	6323.0
186.00	46464.0
187.00	14112.0
188.00	1475.0
188.90	3344.0
189.80	432.0
191.00	1485.0
192.00	4038.0
193.00	4853.0
194.00	993.0
195.00	555.0
196.00	10101.0
197.10	1934.0
197.90	311744.0
198.90	20680.0
200.00	1712.0
201.50	1569.0
202.90	2746.0
204.00	13538.0
205.00	22272.0
206.00	90384.0
207.00	13295.0
208.00	3685.0
209.00	1253.0
210.00	1552.0
211.00	3298.0
212.00	568.0
213.00	320.0
214.90	1091.0
215.90	1997.0
216.90	25760.0
218.00	3392.0
221.00	14446.0
221.90	3714.0
223.00	5310.0
224.00	51224.0
225.00	13217.0
226.10	1643.0
227.00	22744.0
228.00	3229.0
228.90	4287.0
229.90	784.0
231.00	1800.0
232.00	312.0
232.80	310.0
233.10	340.0
234.00	1540.0
234.90	1636.0
235.90	1019.0
237.00	1851.0
239.00	777.0
239.90	704.0
241.00	1283.0
242.00	2670.0
243.00	2968.0
244.00	42192.0
245.00	5390.0
246.00	8844.0
247.00	1766.0
249.00	1612.0
250.90	307.0
252.00	315.0
252.90	867.0
254.10	1806.0
255.00	216832.0

Instrument :
BNA_M
ClientSampleId :
DFTPP68

256.00	31024.0
256.90	2568.0
258.00	12855.0
259.00	2260.0
259.90	387.0
264.00	481.0
264.90	5373.0
265.90	707.0
271.00	1133.0
272.00	724.0
273.00	6725.0
274.00	15988.0
275.00	99096.0
276.00	12741.0
277.00	8273.0
278.00	1046.0
282.90	897.0
284.00	654.0
284.90	1429.0
289.00	356.0
290.10	358.0
291.00	506.0
292.00	492.0
293.00	2345.0
293.90	589.0
295.00	555.0
296.00	31336.0
297.00	4117.0
300.90	369.0
302.00	470.0
303.00	2997.0
304.10	733.0
310.10	349.0
313.00	310.0
314.10	1046.0
315.00	3504.0
316.00	1767.0
321.00	1018.0
321.90	419.0
323.00	9666.0
324.00	1641.0
326.90	1978.0
328.00	622.0
331.90	735.0
333.00	761.0
334.00	6022.0
335.00	1161.0
341.00	1145.0
346.00	2125.0
346.90	429.0
352.00	2696.0
353.00	2166.0
354.00	2760.0
355.10	679.0
365.00	13590.0
366.00	1935.0
371.00	665.0
372.00	5525.0
373.00	1256.0
383.00	1373.0
384.10	467.0
390.00	590.0
390.90	333.0
392.00	407.0
402.00	1643.0
403.00	2403.0
404.00	1091.0
421.00	2187.0
422.00	2054.0
423.00	18112.0
424.10	3291.0
425.00	340.0
441.00	45456.0
442.00	312384.0
443.10	60488.0
444.00	5504.0

Instrument :
BNA_M
ClientSampleId :
DFTPP68

m/z	Abundance
38.00	570.0
39.00	2508.0
44.00	2471.0
49.10	382.0
50.00	10352.0
51.00	37192.0
52.00	1522.0
55.00	420.0
56.00	1281.0
57.00	2869.0
61.00	759.0
62.10	1183.0
63.00	2207.0
63.90	376.0
65.00	1433.0
68.00	1326.0
69.00	71368.0
70.00	304.0
73.10	708.0
74.00	7597.0
75.00	11530.0
76.10	3692.0
77.10	70808.0
78.00	4012.0
79.00	5497.0
80.00	4215.0
81.00	5575.0
81.90	1270.0
83.00	1165.0
85.00	915.0
85.90	1760.0
87.00	771.0
91.00	1552.0
92.00	1343.0
93.00	10025.0
94.00	819.0
96.00	464.0
98.00	7489.0
99.00	5433.0
100.00	335.0
101.00	2497.0
102.90	1083.0
104.00	1932.0
105.00	1710.0
106.00	587.0
107.00	22464.0
108.00	3050.0
109.00	618.0
110.00	35704.0
111.00	5416.0
112.00	864.0
113.00	305.0
115.90	1122.0
117.00	20704.0
118.00	1442.0
122.10	1607.0
122.90	2444.0
124.00	1217.0
125.00	759.0
127.00	65456.0
128.00	5012.0
129.00	29040.0
130.00	2509.0
131.00	496.0
133.90	1007.0
135.00	2527.0
136.10	1123.0
137.10	1413.0
140.00	451.0
141.00	4583.0
141.90	1436.0
142.90	1168.0
146.00	921.0
147.00	2148.0
148.00	6035.0
148.90	1051.0
150.90	463.0
152.00	300.0
153.00	1144.0
154.00	977.0
155.00	2652.0
156.00	3307.0
157.00	506.0
157.90	691.0
158.90	621.0
160.00	1785.0
161.00	2479.0
161.80	497.0
162.10	443.0
165.00	2004.0
166.00	1657.0

Instrument :
BNA_M
ClientSampleId :
DFTPP68

167.00	9377.0
168.00	4834.0
169.10	551.0
170.00	444.0
171.90	793.0
172.90	1310.0
174.00	1852.0
175.00	3434.0
176.00	816.0
177.00	2011.0
178.00	741.0
178.90	6999.0
180.00	4821.0
180.90	2406.0
182.00	422.0
183.90	619.0
185.00	3427.0
186.00	26088.0
187.00	8050.0
187.90	927.0
189.00	1994.0
189.90	304.0
191.00	985.0
191.90	2426.0
193.00	3104.0
194.00	505.0
195.00	363.0
196.00	5251.0
197.10	1309.0
197.90	180992.0
198.90	11907.0
199.90	959.0
201.40	1119.0
203.00	1485.0
204.00	7898.0
205.00	13217.0
206.00	55544.0
207.00	7061.0
207.90	1921.0
208.90	461.0
210.00	783.0
211.00	2130.0
211.70	610.0
214.90	887.0
215.90	1179.0
217.00	14897.0
218.00	1785.0
221.00	8065.0
221.90	2219.0
222.90	3621.0
224.00	31072.0
225.00	8149.0
226.00	586.0
227.00	13613.0
228.00	2097.0
229.00	2602.0
231.00	1321.0
232.90	306.0
234.00	1000.0
234.90	907.0
236.00	680.0
237.00	930.0
239.00	713.0
240.00	470.0
241.00	920.0
242.00	1729.0
243.00	1832.0
244.00	25072.0
245.00	2659.0
245.90	5522.0
247.00	1038.0
249.00	1006.0
252.00	375.0
253.00	660.0
254.00	1124.0
255.00	134528.0
256.00	19384.0
257.00	1198.0
258.00	9005.0
259.00	1293.0
260.00	367.0
263.90	319.0
265.00	3783.0
266.00	497.0
270.90	622.0
272.10	647.0
273.00	3833.0
273.90	10732.0
275.00	64096.0
276.00	7907.0
276.90	5496.0
278.00	826.0
283.10	788.0
283.90	526.0
285.00	950.0

Instrument :
BNA_M
ClientSampleId :
DFTPP68

293.00	1410.0
294.00	422.0
295.00	468.0
296.00	19456.0
297.00	2528.0
300.80	330.0
302.00	433.0
303.00	2287.0
303.90	578.0
314.00	941.0
314.90	2173.0
316.00	1147.0
321.00	616.0
323.00	5820.0
324.00	1121.0
326.90	1242.0
327.90	689.0
331.90	486.0
333.00	587.0
334.00	4567.0
334.90	1222.0
341.00	761.0
345.90	1438.0
352.00	1821.0
353.00	1235.0
354.00	1940.0
355.00	463.0
365.00	10167.0
365.90	1295.0
370.90	453.0
372.00	3057.0
373.10	716.0
383.00	1143.0
389.90	329.0
402.00	1387.0
403.00	2040.0
404.00	929.0
421.00	2037.0
422.00	1762.0
423.00	12241.0
424.00	2621.0
441.00	36504.0
442.10	245632.0
443.10	45224.0
444.10	3903.0

Instrument :
BNA_M
ClientSampleId :
DFTPP68