


Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM052219\
Data File : BM020495.D
Acq On : 22 May 2019 16:43
Operator : JU/SJ
Sample : DFTPP99
Misc :
ALS Vial : 1 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
DFTPP99

Integration File: LSCINT.P

Method : Z:\SVOASRV\HPCHEM1\BNA_M\METHODS\SOM-EPA-BM051619MA.M
Title : SVOA CALIBRATION
Last Update : Thu May 23 07:52:10 2019



AutoFind: Scans 2469, 2470, 2471; Background Corrected with Scan 2464

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	22.9	46570	PASS
68	69	0.00	2	1.9	1701	PASS
69	198	0.00	100	43.0	87251	PASS
70	69	0.00	2	0.3	297	PASS
127	198	10	80	39.9	81064	PASS
197	198	0.00	2	0.5	1033	PASS
198	198	100	100	100.0	202922	PASS
199	198	5	9	6.8	13796	PASS
275	198	10	60	31.4	63693	PASS
365	198	1	100	4.3	8723	PASS
441	198	0.01	100	14.3	29077	PASS
442	198	50	100	95.6	194069	PASS
443	442	15	24	19.1	36978	PASS

Scan 2464 (17.480 min): BM020495.D
DFTPP99

m/z	Abundance
43.00	383.0
43.90	1514.0
69.00	426.0
73.00	476.0
207.00	841.0
281.00	550.0

Instrument :
BNA_M
ClientSampleId :
DFTPP99

m/z	Abundance
36.90	339.0
38.10	702.0
39.10	4176.0
40.00	477.0
41.10	371.0
42.90	307.0
44.00	1954.0
48.90	529.0
50.00	13477.0
51.10	53688.0
52.00	2918.0
55.00	650.0
56.00	1973.0
57.00	4538.0
61.00	915.0
61.90	1256.0
63.00	3460.0
64.10	591.0
65.00	1736.0
66.90	300.0
68.00	1615.0
69.00	97640.0
69.90	473.0
73.00	1066.0
74.00	10074.0
75.00	15106.0
76.00	4764.0
77.10	97528.0
78.00	6722.0
79.00	7503.0
80.00	5532.0
81.00	8029.0
82.00	1950.0
83.00	1876.0
83.80	332.0
84.90	1269.0
86.00	2269.0
87.00	1129.0
87.90	319.0
91.00	1982.0
92.00	2050.0
93.00	12172.0
93.90	666.0
95.00	465.0
96.10	391.0
98.00	9650.0
99.00	6892.0
100.00	613.0
101.00	3729.0
102.90	1268.0
104.00	2508.0
105.00	2362.0
105.90	678.0
107.00	28664.0
108.00	4105.0
109.00	758.0
110.00	47680.0
111.00	7364.0
112.00	782.0
116.10	1548.0
117.00	25168.0
118.00	1970.0
122.00	1682.0
123.00	2684.0
124.00	1592.0
124.90	1475.0
126.20	840.0
127.00	83064.0
128.00	6437.0
129.00	33856.0
130.00	3165.0
130.80	458.0
131.10	423.0
131.90	319.0
133.90	1033.0
135.00	3014.0
136.00	1483.0
136.90	1540.0
138.00	311.0
138.80	303.0
140.00	633.0
140.90	5050.0
142.00	1265.0
142.90	1293.0
144.00	330.0
145.90	947.0
147.00	2408.0
147.90	5797.0
149.10	1161.0
151.00	599.0
151.90	548.0

Instrument :
BNA_M
ClientSampleId :
DFTPP99

153.00	1709.0
153.90	1181.0
155.00	2980.0
156.00	4117.0
157.10	920.0
157.90	809.0
159.00	560.0
160.00	1670.0
161.00	2529.0
161.90	501.0
164.90	2221.0
165.90	1648.0
167.00	10637.0
168.00	4885.0
169.00	761.0
170.00	487.0
170.90	533.0
171.90	1003.0
173.10	1161.0
174.00	2233.0
175.00	4063.0
176.00	1273.0
177.00	2165.0
177.90	706.0
179.00	7599.0
180.00	5278.0
181.00	2645.0
182.10	513.0
183.90	740.0
185.00	3656.0
186.00	28096.0
187.00	8313.0
188.00	1020.0
188.90	1792.0
190.00	351.0
190.80	879.0
192.00	2428.0
193.00	3002.0
193.90	860.0
195.10	387.0
196.00	6086.0
197.00	1322.0
197.90	196800.0
198.90	13514.0
200.00	1244.0
201.50	1163.0
203.00	1533.0
204.00	7656.0
205.00	13708.0
206.00	54720.0
207.00	7768.0
208.00	2012.0
208.90	853.0
210.10	804.0
210.90	2501.0
214.90	458.0
216.00	1367.0
217.00	15567.0
218.00	1709.0
221.00	9083.0
221.90	2214.0
223.00	3455.0
224.00	31360.0
225.00	8071.0
225.90	788.0
227.00	14388.0
227.90	1666.0
229.00	3002.0
230.00	412.0
230.90	1080.0
233.90	890.0
234.90	1055.0
236.00	606.0
236.90	1077.0
238.90	511.0
239.90	378.0
241.00	803.0
242.00	1972.0
242.90	1676.0
244.00	23824.0
245.00	3534.0
246.00	4419.0
246.90	1125.0
249.00	883.0
253.00	421.0
254.10	1265.0
255.00	128344.0
256.00	19056.0
256.90	1710.0
258.00	6899.0
259.00	1030.0
265.00	3271.0
266.00	375.0
270.90	818.0
271.90	470.0

Instrument :
BNA_M
ClientSampleId :
DFTPP99

273.00	3699.0
274.00	9890.0
275.00	56200.0
276.00	8132.0
277.00	5032.0
277.90	817.0
281.00	591.0
282.90	685.0
285.00	912.0
291.00	436.0
292.90	1284.0
294.90	411.0
296.00	16108.0
297.00	2248.0
301.90	377.0
302.90	1820.0
303.90	364.0
314.00	956.0
315.00	1631.0
316.00	1034.0
320.90	528.0
323.00	5026.0
324.00	1096.0
327.00	1081.0
327.90	515.0
332.00	490.0
333.00	503.0
334.00	2956.0
335.00	694.0
341.00	495.0
345.90	1054.0
352.00	1397.0
352.90	871.0
354.10	1282.0
355.00	407.0
364.90	7174.0
365.90	933.0
372.00	2432.0
372.90	678.0
383.00	834.0
390.10	376.0
402.00	902.0
403.00	1295.0
404.00	589.0
421.00	1063.0
422.00	1067.0
423.00	8075.0
424.00	1796.0
441.00	22472.0
442.00	149440.0
443.10	29576.0
444.10	2927.0

Instrument :
BNA_M
ClientSampleId :
DFTPP99

m/z	Abundance
36.90	404.0
37.90	778.0
39.10	4792.0
40.00	532.0
41.10	463.0
42.90	477.0
44.00	2284.0
49.10	540.0
50.00	15549.0
51.00	54304.0
52.00	2911.0
52.90	313.0
55.00	681.0
55.90	2162.0
57.00	5051.0
61.00	1087.0
62.00	1307.0
63.00	3867.0
64.10	451.0
65.10	1737.0
67.10	438.0
68.00	2220.0
69.00	105120.0
70.10	693.0
73.10	1099.0
74.10	10551.0
75.00	17672.0
76.10	4919.0
77.00	105560.0
78.10	7068.0
79.00	8223.0
80.00	5692.0
81.00	8262.0
82.10	1848.0
83.00	1575.0
83.90	311.0
85.00	1483.0
85.90	2396.0
86.90	992.0
88.10	303.0
91.00	2213.0
92.00	2351.0
93.00	14301.0
94.00	777.0
95.10	323.0
95.90	520.0
96.90	391.0
98.00	10341.0
99.00	6831.0
100.00	664.0
101.00	4475.0
102.90	1332.0
104.00	3033.0
105.00	2586.0
106.00	1027.0
107.00	31448.0
108.00	4999.0
109.00	962.0
110.00	53512.0
111.00	8255.0
111.90	1069.0
113.00	450.0
116.10	2201.0
117.00	29600.0
118.00	1921.0
119.90	544.0
121.90	2034.0
123.00	3295.0
124.00	1498.0
125.00	1227.0
126.20	606.0
127.00	97912.0
128.00	7643.0
129.00	39936.0
130.00	3304.0
131.00	710.0
131.90	344.0
133.90	1417.0
135.00	3037.0
136.00	1345.0
137.00	1542.0
138.00	370.0
140.00	504.0
141.00	5706.0
142.00	1750.0
142.90	1338.0
144.00	444.0
145.00	356.0
146.00	1008.0
147.00	2830.0
148.00	7275.0

Instrument :
BNA_M
ClientSampleId :
DFTPP99

149.00	1510.0
150.00	355.0
151.10	637.0
151.90	619.0
153.00	1876.0
154.00	1396.0
155.00	3495.0
156.00	4701.0
156.90	1059.0
157.90	1133.0
159.10	1011.0
159.90	1915.0
161.00	2707.0
162.00	953.0
163.90	310.0
165.00	2607.0
165.90	1838.0
167.00	12030.0
168.00	4890.0
169.00	930.0
170.10	466.0
171.00	602.0
171.90	1182.0
173.00	1446.0
174.00	2964.0
175.00	4910.0
176.00	1363.0
177.00	2218.0
178.10	782.0
179.00	9866.0
180.00	6071.0
181.00	2875.0
181.90	511.0
184.00	926.0
185.00	4510.0
186.00	34952.0
187.00	9226.0
187.90	1373.0
188.90	2725.0
191.00	1032.0
192.00	2926.0
193.00	3861.0
193.90	862.0
194.80	490.0
196.00	7549.0
197.10	1777.0
197.90	244288.0
198.90	16792.0
200.00	1434.0
201.50	1027.0
202.90	2258.0
204.00	11102.0
205.00	17040.0
206.00	69848.0
207.00	10709.0
208.00	2283.0
209.00	923.0
210.00	1163.0
211.00	2672.0
211.90	676.0
215.00	921.0
216.00	1389.0
216.90	19048.0
218.00	2519.0
221.00	11202.0
221.80	2829.0
223.00	3966.0
224.00	40416.0
225.00	9993.0
226.00	1075.0
227.00	16984.0
228.00	2070.0
229.00	3387.0
229.90	570.0
231.00	1438.0
233.00	348.0
234.00	1206.0
234.90	1159.0
235.90	619.0
237.00	1520.0
238.00	325.0
238.80	608.0
239.90	554.0
241.00	754.0
242.00	2407.0
243.00	2497.0
244.00	31080.0
245.00	4003.0
246.00	6190.0
246.90	1315.0
247.90	405.0
248.90	1469.0
250.10	307.0
250.90	336.0
252.00	441.0

Instrument :
BNA_M
ClientSampleId :
DFTPP99

252.90	684.0
254.10	1505.0
255.00	167744.0
256.00	25488.0
256.90	1704.0
258.00	9291.0
259.00	1716.0
260.90	319.0
264.90	4013.0
266.00	586.0
270.90	1164.0
272.00	618.0
273.00	4477.0
274.00	12133.0
275.00	79264.0
276.00	11508.0
277.00	6420.0
278.00	998.0
281.00	407.0
283.00	643.0
283.90	656.0
285.00	1240.0
285.90	358.0
289.10	398.0
289.90	305.0
290.90	404.0
291.80	341.0
292.90	1426.0
293.90	343.0
295.10	611.0
296.00	21968.0
297.00	3239.0
297.90	329.0
301.00	319.0
303.00	2431.0
304.00	608.0
309.10	309.0
313.00	347.0
314.00	1145.0
315.00	2847.0
316.10	1151.0
321.00	693.0
322.00	453.0
323.00	7201.0
324.00	1278.0
326.90	1479.0
328.00	857.0
332.10	627.0
333.10	777.0
334.00	4910.0
335.00	1315.0
341.00	820.0
345.90	1483.0
352.00	2144.0
353.00	1399.0
354.00	2239.0
355.00	421.0
365.00	10783.0
366.00	1511.0
371.00	619.0
372.10	3894.0
373.00	866.0
383.00	1026.0
390.10	558.0
390.90	470.0
392.10	327.0
402.00	1507.0
403.00	2271.0
404.10	417.0
421.00	2024.0
422.00	1768.0
423.00	12421.0
424.00	2804.0
425.10	365.0
441.00	34376.0
442.00	229952.0
443.10	43752.0
444.00	4417.0

Instrument :
BNA_M
ClientSampleId :
DFTPP99

m/z	Abundance
37.90	409.0
39.00	2548.0
40.00	530.0
41.10	344.0
42.90	367.0
44.00	2049.0
49.00	314.0
50.00	9135.0
51.00	31720.0
52.00	1462.0
52.90	327.0
55.00	604.0
56.00	1381.0
57.00	3458.0
61.00	622.0
61.90	675.0
63.10	2075.0
65.00	981.0
66.90	351.0
68.00	1269.0
69.00	60272.0
69.90	420.0
73.00	965.0
74.00	6382.0
75.00	9554.0
76.10	2970.0
77.10	60784.0
78.00	4092.0
79.00	4700.0
80.00	3451.0
81.00	5148.0
82.00	1160.0
83.00	1242.0
83.80	412.0
84.90	952.0
86.00	1697.0
87.10	788.0
91.00	1045.0
92.00	1244.0
93.00	7912.0
93.90	518.0
95.20	408.0
95.90	396.0
97.00	312.0
98.00	5778.0
99.00	4347.0
100.00	364.0
101.10	2347.0
102.90	878.0
104.00	1441.0
105.00	1768.0
106.00	623.0
107.00	19040.0
108.00	2809.0
109.00	679.0
110.00	30488.0
111.00	5067.0
112.00	663.0
116.00	1274.0
117.00	17320.0
117.90	1110.0
122.00	1330.0
123.00	2422.0
123.90	931.0
124.90	924.0
127.00	62216.0
128.00	3875.0
129.00	25520.0
130.00	1860.0
131.00	478.0
133.90	815.0
134.90	1894.0
136.00	955.0
136.90	1270.0
141.00	3647.0
142.00	1236.0
142.90	695.0
145.90	435.0
147.00	2084.0
148.00	4665.0
148.90	841.0
150.00	350.0
151.10	517.0
151.80	434.0
153.00	1211.0
154.00	806.0
155.00	2359.0
156.00	3182.0
157.00	460.0
158.00	780.0
158.80	556.0

Instrument :
BNA_M
ClientSampleId :
DFTPP99

159.90	1098.0
161.00	1723.0
161.90	305.0
164.90	1816.0
165.90	1137.0
167.00	7497.0
168.00	3264.0
169.00	706.0
169.90	318.0
170.90	325.0
171.90	729.0
172.90	1016.0
174.00	2052.0
175.00	3027.0
176.10	909.0
176.90	1408.0
178.00	528.0
178.90	6024.0
180.00	4446.0
180.90	1845.0
181.90	549.0
183.90	496.0
185.00	3343.0
186.00	22576.0
187.00	6815.0
187.90	693.0
189.00	1507.0
191.00	839.0
191.90	2123.0
193.00	2591.0
194.00	528.0
196.00	5401.0
196.80	1098.0
197.90	167680.0
198.90	11084.0
200.00	944.0
201.40	792.0
202.90	1234.0
204.00	6829.0
205.00	11184.0
206.00	48200.0
207.00	6619.0
208.00	2064.0
209.00	869.0
209.90	671.0
211.00	1700.0
211.80	329.0
215.00	746.0
216.00	1160.0
217.00	12902.0
218.00	1693.0
221.00	7874.0
221.90	2064.0
222.90	2687.0
224.00	26448.0
225.00	6727.0
225.90	763.0
227.00	12722.0
227.90	1832.0
229.00	2382.0
229.90	385.0
230.90	1157.0
233.90	1084.0
235.00	760.0
236.00	593.0
237.00	1051.0
239.00	403.0
240.00	391.0
240.90	760.0
242.00	1382.0
243.10	1554.0
244.00	23080.0
245.00	2981.0
246.00	4565.0
246.90	1042.0
248.90	889.0
252.80	497.0
254.10	921.0
255.00	120104.0
256.00	16800.0
256.90	1474.0
258.00	7455.0
258.90	1045.0
261.00	341.0
265.00	3258.0
265.80	437.0
270.90	992.0
272.00	421.0
273.00	3836.0
274.00	9240.0
275.00	55616.0
276.00	7563.0
276.90	5055.0
278.00	611.0
281.10	357.0

Instrument :
BNA_M
ClientSampleId :
DFTPP99

283.00	647.0
283.80	364.0
285.00	816.0
291.00	324.0
292.90	1004.0
294.00	302.0
295.00	416.0
296.00	16201.0
297.00	1958.0
303.00	1987.0
303.90	629.0
314.00	1079.0
315.00	2232.0
316.00	906.0
321.00	535.0
322.10	349.0
323.00	5398.0
324.00	1010.0
326.90	940.0
327.90	502.0
332.00	417.0
333.00	459.0
334.00	3836.0
334.90	885.0
341.00	899.0
341.90	327.0
346.00	1290.0
352.10	1595.0
352.90	1098.0
354.00	1480.0
355.00	404.0
365.00	8213.0
366.00	1142.0
371.00	421.0
372.00	3002.0
373.00	976.0
382.90	557.0
389.90	372.0
390.90	459.0
401.90	1217.0
403.00	1736.0
403.90	589.0
421.00	1670.0
422.00	1385.0
423.00	10841.0
424.10	2266.0
441.00	30384.0
442.10	202816.0
443.00	37608.0
444.00	3245.0

Instrument :
BNA_M
ClientSampleId :
DFTPP99