

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM053023\
 Data File : BM040060.D
 Acq On : 30 May 2023 17:29
 Operator : CG/JU
 Sample : 02961-01
 Misc :
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 H0FP2

Quant Time: May 31 02:02:50 2023
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_M\METHODS\SFAM-EPA-SIM-BM052523.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue May 30 14:54:33 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	8.012	152	6030	0.400	ng/ul	0.00
4) Naphthalene-d8	10.822	136	20071	0.400	ng/ul	0.00
9) Acenaphthene-d10	14.647	164	9371	0.400	ng/ul	0.00
13) Phenanthrene-d10	17.387	188	15313	0.400	ng/ul	0.00
17) Chrysene-d12	21.564	240	7160	0.400	ng/ul	0.00
23) Perylene-d12	24.011	264	7485	0.400	ng/ul	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.380	96	33582	4.783	ng/ul	0.00
6) 2-Methylnaphthalene-d10	12.406	152	5830	0.238	ng/ul	0.00
18) Fluoranthene-d10	19.412	212	7935	0.368	ng/ul	0.00

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM053023\
Data File : BM040060.D
Acq On : 30 May 2023 17:29
Operator : CG/JU
Sample : 02961-01
Misc :
ALS Vial : 10 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
H0FP2

Quant Time: May 31 02:02:50 2023
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_M\METHODS\SFAM-EPA-SIM-BM052523.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue May 30 14:54:33 2023
Response via : Initial Calibration

