

Data Path : Z:\SVOASRV\HPCHEM1\BNA M\DATA\BM060319\
 Data File : BM020675.D
 Acq On : 03 Jun 2019 17:40
 Operator : HP/JU
 Sample : PB120243BL
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 PB120243BL

Quant Time: Jun 04 06:35:42 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA M\METHODS\8270-BM053119.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri May 31 15:48:49 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.83	152	258302	20.00	ng	0.00
21) Naphthalene-d8	10.62	136	1003328	20.00	ng	-0.01
39) Acenaphthene-d10	14.46	164	521899	20.00	ng	-0.01
64) Phenanthrene-d10	17.21	188	1049788	20.00	ng	0.00
76) Chrysene-d12	21.39	240	957304	20.00	ng	-0.01
87) Perylene-d12	23.67	264	1125238	20.00	ng	-0.01
System Monitoring Compounds						
5) 2-Fluorophenol	5.41	112	1854764	126.31	ng	0.00
7) Phenol-d6	6.99	99	2530849	117.05	ng	0.00
23) Nitrobenzene-d5	8.99	82	1542549	79.66	ng	-0.01
42) 2,4,6-Tribromophenol	15.95	330	769724	119.04	ng	-0.01
45) 2-Fluorobiphenyl	13.09	172	3341999	85.13	ng	0.00
79) Terphenyl-d14	19.83	244	4197808	73.35	ng	-0.01

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA M\DATA\BM060319\
Data File : BM020675.D
Acq On : 03 Jun 2019 17:40
Operator : HP/JU
Sample : PB120243BL
Misc :
ALS Vial : 5 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
PB120243BL

Quant Time: Jun 04 06:35:42 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA M\METHODS\8270-BM053119.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri May 31 15:48:49 2019
Response via : Initial Calibration

