

Data Path : Z:\SVOASRV\HPCHEM1\BNA M\DATA\BM060619\
 Data File : BM020775.D
 Acq On : 06 Jun 2019 22:12
 Operator : HP/JU
 Sample : K3201-01
 Misc :
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 C0AB1

Quant Time: Jun 07 06:26:13 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_M\METHODS\SOM-EPA-BM060619MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Fri Jun 07 05:12:28 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.81	152	286017	20.00	ng/ul	0.00
18) Naphthalene-d8	10.60	136	1224400	20.00	ng/ul	0.00
35) Acenaphthene-d10	14.45	164	730639	20.00	ng/ul	0.00
61) Phenanthrene-d10	17.19	188	1667442	20.00	ng/ul	0.00
77) Chrysene-d12	21.37	240	1443245	20.00	ng/ul	-0.01
85) Perylene-d12	23.66	264	1528822	20.00	ng/ul	0.00

System Monitoring Compounds

3) 1,4-Dioxane-d8	3.31	96	22419	3.74	ng/uL	0.00
5) Phenol-d5	6.97	99	434381	19.85	ng/ul	0.00
7) Bis-(2-Chloroethyl)ether-d	7.15	67	277257	21.13	ng/ul	0.00
9) 2-Chlorophenol-d4	7.34	132	373110	21.41	ng/ul	0.00
13) 4-Methylphenol-d8	8.51	113	362540	20.59	ng/ul	0.00
19) Nitrobenzene-d5	8.97	128	180245	21.57	ng/ul	0.00
22) 2-Nitrophenol-d4	9.69	143	187055	21.74	ng/ul	0.00
26) 2,4-Dichlorophenol-d3	10.22	165	384493	21.43	ng/ul	0.00
29) 4-Chloroaniline-d4	10.74	131	457122	21.15	ng/ul	0.00
43) Dimethylphthalate-d6	13.86	166	1297443	24.01	ng/ul	0.00
46) Acenaphthylene-d8	14.14	160	1568254	22.89	ng/ul	0.00
51) 4-Nitrophenol-d4	14.64	143	155233	16.74	ng/ul	0.00
57) Fluorene-d10	15.44	176	1083112	23.58	ng/ul	0.00
62) 4,6-Dinitro-2-methylphenol	15.56	200	112191	13.16	ng/ul	0.00
70) Anthracene-d10	17.29	188	1671015	22.94	ng/ul	0.00
78) Pyrene-d10	19.59	212	1789620	25.65	ng/ul	0.00
89) Benzo(a)pyrene-d12	23.51	264	1683069	24.22	ng/ul	0.00

Target Compounds

					Ovalue
6) Phenol	7.00	94	35882	1.593	ng/ul 98
14) Acetophenone	8.63	105	37215	1.368	ng/ul 91
44) Dimethylphthalate	13.90	163	392917	7.261	ng/ul 100
69) Phenanthrene	17.23	178	116127	1.384	ng/ul 98
76) Fluoranthene	19.25	202	233511	2.428	ng/ul 99
79) Pyrene	19.61	202	214527	2.408	ng/ul 98
82) Benzo(a)anthracene	21.36	228	107900	1.198	ng/ul 96
84) Chrysene	21.40	228	109230	1.308	ng/ul 98
87) Benzo(b)fluoranthene	22.96	252	146093	1.643	ng/ul# 93
90) Benzo(a)pyrene	23.55	252	98331	1.162	ng/ul# 96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA M\DATA\BM060619\
Data File : BM020775.D
Acq On : 06 Jun 2019 22:12
Operator : HP/JU
Sample : K3201-01
Misc :
ALS Vial : 11 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
C0AB1

Quant Time: Jun 07 06:26:13 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_M\METHODS\SOM-EPA-BM060619MA.M
Quant Title : SVOA CALIBRATION
QLast Update : Fri Jun 07 05:12:28 2019
Response via : Initial Calibration

