

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM060623\
 Data File : BM040106.D
 Acq On : 06 Jun 2023 16:30
 Operator : MA/JU
 Sample : 02986-08
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 TP-15-0A

Quant Time: Jun 06 18:02:16 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\8270-BM052323.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Jun 06 16:45:42 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.998	152	270321	20.000	ng	0.00
21) Naphthalene-d8	10.810	136	1156578	20.000	ng	0.00
39) Acenaphthene-d10	14.633	164	666629	20.000	ng	0.00
64) Phenanthrene-d10	17.374	188	1422754	20.000	ng	0.00
76) Chrysene-d12	21.550	240	1153495	20.000	ng	0.00
86) Perylene-d12	23.991	264	1151295	20.000	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.545	112	2441338	138.534	ng	0.00
7) Phenol-d6	7.151	99	2857213	124.289	ng	0.00
23) Nitrobenzene-d5	9.163	82	1875599	87.506	ng	0.00
42) 2,4,6-Tribromophenol	16.121	330	1477459	170.759	ng	0.00
45) 2-Fluorobiphenyl	13.263	172	4729874	83.687	ng	0.00
79) Terphenyl-d14	19.992	244	8820461	118.731	ng	0.00
Target Compounds						
74) Di-n-butylphthalate	18.339	149	2616	3.687	ng	Qvalue # 96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM060623\
Data File : BM040106.D
Acq On : 06 Jun 2023 16:30
Operator : MA/JU
Sample : 02986-08
Misc :
ALS Vial : 5 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
TP-15-0A

Quant Time: Jun 06 18:02:16 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\8270-BM052323.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Jun 06 16:45:42 2023
Response via : Initial Calibration

