

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM060623\
 Data File : BM040111.D
 Acq On : 06 Jun 2023 19:35
 Operator : MA/JU
 Sample : PB153150TB
 Misc :
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 PB153150TB

Quant Time: Jun 07 02:13:25 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\8270-BM052323.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Jun 06 16:45:42 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.993	152	251387	20.000	ng	0.00
21) Naphthalene-d8	10.810	136	1112066	20.000	ng	0.00
39) Acenaphthene-d10	14.633	164	635640	20.000	ng	0.00
64) Phenanthrene-d10	17.374	188	1349403	20.000	ng	0.00
76) Chrysene-d12	21.550	240	1140091	20.000	ng	0.00
86) Perylene-d12	23.991	264	1188040	20.000	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.546	112	2485027	151.634	ng	0.00
7) Phenol-d6	7.151	99	3176275	148.575	ng	0.00
23) Nitrobenzene-d5	9.163	82	1675828	81.315	ng	0.00
42) 2,4,6-Tribromophenol	16.116	330	1222303	148.156	ng	0.00
45) 2-Fluorobiphenyl	13.263	172	4232314	78.534	ng	0.00
79) Terphenyl-d14	19.992	244	7796082	106.175	ng	0.00

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM060623\
Data File : BM040111.D
Acq On : 06 Jun 2023 19:35
Operator : MA/JU
Sample : PB153150TB
Misc :
ALS Vial : 10 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
PB153150TB

Quant Time: Jun 07 02:13:25 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\8270-BM052323.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Jun 06 16:45:42 2023
Response via : Initial Calibration

