

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM060623\
 Data File : BM040118.D
 Acq On : 06 Jun 2023 23:51
 Operator : MA/JU
 Sample : 03018-12
 Misc :
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 278

Quant Time: Jun 07 02:14:54 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\8270-BM052323.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Jun 06 16:45:42 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.993	152	251349	20.000	ng	0.00
21) Naphthalene-d8	10.804	136	1078430	20.000	ng	0.00
39) Acenaphthene-d10	14.628	164	600336	20.000	ng	0.00
64) Phenanthrene-d10	17.375	188	1275853	20.000	ng	0.00
76) Chrysene-d12	21.551	240	982959	20.000	ng	0.00
86) Perylene-d12	23.986	264	907177	20.000	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.546	112	2506097	152.942	ng	0.00
7) Phenol-d6	7.152	99	3042209	142.326	ng	0.00
23) Nitrobenzene-d5	9.163	82	1850283	92.580	ng	0.00
42) 2,4,6-Tribromophenol	16.116	330	1396819	179.266	ng	0.00
45) 2-Fluorobiphenyl	13.257	172	4496020	88.334	ng	0.00
79) Terphenyl-d14	19.992	244	9039698	142.793	ng	0.00
Target Compounds						
74) Di-n-butylphthalate	18.339	149	13398	3.804	ng	Qvalue 97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM060623\
Data File : BM040118.D
Acq On : 06 Jun 2023 23:51
Operator : MA/JU
Sample : 03018-12
Misc :
ALS Vial : 17 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
278

Quant Time: Jun 07 02:14:54 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\8270-BM052323.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Jun 06 16:45:42 2023
Response via : Initial Calibration

