


Data Path : Z:\svoasrv\HPCHEM1\BNA M\Data\BM060818\
 Data File : BM015556.D
 Acq On : 08 Jun 2018 09:10
 Operator : SJ/JU
 Sample : DFTPP56
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 DFTPP56

Integration File: LSCINT.P

Method : Z:\SVOASRV\HPCHEM1\BNA_M\METHODS\SOM-EPA-BM060818MA.M
 Title : SVOA CALIBRATION
 Last Update : Thu Jun 21 11:23:51 2018



AutoFind: Scans 2513, 2514, 2515; Background Corrected with Scan 2507

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	48.1	75757	PASS
68	69	0.00	2	1.9	1356	PASS
69	198	0.00	100	44.8	70629	PASS
70	69	0.00	2	0.4	294	PASS
127	198	10	80	51.5	81093	PASS
197	198	0.00	2	0.7	1052	PASS
198	198	100	100	100.0	157546	PASS
199	198	5	9	6.9	10837	PASS
275	198	10	60	30.0	47280	PASS
365	198	1	100	4.8	7600	PASS
441	198	0.01	100	13.0	20538	PASS
442	198	50	100	85.4	134504	PASS
443	442	15	24	18.9	25397	PASS

Scan 2507 (18.033 min): BM015556.D
DFTPP56

m/z	Abundance
40.00	549.0
41.00	307.0
44.00	1456.0
73.00	884.0
193.00	348.0
207.00	1603.0
207.90	412.0
281.00	500.0

Instrument :
BNA_M
ClientSampleId :
DFTPP56

m/z	Abundance
37.00	670.0
38.00	1438.0
39.10	7201.0
40.00	928.0
40.90	389.0
44.00	1577.0
49.00	1090.0
50.10	23768.0
51.10	88256.0
52.10	4243.0
55.00	676.0
56.00	2903.0
57.00	5002.0
57.90	404.0
60.90	1499.0
62.00	1447.0
63.00	3909.0
64.00	488.0
65.00	1338.0
68.00	1608.0
69.00	79720.0
70.00	457.0
73.00	1611.0
74.00	10037.0
75.00	13431.0
76.10	4898.0
77.10	90168.0
78.00	6853.0
79.00	6739.0
80.00	4437.0
81.00	6972.0
81.90	1472.0
83.00	1324.0
84.00	364.0
84.90	1044.0
86.00	2088.0
86.90	889.0
87.90	460.0
91.00	1375.0
92.00	2269.0
93.00	12757.0
94.00	861.0
96.10	592.0
97.10	325.0
98.00	8767.0
99.00	8441.0
100.00	699.0
101.00	3286.0
103.00	1264.0
104.00	2590.0
105.00	2199.0
106.00	737.0
107.00	28448.0
108.00	3862.0
109.10	654.0
110.00	43864.0
111.00	8207.0
112.00	1074.0
113.10	337.0
116.00	1356.0
117.00	32560.0
118.00	2169.0
119.00	383.0
119.90	417.0
122.00	2129.0
123.00	2944.0
123.90	1547.0
125.00	1217.0
126.20	442.0
127.00	87656.0
128.00	6285.0
129.00	36584.0
130.00	3448.0
131.00	764.0
132.90	522.0
134.00	1053.0
135.00	3525.0
136.00	1345.0
137.00	1836.0
138.00	427.0
140.00	819.0
141.00	5387.0
141.90	1746.0
142.90	993.0
143.90	334.0
146.00	885.0
147.00	3112.0
148.00	7338.0
149.00	1442.0
150.00	307.0
151.00	709.0

Instrument :
BNA_M
ClientSampleId :
DFTPP56

152.10	325.0
153.00	1758.0
153.90	1100.0
155.00	2751.0
156.00	4410.0
156.90	676.0
158.00	1113.0
158.90	871.0
160.00	1665.0
161.00	1941.0
161.90	732.0
165.00	2407.0
166.00	1246.0
167.00	10666.0
168.00	8250.0
169.00	1146.0
170.00	303.0
171.00	342.0
171.90	1035.0
173.00	1240.0
174.00	2285.0
175.00	4213.0
175.90	1076.0
176.90	2128.0
177.90	815.0
179.00	7090.0
180.00	6264.0
180.90	2286.0
181.90	303.0
183.90	648.0
185.00	3807.0
186.00	27568.0
187.00	7939.0
188.00	913.0
188.90	1764.0
190.90	1114.0
192.00	2144.0
193.00	3269.0
194.10	523.0
195.00	338.0
195.90	4944.0
197.10	957.0
197.90	157952.0
198.90	11071.0
199.90	944.0
201.50	680.0
202.90	1686.0
204.00	7305.0
205.00	11474.0
206.00	49176.0
207.00	8050.0
208.00	2230.0
208.90	882.0
210.00	809.0
211.00	1979.0
211.80	326.0
212.10	318.0
215.00	619.0
215.90	1238.0
217.00	13754.0
218.00	1711.0
221.00	6002.0
222.00	1889.0
223.00	2880.0
224.00	26600.0
225.00	7177.0
226.00	1055.0
227.00	13091.0
228.00	1697.0
229.00	2798.0
229.90	348.0
231.00	952.0
233.90	794.0
235.00	950.0
236.00	750.0
237.00	2162.0
238.90	582.0
239.90	523.0
240.90	660.0
241.90	1527.0
243.00	1598.0
244.00	19216.0
245.00	2587.0
246.00	4724.0
247.00	930.0
248.00	300.0
249.00	809.0
251.90	461.0
253.00	437.0
254.10	1026.0
255.00	103896.0
256.00	15683.0
257.00	1563.0
258.00	7310.0
259.00	1439.0

Instrument :
BNA_M
ClientSampleId :
DFTPP56

265.00	2987.0
265.90	344.0
272.00	380.0
273.00	3056.0
274.00	8981.0
275.00	43120.0
276.00	5601.0
276.90	4339.0
277.90	1032.0
281.00	721.0
282.00	367.0
283.00	816.0
283.90	303.0
285.10	784.0
292.00	313.0
292.90	1183.0
295.10	387.0
296.00	16048.0
297.00	2185.0
302.00	318.0
303.00	1585.0
303.90	438.0
313.90	936.0
314.90	1480.0
316.00	812.0
320.90	383.0
323.00	3931.0
324.00	869.0
326.90	1056.0
328.00	446.0
333.10	484.0
334.00	2627.0
335.00	669.0
341.00	605.0
346.00	1087.0
351.90	1434.0
353.00	799.0
353.90	1285.0
355.00	362.0
365.00	6369.0
365.90	831.0
371.00	448.0
372.00	2227.0
373.00	374.0
383.00	575.0
389.90	311.0
401.90	660.0
403.00	1055.0
404.00	435.0
421.00	754.0
421.90	921.0
423.00	6086.0
424.00	2511.0
425.00	391.0
441.00	15855.0
442.00	103480.0
443.00	19912.0
444.00	1961.0

Instrument :
BNA_M
ClientSampleId :
DFTPP56

m/z	Abundance
37.00	678.0
38.10	1443.0
39.10	7740.0
40.00	835.0
41.00	332.0
44.00	1895.0
49.00	960.0
50.10	24776.0
51.10	88192.0
52.00	4631.0
55.00	580.0
56.00	2842.0
57.00	5244.0
61.00	1200.0
62.00	1416.0
63.00	3815.0
64.10	576.0
65.10	1706.0
68.00	1623.0
69.00	83608.0
70.00	427.0
73.10	1813.0
74.00	10191.0
75.00	14845.0
76.10	4822.0
77.00	96848.0
78.00	6790.0
79.00	7449.0
80.00	5279.0
81.00	7143.0
82.00	1635.0
83.00	1509.0
84.10	324.0
85.00	989.0
86.00	2193.0
87.00	1229.0
87.90	375.0
91.00	1820.0
92.00	1869.0
93.00	13132.0
93.90	534.0
96.00	529.0
98.00	10735.0
99.00	8210.0
100.00	630.0
101.00	3948.0
103.00	1281.0
104.00	2698.0
105.00	2824.0
105.90	648.0
107.00	28920.0
108.00	4785.0
109.10	903.0
110.00	48080.0
111.00	8913.0
112.00	1359.0
112.90	479.0
116.00	1780.0
117.00	33728.0
118.00	2436.0
119.80	355.0
122.00	2175.0
123.00	3180.0
124.00	1335.0
125.00	1557.0
126.10	460.0
127.00	94416.0
128.00	7915.0
129.00	41456.0
129.90	3881.0
131.00	867.0
131.90	466.0
132.90	306.0
134.00	1262.0
135.00	3594.0
136.00	1202.0
137.00	2195.0
138.00	466.0
139.90	554.0
141.00	5733.0
142.00	1595.0
143.00	1206.0
145.00	406.0
146.00	1168.0
147.10	3033.0
148.00	7853.0
149.00	1442.0
150.10	325.0
150.90	684.0
153.00	2146.0
154.00	1183.0

Instrument :
BNA_M
ClientSampleId :
DFTPP56

155.00	3091.0
156.00	4399.0
157.10	860.0
158.00	1290.0
159.00	723.0
160.00	2065.0
161.00	3059.0
162.00	805.0
165.00	2593.0
165.90	1888.0
167.00	13208.0
168.00	8800.0
169.00	1188.0
170.00	422.0
170.90	577.0
172.00	1280.0
173.00	1664.0
174.00	2476.0
175.00	4913.0
176.00	1016.0
177.00	2216.0
178.00	832.0
179.00	9527.0
180.00	6404.0
180.90	2919.0
181.90	470.0
183.00	379.0
184.00	831.0
185.00	4268.0
186.00	30248.0
187.00	9473.0
188.00	1029.0
188.90	2192.0
189.80	351.0
190.90	1356.0
192.00	3178.0
193.00	3127.0
194.00	572.0
196.00	5196.0
197.00	1405.0
197.90	188288.0
198.90	13017.0
200.00	1302.0
201.50	1117.0
203.00	1950.0
204.00	8821.0
205.00	14238.0
206.00	57032.0
207.00	9554.0
208.00	2376.0
209.00	918.0
210.00	913.0
211.00	2679.0
211.90	384.0
215.00	839.0
216.00	1519.0
217.00	17640.0
217.90	2178.0
221.00	7534.0
221.90	2081.0
223.00	3680.0
224.00	35288.0
225.00	9469.0
226.00	1133.0
227.00	16800.0
227.90	2406.0
229.00	3069.0
230.00	384.0
231.00	1326.0
234.00	1101.0
234.90	1121.0
235.90	1030.0
237.00	1804.0
237.90	304.0
239.00	772.0
239.90	451.0
241.00	1015.0
242.00	2048.0
243.00	2068.0
244.00	24296.0
245.00	3186.0
246.00	5811.0
246.90	1332.0
248.90	793.0
251.00	316.0
252.90	546.0
253.90	918.0
255.00	135552.0
256.00	19592.0
257.00	1909.0
258.00	9408.0
259.00	1387.0
259.90	318.0
265.00	4344.0
265.90	505.0

Instrument :
BNA_M
ClientSampleId :
DFTPP56

271.00	316.0
272.10	344.0
272.90	4182.0
274.00	10461.0
275.00	57592.0
276.00	7084.0
277.00	6592.0
277.90	1370.0
281.10	483.0
283.00	668.0
284.00	522.0
284.90	1259.0
293.00	1501.0
294.00	457.0
295.10	365.0
296.00	21472.0
297.00	3057.0
302.00	323.0
303.00	1864.0
304.00	550.0
314.00	815.0
315.00	1911.0
316.00	1344.0
321.00	718.0
321.90	432.0
323.00	5487.0
324.00	946.0
326.90	1438.0
328.00	637.0
332.00	489.0
333.00	828.0
334.00	4311.0
335.10	920.0
341.00	637.0
345.90	1322.0
347.00	430.0
352.00	1636.0
353.00	1068.0
354.00	1743.0
355.00	453.0
365.00	9328.0
366.00	1236.0
370.90	378.0
372.10	3192.0
373.10	758.0
383.00	862.0
390.00	425.0
391.00	314.0
402.00	1101.0
403.00	1641.0
403.90	628.0
421.00	1211.0
422.00	1367.0
423.00	9322.0
424.00	3198.0
425.00	466.0
441.00	24320.0
442.00	158016.0
443.00	29784.0
444.00	2336.0

Instrument :
BNA_M
ClientSampleId :
DFTPP56

m/z	Abundance
38.00	919.0
39.00	4598.0
40.00	751.0
41.00	379.0
44.00	1402.0
49.00	736.0
50.10	14891.0
51.10	50824.0
52.00	2270.0
54.90	428.0
56.00	1595.0
57.00	3350.0
60.90	671.0
62.00	927.0
63.00	2146.0
64.00	359.0
65.00	1030.0
68.10	838.0
69.00	48560.0
73.00	1408.0
74.10	6145.0
75.00	8504.0
76.10	2780.0
77.10	56736.0
78.10	3966.0
79.00	4946.0
80.00	3433.0
81.00	4391.0
82.00	906.0
83.00	797.0
84.90	572.0
85.90	1337.0
86.90	641.0
90.90	1034.0
92.00	1183.0
93.00	7847.0
94.00	809.0
96.00	447.0
98.00	6356.0
99.00	5004.0
100.00	422.0
101.00	2182.0
102.90	795.0
104.00	1529.0
105.00	1367.0
106.00	489.0
107.00	18112.0
108.00	2345.0
108.90	468.0
110.00	29760.0
111.00	5129.0
112.00	518.0
115.90	902.0
117.00	21736.0
118.00	1524.0
119.90	343.0
121.90	1096.0
123.00	2011.0
123.90	932.0
125.00	962.0
126.20	419.0
127.00	61208.0
128.00	4710.0
129.00	26624.0
129.90	2055.0
131.00	421.0
133.00	348.0
134.00	829.0
135.00	2192.0
136.00	1049.0
137.00	1391.0
140.00	406.0
141.00	3636.0
141.90	1250.0
143.00	798.0
145.90	830.0
147.00	2175.0
147.90	4962.0
148.90	986.0
150.00	343.0
150.80	394.0
151.30	310.0
153.00	1424.0
154.00	939.0
155.00	2620.0
156.00	2845.0
157.00	602.0
158.00	700.0
158.90	695.0
159.90	1436.0
161.00	1546.0

Instrument :
BNA_M
ClientSampleId :
DFTPP56

161.90	519.0
165.00	1696.0
166.00	1069.0
167.00	8175.0
168.00	5444.0
169.00	760.0
171.90	657.0
172.90	828.0
174.00	1913.0
175.00	3085.0
175.90	858.0
177.00	1571.0
178.00	448.0
179.00	6064.0
180.00	3710.0
181.00	1889.0
184.00	545.0
185.00	3008.0
186.00	21088.0
187.00	6219.0
188.00	791.0
188.90	1459.0
190.00	366.0
190.90	970.0
192.00	2108.0
193.00	2345.0
193.90	567.0
196.00	4033.0
197.00	796.0
197.90	126400.0
199.00	8424.0
200.00	731.0
201.40	563.0
203.00	1233.0
204.00	5859.0
205.00	10318.0
206.00	38248.0
207.10	6985.0
208.00	2121.0
209.00	939.0
210.10	675.0
210.90	1968.0
215.00	571.0
215.90	978.0
217.00	11493.0
218.00	1724.0
221.00	4399.0
221.80	1412.0
223.00	2715.0
224.00	23464.0
225.00	5935.0
226.00	746.0
227.00	11631.0
228.00	1440.0
229.00	2064.0
230.90	831.0
232.90	306.0
234.00	820.0
234.80	660.0
236.00	646.0
237.00	1457.0
238.90	346.0
239.90	424.0
241.00	858.0
242.00	1338.0
243.00	1210.0
244.00	17832.0
245.00	2320.0
245.90	4732.0
246.90	1002.0
249.00	641.0
250.80	312.0
254.10	751.0
255.00	94176.0
256.00	13618.0
257.00	1541.0
258.00	6498.0
258.90	1198.0
265.00	2763.0
266.10	407.0
273.00	2931.0
274.00	7991.0
275.00	41128.0
276.00	5456.0
276.90	4496.0
278.00	919.0
281.00	562.0
283.00	688.0
283.90	395.0
285.00	729.0
293.00	968.0
295.00	377.0
296.00	15311.0
296.90	2273.0
301.80	352.0

Instrument :
BNA_M
ClientSampleId :
DFTPP56

303.00	1855.0
304.00	506.0
314.00	729.0
315.00	1849.0
316.00	993.0
321.00	424.0
323.00	4255.0
324.00	888.0
326.90	1022.0
327.90	585.0
332.00	324.0
333.00	423.0
334.00	2774.0
335.00	988.0
341.00	757.0
346.00	1172.0
352.00	1390.0
353.00	999.0
354.10	1405.0
354.90	555.0
365.00	7104.0
365.90	915.0
370.90	335.0
372.00	2420.0
373.00	630.0
383.00	700.0
390.00	340.0
402.00	1002.0
402.90	1372.0
404.10	551.0
421.00	1445.0
422.00	1196.0
423.00	7737.0
424.00	2347.0
425.00	352.0
441.00	21440.0
442.00	142016.0
443.00	26496.0
444.00	2562.0

Instrument :
BNA_M
ClientSampleId :
DFTPP56