


Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM060821\
 Data File : BM030313.D
 Acq On : 08 Jun 2021 07:03
 Operator : CG/JU
 Sample : DFTPP50
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 DFTPP050

Integration File: LSCINT.P

Method : Z:\SVOASRV\HPCHEM1\BNA_M\METHODS\SFAM-EPA-BM060621.M
 Title : SVOA CALIBRATION
 Last Update : Tue Jun 08 09:51:19 2021



AutoFind: Scans 2475, 2476, 2477; Background Corrected with Scan 2468

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	39.7	195212	PASS
68	69	0.00	2	1.6	3150	PASS
69	198	0.00	100	39.3	192899	PASS
70	69	0.00	2	0.5	918	PASS
127	198	10	80	47.9	235477	PASS
197	198	0.00	2	0.5	2231	PASS
198	198	100	100	100.0	491179	PASS
199	198	5	9	6.8	33200	PASS
275	198	10	60	27.7	136136	PASS
365	198	1	100	4.0	19662	PASS
441	198	0.01	100	11.1	54560	PASS
442	198	50	100	72.1	353920	PASS
443	442	15	24	19.3	68347	PASS

m/z	Abundance
39.00	195.0
40.00	1037.0
41.10	237.0
42.90	261.0
44.00	1010.0
48.90	672.0
50.80	153.0
73.00	1335.0
74.00	169.0
75.00	153.0
78.00	214.0
79.00	211.0
83.90	419.0
85.80	170.0
91.10	162.0
96.10	213.0
119.10	151.0
133.00	227.0
147.00	514.0
149.00	190.0
176.90	150.0
190.90	302.0
192.90	409.0
207.00	2663.0
207.90	571.0
209.00	235.0
221.10	387.0
281.00	808.0
281.90	194.0
341.00	150.0

Instrument :
BNA_M
ClientSampleId :
DFTPP050

m/z	Abundance
36.00	421.0
37.00	938.0
38.10	3244.0
39.00	18856.0
40.00	1859.0
41.10	940.0
43.00	253.0
44.00	1544.0
45.00	630.0
46.80	239.0
48.00	253.0
49.00	2162.0
50.00	69856.0
51.00	262592.0
52.10	13268.0
53.00	475.0
54.90	1334.0
56.00	7464.0
57.00	18264.0
58.00	917.0
58.90	246.0
60.10	200.0
61.00	2954.0
62.00	3679.0
63.00	10260.0
64.00	1519.0
65.10	4480.0
66.10	506.0
66.90	414.0
68.00	4296.0
69.00	255040.0
70.00	1097.0
73.00	3173.0
74.00	26480.0
75.00	40640.0
76.00	14325.0
77.00	297344.0
78.00	21288.0
79.00	17848.0
80.00	14909.0
81.00	19872.0
82.00	5327.0
83.00	4861.0
84.00	933.0
85.00	3542.0
86.00	6334.0
87.00	2463.0
88.00	1139.0
89.00	311.0
91.00	4356.0
92.00	5187.0
93.00	34176.0
94.00	1973.0
95.00	562.0
96.10	1339.0
96.90	406.0
98.00	24792.0
99.00	23184.0
100.10	1821.0
101.00	11320.0
102.00	714.0
103.00	3602.0
104.00	6843.0
105.00	6001.0
106.00	2042.0
107.00	84928.0
108.00	12712.0
109.00	2574.0
110.00	144256.0
111.00	22960.0
112.00	3339.0
113.00	1207.0
114.20	194.0
115.10	490.0
116.00	4869.0
117.00	78744.0
118.00	5952.0
119.10	815.0
120.10	1137.0
121.00	329.0
122.00	5736.0
123.00	8629.0
124.00	3988.0
125.00	3306.0
126.10	1037.0
127.00	295232.0
128.00	23208.0
129.00	111656.0
130.00	9946.0
131.00	2184.0
132.10	1015.0
132.80	737.0

Instrument :
BNA_M
ClientSampleId :
DFTPP050

134.00	3687.0
135.00	9116.0
136.00	3827.0
137.00	6514.0
138.00	922.0
139.10	524.0
140.00	1276.0
141.00	15270.0
142.00	4245.0
143.00	3853.0
144.00	752.0
145.00	985.0
146.00	2177.0
147.00	7478.0
148.00	19528.0
149.00	4468.0
150.00	1125.0
151.20	1776.0
151.90	1109.0
153.00	5049.0
154.00	3459.0
155.00	8339.0
156.00	12601.0
157.00	2227.0
158.00	2523.0
159.00	2484.0
160.00	4888.0
161.00	7065.0
161.90	2124.0
163.00	650.0
164.00	609.0
165.00	5573.0
166.00	5534.0
167.00	32928.0
168.00	22032.0
169.00	3340.0
170.00	1001.0
170.90	1286.0
172.00	2657.0
173.00	3714.0
174.00	6407.0
175.00	12435.0
176.00	3070.0
177.00	5658.0
178.00	1907.0
179.00	23840.0
180.00	18008.0
181.00	7780.0
182.00	1183.0
183.00	476.0
183.90	1637.0
185.00	10955.0
186.00	83040.0
187.00	24016.0
188.00	2977.0
189.00	5593.0
189.90	913.0
191.00	2898.0
192.00	7567.0
193.00	8767.0
194.00	1978.0
195.00	879.0
196.00	17624.0
197.00	2561.0
198.00	580416.0
199.00	40080.0
200.00	3411.0
201.50	2248.0
203.00	4610.0
204.00	22768.0
205.00	37776.0
206.00	156736.0
207.00	23504.0
208.00	6450.0
209.00	2589.0
210.00	2249.0
211.00	6065.0
212.10	1070.0
213.00	465.0
213.90	150.0
215.00	1660.0
216.00	3627.0
217.00	46000.0
218.00	5358.0
218.90	514.0
220.20	1156.0
221.00	23008.0
221.80	5407.0
223.00	9131.0
224.00	90824.0
225.00	22616.0
226.10	2652.0
227.00	37208.0
228.00	5407.0
229.00	7868.0
230.00	1133.0

Instrument :
BNA_M
ClientSampleId :
DFTPP050

231.00	2921.0
232.00	515.0
232.90	725.0
234.00	2131.0
235.00	3311.0
236.00	1856.0
237.00	4885.0
238.00	822.0
239.00	1215.0
240.00	1150.0
241.00	2382.0
242.00	4476.0
243.00	5438.0
244.00	70024.0
245.00	9185.0
246.00	13030.0
247.00	2946.0
248.00	746.0
248.90	2628.0
250.00	504.0
251.00	733.0
252.00	602.0
253.00	1683.0
254.00	2069.0
255.00	346688.0
256.00	49320.0
257.00	5231.0
258.00	21080.0
259.00	3521.0
259.90	756.0
260.90	621.0
263.90	493.0
265.00	8710.0
266.00	1243.0
267.00	400.0
267.90	175.0
269.90	434.0
271.00	733.0
272.00	1109.0
273.00	9930.0
274.00	25984.0
275.00	147328.0
276.00	20224.0
277.00	12857.0
278.00	2493.0
279.00	558.0
281.00	752.0
282.10	416.0
283.00	1366.0
284.00	1220.0
285.00	2252.0
286.00	374.0
288.10	181.0
288.90	602.0
289.80	292.0
290.90	308.0
292.00	582.0
293.00	2884.0
294.00	908.0
295.00	917.0
296.00	44424.0
297.00	6960.0
298.00	382.0
301.00	451.0
302.00	841.0
303.00	4594.0
304.00	1525.0
304.90	404.0
308.10	572.0
309.00	514.0
310.10	515.0
310.90	176.0
313.10	533.0
314.00	1916.0
315.00	5332.0
316.00	2518.0
317.00	399.0
319.90	172.0
321.00	1218.0
322.00	737.0
323.00	13161.0
324.00	3032.0
325.00	288.0
325.90	323.0
326.90	2694.0
327.90	1408.0
328.90	212.0
331.90	815.0
333.10	1202.0
334.00	8182.0
335.00	2253.0
336.10	218.0
338.90	230.0
341.00	1665.0
342.00	382.0
346.00	3173.0

Instrument :
BNA_M
ClientSampleId :
DFTPP050

346.90	616.0
351.00	194.0
352.00	3640.0
353.00	3049.0
354.10	3689.0
355.00	814.0
359.00	231.0
365.00	20296.0
366.00	2974.0
366.80	159.0
369.90	295.0
371.00	1051.0
372.00	6268.0
373.00	1595.0
373.80	186.0
383.00	1598.0
384.00	510.0
385.00	190.0
390.00	978.0
391.00	586.0
392.00	707.0
400.80	351.0
402.00	2699.0
403.00	3356.0
404.00	1393.0
405.00	547.0
410.00	186.0
414.90	253.0
421.00	3014.0
422.10	2900.0
423.00	19560.0
424.00	7534.0
425.10	1069.0
441.00	50352.0
442.10	324032.0
443.10	65312.0
444.00	6241.0
445.00	258.0

Instrument :
BNA_M
ClientSampleId :
DFTPP050

m/z	Abundance
36.00	247.0
37.00	1050.0
38.00	2778.0
39.00	17256.0
40.00	1703.0
41.00	705.0
42.10	168.0
42.90	314.0
44.00	1254.0
45.00	541.0
47.00	184.0
48.00	201.0
49.00	2039.0
50.00	59304.0
51.00	220864.0
52.00	11259.0
53.00	710.0
55.10	958.0
56.00	5609.0
57.00	15011.0
58.00	944.0
58.80	252.0
59.90	198.0
61.00	2787.0
62.00	3072.0
63.00	8363.0
64.00	1074.0
65.00	3893.0
66.10	183.0
67.00	490.0
68.00	3589.0
69.00	219520.0
70.00	1013.0
73.00	2831.0
74.00	24224.0
75.00	36776.0
76.10	12798.0
77.10	257920.0
78.10	18040.0
79.00	15768.0
80.00	12446.0
81.00	17480.0
82.00	4338.0
83.00	4415.0
84.00	730.0
85.00	3169.0
86.00	5584.0
87.00	2146.0
88.00	866.0
89.10	417.0
91.00	4097.0
92.00	4688.0
93.00	30816.0
94.00	2158.0
95.10	557.0
96.10	1456.0
97.00	444.0
98.00	23248.0
99.00	19160.0
100.00	1552.0
101.00	10003.0
101.90	490.0
103.00	2841.0
104.00	6216.0
105.00	6500.0
106.00	1922.0
107.00	77440.0
108.00	12131.0
109.00	1861.0
110.00	130864.0
111.00	20472.0
112.10	2645.0
113.00	935.0
114.10	250.0
114.90	398.0
116.00	4200.0
117.00	74320.0
118.00	5131.0
119.00	883.0
120.00	887.0
121.00	343.0
122.00	4661.0
123.00	7782.0
124.00	3073.0
125.00	3149.0
126.10	1046.0
127.00	274048.0
128.00	21624.0
129.00	105248.0
130.00	8411.0
130.90	1807.0
131.90	856.0

Instrument :
BNA_M
ClientSampleId :
DFTPP050

133.00	537.0
134.00	2822.0
135.00	8378.0
136.00	3283.0
137.00	5010.0
138.00	948.0
139.00	507.0
140.00	1478.0
141.00	13321.0
142.00	4313.0
143.00	3254.0
144.00	991.0
144.90	707.0
146.00	2777.0
147.00	7058.0
148.00	20032.0
149.00	4121.0
150.00	924.0
151.00	1524.0
151.70	1041.0
153.00	4407.0
154.00	3092.0
155.00	8162.0
156.00	12769.0
157.00	2059.0
158.00	2403.0
159.00	2078.0
160.00	4587.0
161.00	7400.0
162.00	1685.0
163.10	680.0
164.10	558.0
165.00	5761.0
166.00	4224.0
167.00	33368.0
168.00	20616.0
169.00	3071.0
170.00	1018.0
170.90	994.0
172.00	2694.0
173.00	3041.0
174.00	7564.0
175.00	12412.0
176.00	3506.0
177.00	4897.0
178.00	1867.0
179.00	22784.0
180.00	17480.0
181.00	7757.0
182.00	1204.0
182.90	535.0
184.00	1771.0
185.00	10799.0
186.00	83216.0
187.00	24104.0
188.00	2881.0
189.00	4848.0
189.90	805.0
191.00	2816.0
192.00	7740.0
193.00	8765.0
194.10	2215.0
195.00	884.0
196.00	17104.0
197.00	2716.0
198.00	581632.0
199.00	38568.0
200.00	3494.0
201.40	2302.0
202.00	898.0
203.00	4707.0
204.00	24296.0
205.00	40568.0
206.00	163904.0
207.10	21680.0
208.00	6212.0
209.00	2257.0
210.10	2641.0
211.00	5454.0
211.90	1259.0
212.90	539.0
215.00	1801.0
216.00	3758.0
217.00	46848.0
218.00	5718.0
219.10	388.0
221.00	23936.0
222.00	5645.0
223.00	11172.0
224.00	90352.0
225.00	22760.0
226.00	2796.0
227.00	38816.0
228.00	5759.0
229.00	8453.0
229.90	1017.0

Instrument :
BNA_M
ClientSampleId :
DFTPP050

231.00	3064.0
232.00	697.0
233.00	568.0
234.00	2226.0
235.00	3228.0
236.00	1911.0
237.00	4161.0
238.00	781.0
239.00	1806.0
240.00	1112.0
241.00	2480.0
242.00	5031.0
243.00	5298.0
244.00	71888.0
245.00	9523.0
246.00	14211.0
247.00	2508.0
248.00	544.0
249.00	2952.0
250.00	420.0
251.00	638.0
251.90	643.0
253.00	1744.0
254.10	2434.0
255.00	376704.0
256.00	53632.0
257.00	4907.0
258.00	22400.0
259.00	3709.0
260.00	615.0
261.00	671.0
263.00	305.0
264.00	563.0
265.00	9918.0
265.90	1374.0
267.00	380.0
269.90	587.0
271.00	786.0
272.00	1054.0
273.00	12100.0
274.00	28480.0
275.00	167168.0
276.00	21688.0
277.00	13981.0
278.00	3277.0
278.90	747.0
281.00	677.0
281.80	528.0
283.00	1704.0
284.10	1284.0
285.00	2228.0
286.10	384.0
287.90	250.0
288.90	660.0
289.90	649.0
290.90	283.0
292.00	650.0
293.00	2979.0
293.90	860.0
294.90	974.0
296.00	50432.0
297.00	6419.0
297.90	433.0
301.00	762.0
302.00	1016.0
303.00	5771.0
304.00	1560.0
305.00	283.0
307.90	789.0
308.90	558.0
310.00	554.0
310.90	165.0
311.90	188.0
312.90	362.0
314.00	2496.0
315.00	6202.0
316.00	2992.0
317.00	536.0
320.00	268.0
321.00	1578.0
322.00	1007.0
323.00	15443.0
324.00	2854.0
325.00	318.0
325.90	330.0
326.90	3248.0
328.00	1347.0
328.90	235.0
332.00	1061.0
333.00	1535.0
334.10	9552.0
335.10	2560.0
336.10	518.0
339.00	320.0
339.90	329.0
341.10	1854.0

Instrument :
BNA_M
ClientSampleId :
DFTPP050

342.10	284.0
346.00	3854.0
347.00	819.0
350.80	234.0
352.00	4950.0
353.00	3434.0
354.00	4244.0
355.00	1227.0
358.90	360.0
364.10	151.0
365.00	23288.0
366.00	3474.0
367.00	202.0
370.00	455.0
371.00	1327.0
372.10	7718.0
373.00	2135.0
374.00	203.0
377.00	191.0
383.00	2312.0
383.90	566.0
385.00	286.0
390.00	1204.0
391.00	810.0
391.90	461.0
401.10	640.0
402.00	2825.0
403.00	4486.0
404.00	1708.0
405.00	447.0
414.90	334.0
421.00	4117.0
422.00	3315.0
423.00	25264.0
424.10	7500.0
425.00	1200.0
438.80	200.0
441.10	65712.0
442.10	433024.0
443.10	81224.0
444.00	7081.0
445.00	391.0

Instrument :
BNA_M
ClientSampleId :
DFTPP050

m/z	Abundance
35.90	238.0
37.00	614.0
38.00	1310.0
39.10	8126.0
40.00	1478.0
41.00	430.0
42.90	251.0
44.00	1367.0
45.00	311.0
49.00	1252.0
50.00	27048.0
51.00	102640.0
52.00	5154.0
53.10	190.0
55.10	551.0
56.00	3211.0
57.00	6950.0
58.00	352.0
59.00	198.0
61.00	1393.0
62.10	1404.0
63.00	4217.0
64.00	490.0
65.00	1892.0
67.00	210.0
68.00	1564.0
69.00	104136.0
70.00	643.0
73.10	2149.0
74.00	11513.0
75.00	17224.0
76.00	5711.0
77.10	123800.0
78.00	8699.0
79.00	7678.0
80.00	6568.0
81.00	8245.0
82.00	2328.0
83.00	2019.0
84.00	521.0
84.90	1480.0
86.00	2913.0
87.00	1045.0
87.90	593.0
89.00	257.0
91.00	1992.0
92.00	2492.0
93.00	14380.0
93.90	868.0
95.00	268.0
96.00	910.0
96.70	257.0
98.00	11196.0
99.00	10249.0
100.00	708.0
101.00	5013.0
102.00	294.0
103.00	1387.0
104.00	3073.0
105.00	2869.0
106.10	846.0
107.00	37608.0
108.00	5527.0
109.10	1122.0
110.00	63048.0
111.00	9889.0
112.00	1456.0
113.00	501.0
115.00	260.0
116.00	2033.0
117.00	35608.0
118.00	2437.0
119.00	392.0
120.00	607.0
121.00	278.0
122.00	2596.0
123.00	4170.0
124.00	2132.0
125.00	1593.0
126.10	464.0
127.00	137152.0
128.00	9360.0
129.00	51928.0
130.00	4496.0
130.90	780.0
132.10	629.0
133.00	531.0
133.90	1884.0
135.00	4426.0
136.00	1903.0
137.10	2278.0
138.00	659.0

Instrument :
BNA_M
ClientSampleId :
DFTPP050

138.90	294.0
140.00	638.0
141.00	7647.0
142.00	2444.0
143.00	1537.0
143.90	515.0
145.00	571.0
146.00	1202.0
147.00	4471.0
148.00	10377.0
149.00	1797.0
150.10	703.0
151.10	930.0
151.90	665.0
153.00	2003.0
154.00	2005.0
155.00	4312.0
156.00	5800.0
156.90	1380.0
158.00	1375.0
159.00	1051.0
160.00	2271.0
161.00	3401.0
161.90	749.0
162.90	425.0
164.00	366.0
164.90	2570.0
166.00	1788.0
167.00	16920.0
168.00	10904.0
169.00	1518.0
169.90	534.0
170.90	695.0
172.00	1561.0
173.00	2070.0
173.90	3178.0
175.00	6462.0
176.00	1716.0
177.00	2907.0
178.00	1058.0
179.00	12043.0
180.00	9108.0
181.00	3554.0
182.00	654.0
182.80	281.0
184.00	1031.0
185.00	6437.0
186.00	43840.0
187.00	12212.0
188.10	1464.0
189.00	2953.0
190.10	309.0
190.90	1405.0
192.00	3740.0
193.00	5173.0
194.00	583.0
195.00	600.0
196.00	8848.0
197.00	1417.0
198.00	311488.0
199.00	20952.0
200.00	1414.0
201.40	1605.0
203.00	2426.0
204.00	11988.0
205.00	20560.0
206.00	86256.0
207.00	12829.0
208.00	3791.0
209.00	1686.0
211.00	3594.0
211.90	564.0
212.90	257.0
215.00	998.0
216.00	2160.0
217.00	24096.0
218.00	2585.0
219.00	307.0
221.00	13434.0
221.80	3165.0
223.00	5158.0
224.00	51848.0
225.00	12504.0
226.00	1544.0
227.00	21680.0
228.00	2827.0
229.00	4448.0
230.00	620.0
231.00	1690.0
232.00	291.0
233.00	380.0
233.90	1475.0
235.00	1922.0
236.00	928.0
237.00	2969.0
238.00	413.0

Instrument :
BNA_M
ClientSampleId :
DFTPP050

239.00	993.0
240.00	727.0
241.00	1246.0
242.00	3206.0
243.00	3216.0
244.00	40744.0
245.00	5276.0
246.00	8189.0
247.00	1408.0
247.90	346.0
249.00	1520.0
249.90	260.0
251.00	350.0
251.90	396.0
252.90	754.0
254.00	1410.0
255.00	216384.0
256.00	31200.0
257.00	2082.0
258.00	12752.0
259.00	1766.0
259.90	345.0
260.90	388.0
264.00	283.0
265.00	5691.0
266.00	903.0
266.90	255.0
267.90	186.0
269.90	426.0
271.00	454.0
271.90	615.0
273.00	6592.0
274.00	17128.0
275.00	93912.0
276.00	12329.0
277.00	8672.0
278.00	1734.0
279.00	452.0
281.00	746.0
282.00	470.0
283.00	985.0
284.00	692.0
285.00	1634.0
286.00	400.0
287.90	161.0
288.90	382.0
289.80	404.0
291.00	274.0
292.00	483.0
293.00	1901.0
294.00	650.0
295.00	836.0
296.00	30600.0
297.00	3974.0
298.10	302.0
301.00	438.0
301.90	557.0
303.00	3765.0
304.00	1166.0
307.90	362.0
309.00	369.0
310.00	477.0
312.90	294.0
314.00	1298.0
315.00	3449.0
316.00	1988.0
316.90	218.0
321.00	981.0
322.00	476.0
323.10	9428.0
324.00	1444.0
326.00	156.0
327.00	1794.0
327.90	930.0
328.90	171.0
332.00	664.0
332.90	1064.0
334.00	6814.0
335.00	1314.0
336.00	189.0
338.90	180.0
341.00	955.0
342.00	351.0
345.90	2312.0
346.90	407.0
351.00	181.0
352.00	3139.0
353.00	2206.0
354.00	3190.0
355.00	801.0
359.00	222.0
365.00	15402.0
366.00	2507.0
369.90	321.0
370.90	888.0
372.00	5425.0

Instrument :
BNA_M
ClientSampleId :
DFTPP050

373.00	1030.0
374.00	167.0
383.00	1384.0
383.90	510.0
390.00	505.0
391.00	302.0
392.00	355.0
401.00	366.0
402.00	1834.0
403.00	3230.0
404.00	899.0
405.00	261.0
415.00	207.0
421.00	2448.0
422.00	2423.0
423.10	17800.0
424.00	4886.0
425.00	617.0
439.90	153.0
441.10	47616.0
442.10	304704.0
443.10	58504.0
444.10	5490.0
445.00	319.0

Instrument :
BNA_M
ClientSampleId :
DFTPP050