

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM060823\
 Data File : BM040178.D
 Acq On : 09 Jun 2023 04:09
 Operator : MA/JU
 Sample : 03047-07
 Misc :
 ALS Vial : 29 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 MW-41

Quant Time: Jun 09 04:56:27 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\8270-BM060823.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Jun 08 23:25:50 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.986	152	309905	20.000	ng	0.00
21) Naphthalene-d8	10.804	136	1316416	20.000	ng	0.00
39) Acenaphthene-d10	14.627	164	780113	20.000	ng	0.00
64) Phenanthrene-d10	17.368	188	1682571	20.000	ng	0.00
76) Chrysene-d12	21.544	240	1566761	20.000	ng	0.00
86) Perylene-d12	23.979	264	1915449	20.000	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.534	112	555035	26.742	ng	0.00
7) Phenol-d6	7.139	99	455130	16.070	ng	0.00
23) Nitrobenzene-d5	9.151	82	834981	34.331	ng	0.00
42) 2,4,6-Tribromophenol	16.109	330	617598	52.619	ng	0.00
45) 2-Fluorobiphenyl	13.251	172	2090273	35.415	ng	0.00
79) Terphenyl-d14	19.986	244	3650542	42.951	ng	0.00

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM060823\
Data File : BM040178.D
Acq On : 09 Jun 2023 04:09
Operator : MA/JU
Sample : 03047-07
Misc :
ALS Vial : 29 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
MW-41

Quant Time: Jun 09 04:56:27 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\8270-BM060823.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Jun 08 23:25:50 2023
Response via : Initial Calibration

