

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM060823\
 Data File : BM040173.D
 Acq On : 09 Jun 2023 01:09
 Operator : MA/JU
 Sample : 03039-07
 Misc :
 ALS Vial : 24 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 MW-34

Quant Time: Jun 09 02:26:01 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\8270-BM060823.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Jun 08 23:25:50 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.987	152	245363	20.000	ng	0.00
21) Naphthalene-d8	10.804	136	1061919	20.000	ng	0.00
39) Acenaphthene-d10	14.627	164	574950	20.000	ng	0.00
64) Phenanthrene-d10	17.368	188	1238982	20.000	ng	0.00
76) Chrysene-d12	21.544	240	1093386	20.000	ng	0.00
86) Perylene-d12	23.980	264	1309039	20.000	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.534	112	939680	57.184	ng	0.00
7) Phenol-d6	7.140	99	713640	31.827	ng	0.00
23) Nitrobenzene-d5	9.157	82	1660948	84.658	ng	0.00
42) 2,4,6-Tribromophenol	16.110	330	1132071	130.868	ng	0.00
45) 2-Fluorobiphenyl	13.251	172	3888914	89.400	ng	0.00
79) Terphenyl-d14	19.986	244	7254933	122.315	ng	0.00


Target Compounds	Qvalue


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM060823\
Data File : BM040173.D
Acq On : 09 Jun 2023 01:09
Operator : MA/JU
Sample : 03039-07
Misc :
ALS Vial : 24 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
MW-34

Quant Time: Jun 09 02:26:01 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\8270-BM060823.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Jun 08 23:25:50 2023
Response via : Initial Calibration









#1
1,4-Dichlorobenzene-d4
Concen: 20.000 ng
RT: 7.987 min Scan# 79
Delta R.T. -0.006 min
Lab File: BM040173.D
Acq: 09 Jun 2023 01:09




Instrument :
BNA_M
ClientSampleId :
MW-34


Tgt Ion:152 Resp: 245363
Ion Ratio Lower Upper
152 100
150 157.0 121.4 182.0
115 61.6 48.2 72.4



#5
2-Fluorophenol
Concen: 57.184 ng
RT: 5.534 min Scan# 382
Delta R.T. -0.000 min
Lab File: BM040173.D
Acq: 09 Jun 2023 01:09

Tgt Ion:112 Resp: 939680
Ion Ratio Lower Upper
112 100
64 52.3 43.8 65.6
63 26.6 22.8 34.2









#7
Phenol-d6
Concen: 31.827 ng
RT: 7.140 min Scan# 61
Delta R.T. -0.006 min
Lab File: BM040173.D
Acq: 09 Jun 2023 01:09




Instrument :
BNA_M
ClientSampleId :
MW-34


Tgt Ion: 99 Resp: 713640
Ion Ratio Lower Upper
99 100
42 15.9 13.4 20.2
71 28.3 24.3 36.5



#21
Naphthalene-d8
Concen: 20.000 ng
RT: 10.804 min Scan# 1278
Delta R.T. 0.000 min
Lab File: BM040173.D
Acq: 09 Jun 2023 01:09

Tgt Ion:136 Resp: 1061919
Ion Ratio Lower Upper
136 100
137 10.6 8.8 13.2
54 6.7 6.2 9.4
68 5.9 5.0 7.4









#23
 Nitrobenzene-d5
 Concen: 84.658 ng
 RT: 9.157 min Scan# 998
 Delta R.T. -0.000 min
 Lab File: BM040173.D
 Acq: 09 Jun 2023 01:09




Instrument :
 BNA_M
 ClientSampleId :
 MW-34


Tgt Ion: 82 Resp: 1660948
 Ion Ratio Lower Upper
 82 100
 128 46.5 36.0 54.0
 54 46.0 38.5 57.7



#39
 Acenaphthene-d10
 Concen: 20.000 ng
 RT: 14.627 min Scan# 1928
 Delta R.T. 0.000 min
 Lab File: BM040173.D
 Acq: 09 Jun 2023 01:09

Tgt Ion: 164 Resp: 574950
 Ion Ratio Lower Upper
 164 100
 162 103.0 80.4 120.6
 160 48.0 35.9 53.9









#42
2,4,6-Tribromophenol
Concen: 130.868 ng
RT: 16.110 min Scan# 2181
Delta R.T. -0.006 min
Lab File: BM040173.D
Acq: 09 Jun 2023 01:09




Instrument :
BNA_M
ClientSampleId :
MW-34


Tgt Ion:330 Resp: 1132071
Ion Ratio Lower Upper
330 100
332 96.6 77.1 115.7
141 37.9 31.6 47.4



#45
2-Fluorobiphenyl
Concen: 89.400 ng
RT: 13.251 min Scan# 1694
Delta R.T. -0.006 min
Lab File: BM040173.D
Acq: 09 Jun 2023 01:09

Tgt Ion:172 Resp: 3888914
Ion Ratio Lower Upper
172 100
171 35.9 28.8 43.2
170 24.3 19.8 29.6





#64

Phenanthrene-d10

Concen: 20.000 ng

RT: 17.368 min Scan# 21

Delta R.T. -0.000 min

Lab File: BM040173.D

Acq: 09 Jun 2023 01:09

Instrument :

BNA_M

ClientSampleId :

MW-34





Tgt Ion:188 Resp: 1238982

Ion Ratio Lower Upper

188 100

94 10.4 8.9 13.3

80 11.4 9.6 14.4



#76

Chrysene-d12

Concen: 20.000 ng

RT: 21.544 min Scan# 3104

Delta R.T. -0.006 min

Lab File: BM040173.D

Acq: 09 Jun 2023 01:09




Tgt Ion:240 Resp: 1093386


Ion Ratio Lower Upper

240 100

120 13.1 9.9 14.9

236 26.6 21.1 31.7









#79
 Terphenyl-d14
 Concen: 122.315 ng
 RT: 19.986 min Scan# 2839
 Delta R.T. -0.000 min
 Lab File: BM040173.D
 Acq: 09 Jun 2023 01:09

Instrument :
 BNA_M
 ClientSampleId :
 MW-34

Tgt Ion:244 Resp: 7254933
 Ion Ratio Lower Upper
 244 100
 212 6.9 5.7 8.5
 122 9.6 7.9 11.9



#86
 Perylene-d12
 Concen: 20.000 ng
 RT: 23.980 min Scan# 3518
 Delta R.T. -0.006 min
 Lab File: BM040173.D
 Acq: 09 Jun 2023 01:09

Tgt Ion:264 Resp: 1309039
 Ion Ratio Lower Upper
 264 100
 260 23.6 19.1 28.7
 265 21.7 17.7 26.5

