

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM060921\
 Data File : BM030342.D
 Acq On : 09 Jun 2021 14:53
 Operator : CG/JU
 Sample : M2613-09
 Misc :
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 18-B12(118-122)

Quant Time: Jun 09 15:28:29 2021
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\8270-BM060921.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Jun 08 20:51:08 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.851	152	175516	20.000	ng	0.00
21) Naphthalene-d8	10.639	136	743370	20.000	ng	# 0.00
39) Acenaphthene-d10	14.474	164	514688	20.000	ng	0.00
64) Phenanthrene-d10	17.204	188	1212327	20.000	ng	# 0.00
76) Chrysene-d12	21.368	240	1407761	20.000	ng	0.00
86) Perylene-d12	23.650	264	1282920	20.000	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.428	112	1569874	145.404	ng	0.00
7) Phenol-d6	7.010	99	2119827	135.629	ng	0.00
23) Nitrobenzene-d5	9.004	82	1289639	84.518	ng	0.00
42) 2,4,6-Tribromophenol	15.957	330	924224	142.824	ng	0.00
45) 2-Fluorobiphenyl	13.098	172	3162146	80.795	ng	0.00
79) Terphenyl-d14	19.821	244	6113314	76.275	ng	0.00
Target Compounds						
84) Bis(2-ethylhexyl)phtha...	21.280	149	19949	4.548	ng	Qvalue 98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM060921\
Data File : BM030342.D
Acq On : 09 Jun 2021 14:53
Operator : CG/JU
Sample : M2613-09
Misc :
ALS Vial : 11 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
18-B12(118-122)

Quant Time: Jun 09 15:28:29 2021
Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\8270-BM060921.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Jun 08 20:51:08 2021
Response via : Initial Calibration

