



2-Nitrophenol




Response = $2.010 \times 10^{-1} \times \text{Amt} - 3.249 \times 10^{-2}$
 Coef of Det (r^2) = 0.992135 Curve Fit: Linear
 Method Name: Z:\svoasrv\HPCHEM1\BNA M\Methods\8270-BM060921.M
 Calibration Table Last Updated: Tue Jun 08 20:51:08 2021

Benzoic acid



$\text{Response} = 2.639\text{e-}001 * \text{Amt} - 6.582\text{e-}002$
 Coef of Det (r^2) = 0.992517 Curve Fit: Linear
 Method Name: Z:\svoasrv\HPCHEM1\BNA M\Methods\8270-BM060921.M
 Calibration Table Last Updated: Tue Jun 08 20:51:08 2021


Caprolactam



Response = 1.343e-001 * Amt - 4.011e-002
 Coef of Det (r^2) = 0.992979 Curve Fit: Linear
 Method Name: Z:\svoasrv\HPCHEM1\BNA M\Methods\8270-BM060921.M
 Calibration Table Last Updated: Tue Jun 08 20:51:08 2021

2-Nitroaniline

Response Ratio



$$\text{Response} = 4.354\text{e-}001 * \text{Amt} - 8.753\text{e-}002$$


Coef of Det (r^2) = 0.993518 Curve Fit: Linear

Method Name: Z:\svoasrv\HPCHEM1\BNA M\Methods\8270-BM060921.M

Calibration Table Last Updated: Tue Jun 08 20:51:08 2021

3-Nitroaniline

Response Ratio



$$\text{Response} = 4.268\text{e-}001 * \text{Amt} - 8.455\text{e-}002$$


Coef of Det (r^2) = 0.991242 Curve Fit: Linear

Method Name: Z:\svoasrv\HPCHEM1\BNA M\Methods\8270-BM060921.M

Calibration Table Last Updated: Tue Jun 08 20:51:08 2021

2,4-Dinitrophenol


Response Ratio



$R = 3.342e-002 A^2 + 1.201e-001 A + 5.156e-004$
 Coef of Det (r^2) = 0.998806 Curve Fit: Quadratic
 Method Name: Z:\svoasrv\HPCHEM1\BNA M\Methods\8270-BM060921.M
 Calibration Table Last Updated: Tue Jun 08 20:51:08 2021

2,4-Dinitrotoluene

Response Ratio



$$\text{Response} = 5.290\text{e-}001 * \text{Amt} - 7.922\text{e-}002$$


Coef of Det (r²) = 0.993210 Curve Fit: Linear

Method Name: Z:\svoasrv\HPCHEM1\BNA M\Methods\8270-BM060921.M

Calibration Table Last Updated: Tue Jun 08 20:51:08 2021

4-Nitroaniline

Response Ratio




$$\text{Response} = 4.583\text{e-}001 * \text{Amt} - 1.057\text{e-}001$$

Coef of Det (r^2) = 0.990656 Curve Fit: Linear

Method Name: Z:\svoasrv\HPCHEM1\BNA M\Methods\8270-BM060921.M

Calibration Table Last Updated: Tue Jun 08 20:51:08 2021


4,6-Dinitro-2-methylphenol



$R = 1.823e-002 A^2 + 9.558e-002 A - 1.589e-003$
 Coef of Det (r^2) = 0.997947 Curve Fit: Quadratic
 Method Name: Z:\svoasrv\HPCHEM1\BNA M\Methods\8270-BM060921.M
 Calibration Table Last Updated: Tue Jun 08 20:51:08 2021

Butylbenzylphthalate

Response Ratio



$$\text{Response} = 6.897\text{e-}001 * \text{Amt} - 1.485\text{e-}001$$


Coef of Det (r^2) = 0.994222 Curve Fit: Linear

Method Name: Z:\svoasrv\HPCHEM1\BNA M\Methods\8270-BM060921.M

Calibration Table Last Updated: Tue Jun 08 20:51:08 2021

Bis(2-ethylhexyl)phthalate


Response Ratio



Response = 1.037e+000 * Amt - 2.216e-001
 Coef of Det (r^2) = 0.995132 Curve Fit: Linear
 Method Name: Z:\svoasrv\HPCHEM1\BNA M\Methods\8270-BM060921.M
 Calibration Table Last Updated: Tue Jun 08 20:51:08 2021

Di-n-octyl phthalate

Response Ratio



$$\text{Response} = 1.593\text{e}+000 * \text{Amt} - 3.318\text{e}-001$$

Coef of Det (r^2) = 0.993344 Curve Fit: Linear

Method Name: Z:\svoasrv\HPCHEM1\BNA M\Methods\8270-BM060921.M

Calibration Table Last Updated: Tue Jun 08 20:51:08 2021