


Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM060921\
 Data File : BM030345.D
 Acq On : 09 Jun 2021 16:48
 Operator : CG/JU
 Sample : DFTPP
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 DFTPP

Integration File: rteint.p

Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\8270-BM060921.M
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 Last Update : Tue Jun 08 20:51:08 2021






AutoFind: Scans 2474, 2475, 2476; Background Corrected with Scan 2467

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	41.6	123811	PASS
68	69	0.00	2	1.7	2037	PASS
69	198	0.00	100	41.0	121931	PASS
70	69	0.00	2	0.4	507	PASS
127	198	10	80	48.6	144477	PASS
197	198	0.00	2	0.5	1429	PASS
198	198	100	100	100.0	297280	PASS
199	198	5	9	6.8	20219	PASS
275	198	10	60	27.3	81072	PASS
365	198	1	100	3.8	11156	PASS
441	198	0.01	100	10.1	30112	PASS
442	442	50	100	100.0	188693	PASS
443	442	15	24	19.6	36941	PASS

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM060921\
 Data File : BM030345.D
 Acq On : 09 Jun 2021 16:48
 Operator : CG/JU
 Sample : DFTPP
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 DFTPP

Quant Time: Jun 09 17:25:15 2021
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\8270-BM060921.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Jun 08 20:51:08 2021
 Response via : Initial Calibration



TIC: BM030345.D\data.ms

(70) Pentachlorophenol (C)

16.857min (+ 0.001) 167685.70 ng




response 399435

Ion	Exp%	Act%
265.70	100.00	100.00
268.00	66.00	64.81
264.00	64.90	64.01
0.00	0.00	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM060921\
Data File : BM030345.D
Acq On : 09 Jun 2021 16:48
Operator : CG/JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
DFTPP

Quant Time: Jun 09 17:25:15 2021
Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\8270-BM060921.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Jun 08 20:51:08 2021
Response via : Initial Calibration



TIC: BM030345.D\data.ms

(77) Benzidine

19.451min (+ 0.012) 61728.83 ng

response 990626

Ion	Exp%	Act%
184.00	100.00	100.00
185.00	13.60	14.15
183.00	12.30	11.62
0.00	0.00	0.00

DDT Breakdown

Instrument :
BNA_M
ClientSampleId :
DFTPP

Date	Instrument Name	DFTPP Data File
6/9/2021	BNA_M	<u>BM030345.D</u>
Compound Name	Response	Retention Time
DDT	1310995	20.66
DDD	7334	20.23
DDE	1596	19.72
SUM(DDD+DDE)	SUM(DDT+DDD+DDE)	% Breakdown Of DDT
8930	1319925	0.68