

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM061221\
 Data File : BM030411.D
 Acq On : 12 Jun 2021 12:46
 Operator : CG/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 SSTDCCC040

Quant Time: Jun 14 00:58:24 2021
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\8270-BM060921.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Jun 11 14:10:50 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.834	152	179629	20.000	ng	0.00
21) Naphthalene-d8	10.622	136	756171	20.000	ng	0.00
39) Acenaphthene-d10	14.457	164	534878	20.000	ng	0.00
64) Phenanthrene-d10	17.192	188	1156320	20.000	ng	0.00
76) Chrysene-d12	21.362	240	1367000	20.000	ng	0.00
86) Perylene-d12	23.639	264	1449091	20.000	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.416	112	819086	74.128	ng	0.00
7) Phenol-d6	6.998	99	1230915	76.952	ng	0.00
23) Nitrobenzene-d5	8.992	82	1140455	73.476	ng	0.00
42) 2,4,6-Tribromophenol	15.945	330	596194	88.655	ng	0.00
45) 2-Fluorobiphenyl	13.080	172	2995917	73.658	ng	0.00
79) Terphenyl-d14	19.809	244	5383738	69.175	ng	0.00
Target Compounds						
						Qvalue
2) 1,4-Dioxane	3.340	88	164328	34.402	ng	90
3) Pyridine	3.740	79	463095	37.978	ng	# 87
4) n-Nitrosodimethylamine	3.646	42	207640	33.544	ng	# 80
6) Aniline	7.163	93	682266	37.376	ng	95
8) 2-Chlorophenol	7.398	128	490222	37.800	ng	98
9) Benzaldehyde	6.981	77	270406	40.473	ng	91
10) Phenol	7.022	94	600829	37.182	ng	92
11) bis(2-Chloroethyl)ether	7.257	93	456132	35.826	ng	96
12) 1,3-Dichlorobenzene	7.722	146	521654	36.509	ng	98
13) 1,4-Dichlorobenzene	7.869	146	528461	36.293	ng	99
14) 1,2-Dichlorobenzene	8.187	146	519004	36.641	ng	100
15) Benzyl Alcohol	8.075	79	424625	37.826	ng	93
16) 2,2'-oxybis(1-Chloropr...	8.363	45	658996	32.166	ng	91
17) 2-Methylphenol	8.269	107	405134	38.688	ng	93
18) Hexachloroethane	8.910	117	183102	34.988	ng	98
19) n-Nitroso-di-n-propyla...	8.640	70	395170	37.209	ng	98
20) 3+4-Methylphenols	8.598	107	557557	39.134	ng	96
22) Acetophenone	8.651	105	756065	37.423	ng	99
24) Nitrobenzene	9.034	77	570797	35.253	ng	95
25) Isophorone	9.557	82	1076099	36.173	ng	99
26) 2-Nitrophenol	9.739	139	272191	39.041	ng	93
27) 2,4-Dimethylphenol	9.798	122	430550	37.390	ng	97
28) bis(2-Chloroethoxy)met...	10.034	93	586187	34.656	ng	100
29) 2,4-Dichlorophenol	10.269	162	512672	39.715	ng	98
30) 1,2,4-Trichlorobenzene	10.486	180	558181	37.882	ng	98
31) Naphthalene	10.675	128	1593062	36.166	ng	99
32) Benzoic acid	9.939	122	300426	35.098	ng	91
33) 4-Chloroaniline	10.781	127	688531	38.409	ng	97
34) Hexachlorobutadiene	10.957	225	337999	38.428	ng	97
35) Caprolactam	11.569	113	167747	39.002	ng	# 87
36) 4-Chloro-3-methylphenol	11.898	107	527438	39.007	ng	99
37) 2-Methylnaphthalene	12.281	142	1214489	37.578	ng	98
38) 1-Methylnaphthalene	12.504	142	1152073	37.426	ng	99
40) 1,2,4,5-Tetrachloroben...	12.651	216	666723	38.672	ng	98
41) Hexachlorocyclopentadiene	12.633	237	336383	35.585	ng	99
43) 2,4,6-Trichlorophenol	12.892	196	465881	39.971	ng	100

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM061221\
 Data File : BM030411.D
 Acq On : 12 Jun 2021 12:46
 Operator : CG/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1


Instrument :
 BNA_M
 ClientSampleId :
 SSTDCCC040


Quant Time: Jun 14 00:58:24 2021
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\8270-BM060921.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Jun 11 14:10:50 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
44) 2,4,5-Trichlorophenol	12.957	196	505061	39.505	ng	99
46) 1,1'-Biphenyl	13.292	154	1568357	36.208	ng	99
47) 2-Chloronaphthalene	13.333	162	1254856	35.841	ng	98
48) 2-Nitroaniline	13.539	65	339429	33.172	ng	93
49) Acenaphthylene	14.180	152	1968481	36.701	ng	99
50) Dimethylphthalate	13.916	163	1549746	36.407	ng	99
51) 2,6-Dinitrotoluene	14.033	165	359671	38.330	ng	93
52) Acenaphthene	14.522	154	1215360	35.475	ng	98
53) 3-Nitroaniline	14.363	138	360564	35.551	ng	97
54) 2,4-Dinitrophenol	14.569	184	206555	40.928	ng	85
55) Dibenzofuran	14.857	168	1987863	36.466	ng	99
56) 4-Nitrophenol	14.663	139	310388	38.954	ng	90
57) 2,4-Dinitrotoluene	14.822	165	501308	38.431	ng	92
58) Fluorene	15.504	166	1669131	37.114	ng	100
59) 2,3,4,6-Tetrachlorophenol	15.080	232	462940	41.838	ng	# 94
60) Diethylphthalate	15.280	149	1548464	36.389	ng	99
61) 4-Chlorophenyl-phenyle...	15.498	204	831470	37.514	ng	98
62) 4-Nitroaniline	15.521	138	390624	36.485	ng	87
63) Azobenzene	15.792	77	1468749	34.173	ng	95
65) 4,6-Dinitro-2-methylph...	15.580	198	311600	40.828	ng	97
66) n-Nitrosodiphenylamine	15.710	169	1435346	35.694	ng	99
67) 4-Bromophenyl-phenylether	16.392	248	501088	37.186	ng	94
68) Hexachlorobenzene	16.510	284	581661	38.068	ng	94
69) Atrazine	16.663	200	483908	44.933	ng	95
70) Pentachlorophenol	16.845	266	382377	42.758	ng	98
71) Phenanthrene	17.239	178	2655850	36.501	ng	100
72) Anthracene	17.327	178	2633363	37.275	ng	100
73) Carbazole	17.598	167	2542503	37.738	ng	100
74) Di-n-butylphthalate	18.157	149	2665141	37.944	ng	99
75) Fluoranthene	19.245	202	3301747	39.978	ng	98
77) Benzidine	19.427	184	1276597	52.780	ng	99
78) Pyrene	19.609	202	3465656	33.457	ng	99
80) Butylbenzylphthalate	20.498	149	1286025	31.587	ng	97
81) Benzo(a)anthracene	21.345	228	3668075	36.868	ng	99
82) 3,3'-Dichlorobenzidine	21.274	252	1217733	40.431	ng	99
83) Chrysene	21.398	228	3561897	36.839	ng	99
84) Bis(2-ethylhexyl)phtha...	21.274	149	1970249	32.080	ng	100
85) Di-n-octyl phthalate	22.156	149	3338816	34.823	ng	99
87) Indeno(1,2,3-cd)pyrene	25.956	276	4191290	38.027	ng	99
88) Benzo(b)fluoranthene	22.950	252	3948066	35.998	ng	99
89) Benzo(k)fluoranthene	22.997	252	3700945	34.550	ng	99
90) Benzo(a)pyrene	23.539	252	3579953	36.913	ng	99
91) Dibenzo(a,h)anthracene	25.968	278	3599443	37.877	ng	99
92) Benzo(g,h,i)perylene	26.668	276	3400074	37.616	ng	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Quant Time: Jun 14 00:58:24 2021
Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\8270-BM060921.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Jun 11 14:10:50 2021
Response via : Initial Calibration









#1
1,4-Dichlorobenzene-d4
Concen: 20.000 ng
RT: 7.834 min Scan# 807
Delta R.T. -0.000 min
Lab File: BM030411.D
Acq: 12 Jun 2021 12:46




Instrument :
BNA_M
ClientSampleId :
SSTDCCC040


Tgt Ion:152 Resp: 179629
Ion Ratio Lower Upper
152 100
150 155.3 125.9 188.9
115 55.6 44.6 67.0



#2
1,4-Dioxane
Concen: 34.402 ng
RT: 3.340 min Scan# 43
Delta R.T. 0.000 min
Lab File: BM030411.D
Acq: 12 Jun 2021 12:46

Tgt Ion: 88 Resp: 164328
Ion Ratio Lower Upper
88 100
58 76.9 53.6 80.4
43 30.2 25.8 38.6









#3
Pyridine
Concen: 37.978 ng
RT: 3.740 min Scan# 111
Delta R.T. 0.000 min
Lab File: BM030411.D
Acq: 12 Jun 2021 12:46




Instrument :
BNA_M
ClientSampleId :
SSTDCCC040


Tgt Ion: 79 Resp: 463095
Ion Ratio Lower Upper
79 100
52 72.4 48.2 72.2#
51 39.2 27.9 41.9



#4
n-Nitrosodimethylamine
Concen: 33.544 ng
RT: 3.646 min Scan# 95
Delta R.T. 0.000 min
Lab File: BM030411.D
Acq: 12 Jun 2021 12:46

Tgt Ion: 42 Resp: 207640
Ion Ratio Lower Upper
42 100
74 126.8 84.1 126.1#
44 6.8 4.5 6.7#









#5
2-Fluorophenol
Concen: 74.128 ng
RT: 5.416 min Scan# 396
Delta R.T. 0.000 min
Lab File: BM030411.D
Acq: 12 Jun 2021 12:46




Instrument :
BNA_M
ClientSampleId :
SSTDCCC040


Tgt Ion:112 Resp: 819086
Ion Ratio Lower Upper
112 100
64 57.6 47.1 70.7
63 29.9 27.3 40.9



#6
Aniline
Concen: 37.376 ng
RT: 7.163 min Scan# 693
Delta R.T. -0.006 min
Lab File: BM030411.D
Acq: 12 Jun 2021 12:46

Tgt Ion: 93 Resp: 682266
Ion Ratio Lower Upper
93 100
66 37.1 32.7 49.1
65 19.7 16.8 25.2









#7
Phenol-d6
Concen: 76.952 ng
RT: 6.998 min Scan# 665
Delta R.T. 0.000 min
Lab File: BM030411.D
Acq: 12 Jun 2021 12:46

Instrument :
BNA_M
ClientSampleId :
SSTDCCC040

Tgt Ion: 99 Resp: 1230915




Ion	Ratio	Lower	Upper
99	100		
42	19.0	18.0	27.0
71	33.6	30.9	46.3




#8
2-Chlorophenol
Concen: 37.800 ng
RT: 7.398 min Scan# 733
Delta R.T. 0.000 min
Lab File: BM030411.D
Acq: 12 Jun 2021 12:46

Tgt Ion: 128 Resp: 490222

Ion	Ratio	Lower	Upper
128	100		
130	31.5	13.5	53.5
64	46.0	25.7	65.7









#9
Benzaldehyde
Concen: 40.473 ng
RT: 6.981 min Scan# 662
Delta R.T. 0.000 min
Lab File: BM030411.D
Acq: 12 Jun 2021 12:46




Instrument :
BNA_M
ClientSampleId :
SSTDCCC040


Tgt Ion: 77 Resp: 270406
Ion Ratio Lower Upper
77 100
105 100.6 71.2 111.2
106 96.8 68.6 108.6



#10
Phenol
Concen: 37.182 ng
RT: 7.022 min Scan# 669
Delta R.T. -0.006 min
Lab File: BM030411.D
Acq: 12 Jun 2021 12:46

Tgt Ion: 94 Resp: 600829
Ion Ratio Lower Upper
94 100
65 28.2 11.9 51.9
66 37.0 22.6 62.6









#11
bis(2-Chloroethyl)ether
Concen: 35.826 ng
RT: 7.257 min Scan# 709
Delta R.T. -0.006 min
Lab File: BM030411.D
Acq: 12 Jun 2021 12:46




Instrument :
BNA_M
ClientSampleId :
SSTDCCC040


Tgt Ion: 93 Resp: 456132
Ion Ratio Lower Upper
93 100
63 78.4 54.3 94.3
95 33.1 12.7 52.7



#12
1,3-Dichlorobenzene
Concen: 36.509 ng
RT: 7.722 min Scan# 788
Delta R.T. -0.006 min
Lab File: BM030411.D
Acq: 12 Jun 2021 12:46

Tgt Ion: 146 Resp: 521654
Ion Ratio Lower Upper
146 100
148 65.3 52.2 78.4
75 29.6 25.8 38.8









#13
1,4-Dichlorobenzene
Concen: 36.293 ng
RT: 7.869 min Scan# 813
Delta R.T. 0.000 min
Lab File: BM030411.D
Acq: 12 Jun 2021 12:46




Instrument :
BNA_M
ClientSampleId :
SSTDCCC040


Tgt Ion:146 Resp: 528461
Ion Ratio Lower Upper
146 100
148 64.3 51.8 77.8
111 37.9 31.0 46.6



#14
1,2-Dichlorobenzene
Concen: 36.641 ng
RT: 8.187 min Scan# 867
Delta R.T. 0.000 min
Lab File: BM030411.D
Acq: 12 Jun 2021 12:46

Tgt Ion:146 Resp: 519004
Ion Ratio Lower Upper
146 100
148 65.1 51.7 77.5
111 40.7 32.6 49.0









#15
Benzyl Alcohol
Concen: 37.826 ng
RT: 8.075 min Scan# 848
Delta R.T. 0.000 min
Lab File: BM030411.D
Acq: 12 Jun 2021 12:46




Instrument :
BNA_M
ClientSampleId :
SSTDCCC040


Tgt Ion: 79 Resp: 424625
Ion Ratio Lower Upper
79 100
108 79.3 57.0 85.4
77 65.6 54.6 81.8



#16
2,2'-oxybis(1-Chloropropane)
Concen: 32.166 ng
RT: 8.363 min Scan# 897
Delta R.T. -0.006 min
Lab File: BM030411.D
Acq: 12 Jun 2021 12:46

Tgt Ion: 45 Resp: 658996
Ion Ratio Lower Upper
45 100
77 14.0 0.0 38.0
79 10.8 0.0 34.1









#17
2-Methylphenol
Concen: 38.688 ng
RT: 8.269 min Scan# 881
Delta R.T. 0.000 min
Lab File: BM030411.D
Acq: 12 Jun 2021 12:46

Instrument :
BNA_M
ClientSampleId :
SSTDCCC040

Tgt Ion:107 Resp: 405134




Ion	Ratio	Lower	Upper
107	100		
108	111.1	85.0	127.6
77	45.4	41.3	61.9
79	43.5	40.7	61.1




#18
Hexachloroethane
Concen: 34.988 ng
RT: 8.910 min Scan# 990
Delta R.T. -0.006 min
Lab File: BM030411.D
Acq: 12 Jun 2021 12:46

Tgt Ion:117 Resp: 183102

Ion	Ratio	Lower	Upper
117	100		
119	97.0	76.2	114.2
201	110.8	89.8	134.6









#19
n-Nitroso-di-n-propylamine
Concen: 37.209 ng
RT: 8.640 min Scan# 944
Delta R.T. 0.000 min
Lab File: BM030411.D
Acq: 12 Jun 2021 12:46

Instrument :
BNA_M
ClientSampleId :
SSTDCCC040

Tgt Ion: 70 Resp: 395170




Ion	Ratio	Lower	Upper
70	100		
42	61.8	51.4	77.0
101	10.3	8.1	12.1
130	22.7	17.8	26.6




#20
3+4-Methylphenols
Concen: 39.134 ng
RT: 8.598 min Scan# 937
Delta R.T. 0.000 min
Lab File: BM030411.D
Acq: 12 Jun 2021 12:46

Tgt Ion: 107 Resp: 557557

Ion	Ratio	Lower	Upper
107	100		
108	89.7	69.9	109.9
77	32.9	17.3	57.3
79	28.3	13.3	53.3









#21
Naphthalene-d8
Concen: 20.000 ng
RT: 10.622 min Scan# 1281
Delta R.T. -0.006 min
Lab File: BM030411.D
Acq: 12 Jun 2021 12:46

Instrument :
BNA_M
ClientSampleId :
SSTDCCC040

Tgt Ion:136 Resp: 756171




Ion	Ratio	Lower	Upper
136	100		
137	11.1	8.4	12.6
54	8.0	5.5	8.3
68	6.0	4.2	6.2




#22
Acetophenone
Concen: 37.423 ng
RT: 8.651 min Scan# 946
Delta R.T. -0.006 min
Lab File: BM030411.D
Acq: 12 Jun 2021 12:46

Tgt Ion:105 Resp: 756065

Ion	Ratio	Lower	Upper
105	100		
71	4.7	3.4	5.0
51	28.0	22.2	33.4
120	22.3	17.6	26.4









#23
Nitrobenzene-d5
Concen: 73.476 ng
RT: 8.992 min Scan# 1004
Delta R.T. 0.000 min
Lab File: BM030411.D
Acq: 12 Jun 2021 12:46




Instrument :
BNA_M
ClientSampleId :
SSTDCCC040


Tgt Ion: 82 Resp: 1140455
Ion Ratio Lower Upper
82 100
128 42.6 31.1 46.7
54 49.9 37.3 55.9



#24
Nitrobenzene
Concen: 35.253 ng
RT: 9.034 min Scan# 1011
Delta R.T. 0.000 min
Lab File: BM030411.D
Acq: 12 Jun 2021 12:46

Tgt Ion: 77 Resp: 570797
Ion Ratio Lower Upper
77 100
123 46.1 34.0 51.0
65 13.2 11.0 16.4









#25
Isophorone
Concen: 36.173 ng
RT: 9.557 min Scan# 1100
Delta R.T. 0.000 min
Lab File: BM030411.D
Acq: 12 Jun 2021 12:46




Instrument :
BNA_M
ClientSampleId :
SSTDCCC040


Tgt Ion: 82 Resp: 1076099
Ion Ratio Lower Upper
82 100
95 6.9 6.3 9.5
138 18.5 14.5 21.7



#26
2-Nitrophenol
Concen: 39.041 ng
RT: 9.739 min Scan# 1131
Delta R.T. 0.000 min
Lab File: BM030411.D
Acq: 12 Jun 2021 12:46

Tgt Ion: 139 Resp: 272191
Ion Ratio Lower Upper
139 100
109 29.4 26.7 40.1
65 47.7 41.9 62.9









#27
2,4-Dimethylphenol
Concen: 37.390 ng
RT: 9.798 min Scan# 1141
Delta R.T. 0.000 min
Lab File: BM030411.D
Acq: 12 Jun 2021 12:46




Instrument :
BNA_M
ClientSampleId :
SSTDCCC040


Tgt Ion:122 Resp: 430550
Ion Ratio Lower Upper
122 100
107 114.0 95.0 142.6
121 59.2 48.1 72.1



#28
bis(2-Chloroethoxy)methane
Concen: 34.656 ng
RT: 10.034 min Scan# 1181
Delta R.T. 0.000 min
Lab File: BM030411.D
Acq: 12 Jun 2021 12:46

Tgt Ion: 93 Resp: 586187
Ion Ratio Lower Upper
93 100
95 32.6 26.2 39.2
123 11.5 9.3 13.9









#29
2,4-Dichlorophenol
Concen: 39.715 ng
RT: 10.269 min Scan# 1221
Delta R.T. 0.000 min
Lab File: BM030411.D
Acq: 12 Jun 2021 12:46




Instrument :
BNA_M
ClientSampleId :
SSTDCCC040


Tgt Ion:162 Resp: 512672
Ion Ratio Lower Upper
162 100
164 65.4 46.5 86.5
98 33.6 12.1 52.1



#30
1,2,4-Trichlorobenzene
Concen: 37.882 ng
RT: 10.486 min Scan# 1258
Delta R.T. 0.000 min
Lab File: BM030411.D
Acq: 12 Jun 2021 12:46

Tgt Ion:180 Resp: 558181
Ion Ratio Lower Upper
180 100
182 97.1 76.3 114.5
145 27.4 22.3 33.5









#31
Naphthalene
Concen: 36.166 ng
RT: 10.675 min Scan# 1290
Delta R.T. 0.000 min
Lab File: BM030411.D
Acq: 12 Jun 2021 12:46




Instrument :
BNA_M
ClientSampleId :
SSTDCCC040


Tgt Ion:128 Resp: 1593062
Ion Ratio Lower Upper
128 100
129 11.0 9.0 13.6
127 13.1 10.8 16.2



#32
Benzoic acid
Concen: 35.098 ng
RT: 9.939 min Scan# 1165
Delta R.T. 0.006 min
Lab File: BM030411.D
Acq: 12 Jun 2021 12:46

Tgt Ion:122 Resp: 300426
Ion Ratio Lower Upper
122 100
105 123.4 107.3 147.3
77 91.6 88.0 128.0









#33
4-Chloroaniline
Concen: 38.409 ng
RT: 10.781 min Scan# 1308
Delta R.T. 0.000 min
Lab File: BM030411.D
Acq: 12 Jun 2021 12:46

Instrument :
BNA_M
ClientSampleId :
SSTDCCC040

Tgt Ion:127 Resp: 688531




Ion	Ratio	Lower	Upper
127	100		
129	32.7	26.8	40.2
65	28.9	25.2	37.8
92	18.2	14.8	22.2




#34
Hexachlorobutadiene
Concen: 38.428 ng
RT: 10.957 min Scan# 1338
Delta R.T. -0.006 min
Lab File: BM030411.D
Acq: 12 Jun 2021 12:46

Tgt Ion:225 Resp: 337999

Ion	Ratio	Lower	Upper
225	100		
223	63.1	52.1	78.1
227	64.7	53.8	80.6









#35
Caprolactam
Concen: 39.002 ng
RT: 11.569 min Scan# 1442
Delta R.T. 0.000 min
Lab File: BM030411.D
Acq: 12 Jun 2021 12:46




Instrument :
BNA_M
ClientSampleId :
SSTDCCC040


Tgt Ion:113 Resp: 167747
Ion Ratio Lower Upper
113 100
55 173.4 132.0 172.0#
56 125.1 96.3 136.3



#36
4-Chloro-3-methylphenol
Concen: 39.007 ng
RT: 11.898 min Scan# 1498
Delta R.T. 0.000 min
Lab File: BM030411.D
Acq: 12 Jun 2021 12:46

Tgt Ion:107 Resp: 527438
Ion Ratio Lower Upper
107 100
144 27.7 22.3 33.5
142 86.0 69.3 103.9









#37
2-Methylnaphthalene
Concen: 37.578 ng
RT: 12.281 min Scan# 1563
Delta R.T. -0.006 min
Lab File: BM030411.D
Acq: 12 Jun 2021 12:46




Instrument :
BNA_M
ClientSampleId :
SSTDCCC040


Tgt Ion:142 Resp: 1214489
Ion Ratio Lower Upper
142 100
141 88.6 71.7 107.5
115 34.4 29.1 43.7



#38
1-Methylnaphthalene
Concen: 37.426 ng
RT: 12.504 min Scan# 1601
Delta R.T. 0.000 min
Lab File: BM030411.D
Acq: 12 Jun 2021 12:46

Tgt Ion:142 Resp: 1152073
Ion Ratio Lower Upper
142 100
141 91.7 74.2 111.4
116 3.7 3.0 4.6









#39
Acenaphthene-d10
Concen: 20.000 ng
RT: 14.457 min Scan# 1933
Delta R.T. -0.006 min
Lab File: BM030411.D
Acq: 12 Jun 2021 12:46




Instrument :
BNA_M
ClientSampleId :
SSTDCCC040


Tgt Ion:164 Resp: 534878
Ion Ratio Lower Upper
164 100
162 101.0 83.4 125.0
160 45.3 37.2 55.8



#40
1,2,4,5-Tetrachlorobenzene
Concen: 38.672 ng
RT: 12.651 min Scan# 1626
Delta R.T. 0.000 min
Lab File: BM030411.D
Acq: 12 Jun 2021 12:46

Tgt Ion:216 Resp: 666723
Ion Ratio Lower Upper
216 100
214 77.9 63.0 94.4
179 19.6 17.6 26.4
108 16.4 14.9 22.3









#41
Hexachlorocyclopentadiene
Concen: 35.585 ng
RT: 12.633 min Scan# 1623
Delta R.T. 0.000 min
Lab File: BM030411.D
Acq: 12 Jun 2021 12:46




Instrument :
BNA_M
ClientSampleId :
SSTDCCC040


Tgt Ion:237 Resp: 336383
Ion Ratio Lower Upper
237 100
235 61.5 42.4 82.4
272 11.9 0.0 31.7



#42
2,4,6-Tribromophenol
Concen: 88.655 ng
RT: 15.945 min Scan# 2186
Delta R.T. 0.000 min
Lab File: BM030411.D
Acq: 12 Jun 2021 12:46

Tgt Ion:330 Resp: 596194
Ion Ratio Lower Upper
330 100
332 96.1 77.0 115.4
141 33.8 30.2 45.2









#43
2,4,6-Trichlorophenol
Concen: 39.971 ng
RT: 12.892 min Scan# 1667
Delta R.T. 0.000 min
Lab File: BM030411.D
Acq: 12 Jun 2021 12:46




Instrument :
BNA_M
ClientSampleId :
SSTDCCC040


Tgt Ion:196 Resp: 465881
Ion Ratio Lower Upper
196 100
198 97.4 77.6 116.4
200 30.8 24.6 37.0



#44
2,4,5-Trichlorophenol
Concen: 39.505 ng
RT: 12.957 min Scan# 1678
Delta R.T. 0.000 min
Lab File: BM030411.D
Acq: 12 Jun 2021 12:46

Tgt Ion:196 Resp: 505061
Ion Ratio Lower Upper
196 100
198 96.1 77.0 115.4
97 45.7 36.6 55.0
132 24.1 18.2 27.2









#45
2-Fluorobiphenyl
Concen: 73.658 ng
RT: 13.080 min Scan# 1699
Delta R.T. -0.006 min
Lab File: BM030411.D
Acq: 12 Jun 2021 12:46




Instrument :
BNA_M
ClientSampleId :
SSTDCCC040


Tgt Ion:172 Resp: 2995917
Ion Ratio Lower Upper
172 100
171 35.7 28.6 43.0
170 23.5 18.4 27.6



#46
1,1'-Biphenyl
Concen: 36.208 ng
RT: 13.292 min Scan# 1735
Delta R.T. 0.000 min
Lab File: BM030411.D
Acq: 12 Jun 2021 12:46

Tgt Ion:154 Resp: 1568357
Ion Ratio Lower Upper
154 100
153 40.3 20.7 60.7
76 12.2 0.0 31.0









#47
2-Chloronaphthalene
Concen: 35.841 ng
RT: 13.333 min Scan# 1742
Delta R.T. -0.006 min
Lab File: BM030411.D
Acq: 12 Jun 2021 12:46




Instrument :
BNA_M
ClientSampleId :
SSTDCCC040


Tgt Ion:162 Resp: 1254856
Ion Ratio Lower Upper
162 100
127 35.6 27.1 40.7
164 32.8 26.1 39.1



#48
2-Nitroaniline
Concen: 33.172 ng
RT: 13.539 min Scan# 1777
Delta R.T. 0.000 min
Lab File: BM030411.D
Acq: 12 Jun 2021 12:46

Tgt Ion: 65 Resp: 339429
Ion Ratio Lower Upper
65 100
92 71.4 55.5 83.3
138 117.2 85.7 128.5









#49
Acenaphthylene
Concen: 36.701 ng
RT: 14.180 min Scan# 1886
Delta R.T. 0.000 min
Lab File: BM030411.D
Acq: 12 Jun 2021 12:46




Instrument :
BNA_M
ClientSampleId :
SSTDCCC040


Tgt Ion:152 Resp: 1968481
Ion Ratio Lower Upper
152 100
151 19.9 16.3 24.5
153 13.2 10.2 15.4



#50
Dimethylphthalate
Concen: 36.407 ng
RT: 13.916 min Scan# 1841
Delta R.T. -0.006 min
Lab File: BM030411.D
Acq: 12 Jun 2021 12:46

Tgt Ion:163 Resp: 1549746
Ion Ratio Lower Upper
163 100
194 4.1 3.3 4.9
164 10.4 7.9 11.9









#51
2,6-Dinitrotoluene
Concen: 38.330 ng
RT: 14.033 min Scan# 1861
Delta R.T. 0.000 min
Lab File: BM030411.D
Acq: 12 Jun 2021 12:46




Instrument :
BNA_M
ClientSampleId :
SSTDCCC040


Tgt Ion:165 Resp: 359671
Ion Ratio Lower Upper
165 100
63 52.4 45.3 67.9
89 52.0 46.3 69.5



#52
Acenaphthene
Concen: 35.475 ng
RT: 14.522 min Scan# 1944
Delta R.T. -0.006 min
Lab File: BM030411.D
Acq: 12 Jun 2021 12:46

Tgt Ion:154 Resp: 1215360
Ion Ratio Lower Upper
154 100
153 112.9 92.2 138.2
152 53.5 43.4 65.2









#53
3-Nitroaniline
Concen: 35.551 ng
RT: 14.363 min Scan# 1917
Delta R.T. 0.000 min
Lab File: BM030411.D
Acq: 12 Jun 2021 12:46




Instrument :
BNA_M
ClientSampleId :
SSTDCCC040


Tgt Ion:138 Resp: 360564
Ion Ratio Lower Upper
138 100
108 9.0 7.7 11.5
92 112.8 92.6 138.8



#54
2,4-Dinitrophenol
Concen: 40.928 ng
RT: 14.569 min Scan# 1952
Delta R.T. 0.000 min
Lab File: BM030411.D
Acq: 12 Jun 2021 12:46

Tgt Ion:184 Resp: 206555
Ion Ratio Lower Upper
184 100
63 56.8 54.2 81.2
154 56.9 56.1 84.1









#55
Dibenzenofuran
Concen: 36.466 ng
RT: 14.857 min Scan# 2001
Delta R.T. 0.000 min
Lab File: BM030411.D
Acq: 12 Jun 2021 12:46




Instrument :
BNA_M
ClientSampleId :
SSTDCCC040


Tgt Ion:168 Resp: 1987863
Ion Ratio Lower Upper
168 100
139 35.7 29.4 44.0
169 13.3 10.8 16.2



#56
4-Nitrophenol
Concen: 38.954 ng
RT: 14.663 min Scan# 1968
Delta R.T. 0.000 min
Lab File: BM030411.D
Acq: 12 Jun 2021 12:46

Tgt Ion:139 Resp: 310388
Ion Ratio Lower Upper
139 100
109 73.0 64.5 104.5
65 91.8 78.9 118.9









#57
2,4-Dinitrotoluene
Concen: 38.431 ng
RT: 14.822 min Scan# 1995
Delta R.T. 0.000 min
Lab File: BM030411.D
Acq: 12 Jun 2021 12:46

Instrument :
BNA_M
ClientSampleId :
SSTDCCC040

Tgt Ion:165 Resp: 501308




Ion	Ratio	Lower	Upper
165	100		
63	37.9	35.0	52.4
89	63.1	55.2	82.8
182	2.9	2.6	3.8




#58
Fluorene
Concen: 37.114 ng
RT: 15.504 min Scan# 2111
Delta R.T. 0.000 min
Lab File: BM030411.D
Acq: 12 Jun 2021 12:46

Tgt Ion:166 Resp: 1669131

Ion	Ratio	Lower	Upper
166	100		
165	98.5	79.0	118.6
167	13.5	10.8	16.2









#59
2,3,4,6-Tetrachlorophenol
Concen: 41.838 ng
RT: 15.080 min Scan# 2039
Delta R.T. 0.000 min
Lab File: BM030411.D
Acq: 12 Jun 2021 12:46

Instrument :
BNA_M
ClientSampleId :
SSTDCCC040

Tgt Ion:232 Resp: 462940




Ion	Ratio	Lower	Upper
232	100		
131	39.4	34.4	51.6
130	2.1	2.2	3.2#
166	28.7	26.2	39.4




#60
Diethylphthalate
Concen: 36.389 ng
RT: 15.280 min Scan# 2073
Delta R.T. 0.000 min
Lab File: BM030411.D
Acq: 12 Jun 2021 12:46

Tgt Ion:149 Resp: 1548464

Ion	Ratio	Lower	Upper
149	100		
177	22.3	17.9	26.9
150	12.3	10.3	15.5









#61
4-Chlorophenyl-phenylether
Concen: 37.514 ng
RT: 15.498 min Scan# 2110
Delta R.T. 0.000 min
Lab File: BM030411.D
Acq: 12 Jun 2021 12:46




Instrument :
BNA_M
ClientSampleId :
SSTDCCC040


Tgt Ion:204 Resp: 831470
Ion Ratio Lower Upper
204 100
206 32.6 25.5 38.3
141 52.6 40.3 60.5



#62
4-Nitroaniline
Concen: 36.485 ng
RT: 15.521 min Scan# 2114
Delta R.T. 0.000 min
Lab File: BM030411.D
Acq: 12 Jun 2021 12:46

Tgt Ion:138 Resp: 390624
Ion Ratio Lower Upper
138 100
92 51.0 33.4 73.4
108 78.3 75.5 115.5









#63
Azobenzene
Concen: 34.173 ng
RT: 15.792 min Scan# 2160
Delta R.T. 0.000 min
Lab File: BM030411.D
Acq: 12 Jun 2021 12:46




Instrument :
BNA_M
ClientSampleId :
SSTDCCC040


Tgt Ion: 77 Resp: 1468749
Ion Ratio Lower Upper
77 100
182 29.6 6.8 46.8
105 19.7 0.0 36.6
51 30.2 11.5 51.5



#64
Phenanthrene-d10
Concen: 20.000 ng
RT: 17.192 min Scan# 2398
Delta R.T. -0.006 min
Lab File: BM030411.D
Acq: 12 Jun 2021 12:46

Tgt Ion:188 Resp: 1156320
Ion Ratio Lower Upper
188 100
94 8.8 6.0 9.0
80 8.9 6.6 9.8









#65
4,6-Dinitro-2-methylphenol
Concen: 40.828 ng
RT: 15.580 min Scan# 2124
Delta R.T. 0.000 min
Lab File: BM030411.D
Acq: 12 Jun 2021 12:46




Instrument :
BNA_M
ClientSampleId :
SSTDCCC040


Tgt Ion:198 Resp: 311600
Ion Ratio Lower Upper
198 100
51 40.9 18.9 58.9
105 39.6 17.6 57.6



#66
n-Nitrosodiphenylamine
Concen: 35.694 ng
RT: 15.710 min Scan# 2146
Delta R.T. 0.000 min
Lab File: BM030411.D
Acq: 12 Jun 2021 12:46

Tgt Ion:169 Resp: 1435346
Ion Ratio Lower Upper
169 100
168 67.1 54.2 81.2
167 36.0 29.0 43.6









#67
4-Bromophenyl-phenylether
Concen: 37.186 ng
RT: 16.392 min Scan# 2262
Delta R.T. 0.000 min
Lab File: BM030411.D
Acq: 12 Jun 2021 12:46




Instrument :
BNA_M
ClientSampleId :
SSTDCCC040


Tgt Ion:248 Resp: 501088
Ion Ratio Lower Upper
248 100
250 96.7 77.9 116.9
141 60.1 56.2 84.2



#68
Hexachlorobenzene
Concen: 38.068 ng
RT: 16.510 min Scan# 2282
Delta R.T. 0.000 min
Lab File: BM030411.D
Acq: 12 Jun 2021 12:46

Tgt Ion:284 Resp: 581661
Ion Ratio Lower Upper
284 100
142 31.0 29.0 43.4
249 31.1 26.0 39.0









#69
Atrazine
Concen: 44.933 ng
RT: 16.663 min Scan# 2308
Delta R.T. 0.006 min
Lab File: BM030411.D
Acq: 12 Jun 2021 12:46




Instrument :
BNA_M
ClientSampleId :
SSTDCCC040


Tgt Ion:	200	Resp:	483908
Ion Ratio	Lower	Upper	
200	100		
173	24.6	3.8	43.8
215	48.8	24.8	64.8



#70
Pentachlorophenol
Concen: 42.758 ng
RT: 16.845 min Scan# 2339
Delta R.T. 0.000 min
Lab File: BM030411.D
Acq: 12 Jun 2021 12:46

Tgt Ion:	266	Resp:	382377
Ion Ratio	Lower	Upper	
266	100		
268	64.0	52.8	79.2
264	63.2	51.9	77.9









#71
Phenanthrene
Concen: 36.501 ng
RT: 17.239 min Scan# 2406
Delta R.T. 0.000 min
Lab File: BM030411.D
Acq: 12 Jun 2021 12:46




Instrument :
BNA_M
ClientSampleId :
SSTDCCC040


Tgt Ion:178 Resp: 2655850
Ion Ratio Lower Upper
178 100
176 19.5 15.4 23.0
179 15.3 12.2 18.4



#72
Anthracene
Concen: 37.275 ng
RT: 17.327 min Scan# 2421
Delta R.T. 0.000 min
Lab File: BM030411.D
Acq: 12 Jun 2021 12:46

Tgt Ion:178 Resp: 2633363
Ion Ratio Lower Upper
178 100
176 18.9 15.0 22.4
179 15.6 12.5 18.7









#73
Carbazole
Concen: 37.738 ng
RT: 17.598 min Scan# 2467
Delta R.T. 0.000 min
Lab File: BM030411.D
Acq: 12 Jun 2021 12:46




Instrument :
BNA_M
ClientSampleId :
SSTDCCC040


Tgt Ion:167 Resp: 2542503
Ion Ratio Lower Upper
167 100
166 21.9 17.4 26.0
139 12.0 9.8 14.6



#74
Di-n-butylphthalate
Concen: 37.944 ng
RT: 18.157 min Scan# 2562
Delta R.T. 0.000 min
Lab File: BM030411.D
Acq: 12 Jun 2021 12:46

Tgt Ion:149 Resp: 2665141
Ion Ratio Lower Upper
149 100
150 9.2 7.2 10.8
104 4.9 4.3 6.5









#75
Fluoranthene
Concen: 39.978 ng
RT: 19.245 min Scan# 2747
Delta R.T. 0.000 min
Lab File: BM030411.D
Acq: 12 Jun 2021 12:46




Instrument :
BNA_M
ClientSampleId :
SSTDCCC040


Tgt Ion:202 Resp: 3301747
Ion Ratio Lower Upper
202 100
101 10.2 0.0 28.6
203 17.8 0.0 37.3



#76
Chrysene-d12
Concen: 20.000 ng
RT: 21.362 min Scan# 3107
Delta R.T. 0.006 min
Lab File: BM030411.D
Acq: 12 Jun 2021 12:46

Tgt Ion:240 Resp: 1367000
Ion Ratio Lower Upper
240 100
120 8.8 7.0 10.6
236 26.1 21.0 31.4









#77
Benzidine
Concen: 52.780 ng
RT: 19.427 min Scan# 2778
Delta R.T. -0.006 min
Lab File: BM030411.D
Acq: 12 Jun 2021 12:46




Instrument :
BNA_M
ClientSampleId :
SSTDCCC040


Tgt Ion:184 Resp: 1276597
Ion Ratio Lower Upper
184 100
185 14.2 10.9 16.3
183 11.7 9.8 14.8



#78
Pyrene
Concen: 33.457 ng
RT: 19.609 min Scan# 2809
Delta R.T. 0.000 min
Lab File: BM030411.D
Acq: 12 Jun 2021 12:46

Tgt Ion:202 Resp: 3465656
Ion Ratio Lower Upper
202 100
200 21.3 17.3 25.9
203 17.9 13.9 20.9









#79
Terphenyl-d14
Concen: 69.175 ng
RT: 19.809 min Scan# 2843
Delta R.T. 0.000 min
Lab File: BM030411.D
Acq: 12 Jun 2021 12:46




Instrument :
BNA_M
ClientSampleId :
SSTDCCC040


Tgt Ion:244 Resp: 5383738
Ion Ratio Lower Upper
244 100
212 7.6 6.3 9.5
122 10.0 6.9 10.3



#80
Butylbenzylphthalate
Concen: 31.587 ng
RT: 20.498 min Scan# 2960
Delta R.T. 0.000 min
Lab File: BM030411.D
Acq: 12 Jun 2021 12:46

Tgt Ion:149 Resp: 1286025
Ion Ratio Lower Upper
149 100
91 69.3 57.4 86.0
206 22.1 16.6 25.0









#81
Benzo(a)anthracene
Concen: 36.868 ng
RT: 21.345 min Scan# 3104
Delta R.T. 0.000 min
Lab File: BM030411.D
Acq: 12 Jun 2021 12:46




Instrument :
BNA_M
ClientSampleId :
SSTDCCC040


Tgt Ion:228 Resp: 3668075
Ion Ratio Lower Upper
228 100
226 28.2 22.3 33.5
229 20.1 15.7 23.5



#82
3,3'-Dichlorobenzidine
Concen: 40.431 ng
RT: 21.274 min Scan# 3092
Delta R.T. 0.000 min
Lab File: BM030411.D
Acq: 12 Jun 2021 12:46

Tgt Ion:252 Resp: 1217733
Ion Ratio Lower Upper
252 100
254 63.9 51.5 77.3
126 10.2 8.6 13.0









#83
Chrysene
Concen: 36.839 ng
RT: 21.398 min Scan# 3113
Delta R.T. 0.000 min
Lab File: BM030411.D
Acq: 12 Jun 2021 12:46




Instrument :
BNA_M
ClientSampleId :
SSTDCCC040


Tgt Ion:228 Resp: 3561897
Ion Ratio Lower Upper
228 100
226 31.3 24.6 36.8
229 20.0 15.5 23.3



#84
Bis(2-ethylhexyl)phthalate
Concen: 32.080 ng
RT: 21.274 min Scan# 3092
Delta R.T. 0.000 min
Lab File: BM030411.D
Acq: 12 Jun 2021 12:46

Tgt Ion:149 Resp: 1970249
Ion Ratio Lower Upper
149 100
167 28.1 22.6 33.8
279 4.7 3.8 5.8









#85
Di-n-octyl phthalate
Concen: 34.823 ng
RT: 22.156 min Scan# 3242
Delta R.T. 0.000 min
Lab File: BM030411.D
Acq: 12 Jun 2021 12:46




Instrument :
BNA_M
ClientSampleId :
SSTDCCC040


Tgt Ion:149 Resp: 3338816
Ion Ratio Lower Upper
149 100
167 1.6 1.3 1.9
43 8.7 7.4 11.0



#86
Perylene-d12
Concen: 20.000 ng
RT: 23.639 min Scan# 3494
Delta R.T. 0.006 min
Lab File: BM030411.D
Acq: 12 Jun 2021 12:46

Tgt Ion:264 Resp: 1449091
Ion Ratio Lower Upper
264 100
260 23.7 19.8 29.8
265 21.4 17.7 26.5









#87
Indeno(1,2,3-cd)pyrene
Concen: 38.027 ng
RT: 25.956 min Scan# 3888
Delta R.T. 0.006 min
Lab File: BM030411.D
Acq: 12 Jun 2021 12:46




Instrument :
BNA_M
ClientSampleId :
SSTDCCC040


Tgt Ion:276 Resp: 4191290
Ion Ratio Lower Upper
276 100
138 22.2 17.1 25.7
277 25.4 20.3 30.5



#88
Benzo(b)fluoranthene
Concen: 35.998 ng
RT: 22.950 min Scan# 3377
Delta R.T. 0.000 min
Lab File: BM030411.D
Acq: 12 Jun 2021 12:46

Tgt Ion:252 Resp: 3948066
Ion Ratio Lower Upper
252 100
253 22.3 17.4 26.0
125 8.9 6.8 10.2









#89
Benzo(k)fluoranthene
Concen: 34.550 ng
RT: 22.997 min Scan# 3385
Delta R.T. 0.006 min
Lab File: BM030411.D
Acq: 12 Jun 2021 12:46




Instrument :
BNA_M
ClientSampleId :
SSTDCCC040


Tgt Ion:252 Resp: 3700945
Ion Ratio Lower Upper
252 100
253 22.0 17.4 26.2
125 8.7 6.2 9.2



#90
Benzo(a)pyrene
Concen: 36.913 ng
RT: 23.539 min Scan# 3477
Delta R.T. 0.000 min
Lab File: BM030411.D
Acq: 12 Jun 2021 12:46

Tgt Ion:252 Resp: 3579953
Ion Ratio Lower Upper
252 100
253 21.7 17.7 26.5
125 9.3 7.3 10.9









#91
Dibenzo(a,h)anthracene
Concen: 37.877 ng
RT: 25.968 min Scan# 3890
Delta R.T. 0.006 min
Lab File: BM030411.D
Acq: 12 Jun 2021 12:46

Instrument :
BNA_M
ClientSampleId :
SSTDCCC040

Tgt Ion:278 Resp: 3599443
Ion Ratio Lower Upper
278 100
139 14.1 11.0 16.4
279 23.6 18.7 28.1



#92
Benzo(g,h,i)perylene
Concen: 37.616 ng
RT: 26.668 min Scan# 4009
Delta R.T. 0.012 min
Lab File: BM030411.D
Acq: 12 Jun 2021 12:46

Tgt Ion:276 Resp: 3400074
Ion Ratio Lower Upper
276 100
277 23.9 19.4 29.2
138 18.8 14.4 21.6

