

Data Path : Z:\SVOASRV\HPCHEM1\BNA M\DATA\BM061419\
 Data File : BM020949.D
 Acq On : 14 Jun 2019 18:44
 Operator : HP/JU
 Sample : PB120635BL
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 SBLK35

Quant Time: Jun 14 23:46:22 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_M\METHODS\SOM-EPA-BM061419MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Fri Jun 14 15:40:58 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.80	152	213921	20.00	ng/ul	0.00
18) Naphthalene-d8	10.59	136	1000994	20.00	ng/ul	0.00
35) Acenaphthene-d10	14.43	164	682988	20.00	ng/ul	0.00
61) Phenanthrene-d10	17.18	188	1609870	20.00	ng/ul	0.00
77) Chrysene-d12	21.36	240	1280464	20.00	ng/ul	0.00
85) Perylene-d12	23.64	264	1293011	20.00	ng/ul	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
3) 1,4-Dioxane-d8	3.30	96	37735	8.38	ng/uL	0.00
5) Phenol-d5	6.96	99	731394	39.28	ng/ul	0.00
7) Bis-(2-Chloroethyl)ether-d	7.13	67	489065	44.09	ng/ul	0.00
9) 2-Chlorophenol-d4	7.33	132	625898	41.37	ng/ul	0.00
13) 4-Methylphenol-d8	8.50	113	637601	41.02	ng/ul	0.00
19) Nitrobenzene-d5	8.96	128	341717	42.04	ng/ul	0.00
22) 2-Nitrophenol-d4	9.67	143	395178	43.80	ng/ul	0.00
26) 2,4-Dichlorophenol-d3	10.20	165	701765	38.25	ng/ul	0.00
29) 4-Chloroaniline-d4	10.73	131	993727	51.93	ng/ul	0.00
43) Dimethylphthalate-d6	13.85	166	2709648	43.91	ng/ul	0.00
46) Acenaphthylene-d8	14.13	160	3147656	41.67	ng/ul	0.00
51) 4-Nitrophenol-d4	14.64	143	405121	40.36	ng/ul	0.00
57) Fluorene-d10	15.43	176	2276032	42.40	ng/ul	0.00
62) 4,6-Dinitro-2-methylphenol	15.55	200	348807	35.25	ng/ul	0.00
70) Anthracene-d10	17.28	188	3568381	42.65	ng/ul	0.00
78) Pyrene-d10	19.57	212	3689627	48.02	ng/ul	0.00
89) Benzo(a)pyrene-d12	23.49	264	3344601	43.84	ng/ul	0.00

Target Compounds	Ovalue
------------------	--------

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA M\DATA\BM061419\
Data File : BM020949.D
Acq On : 14 Jun 2019 18:44
Operator : HP/JU
Sample : PB120635BL
Misc :
ALS Vial : 5 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
SBLK35

Quant Time: Jun 14 23:46:22 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_M\METHODS\SOM-EPA-BM061419MA.M
Quant Title : SVOA CALIBRATION
QLast Update : Fri Jun 14 15:40:58 2019
Response via : Initial Calibration

