

Data Path : Z:\SVOASRV\HPCHEM1\BNA M\DATA\BM061419\
 Data File : BM020955.D
 Acq On : 14 Jun 2019 22:24
 Operator : HP/JU
 Sample : K3353-01
 Misc :
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 C0JS8

Quant Time: Jun 14 23:00:57 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_M\METHODS\SOM-EPA-BM061419MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Fri Jun 14 15:40:58 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.80	152	267117	20.00	ng/ul	0.00
18) Naphthalene-d8	10.59	136	1211592	20.00	ng/ul	0.00
35) Acenaphthene-d10	14.43	164	829413	20.00	ng/ul	0.00
61) Phenanthrene-d10	17.18	188	2067980	20.00	ng/ul	0.00
77) Chrysene-d12	21.36	240	2217063	20.00	ng/ul	0.00
85) Perylene-d12	23.64	264	2534547	20.00	ng/ul	0.00

System Monitoring Compounds

3) 1,4-Dioxane-d8	3.30	96	12951	2.30	ng/uL	0.00
5) Phenol-d5	6.96	99	262422	11.29	ng/ul	0.00
7) Bis-(2-Chloroethyl)ether-d	7.13	67	444056	32.06	ng/ul	0.00
9) 2-Chlorophenol-d4	7.33	132	528687	27.98	ng/ul	0.00
13) 4-Methylphenol-d8	8.50	113	426699	21.99	ng/ul	0.00
19) Nitrobenzene-d5	8.96	128	314942	32.01	ng/ul	0.00
22) 2-Nitrophenol-d4	9.68	143	364694	33.39	ng/ul	0.00
26) 2,4-Dichlorophenol-d3	10.20	165	679173	30.58	ng/ul	0.00
29) 4-Chloroaniline-d4	10.73	131	886873	38.29	ng/ul	0.00
43) Dimethylphthalate-d6	13.85	166	2759603	36.82	ng/ul	0.00
46) Acenaphthylene-d8	14.13	160	3102257	33.82	ng/ul	0.00
51) 4-Nitrophenol-d4	14.64	143	142175	11.66	ng/ul	0.00
57) Fluorene-d10	15.43	176	2342919	35.94	ng/ul	0.00
62) 4,6-Dinitro-2-methylphenol	15.55	200	259054	20.38	ng/ul	0.00
70) Anthracene-d10	17.28	188	3907897	36.36	ng/ul	0.00
78) Pyrene-d10	19.57	212	4572146	34.37	ng/ul	0.00
89) Benzo(a)pyrene-d12	23.50	264	5588914	37.37	ng/ul	0.00

Target Compounds Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA M\DATA\BM061419\
Data File : BM020955.D
Acq On : 14 Jun 2019 22:24
Operator : HP/JU
Sample : K3353-01
Misc :
ALS Vial : 11 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
C0JS8

Quant Time: Jun 14 23:00:57 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_M\METHODS\SOM-EPA-BM061419MA.M
Quant Title : SVOA CALIBRATION
QLast Update : Fri Jun 14 15:40:58 2019
Response via : Initial Calibration

