



Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM061423\
 Data File : BM040234.D
 Acq On : 14 Jun 2023 12:04
 Operator : MA/JU
 Sample : DFTPP127
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 DFTPP127

Integration File: rteint.p

Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\SFAM-EPA-SIM-BM061423.M
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 Last Update : Thu Jun 15 00:31:16 2023



AutoFind: Scans 2465, 2466, 2467; Background Corrected with Scan 2458

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	34.5	339443	PASS
68	69	0.00	2	1.6	6080	PASS
69	198	0.00	100	38.0	374120	PASS
70	69	0.00	2	0.6	2258	PASS
127	198	10	80	46.0	452843	PASS
197	198	0.00	2	0.3	2820	PASS
198	198	100	100	100.0	983403	PASS
199	198	5	9	6.9	68091	PASS
275	198	10	60	25.6	251264	PASS
365	198	1	100	3.2	31800	PASS
441	198	0.01	100	13.7	135173	PASS
442	198	50	100	86.3	848661	PASS
443	442	15	24	19.5	165784	PASS

m/z	Abundance
35.90	863.0
37.00	1332.0
37.90	409.0
38.90	609.0
39.90	2551.0
40.90	1800.0
42.00	843.0
43.00	1167.0
44.00	10673.0
44.90	539.0
46.00	211.0
47.00	6093.0
47.90	3220.0
49.00	32392.0
50.00	1569.0
51.00	10189.0
51.90	195.0
53.10	199.0
54.00	165.0
54.90	580.0
56.00	357.0
56.90	270.0
58.00	221.0
58.90	290.0
59.90	271.0
61.00	334.0
64.90	247.0
67.00	440.0
68.90	323.0
69.80	316.0
71.00	285.0
71.80	173.0
73.00	2683.0
74.10	364.0
75.00	451.0
75.80	278.0
76.90	624.0
78.00	441.0
79.00	335.0
80.90	291.0
82.00	402.0
82.90	1084.0
83.90	23152.0
84.90	859.0
85.90	15539.0
86.90	394.0
87.90	3004.0
89.00	308.0
90.90	460.0
95.90	1609.0
97.00	501.0
97.90	167.0
102.80	470.0
104.00	161.0
105.00	184.0
111.00	158.0
114.90	318.0
116.90	204.0
118.90	364.0
123.00	158.0
128.90	272.0
132.90	1381.0
133.90	201.0
135.00	557.0
147.00	783.0
148.00	237.0
148.70	236.0
160.90	189.0
162.80	428.0
165.00	180.0
166.90	194.0
176.00	207.0
176.90	801.0
178.90	287.0
188.90	168.0
190.90	1529.0
191.90	343.0
193.00	1004.0
194.00	200.0
194.80	176.0
207.00	12711.0
208.00	2871.0
208.90	1714.0
209.80	179.0
220.90	380.0
248.70	219.0
250.80	174.0
264.80	170.0
266.90	502.0
281.00	2067.0
282.00	722.0
283.10	371.0

Instrument :
BNA_M
ClientSampleId :
DFTPP127

340.90203.0

354.90207.0

Instrument :
BNA_M
ClientSampleId :
DFTPP127

m/z	Abundance
35.00	1131.0
37.00	2638.0
38.00	4705.0
39.00	27192.0
40.00	3306.0
41.00	2370.0
42.00	897.0
43.00	1134.0
44.00	10211.0
44.90	1282.0
47.00	5822.0
48.00	3339.0
49.00	31112.0
50.00	98256.0
51.00	389312.0
52.00	20744.0
53.00	1214.0
53.90	165.0
55.00	2083.0
56.00	11334.0
57.00	25328.0
58.00	1115.0
59.00	585.0
59.90	677.0
61.00	5281.0
62.00	5749.0
63.00	16432.0
64.00	2314.0
65.00	8236.0
66.00	600.0
66.90	659.0
68.00	6607.0
69.00	414400.0
70.00	2549.0
71.00	512.0
72.00	479.0
73.00	5526.0
74.00	43528.0
75.00	68304.0
76.00	24120.0
77.00	502144.0
78.00	35216.0
79.00	30416.0
80.00	22872.0
81.00	32776.0
82.00	8336.0
83.00	8338.0
83.90	25264.0
85.00	6281.0
85.90	23992.0
87.00	5232.0
87.90	4233.0
88.90	1113.0
91.00	6971.0
92.00	8569.0
93.00	52816.0
93.90	3881.0
94.90	942.0
96.00	3966.0
97.00	1443.0
98.00	42824.0
99.00	38216.0
100.00	2937.0
101.00	18840.0
101.90	1217.0
103.00	5715.0
104.00	10152.0
105.00	10668.0
106.00	3519.0
107.00	142336.0
108.00	22744.0
109.00	3342.0
110.00	253120.0
111.00	40104.0
112.00	4766.0
113.00	1637.0
113.90	296.0
115.00	863.0
116.00	7072.0
117.00	127328.0
118.00	10015.0
118.90	1631.0
120.00	1757.0
120.80	618.0
121.90	8295.0
123.00	13835.0
124.00	6531.0
124.90	6546.0
126.00	1320.0
127.00	475008.0
128.00	36392.0
129.00	178496.0

Instrument :
BNA_M
ClientSampleId :
DFTPP127

130.00	15523.0
130.90	3785.0
131.90	1331.0
133.00	1977.0
134.00	5099.0
135.00	15424.0
135.90	5617.0
137.00	8824.0
137.90	1890.0
138.80	758.0
139.90	2152.0
141.00	21896.0
142.00	7454.0
143.00	4945.0
144.00	1708.0
145.00	1357.0
146.00	4065.0
147.00	13114.0
147.90	28848.0
148.90	7587.0
149.90	2079.0
151.10	3178.0
151.90	1953.0
152.90	7278.0
154.00	5941.0
155.00	13186.0
156.00	18520.0
157.00	3666.0
157.90	3902.0
158.90	3036.0
160.00	6688.0
160.90	12116.0
162.00	3427.0
162.90	1265.0
164.00	1184.0
165.00	8454.0
166.00	6999.0
167.00	53376.0
168.00	36744.0
169.00	4990.0
170.00	1432.0
171.00	1844.0
171.90	4253.0
173.00	5678.0
174.00	10287.0
175.00	20016.0
176.00	5044.0
176.90	8599.0
177.90	2982.0
178.90	34792.0
180.00	27872.0
181.00	11575.0
182.00	1599.0
182.90	1107.0
183.90	2828.0
185.00	16456.0
186.00	133568.0
187.00	39064.0
188.00	4614.0
189.00	7642.0
189.90	1674.0
190.90	5640.0
192.00	11975.0
193.00	14291.0
194.00	2833.0
194.90	1278.0
196.00	24488.0
196.90	3894.0
197.90	971712.0
198.90	68864.0
200.00	5006.0
201.40	3696.0
203.00	6343.0
204.00	36200.0
205.00	59072.0
206.00	250752.0
207.00	48440.0
208.00	11083.0
209.00	4765.0
210.00	3720.0
211.00	9685.0
211.90	1812.0
212.90	706.0
213.90	327.0
215.00	2531.0
215.90	5136.0
216.90	64528.0
217.90	9202.0
218.90	1031.0
220.10	1269.0
221.00	37800.0
221.80	8489.0
222.90	14734.0
224.00	139840.0
225.00	34848.0
226.00	3678.0

Instrument :
BNA_M
ClientSampleId :
DFTPP127

226.90	56280.0
228.00	8795.0
228.90	11737.0
229.90	1653.0
231.00	5418.0
231.90	952.0
232.90	1006.0
233.90	3207.0
234.90	3692.0
235.90	2514.0
237.00	7088.0
237.90	972.0
238.90	2349.0
239.90	2223.0
241.00	3297.0
242.00	7275.0
243.00	8036.0
244.00	110736.0
245.00	15214.0
245.90	20224.0
247.00	4048.0
247.90	797.0
248.90	4091.0
250.00	768.0
250.90	970.0
251.90	794.0
252.90	1758.0
253.90	2414.0
255.00	525120.0
256.00	76896.0
256.90	6737.0
258.00	29768.0
258.90	4486.0
259.90	751.0
260.90	683.0
263.00	328.0
263.90	831.0
265.00	12266.0
266.00	2067.0
266.90	519.0
267.90	178.0
268.90	177.0
269.90	539.0
270.80	910.0
271.90	1414.0
272.90	15772.0
274.00	40952.0
275.00	223872.0
276.00	31008.0
277.00	19944.0
277.90	4372.0
279.00	662.0
281.00	2753.0
282.00	1145.0
283.00	2988.0
283.90	1656.0
285.00	3933.0
285.90	653.0
288.00	313.0
288.90	747.0
290.10	516.0
290.90	455.0
291.90	800.0
292.90	4628.0
294.00	1522.0
295.00	1291.0
295.90	64320.0
297.00	9688.0
297.90	649.0
299.00	161.0
300.90	866.0
301.90	1209.0
303.00	7234.0
303.90	1901.0
304.90	409.0
308.00	888.0
308.90	743.0
309.90	723.0
310.80	379.0
311.90	195.0
312.90	438.0
314.00	2582.0
314.90	6978.0
316.00	4276.0
316.90	752.0
317.90	156.0
319.90	231.0
321.00	2036.0
322.00	932.0
323.00	19928.0
324.00	4197.0
325.00	559.0
325.80	313.0
327.00	3911.0
327.90	2284.0
328.90	655.0

Instrument :
BNA_M
ClientSampleId :
DFTPP127

331.90	1570.0
332.90	2043.0
334.00	12825.0
335.00	2946.0
336.00	346.0
338.90	453.0
339.90	348.0
341.00	2518.0
342.00	653.0
345.90	4466.0
346.90	1092.0
351.00	321.0
352.00	5948.0
353.00	3847.0
354.10	5771.0
355.00	1429.0
355.90	177.0
359.00	577.0
360.70	180.0
363.90	249.0
364.90	25168.0
365.90	3987.0
366.80	231.0
369.90	464.0
370.90	1403.0
372.00	9347.0
373.00	2099.0
373.80	259.0
376.90	201.0
383.00	2836.0
384.00	796.0
384.90	271.0
390.00	1181.0
390.90	881.0
391.90	489.0
393.00	177.0
401.00	724.0
402.00	4043.0
403.00	5261.0
404.00	2017.0
405.00	673.0
410.00	244.0
415.00	338.0
421.00	4176.0
422.00	4940.0
423.00	29968.0
424.00	11363.0
425.00	1918.0
439.00	179.0
439.90	190.0
441.00	98320.0
442.00	611584.0
443.00	124488.0
444.00	11683.0
445.00	761.0

Instrument :
BNA_M
ClientSampleId :
DFTPP127

m/z	Abundance
36.00	1460.0
36.90	2515.0
38.00	4690.0
39.00	28712.0
40.00	3206.0
41.00	2760.0
41.90	1027.0
43.00	1316.0
44.00	9010.0
45.00	1120.0
46.10	215.0
47.00	5338.0
48.00	2942.0
49.00	30592.0
50.00	101504.0
51.00	398592.0
52.00	19816.0
53.00	826.0
55.00	2545.0
56.00	11781.0
57.00	27280.0
58.00	1311.0
59.00	543.0
60.00	681.0
61.00	5633.0
62.00	5683.0
63.00	17416.0
64.00	2341.0
65.00	8408.0
66.00	726.0
67.00	772.0
68.00	7264.0
69.00	432128.0
70.00	2337.0
70.90	407.0
71.90	596.0
73.00	5434.0
74.00	45312.0
75.00	73424.0
76.00	25576.0
77.00	528000.0
78.00	36464.0
79.00	31304.0
80.00	24224.0
81.00	35680.0
82.00	8956.0
83.00	9592.0
83.90	24752.0
85.00	6964.0
85.90	23864.0
87.00	5281.0
87.90	4306.0
89.00	979.0
89.90	197.0
91.00	7424.0
92.00	8812.0
93.00	56416.0
94.00	4601.0
95.00	1087.0
96.00	4245.0
96.90	1196.0
98.00	42976.0
99.00	38120.0
99.90	3569.0
101.00	19232.0
101.90	1292.0
102.90	5607.0
104.00	12139.0
105.00	11296.0
106.00	4051.0
107.00	157504.0
108.00	23264.0
109.00	3824.0
110.00	281984.0
111.00	41464.0
112.00	4809.0
113.00	1912.0
114.00	474.0
115.00	873.0
116.00	7319.0
117.00	135872.0
118.00	10039.0
119.00	1972.0
120.00	1682.0
120.90	887.0
122.00	9521.0
122.90	13934.0
124.00	6439.0
125.00	6471.0
126.00	1118.0
127.00	530304.0
128.00	39392.0

Instrument :
BNA_M
ClientSampleId :
DFTPP127

129.00	200192.0
130.00	17488.0
131.00	3339.0
132.00	1412.0
133.00	2330.0
134.00	6316.0
135.00	16832.0
135.90	6810.0
137.00	9147.0
137.90	2116.0
139.00	963.0
140.00	2187.0
141.00	24880.0
142.00	8220.0
143.00	6166.0
143.90	1499.0
145.00	1541.0
145.90	4720.0
147.00	14857.0
147.90	34680.0
149.00	8065.0
150.00	2097.0
151.00	3354.0
151.80	2226.0
152.90	8547.0
154.00	6048.0
155.00	16253.0
156.00	22792.0
157.00	4150.0
158.00	4710.0
158.90	4265.0
160.00	8118.0
161.00	13154.0
161.90	3601.0
162.90	1109.0
163.90	1565.0
165.00	9361.0
166.00	7816.0
167.00	60464.0
168.00	41816.0
169.00	6047.0
170.00	1572.0
170.90	2845.0
171.90	5139.0
173.00	6807.0
174.00	12307.0
175.00	22784.0
176.00	6246.0
176.90	9902.0
178.00	3361.0
178.90	41384.0
180.00	30480.0
181.00	14278.0
182.00	2017.0
182.90	1109.0
183.90	2956.0
185.00	20296.0
186.00	162112.0
187.00	45176.0
188.00	5497.0
189.00	8673.0
189.90	1645.0
190.90	5982.0
192.00	13796.0
193.00	16375.0
194.00	3343.0
195.00	1691.0
196.00	31824.0
196.80	4566.0
197.90	1170432.0
198.90	82616.0
200.00	6246.0
201.50	5037.0
203.00	7995.0
204.00	43320.0
205.00	72104.0
206.00	306816.0
207.00	57120.0
208.00	13758.0
208.90	4873.0
210.10	4867.0
211.00	12381.0
211.80	2359.0
213.00	897.0
213.80	327.0
215.00	3102.0
216.00	7028.0
216.90	80032.0
218.00	10520.0
219.00	1078.0
220.00	1549.0
221.00	49280.0
221.80	10322.0
223.00	18448.0
224.00	173824.0
225.00	44544.0

Instrument :
BNA_M
ClientSampleId :
DFTPP127

226.00	5410.0
227.00	69528.0
228.00	9244.0
229.00	14928.0
230.00	1494.0
231.00	5616.0
231.90	1137.0
232.90	1200.0
233.90	4755.0
235.00	5197.0
235.90	3420.0
237.00	8055.0
237.90	1007.0
239.00	2782.0
240.00	2236.0
241.00	3412.0
242.00	10181.0
243.00	9626.0
244.00	143360.0
245.00	17352.0
246.00	25544.0
247.00	5174.0
247.90	1137.0
248.90	5186.0
249.90	1148.0
251.00	1028.0
252.10	995.0
253.00	2913.0
254.00	2997.0
255.00	693120.0
256.00	102664.0
257.00	8109.0
258.00	39936.0
259.00	6335.0
259.90	928.0
260.90	921.0
261.90	200.0
262.90	439.0
263.90	1289.0
265.00	15969.0
266.00	2125.0
266.90	753.0
267.90	278.0
269.90	780.0
270.90	1860.0
271.90	1827.0
273.00	22696.0
274.00	55224.0
275.00	305728.0
276.00	39680.0
277.00	26840.0
278.00	5143.0
279.00	1122.0
280.00	206.0
281.00	2917.0
282.00	1586.0
283.00	3420.0
284.00	1762.0
285.00	4552.0
285.90	969.0
288.00	306.0
288.90	1120.0
289.90	708.0
290.90	736.0
291.90	1301.0
293.00	6538.0
293.90	1927.0
295.00	1749.0
296.00	88744.0
297.00	12889.0
298.00	923.0
298.90	222.0
300.90	1161.0
302.00	1347.0
303.00	10214.0
304.00	2391.0
305.00	486.0
305.90	173.0
307.00	182.0
307.90	1123.0
308.90	883.0
309.90	1046.0
310.90	301.0
312.00	337.0
313.00	1046.0
314.00	4295.0
315.00	9276.0
315.90	5178.0
316.90	1045.0
319.90	406.0
321.00	2622.0
322.00	1484.0
323.00	28168.0
324.00	5751.0
325.00	559.0
325.90	503.0

Instrument :
BNA_M
ClientSampleId :
DFTPP127

326.90	5850.0
328.00	3283.0
329.00	708.0
330.90	172.0
332.00	2143.0
333.00	2553.0
334.00	19384.0
335.00	5214.0
336.00	509.0
339.00	480.0
339.80	424.0
341.00	3721.0
342.00	1072.0
342.90	305.0
345.90	6869.0
346.90	1393.0
347.90	190.0
350.00	177.0
351.00	459.0
352.00	8603.0
353.00	5938.0
354.00	9170.0
355.00	2325.0
355.90	296.0
357.90	209.0
358.90	679.0
359.80	213.0
360.80	219.0
365.00	38024.0
366.00	5990.0
367.00	448.0
369.90	861.0
371.00	2440.0
372.00	15034.0
373.00	3406.0
374.10	326.0
376.90	420.0
383.00	4508.0
384.00	829.0
384.90	465.0
389.00	184.0
390.00	1886.0
391.00	1505.0
391.80	792.0
394.80	173.0
396.90	177.0
400.90	1222.0
402.00	5745.0
403.00	8823.0
404.00	3313.0
405.00	761.0
410.00	202.0
415.00	509.0
415.90	183.0
421.00	7743.0
422.00	7318.0
423.00	48944.0
424.00	15228.0
425.00	1918.0
439.00	259.0
441.00	162176.0
442.00	1029696.0
443.00	202176.0
444.00	18864.0
445.00	1345.0

Instrument :
BNA_M
ClientSampleId :
DFTPP127

m/z	Abundance
36.00	1327.0
37.00	2149.0
38.00	3026.0
39.00	18840.0
40.00	2704.0
41.00	2076.0
41.90	937.0
43.00	1317.0
44.00	10414.0
45.00	1051.0
46.10	167.0
47.00	6204.0
48.00	3102.0
49.00	31712.0
50.00	65424.0
51.00	260992.0
52.00	13284.0
53.00	704.0
55.00	1898.0
55.90	7647.0
57.00	17552.0
58.00	873.0
58.90	604.0
59.90	394.0
61.00	3467.0
62.00	3905.0
63.00	10481.0
64.00	1553.0
65.00	5146.0
66.00	560.0
67.00	743.0
68.00	4368.0
69.00	276800.0
70.00	1888.0
71.00	367.0
71.90	350.0
73.00	5127.0
74.00	28800.0
75.00	45312.0
76.00	16432.0
77.00	338880.0
78.00	23224.0
79.00	20824.0
80.00	16282.0
81.00	22200.0
82.00	5873.0
83.00	6148.0
83.90	25568.0
85.00	5275.0
85.90	21488.0
87.00	3384.0
87.90	4169.0
89.00	844.0
89.90	173.0
91.00	5389.0
92.00	5901.0
93.00	36376.0
94.00	2205.0
94.90	656.0
96.00	3153.0
97.00	729.0
98.00	29104.0
99.00	24392.0
100.00	1881.0
101.00	13508.0
102.00	909.0
102.90	3964.0
103.90	8076.0
105.00	7542.0
106.00	2174.0
107.00	102536.0
108.00	15549.0
109.00	2613.0
110.00	181056.0
111.00	26928.0
111.90	3036.0
113.00	1402.0
113.70	282.0
115.00	702.0
116.00	5301.0
117.00	89896.0
118.00	7543.0
119.00	1175.0
119.90	1249.0
121.00	679.0
122.00	5895.0
123.00	10379.0
123.90	4873.0
125.00	4209.0
125.90	932.0
127.00	353216.0
128.00	24896.0

Instrument :
BNA_M
ClientSampleId :
DFTPP127

128.90	135168.0
130.00	11343.0
130.90	2827.0
131.90	1163.0
133.00	1909.0
133.90	4196.0
135.00	11359.0
136.00	3821.0
137.00	6995.0
138.00	1101.0
139.00	725.0
139.90	1789.0
140.90	17552.0
142.00	5883.0
143.00	4056.0
144.00	1081.0
145.00	851.0
145.90	3197.0
147.00	9634.0
148.00	23376.0
149.00	5415.0
150.00	1585.0
151.00	2308.0
151.90	1310.0
153.00	5879.0
153.90	4822.0
155.00	10820.0
156.00	15304.0
157.00	3045.0
157.90	3564.0
158.90	2800.0
159.90	5426.0
160.90	8471.0
161.90	2543.0
163.00	995.0
163.90	732.0
164.90	6588.0
165.90	5313.0
167.00	42624.0
168.00	27592.0
168.90	3464.0
169.90	1195.0
171.00	1464.0
171.90	3203.0
172.90	4225.0
174.00	7735.0
175.00	15611.0
176.00	4011.0
176.90	7609.0
177.90	2412.0
178.90	29320.0
180.00	21968.0
181.00	9215.0
181.90	1706.0
183.00	799.0
183.90	2024.0
185.00	14154.0
186.00	108184.0
187.00	31448.0
188.00	3733.0
188.90	6303.0
189.90	1226.0
190.90	4762.0
192.00	10875.0
193.00	11254.0
194.00	2677.0
194.90	1160.0
196.00	19664.0
197.90	808064.0
198.90	52792.0
200.00	4130.0
201.40	3122.0
202.90	5155.0
204.00	30664.0
205.00	49448.0
206.00	207808.0
207.00	42056.0
208.00	10454.0
209.00	4336.0
210.00	3114.0
211.00	8469.0
211.80	1589.0
213.00	774.0
213.90	157.0
215.00	2478.0
216.00	4505.0
216.90	57680.0
218.00	6660.0
218.90	869.0
220.20	1070.0
221.00	32632.0
221.90	7929.0
223.00	12361.0
224.00	125096.0
225.00	29128.0
226.00	3223.0

Instrument :
BNA_M
ClientSampleId :
DFTPP127

227.00	50600.0
228.00	6707.0
228.90	10886.0
229.90	1653.0
230.90	4470.0
232.00	585.0
233.00	736.0
234.00	3086.0
234.90	3657.0
235.90	2521.0
237.00	4484.0
238.00	562.0
238.90	1819.0
239.90	1554.0
240.90	2648.0
242.00	6315.0
243.00	7249.0
244.00	100616.0
245.00	13551.0
246.00	18712.0
247.00	3640.0
247.90	826.0
249.00	3768.0
249.90	700.0
250.90	1153.0
252.00	800.0
252.90	1920.0
253.90	2317.0
255.00	495552.0
256.00	72568.0
257.00	5802.0
258.00	28352.0
258.90	5070.0
260.00	946.0
260.90	887.0
263.00	254.0
263.90	846.0
265.00	13377.0
265.90	1743.0
266.90	688.0
268.00	261.0
268.90	239.0
269.80	486.0
271.00	1528.0
272.00	1385.0
272.90	15882.0
274.00	40840.0
275.00	224192.0
276.00	30824.0
276.90	18664.0
278.00	3456.0
278.90	646.0
280.00	158.0
281.00	3037.0
281.90	1138.0
283.00	2633.0
284.00	1621.0
285.00	3538.0
286.00	473.0
287.90	294.0
288.90	903.0
290.00	815.0
290.90	445.0
291.90	776.0
292.90	5222.0
294.00	1193.0
294.90	1227.0
295.90	70440.0
296.90	8867.0
297.90	788.0
301.00	829.0
301.90	1135.0
303.00	7993.0
303.90	1895.0
305.00	314.0
308.00	1154.0
309.00	543.0
309.90	1006.0
310.90	265.0
312.00	229.0
312.90	366.0
313.90	2996.0
314.90	7516.0
316.00	4195.0
316.90	678.0
319.90	175.0
321.00	2755.0
322.00	903.0
323.00	22704.0
324.00	3889.0
325.00	529.0
326.00	590.0
326.90	4856.0
327.90	2708.0
328.90	574.0
332.00	1560.0

Instrument :
BNA_M
ClientSampleId :
DFTPP127

332.90	2557.0
334.00	15445.0
335.00	3043.0
336.00	491.0
339.00	297.0
340.00	454.0
341.00	2695.0
342.00	656.0
343.00	158.0
345.90	5847.0
346.90	1034.0
349.90	180.0
350.90	275.0
352.00	7076.0
353.00	4152.0
354.00	7662.0
355.00	1307.0
356.00	316.0
356.90	204.0
357.90	192.0
359.00	382.0
359.80	204.0
364.00	156.0
364.90	32208.0
365.90	4389.0
366.80	299.0
370.00	720.0
371.00	1777.0
372.00	11978.0
372.90	2799.0
373.90	167.0
376.90	363.0
383.00	3498.0
383.90	964.0
385.00	430.0
390.00	1564.0
390.90	1223.0
392.00	724.0
394.90	208.0
400.90	967.0
402.00	4851.0
403.00	6808.0
404.00	2149.0
404.90	525.0
406.10	172.0
410.10	271.0
414.90	437.0
420.10	266.0
421.00	6404.0
422.00	5884.0
423.00	44504.0
424.00	12013.0
425.00	1481.0
441.00	145024.0
442.00	904704.0
443.00	170688.0
444.00	15425.0
445.00	491.0

Instrument :
BNA_M
ClientSampleId :
DFTPP127