

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM061523\
 Data File : BM040274.D
 Acq On : 16 Jun 2023 00:49
 Operator : MA/JU
 Sample : SSTDCCC0.4EC
 Misc :
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 SSTD0.4120

Quant Time: Jun 16 01:20:18 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\SFAM-EPA-SIM-BM061423.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Jun 15 00:31:16 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.970	152	9052	0.400	ng/ul	0.00
4) Naphthalene-d8	10.782	136	31960	0.400	ng/ul	0.00
9) Acenaphthene-d10	14.608	164	15264	0.400	ng/ul	0.00
13) Phenanthrene-d10	17.352	188	28057	0.400	ng/ul	0.00
17) Chrysene-d12	21.525	240	16761	0.400	ng/ul	0.00
23) Perylene-d12	23.948	264	16091	0.400	ng/ul	-0.01
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.350	96	5251	0.470	ng/ul	0.00
6) 2-Methylnaphthalene-d10	12.366	152	15684	0.387	ng/ul	0.00
18) Fluoranthene-d10	19.374	212	20968	0.502	ng/ul	0.00
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.384	88	5351	0.456	ng/ul#	83
5) Naphthalene	10.832	128	30418	0.402	ng/ul	100
7) 2-Methylnaphthalene	12.437	142	18567	0.395	ng/ul	99
8) 1-Methylnaphthalene	12.657	142	18893	0.383	ng/ul	99
10) Acenaphthylene	14.331	152	22542	0.422	ng/ul	100
11) Acenaphthene	14.668	153	19648	0.429	ng/ul	98
12) Fluorene	15.654	166	20996	0.412	ng/ul	99
14) Pentachlorophenol	17.002	266	2005	0.350	ng/ul	99
15) Phenanthrene	17.390	178	30362	0.412	ng/ul	99
16) Anthracene	17.483	178	25418	0.407	ng/ul	99
19) Fluoranthene	19.406	202	28327	0.496	ng/ul	99
20) Pyrene	19.764	202	29001	0.489	ng/ul	95
21) Benzo(a)anthracene	21.510	228	21060	0.426	ng/ul	100
22) Chrysene	21.560	228	23264	0.445	ng/ul	100
24) Benzo(b)fluoranthene	23.209	252	24305	0.410	ng/ul	95
25) Benzo(k)fluoranthene	23.255	252	23664	0.424	ng/ul#	96
26) Benzo(a)pyrene	23.840	252	20190	0.415	ng/ul#	92
27) Indeno(1,2,3-cd)pyrene	26.463	276	26541	0.431	ng/ul#	95
28) Dibenzo(a,h)anthracene	26.483	278	21241	0.429	ng/ul	97
29) Benzo(g,h,i)perylene	27.237	276	22817	0.443	ng/ul	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM061523\
Data File : BM040274.D
Acq On : 16 Jun 2023 00:49
Operator : MA/JU
Sample : SSTDCCC0.4EC
Misc :
ALS Vial : 22 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
SSTD0.4120

Quant Time: Jun 16 01:20:18 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\SFAM-EPA-SIM-BM061423.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Jun 15 00:31:16 2023
Response via : Initial Calibration

