

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM061622\
 Data File : BM035480.D
 Acq On : 16 Jun 2022 14:40
 Operator : CG/JU
 Sample : SSTD02049
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 SSTD020049

Quant Time: Jun 17 01:04:45 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\SFAM-EPA-BM061622.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Fri Jun 17 00:59:32 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.816	152	243722	20.000	ng/u1	0.00
20) Naphthalene-d8	10.610	136	1017556	20.000	ng/u1	0.00
38) Acenaphthene-d10	14.457	164	583379	20.000	ng/u1	0.00
64) Phenanthrene-d10	17.204	188	1160528	20.000	ng/u1	0.00
79) Chrysene-d12	21.392	240	1194038	20.000	ng/u1	0.00
88) Perylene-d12	23.733	264	1308570	20.000	ng/u1	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.269	96	47329	8.024	ng/uL	0.00
4) Pyridine-d5	3.687	84	280274	18.419	ng/u1	0.00
7) Phenol-d5	6.998	99	344592	19.185	ng/u1	0.00
9) Bis-(2-Chloroethyl)eth...	7.145	67	217716	19.211	ng/u1	0.00
11) 2-Chlorophenol-d4	7.351	132	311243	19.546	ng/u1	0.00
15) 4-Methylphenol-d8	8.539	113	283379	18.781	ng/u1	0.00
21) Nitrobenzene-d5	8.975	128	143459	17.982	ng/u1	0.00
24) 2-Nitrophenol-d4	9.698	143	156601	17.531	ng/u1	0.00
28) 2,4-Dichlorophenol-d3	10.251	165	263241	16.568	ng/u1	0.00
31) 4-Chloroaniline-d4	10.757	131	397802	18.157	ng/u1	0.00
46) Dimethylphthalate-d6	13.874	166	752537	17.706	ng/u1	0.00
49) Acenaphthylene-d8	14.151	160	943090	18.566	ng/u1	0.00
54) 4-Nitrophenol-d4	14.704	143	97232	12.903	ng/u1	0.00
60) Fluorene-d10	15.451	176	652072	18.134	ng/u1	0.00
65) 4,6-Dinitro-2-methylph...	15.592	200	119540	15.952	ng/u1	0.00
73) Anthracene-d10	17.304	188	991983	18.364	ng/u1	0.00
81) Pyrene-d10	19.598	212	1108342	17.428	ng/u1	0.00
92) Benzo(a)pyrene-d12	23.580	264	1202908	18.138	ng/u1	0.00
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.299	88	48092	7.846	ng/uL	100
5) Pyridine	3.704	79	297431	18.860	ng/u1	100
6) Benzaldehyde	6.963	77	197616	22.173	ng/u1	100
8) Phenol	7.028	94	371895	19.686	ng/u1	100
10) Bis(2-Chloroethyl)ether	7.240	93	289733	19.787	ng/u1	100
12) 2-Chlorophenol	7.381	128	324169	19.927	ng/u1	100
13) 2-Methylphenol	8.275	108	271707	19.162	ng/u1	100
14) 2,2'-oxybis(1-Chloropr...	8.345	45	435422	20.932	ng/u1	100
16) Acetophenone	8.645	105	445341	19.055	ng/u1	100
17) N-Nitroso-di-n-propyla...	8.628	70	215659	19.463	ng/u1	100
18) 4-Methylphenol	8.604	108	292223	19.285	ng/u1	100
19) Hexachloroethane	8.886	117	126614	17.905	ng/u1	100
22) Nitrobenzene	9.016	77	313617	16.439	ng/u1	100
23) Isophorone	9.545	82	573527	17.034	ng/u1	100
25) 2-Nitrophenol	9.733	139	170527	17.832	ng/u1	100
26) 2,4-Dimethylphenol	9.804	107	320396	16.998	ng/u1	100
27) Bis(2-Chloroethoxy)met...	10.028	93	373664	18.438	ng/u1	100
29) 2,4-Dichlorophenol	10.275	162	269613	16.829	ng/u1	100
30) Naphthalene	10.657	128	979756	18.615	ng/u1	100
32) 4-Chloroaniline	10.786	127	392982	17.937	ng/u1	100
33) Hexachlorobutadiene	10.945	225	167312	14.362	ng/u1	100
34) Caprolactam	11.575	113	79787	17.285	ng/u1	100
35) 4-Chloro-3-methylphenol	11.927	107	275962	17.158	ng/u1	100
36) 2-Methylnaphthalene	12.280	142	636366	18.050	ng/u1	100

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM061622\
 Data File : BM035480.D
 Acq On : 16 Jun 2022 14:40
 Operator : CG/JU
 Sample : SSTD02049
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 SSTD020049

Quant Time: Jun 17 01:04:45 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\SFAM-EPA-BM061622.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Fri Jun 17 00:59:32 2022
 Response via : Initial Calibration


Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
37) 1-Methylnaphthalene	12.498	142	640060	18.112	ng/ul	100
39) 1,2,4,5-Tetrachloroben...	12.651	216	311284	16.594	ng/ul	100
40) Hexachlorocyclopentadiene	12.627	237	87064	6.509	ng/ul	100
41) 2,4,6-Trichlorophenol	12.904	196	193624	16.162	ng/ul	100
42) 2,4,5-Trichlorophenol	12.992	196	202624	15.935	ng/ul	100
43) 1,1'-Biphenyl	13.292	154	830544	18.389	ng/ul	100
44) 2-Chloronaphthalene	13.333	162	637148	18.192	ng/ul	100
45) 2-Nitroaniline	13.551	65	165361	16.803	ng/ul	100
47) Dimethylphthalate	13.921	163	753217	17.700	ng/ul	100
48) 2,6-Dinitrotoluene	14.045	165	151135	17.831	ng/ul	100
50) Acenaphthylene	14.174	152	1038068	19.015	ng/ul	100
51) 3-Nitroaniline	14.380	138	146980	18.083	ng/ul	100
52) Acenaphthene	14.521	153	681270	18.738	ng/ul	100
53) 2,4-Dinitrophenol	14.610	184	59999	10.805	ng/ul	100
55) 4-Nitrophenol	14.716	109	63513	8.985	ng/ul	100
56) Dibenzofuran	14.857	168	928226	18.255	ng/ul	100
57) 2,4-Dinitrotoluene	14.839	165	221970	18.650	ng/ul	100
58) 2,3,4,6-Tetrachlorophenol	15.098	232	165239	16.040	ng/ul	100
59) Diethylphthalate	15.280	149	764556	17.742	ng/ul	100
61) Fluorene	15.504	166	758428	18.332	ng/ul	100
62) 4-Chlorophenyl-phenyle...	15.498	204	357076	17.002	ng/ul	100
63) 4-Nitroaniline	15.545	138	136879	18.587	ng/ul	100
66) 4,6-Dinitro-2-methylph...	15.604	198	120781	16.080	ng/ul	100
67) N-Nitrosodiphenylamine	15.715	169	643691	18.335	ng/ul	100
68) 4-Bromophenyl-phenylether	16.392	248	208115	16.932	ng/ul	100
69) Hexachlorobenzene	16.515	284	240937	17.166	ng/ul	100
70) Atrazine	16.674	200	218911	15.981	ng/ul	100
71) Pentachlorophenol	16.874	266	114861	12.246	ng/ul	100
72) Phenanthrene	17.245	178	1174280	18.687	ng/ul	100
74) Anthracene	17.339	178	1206952	18.848	ng/ul	100
75) 1,2,3,4-Tetrachloroben...	13.257	216	325319	16.092	ng/ul	100
76) Pentachlorobenzene	14.780	250	299116	17.224	ng/ul	100
77) Carbazole	17.615	167	1083179	19.114	ng/ul	100
78) Di-n-butylphthalate	18.174	149	1335780	16.471	ng/ul	100
80) Fluoranthene	19.268	202	1344090	17.595	ng/ul	100
82) Pyrene	19.627	202	1417271	18.025	ng/ul	100
83) Butylbenzylphthalate	20.527	149	600014	17.815	ng/ul	100
84) 3,3'-Dichlorobenzidine	21.315	252	456330	18.590	ng/ul	100
85) Benzo(a)anthracene	21.374	228	1359335	17.820	ng/ul	100
86) Bis(2-ethylhexyl)phtha...	21.309	149	897658	17.725	ng/ul	100
87) Chrysene	21.427	228	1369483	18.413	ng/ul	100
89) Di-n-octyl phthalate	22.209	149	1574908	16.786	ng/ul	100
90) Benzo(b)fluoranthene	23.021	252	1449784	17.036	ng/ul	100
91) Benzo(k)fluoranthene	23.068	252	1434615	17.745	ng/ul	100
93) Benzo(a)pyrene	23.627	252	1489133	17.959	ng/ul	100
94) Indeno(1,2,3-cd)pyrene	26.150	276	1754825	18.417	ng/ul	100
95) Dibenzo(a,h)anthracene	26.162	278	1505839	18.399	ng/ul	100
96) Benzo(g,h,i)perylene	26.897	276	1503159	18.965	ng/ul	100


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM061622\
 Data File : BM035480.D
 Acq On : 16 Jun 2022 14:40
 Operator : CG/JU
 Sample : SST020049
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 SST020049

Quant Time: Jun 17 01:04:45 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\SFAM-EPA-BM061622.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Fri Jun 17 00:59:32 2022
 Response via : Initial Calibration









#2
1,4-Dioxane
Concen: 7.846 ng/uL
RT: 3.299 min Scan# 30
Delta R.T. 0.000 min
Lab File: BM035480.D
Acq: 16 Jun 2022 14:40




Instrument :
BNA_M
ClientSampleId :
SSTD020049


Tgt Ion: 88 Resp: 48092
Ion Ratio Lower Upper
88 100
43 34.8 27.8 41.8
58 77.3 61.8 92.8



#5
Pyridine
Concen: 18.860 ng/uL
RT: 3.704 min Scan# 105
Delta R.T. 0.000 min
Lab File: BM035480.D
Acq: 16 Jun 2022 14:40

Tgt Ion: 79 Resp: 297431
Ion Ratio Lower Upper
79 100
52 84.7 67.8 101.6
51 39.2 31.4 47.0









#6
Benzaldehyde
Concen: 22.173 ng/ul
RT: 6.963 min Scan# 659
Delta R.T. 0.000 min
Lab File: BM035480.D
Acq: 16 Jun 2022 14:40




Instrument :
BNA_M
ClientSampleId :
SSTD020049


Tgt Ion: 77 Resp: 197616
Ion Ratio Lower Upper
77 100
105 105.8 84.6 127.0
106 100.4 80.3 120.5



#8
Phenol
Concen: 19.686 ng/ul
RT: 7.028 min Scan# 670
Delta R.T. 0.000 min
Lab File: BM035480.D
Acq: 16 Jun 2022 14:40

Tgt Ion: 94 Resp: 371895
Ion Ratio Lower Upper
94 100
65 26.6 21.3 31.9
66 34.2 27.4 41.0





#10

Bis(2-Chloroethyl)ether

Concen: 19.787 ng/ul

RT: 7.240 min Scan# 706

Delta R.T. 0.000 min

Lab File: BM035480.D

Acq: 16 Jun 2022 14:40

Instrument :

BNA_M

ClientSampleId :

SSTD020049





Tgt Ion: 93 Resp: 289733

Ion Ratio Lower Upper

93 100

63 80.0 64.0 96.0

95 32.2 25.8 38.6



#12

2-Chlorophenol

Concen: 19.927 ng/ul

RT: 7.381 min Scan# 730

Delta R.T. 0.000 min

Lab File: BM035480.D

Acq: 16 Jun 2022 14:40




Tgt Ion:128 Resp: 324169


Ion Ratio Lower Upper

128 100

64 46.1 36.9 55.3

130 32.0 25.6 38.4









#13
2-Methylphenol
Concen: 19.162 ng/ul
RT: 8.275 min Scan# 882
Delta R.T. 0.000 min
Lab File: BM035480.D
Acq: 16 Jun 2022 14:40




Instrument :
BNA_M
ClientSampleId :
SSTD020049


Tgt Ion:108 Resp: 271707
Ion Ratio Lower Upper
108 100
107 93.7 75.0 112.4



#14
2,2'-oxybis(1-Chloropropane)
Concen: 20.932 ng/ul
RT: 8.345 min Scan# 894
Delta R.T. 0.000 min
Lab File: BM035480.D
Acq: 16 Jun 2022 14:40

Tgt Ion: 45 Resp: 435422
Ion Ratio Lower Upper
45 100
77 13.2 10.6 15.8
79 10.4 8.3 12.5









#16
Acetophenone
Concen: 19.055 ng/ul
RT: 8.645 min Scan# 945
Delta R.T. 0.000 min
Lab File: BM035480.D
Acq: 16 Jun 2022 14:40




Instrument :
BNA_M
ClientSampleId :
SSTD020049


Tgt Ion:105 Resp: 445341
Ion Ratio Lower Upper
105 100
77 74.7 59.8 89.6
51 27.6 22.1 33.1



#17
N-Nitroso-di-n-propylamine
Concen: 19.463 ng/ul
RT: 8.628 min Scan# 942
Delta R.T. 0.000 min
Lab File: BM035480.D
Acq: 16 Jun 2022 14:40

Tgt Ion: 70 Resp: 215659
Ion Ratio Lower Upper
70 100
42 62.0 49.6 74.4
101 10.1 8.1 12.1
130 23.0 18.4 27.6









#18
4-Methylphenol
Concen: 19.285 ng/ul
RT: 8.604 min Scan# 911
Delta R.T. 0.000 min
Lab File: BM035480.D
Acq: 16 Jun 2022 14:40




Instrument :
BNA_M
ClientSampleId :
SSTD020049


Tgt Ion:108 Resp: 292223
Ion Ratio Lower Upper
108 100
107 112.9 90.3 135.5



#19
Hexachloroethane
Concen: 17.905 ng/ul
RT: 8.886 min Scan# 986
Delta R.T. 0.000 min
Lab File: BM035480.D
Acq: 16 Jun 2022 14:40

Tgt Ion:117 Resp: 126614
Ion Ratio Lower Upper
117 100
201 102.2 81.8 122.6
199 64.5 51.6 77.4









#22
Nitrobenzene
Concen: 16.439 ng/ul
RT: 9.016 min Scan# 1008
Delta R.T. 0.000 min
Lab File: BM035480.D
Acq: 16 Jun 2022 14:40




Instrument :
BNA_M
ClientSampleId :
SSTD020049


Tgt Ion: 77 Resp: 313617
Ion Ratio Lower Upper
77 100
123 46.7 37.4 56.0
65 13.5 10.8 16.2



#23
Isophorone
Concen: 17.034 ng/ul
RT: 9.545 min Scan# 1098
Delta R.T. 0.000 min
Lab File: BM035480.D
Acq: 16 Jun 2022 14:40

Tgt Ion: 82 Resp: 573527
Ion Ratio Lower Upper
82 100
95 6.9 5.5 8.3
138 19.1 15.3 22.9









#25
2-Nitrophenol
Concen: 17.832 ng/ul
RT: 9.733 min Scan# 1130
Delta R.T. 0.000 min
Lab File: BM035480.D
Acq: 16 Jun 2022 14:40




Instrument :
BNA_M
ClientSampleId :
SSTD020049


Tgt Ion:139 Resp: 170527
Ion Ratio Lower Upper
139 100
65 49.7 39.8 59.6
109 27.8 22.2 33.4



#26
2,4-Dimethylphenol
Concen: 16.998 ng/ul
RT: 9.804 min Scan# 1142
Delta R.T. 0.000 min
Lab File: BM035480.D
Acq: 16 Jun 2022 14:40

Tgt Ion:107 Resp: 320396
Ion Ratio Lower Upper
107 100
121 53.4 42.7 64.1
122 89.4 71.5 107.3





#27

Bis(2-Chloroethoxy)methane

Concen: 18.438 ng/ul

RT: 10.028 min Scan# 1180

Delta R.T. 0.000 min

Lab File: BM035480.D

Acq: 16 Jun 2022 14:40

Instrument :

BNA_M

ClientSampleId :

SSTD020049





Tgt Ion: 93 Resp: 373664

Ion Ratio Lower Upper

93 100

95 32.6 26.1 39.1

123 11.7 9.4 14.0



#29

2,4-Dichlorophenol

Concen: 16.829 ng/ul

RT: 10.275 min Scan# 1222

Delta R.T. 0.000 min

Lab File: BM035480.D

Acq: 16 Jun 2022 14:40




Tgt Ion:162 Resp: 269613


Ion Ratio Lower Upper

162 100

164 64.7 51.8 77.6

98 36.9 29.5 44.3









#30
Naphthalene
Concen: 18.615 ng/ul
RT: 10.657 min Scan# 1287
Delta R.T. 0.000 min
Lab File: BM035480.D
Acq: 16 Jun 2022 14:40




Instrument :
BNA_M
ClientSampleId :
SSTD020049


Tgt Ion:128 Resp: 979756
Ion Ratio Lower Upper
128 100
129 11.0 8.8 13.2
127 13.3 10.6 16.0



#32
4-Chloroaniline
Concen: 17.937 ng/ul
RT: 10.786 min Scan# 1309
Delta R.T. 0.000 min
Lab File: BM035480.D
Acq: 16 Jun 2022 14:40

Tgt Ion:127 Resp: 392982
Ion Ratio Lower Upper
127 100
129 31.6 25.3 37.9









#33
Hexachlorobutadiene
Concen: 14.362 ng/ul
RT: 10.945 min Scan# 1336
Delta R.T. 0.000 min
Lab File: BM035480.D
Acq: 16 Jun 2022 14:40




Instrument :
BNA_M
ClientSampleId :
SSTD020049


Tgt Ion:225 Resp: 167312
Ion Ratio Lower Upper
225 100
223 62.5 50.0 75.0
227 62.4 49.9 74.9



#34
Caprolactam
Concen: 17.285 ng/ul
RT: 11.575 min Scan# 1443
Delta R.T. 0.000 min
Lab File: BM035480.D
Acq: 16 Jun 2022 14:40

Tgt Ion:113 Resp: 79787
Ion Ratio Lower Upper
113 100
55 187.8 150.2 225.4
56 127.6 102.1 153.1









#35
4-Chloro-3-methylphenol
Concen: 17.158 ng/ul
RT: 11.927 min Scan# 1503
Delta R.T. 0.000 min
Lab File: BM035480.D
Acq: 16 Jun 2022 14:40




Instrument :
BNA_M
ClientSampleId :
SSTD020049


Tgt Ion:107 Resp: 275962
Ion Ratio Lower Upper
107 100
144 25.4 20.3 30.5
142 80.5 64.4 96.6



#36
2-Methylnaphthalene
Concen: 18.050 ng/ul
RT: 12.280 min Scan# 1563
Delta R.T. 0.000 min
Lab File: BM035480.D
Acq: 16 Jun 2022 14:40

Tgt Ion:142 Resp: 636366
Ion Ratio Lower Upper
142 100
141 88.4 70.7 106.1





#37

1-Methylnaphthalene

Concen: 18.112 ng/ul

RT: 12.498 min Scan# 1

Delta R.T. 0.000 min

Lab File: BM035480.D

Acq: 16 Jun 2022 14:40

Instrument :

BNA_M

ClientSampleId :

SSTD020049





Tgt Ion:142 Resp: 640060

Ion Ratio Lower Upper

142 100

141 94.0 75.2 112.8

116 3.2 2.6 3.8



#39

1,2,4,5-Tetrachlorobenzene

Concen: 16.594 ng/ul

RT: 12.651 min Scan# 1626

Delta R.T. 0.000 min

Lab File: BM035480.D

Acq: 16 Jun 2022 14:40

Tgt Ion:216 Resp: 311284




Ion Ratio Lower Upper


216 100

214 78.0 62.4 93.6

179 21.6 17.3 25.9

108 17.5 14.0 21.0









#40
Hexachlorocyclopentadiene
Concen: 6.509 ng/ul
RT: 12.627 min Scan# 1622
Delta R.T. 0.000 min
Lab File: BM035480.D
Acq: 16 Jun 2022 14:40




Instrument :
BNA_M
ClientSampleId :
SSTD020049


Tgt Ion:237 Resp: 87064
Ion Ratio Lower Upper
237 100
235 60.5 48.4 72.6
272 14.5 11.6 17.4



#41
2,4,6-Trichlorophenol
Concen: 16.162 ng/ul
RT: 12.904 min Scan# 1669
Delta R.T. 0.000 min
Lab File: BM035480.D
Acq: 16 Jun 2022 14:40

Tgt Ion:196 Resp: 193624
Ion Ratio Lower Upper
196 100
198 97.2 77.8 116.6
200 31.3 25.0 37.6









#42
2,4,5-Trichlorophenol
Concen: 15.935 ng/ul
RT: 12.992 min Scan# 1
Delta R.T. 0.000 min
Lab File: BM035480.D
Acq: 16 Jun 2022 14:40




Instrument :
BNA_M
ClientSampleId :
SSTD020049


Tgt Ion:196 Resp: 202624
Ion Ratio Lower Upper
196 100
198 96.0 76.8 115.2
200 31.1 24.9 37.3



#43
1,1'-Biphenyl
Concen: 18.389 ng/ul
RT: 13.292 min Scan# 1735
Delta R.T. 0.000 min
Lab File: BM035480.D
Acq: 16 Jun 2022 14:40

Tgt Ion:154 Resp: 830544
Ion Ratio Lower Upper
154 100
153 40.5 32.4 48.6
76 12.6 10.1 15.1









#44
2-Chloronaphthalene
Concen: 18.192 ng/ul
RT: 13.333 min Scan# 1742
Delta R.T. 0.000 min
Lab File: BM035480.D
Acq: 16 Jun 2022 14:40




Instrument :
BNA_M
ClientSampleId :
SSTD020049


Tgt Ion:162 Resp: 637148
Ion Ratio Lower Upper
162 100
164 31.7 25.4 38.0
127 39.5 31.6 47.4



#45
2-Nitroaniline
Concen: 16.803 ng/ul
RT: 13.551 min Scan# 1779
Delta R.T. 0.000 min
Lab File: BM035480.D
Acq: 16 Jun 2022 14:40

Tgt Ion: 65 Resp: 165361
Ion Ratio Lower Upper
65 100
92 68.1 54.5 81.7
138 114.0 91.2 136.8









#47
Dimethylphthalate
Concen: 17.700 ng/ul
RT: 13.921 min Scan# 1842
Delta R.T. 0.000 min
Lab File: BM035480.D
Acq: 16 Jun 2022 14:40




Instrument :
BNA_M
ClientSampleId :
SSTD020049


Tgt Ion:163 Resp: 753217
Ion Ratio Lower Upper
163 100
194 4.0 3.2 4.8
164 9.9 7.9 11.9



#48
2,6-Dinitrotoluene
Concen: 17.831 ng/ul
RT: 14.045 min Scan# 1863
Delta R.T. 0.000 min
Lab File: BM035480.D
Acq: 16 Jun 2022 14:40

Tgt Ion:165 Resp: 151135
Ion Ratio Lower Upper
165 100
89 53.1 42.5 63.7
121 21.5 17.2 25.8









#50
Acenaphthylene
Concen: 19.015 ng/ul
RT: 14.174 min Scan# 1885
Delta R.T. 0.000 min
Lab File: BM035480.D
Acq: 16 Jun 2022 14:40




Instrument :
BNA_M
ClientSampleId :
SSTD020049


Tgt Ion:152 Resp: 1038068
Ion Ratio Lower Upper
152 100
151 20.1 16.1 24.1
153 13.2 10.6 15.8



#51
3-Nitroaniline
Concen: 18.083 ng/ul
RT: 14.380 min Scan# 1920
Delta R.T. 0.000 min
Lab File: BM035480.D
Acq: 16 Jun 2022 14:40

Tgt Ion:138 Resp: 146980
Ion Ratio Lower Upper
138 100
108 10.1 8.1 12.1
92 113.0 90.4 135.6









#52
Acenaphthene
Concen: 18.738 ng/ul
RT: 14.521 min Scan# 1944
Delta R.T. 0.000 min
Lab File: BM035480.D
Acq: 16 Jun 2022 14:40




Instrument :
BNA_M
ClientSampleId :
SSTD020049


Tgt Ion:153 Resp: 681270
Ion Ratio Lower Upper
153 100
152 46.4 37.1 55.7
154 89.1 71.3 106.9



#53
2,4-Dinitrophenol
Concen: 10.805 ng/ul
RT: 14.610 min Scan# 1959
Delta R.T. 0.000 min
Lab File: BM035480.D
Acq: 16 Jun 2022 14:40

Tgt Ion:184 Resp: 59999
Ion Ratio Lower Upper
184 100
63 67.8 54.2 81.4
154 61.6 49.3 73.9









#55
4-Nitrophenol
Concen: 8.985 ng/ul
RT: 14.716 min Scan# 1977
Delta R.T. 0.000 min
Lab File: BM035480.D
Acq: 16 Jun 2022 14:40




Instrument :
BNA_M
ClientSampleId :
SSTD020049


Tgt Ion:109 Resp: 63513
Ion Ratio Lower Upper
109 100
139 156.1 124.9 187.3
65 152.8 122.2 183.4



#56
Dibenzofuran
Concen: 18.255 ng/ul
RT: 14.857 min Scan# 2001
Delta R.T. 0.000 min
Lab File: BM035480.D
Acq: 16 Jun 2022 14:40

Tgt Ion:168 Resp: 928226
Ion Ratio Lower Upper
168 100
139 37.6 30.1 45.1









#57
2,4-Dinitrotoluene
Concen: 18.650 ng/ul
RT: 14.839 min Scan# 1998
Delta R.T. 0.000 min
Lab File: BM035480.D
Acq: 16 Jun 2022 14:40




Instrument :
BNA_M
ClientSampleId :
SSTD020049


Tgt Ion:165 Resp: 221970
Ion Ratio Lower Upper
165 100
63 44.9 35.9 53.9
182 3.1 2.5 3.7



#58
2,3,4,6-Tetrachlorophenol
Concen: 16.040 ng/ul
RT: 15.098 min Scan# 2042
Delta R.T. 0.000 min
Lab File: BM035480.D
Acq: 16 Jun 2022 14:40

Tgt Ion:232 Resp: 165239
Ion Ratio Lower Upper
232 100
131 45.9 36.7 55.1
130 2.1 1.7 2.5
166 30.5 24.4 36.6





#59

Diethylphthalate

Concen: 17.742 ng/ul

RT: 15.280 min Scan# 2073

Delta R.T. 0.000 min

Lab File: BM035480.D

Acq: 16 Jun 2022 14:40

Instrument :

BNA_M

ClientSampleId :

SSTD020049





Tgt Ion:149 Resp: 764556

Ion Ratio Lower Upper

149 100

177 21.0 16.8 25.2

150 12.8 10.2 15.4



#61

Fluorene

Concen: 18.332 ng/ul

RT: 15.504 min Scan# 2111

Delta R.T. 0.000 min

Lab File: BM035480.D

Acq: 16 Jun 2022 14:40




Tgt Ion:166 Resp: 758428


Ion Ratio Lower Upper

166 100

165 96.5 77.2 115.8

167 13.7 11.0 16.4





#62

4-Chlorophenyl-phenylether

Concen: 17.002 ng/ul

RT: 15.498 min Scan# 2110

Delta R.T. 0.000 min

Lab File: BM035480.D

Acq: 16 Jun 2022 14:40

Instrument :

BNA_M

ClientSampleId :

SSTD020049





Tgt Ion:204 Resp: 357076

Ion Ratio Lower Upper

204 100

206 32.4 25.9 38.9

141 63.2 50.6 75.8



#63

4-Nitroaniline

Concen: 18.587 ng/ul

RT: 15.545 min Scan# 2118

Delta R.T. 0.000 min

Lab File: BM035480.D

Acq: 16 Jun 2022 14:40




Tgt Ion:138 Resp: 136879


Ion Ratio Lower Upper

138 100

92 49.0 39.2 58.8

108 64.5 51.6 77.4









#66
4,6-Dinitro-2-methylphenol
Concen: 16.080 ng/ul
RT: 15.604 min Scan# 2128
Delta R.T. 0.000 min
Lab File: BM035480.D
Acq: 16 Jun 2022 14:40




Instrument :
BNA_M
ClientSampleId :
SSTD020049


Tgt Ion:198 Resp: 120781
Ion Ratio Lower Upper
198 100
182 5.3 4.2 6.4
77 24.9 19.9 29.9



#67
N-Nitrosodiphenylamine
Concen: 18.335 ng/ul
RT: 15.715 min Scan# 2147
Delta R.T. 0.000 min
Lab File: BM035480.D
Acq: 16 Jun 2022 14:40

Tgt Ion:169 Resp: 643691
Ion Ratio Lower Upper
169 100
168 67.2 53.8 80.6
167 35.9 28.7 43.1









#68
4-Bromophenyl-phenylether
Concen: 16.932 ng/ul
RT: 16.392 min Scan# 2262
Delta R.T. 0.000 min
Lab File: BM035480.D
Acq: 16 Jun 2022 14:40




Instrument :
BNA_M
ClientSampleId :
SSTD020049


Tgt Ion:248 Resp: 208115
Ion Ratio Lower Upper
248 100
250 96.6 77.3 115.9
141 74.8 59.8 89.8



#69
Hexachlorobenzene
Concen: 17.166 ng/ul
RT: 16.515 min Scan# 2283
Delta R.T. 0.000 min
Lab File: BM035480.D
Acq: 16 Jun 2022 14:40

Tgt Ion:284 Resp: 240937
Ion Ratio Lower Upper
284 100
142 35.7 28.6 42.8
249 30.8 24.6 37.0





#70

Atrazine

Concen: 15.981 ng/ul

RT: 16.674 min Scan# 21

Delta R.T. 0.000 min

Lab File: BM035480.D

Acq: 16 Jun 2022 14:40

Instrument :

BNA_M

ClientSampleId :

SSTD020049





Tgt Ion:200 Resp: 218911

Ion Ratio Lower Upper

200 100

173 26.6 21.3 31.9

215 48.8 39.0 58.6



#71

Pentachlorophenol

Concen: 12.246 ng/ul

RT: 16.874 min Scan# 2344

Delta R.T. 0.000 min

Lab File: BM035480.D

Acq: 16 Jun 2022 14:40




Tgt Ion:266 Resp: 114861


Ion Ratio Lower Upper

266 100

264 61.9 49.5 74.3

268 63.7 51.0 76.4









#72
Phenanthrene
Concen: 18.687 ng/ul
RT: 17.245 min Scan# 2407
Delta R.T. 0.000 min
Lab File: BM035480.D
Acq: 16 Jun 2022 14:40




Instrument :
BNA_M
ClientSampleId :
SSTD020049


Tgt Ion:178 Resp: 1174280
Ion Ratio Lower Upper
178 100
179 15.2 12.2 18.2
176 18.6 14.9 22.3



#74
Anthracene
Concen: 18.848 ng/ul
RT: 17.339 min Scan# 2423
Delta R.T. 0.000 min
Lab File: BM035480.D
Acq: 16 Jun 2022 14:40

Tgt Ion:178 Resp: 1206952
Ion Ratio Lower Upper
178 100
179 15.2 12.2 18.2
176 18.0 14.4 21.6









#75
1,2,3,4-Tetrachlorobenzene
Concen: 16.092 ng/uL
RT: 13.257 min Scan# 1
Delta R.T. 0.000 min
Lab File: BM035480.D
Acq: 16 Jun 2022 14:40




Instrument :
BNA_M
ClientSampleId :
SSTD020049


Tgt Ion:216 Resp: 325319
Ion Ratio Lower Upper
216 100
214 76.1 60.9 91.3
218 46.9 37.5 56.3



#76
Pentachlorobenzene
Concen: 17.224 ng/uL
RT: 14.780 min Scan# 1988
Delta R.T. 0.000 min
Lab File: BM035480.D
Acq: 16 Jun 2022 14:40

Tgt Ion:250 Resp: 299116
Ion Ratio Lower Upper
250 100
248 63.0 50.4 75.6
252 63.9 51.1 76.7









#77
 Carbazole
 Concen: 19.114 ng/ul
 RT: 17.615 min Scan# 2470
 Delta R.T. 0.000 min
 Lab File: BM035480.D
 Acq: 16 Jun 2022 14:40




Instrument :
 BNA_M
 ClientSampleId :
 SSTD020049


Tgt Ion:167 Resp: 1083179
 Ion Ratio Lower Upper
 167 100
 166 21.7 17.4 26.0
 139 12.6 10.1 15.1



#78
 Di-n-butylphthalate
 Concen: 16.471 ng/ul
 RT: 18.174 min Scan# 2565
 Delta R.T. 0.000 min
 Lab File: BM035480.D
 Acq: 16 Jun 2022 14:40

Tgt Ion:149 Resp: 1335780
 Ion Ratio Lower Upper
 149 100
 150 9.1 7.3 10.9
 104 4.6 3.7 5.5









#80
Fluoranthene
Concen: 17.595 ng/ul
RT: 19.268 min Scan# 2751
Delta R.T. 0.000 min
Lab File: BM035480.D
Acq: 16 Jun 2022 14:40




Instrument :
BNA_M
ClientSampleId :
SSTD020049


Tgt Ion:202 Resp: 1344090
Ion Ratio Lower Upper
202 100
101 12.6 10.1 15.1
100 9.7 7.8 11.6



#82
Pyrene
Concen: 18.025 ng/ul
RT: 19.627 min Scan# 2812
Delta R.T. 0.000 min
Lab File: BM035480.D
Acq: 16 Jun 2022 14:40

Tgt Ion:202 Resp: 1417271
Ion Ratio Lower Upper
202 100
101 14.9 11.9 17.9
100 12.8 10.2 15.4









#83
 Butylbenzylphthalate
 Concen: 17.815 ng/ul
 RT: 20.527 min Scan# 2965
 Delta R.T. 0.000 min
 Lab File: BM035480.D
 Acq: 16 Jun 2022 14:40




Instrument :
 BNA_M
 ClientSampleId :
 SSTD020049


Tgt Ion:149 Resp: 600014
 Ion Ratio Lower Upper
 149 100
 91 66.5 53.2 79.8
 206 20.0 16.0 24.0



#84
 3,3'-Dichlorobenzidine
 Concen: 18.590 ng/ul
 RT: 21.315 min Scan# 3099
 Delta R.T. 0.000 min
 Lab File: BM035480.D
 Acq: 16 Jun 2022 14:40

Tgt Ion:252 Resp: 456330
 Ion Ratio Lower Upper
 252 100
 254 65.5 52.4 78.6
 126 13.5 10.8 16.2









#85
Benzo(a)anthracene
Concen: 17.820 ng/ul
RT: 21.374 min Scan# 3109
Delta R.T. 0.000 min
Lab File: BM035480.D
Acq: 16 Jun 2022 14:40




Instrument :
BNA_M
ClientSampleId :
SSTD020049


Tgt Ion:228 Resp: 1359335
Ion Ratio Lower Upper
228 100
229 19.5 15.6 23.4
226 27.1 21.7 32.5



#86
Bis(2-ethylhexyl)phthalate
Concen: 17.725 ng/ul
RT: 21.309 min Scan# 3098
Delta R.T. 0.000 min
Lab File: BM035480.D
Acq: 16 Jun 2022 14:40

Tgt Ion:149 Resp: 897658
Ion Ratio Lower Upper
149 100
167 26.9 21.5 32.3
279 4.3 3.4 5.2









#87
Chrysene
Concen: 18.413 ng/ul
RT: 21.427 min Scan# 3118
Delta R.T. 0.000 min
Lab File: BM035480.D
Acq: 16 Jun 2022 14:40




Instrument :
BNA_M
ClientSampleId :
SSTD020049


Tgt Ion:228 Resp: 1369483
Ion Ratio Lower Upper
228 100
226 29.6 23.7 35.5
229 19.4 15.5 23.3



#89
Di-n-octyl phthalate
Concen: 16.786 ng/ul
RT: 22.209 min Scan# 3251
Delta R.T. 0.000 min
Lab File: BM035480.D
Acq: 16 Jun 2022 14:40

Tgt Ion:149 Resp: 1574908









#90
Benzo(b)fluoranthene
Concen: 17.036 ng/ul
RT: 23.021 min Scan# 3389
Delta R.T. 0.000 min
Lab File: BM035480.D
Acq: 16 Jun 2022 14:40




Instrument :
BNA_M
ClientSampleId :
SSTD020049


Tgt Ion:252 Resp: 1449784
Ion Ratio Lower Upper
252 100
253 21.9 17.5 26.3
125 13.4 10.7 16.1



#91
Benzo(k)fluoranthene
Concen: 17.745 ng/ul
RT: 23.068 min Scan# 3397
Delta R.T. 0.000 min
Lab File: BM035480.D
Acq: 16 Jun 2022 14:40

Tgt Ion:252 Resp: 1434615
Ion Ratio Lower Upper
252 100
253 21.7 17.4 26.0
125 13.2 10.6 15.8









#93
Benzo(a)pyrene
Concen: 17.959 ng/ul
RT: 23.627 min Scan# 3492
Delta R.T. 0.000 min
Lab File: BM035480.D
Acq: 16 Jun 2022 14:40




Instrument :
BNA_M
ClientSampleId :
SSTD020049


Tgt Ion:252 Resp: 1489133
Ion Ratio Lower Upper
252 100
253 21.6 17.3 25.9
125 14.8 11.8 17.8



#94
Indeno(1,2,3-cd)pyrene
Concen: 18.417 ng/ul
RT: 26.150 min Scan# 3921
Delta R.T. 0.000 min
Lab File: BM035480.D
Acq: 16 Jun 2022 14:40

Tgt Ion:276 Resp: 1754825
Ion Ratio Lower Upper
276 100
138 32.0 25.6 38.4
277 25.1 20.1 30.1









#95
 Dibenzo(a,h)anthracene
 Concen: 18.399 ng/ul
 RT: 26.162 min Scan# 3923
 Delta R.T. 0.000 min
 Lab File: BM035480.D
 Acq: 16 Jun 2022 14:40

Instrument :
 BNA_M
 ClientSampleId :
 SSTD020049

Tgt Ion:278 Resp: 1505839

Ion	Ratio	Lower	Upper
278	100		
139	20.2	16.2	24.2
279	23.4	18.7	28.1



#96
 Benzo(g,h,i)perylene
 Concen: 18.965 ng/ul
 RT: 26.897 min Scan# 4048
 Delta R.T. 0.000 min
 Lab File: BM035480.D
 Acq: 16 Jun 2022 14:40

Tgt Ion:276 Resp: 1503159

Ion	Ratio	Lower	Upper
276	100		
138	26.6	21.3	31.9
277	23.3	18.6	28.0

