


Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM061622\
 Data File : BM035477.D
 Acq On : 16 Jun 2022 12:13
 Operator : CG/JU
 Sample : DFTPP055
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 DFTPP055

Integration File: LSCINT.P

Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\SFAM-EPA-BM061622.M
 Title : SVOA CALIBRATION
 Last Update : Fri Jun 17 01:58:49 2022



AutoFind: Scans 2476, 2477, 2478; Background Corrected with Scan 2469

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	35.3	339610	PASS
68	69	0.00	2	1.8	6010	PASS
69	198	0.00	100	35.4	340876	PASS
70	69	0.00	2	0.5	1782	PASS
127	198	10	80	47.6	457835	PASS
197	198	0.00	2	0.4	4225	PASS
198	198	100	100	100.0	962709	PASS
199	198	5	9	6.8	65445	PASS
275	198	10	60	27.4	263552	PASS
365	198	1	100	3.4	32464	PASS
441	198	0.01	100	14.3	137640	PASS
442	198	50	100	93.3	897941	PASS
443	442	15	24	19.4	173845	PASS

m/z	Abundance
39.00	334.0
40.00	1629.0
41.00	930.0
42.00	639.0
43.00	2938.0
44.00	4402.0
45.00	2216.0
48.90	466.0
50.00	164.0
51.20	187.0
55.10	582.0
56.00	511.0
57.00	351.0
60.10	1646.0
64.90	190.0
66.90	173.0
69.00	201.0
71.00	215.0
73.00	1438.0
75.10	195.0
77.10	285.0
78.00	216.0
81.10	167.0
83.10	249.0
84.00	582.0
85.90	266.0
86.90	248.0
90.90	203.0
94.90	167.0
96.00	457.0
97.10	281.0
105.00	176.0
111.10	161.0
132.90	284.0
134.90	194.0
147.00	360.0
149.10	157.0
177.00	237.0
190.90	326.0
193.00	320.0
207.00	2968.0
207.90	631.0
209.00	338.0
221.00	235.0
267.00	241.0
281.00	797.0
281.90	205.0
282.90	150.0
355.10	181.0

Instrument :
BNA_M
ClientSampleId :
DFTPP055

m/z	Abundance
36.00	337.0
37.00	1679.0
38.00	5276.0
39.00	31904.0
40.00	2776.0
41.00	1351.0
42.00	614.0
43.00	2655.0
44.00	3457.0
45.00	2524.0
47.00	167.0
48.00	355.0
49.10	3127.0
50.00	111872.0
51.00	440256.0
52.00	21664.0
53.00	984.0
55.00	2747.0
56.00	13275.0
57.00	31664.0
58.00	1388.0
59.00	426.0
60.00	1655.0
61.00	5572.0
62.00	6327.0
63.00	17656.0
64.00	2311.0
65.10	8809.0
66.10	541.0
67.00	853.0
68.10	7312.0
69.00	437696.0
70.00	2037.0
70.90	523.0
73.10	5088.0
74.00	44088.0
75.00	69592.0
76.10	23008.0
77.00	484352.0
78.10	32888.0
79.00	32552.0
80.00	26776.0
81.00	35384.0
82.00	8416.0
83.00	7996.0
84.00	1113.0
85.00	7105.0
86.00	9684.0
87.00	4396.0
88.00	1576.0
88.90	870.0
91.00	8311.0
92.00	8715.0
93.00	54520.0
94.00	3655.0
95.00	1012.0
96.00	3113.0
97.00	1104.0
98.00	43824.0
99.00	34384.0
100.00	2571.0
101.00	20232.0
102.00	1272.0
103.00	7036.0
104.00	11854.0
105.00	11425.0
106.00	3641.0
107.00	152832.0
108.00	23376.0
109.10	4545.0
110.00	296384.0
111.00	43928.0
112.00	5598.0
113.00	1514.0
113.90	287.0
114.90	686.0
116.00	8635.0
117.00	116048.0
118.00	8888.0
119.00	1292.0
120.00	1663.0
121.10	877.0
122.00	9440.0
123.00	15853.0
124.00	7186.0
125.00	6743.0
127.00	563456.0
128.00	42320.0
129.00	214336.0
130.00	17496.0
131.00	3306.0
131.90	1877.0

Instrument :
BNA_M
ClientSampleId :
DFTPP055

132.70	888.0
134.00	5980.0
135.00	17448.0
136.00	6749.0
137.00	9555.0
137.80	2105.0
139.00	1267.0
140.00	2768.0
141.00	27752.0
142.00	8425.0
143.00	5767.0
144.10	1763.0
145.00	1553.0
146.00	5201.0
147.00	15007.0
148.00	30328.0
149.00	6480.0
150.10	1925.0
151.00	3811.0
151.90	2479.0
153.00	8965.0
154.00	6649.0
155.00	15518.0
156.00	21672.0
157.00	4093.0
157.90	5173.0
159.00	4080.0
160.00	9352.0
161.00	12799.0
162.00	3411.0
163.00	1356.0
163.90	1427.0
165.00	10144.0
166.00	7912.0
167.00	51984.0
168.00	26528.0
169.00	5166.0
170.10	2138.0
171.00	1891.0
172.00	5010.0
173.00	6609.0
174.00	11114.0
175.00	21320.0
176.00	6206.0
177.00	9952.0
178.00	3322.0
179.00	38856.0
180.00	28872.0
181.00	13236.0
182.00	1977.0
182.90	1130.0
184.00	3520.0
185.00	19040.0
186.00	149440.0
187.00	42224.0
188.00	4831.0
189.00	9605.0
190.00	1577.0
191.00	4680.0
192.00	12999.0
193.00	13893.0
194.00	3075.0
195.00	2214.0
196.00	35432.0
197.00	6238.0
198.00	1111040.0
199.00	77896.0
200.00	6061.0
201.60	5663.0
203.00	7448.0
204.00	37280.0
205.10	65016.0
206.10	280384.0
207.10	39560.0
208.00	9065.0
209.00	3434.0
210.10	4872.0
211.00	10273.0
213.00	721.0
214.10	266.0
215.00	3019.0
216.00	5259.0
217.00	69816.0
218.00	8895.0
219.00	1117.0
221.10	47952.0
221.70	11255.0
223.00	17176.0
224.00	149952.0
225.00	38608.0
226.00	4718.0
227.00	58944.0
228.00	8852.0
229.00	13885.0
230.00	1978.0
231.00	5582.0

Instrument :
BNA_M
ClientSampleId :
DFTPP055

232.00	1078.0
233.00	1271.0
234.00	4025.0
235.00	5020.0
236.00	2951.0
237.00	5550.0
238.00	821.0
239.00	2374.0
240.00	1717.0
241.00	3482.0
242.00	8147.0
243.00	7826.0
244.10	120328.0
245.00	15863.0
246.00	22664.0
247.00	4689.0
248.00	890.0
249.00	4416.0
250.00	884.0
250.90	1071.0
252.00	1228.0
253.00	2125.0
254.10	5188.0
255.00	614528.0
256.00	93232.0
257.00	7335.0
258.00	31968.0
259.00	5275.0
260.00	1017.0
261.00	1050.0
261.90	186.0
263.00	439.0
264.00	867.0
265.00	13510.0
266.00	2014.0
266.90	352.0
268.00	185.0
269.00	159.0
270.00	508.0
271.00	1246.0
272.00	1568.0
273.00	17320.0
274.00	47520.0
275.00	286208.0
276.00	38704.0
277.00	22392.0
278.00	3588.0
279.00	898.0
279.80	212.0
281.10	1105.0
281.90	759.0
283.10	2625.0
284.00	1347.0
285.10	3702.0
286.00	905.0
287.90	333.0
289.00	829.0
290.00	732.0
290.80	484.0
292.10	939.0
293.00	4853.0
294.10	1423.0
294.90	1631.0
296.00	71096.0
297.00	10338.0
297.90	603.0
298.80	209.0
300.90	697.0
302.00	1415.0
303.00	8540.0
304.10	2214.0
304.90	359.0
308.00	1521.0
309.00	535.0
310.00	1099.0
311.00	214.0
312.10	255.0
312.90	734.0
314.00	3825.0
315.00	8094.0
316.00	4417.0
317.00	700.0
320.00	295.0
321.00	2287.0
322.00	1177.0
323.10	25144.0
324.00	4545.0
325.10	501.0
326.00	584.0
327.00	4526.0
328.00	2254.0
329.00	522.0
331.20	228.0
332.00	1756.0
333.00	2328.0
334.10	14820.0

Instrument :
BNA_M
ClientSampleId :
DFTPP055

335.00	4421.0
336.00	536.0
338.90	238.0
340.10	279.0
341.10	2737.0
342.00	715.0
346.00	4732.0
347.00	872.0
349.90	159.0
351.10	403.0
352.10	7205.0
353.00	4905.0
354.10	8196.0
355.10	1508.0
356.00	277.0
358.00	155.0
359.10	652.0
364.10	309.0
365.00	32152.0
366.00	3919.0
367.00	447.0
370.00	561.0
371.00	1726.0
372.10	13436.0
373.00	3236.0
374.10	336.0
377.00	400.0
383.00	3141.0
384.00	1141.0
384.90	417.0
390.00	1730.0
391.10	1208.0
392.10	1064.0
397.00	168.0
401.10	535.0
402.00	4567.0
403.00	7505.0
404.00	2800.0
404.90	495.0
409.90	191.0
415.00	300.0
421.10	6683.0
422.10	6751.0
423.10	47504.0
424.10	12289.0
425.00	1321.0
437.90	151.0
439.00	349.0
440.10	851.0
441.10	128768.0
442.10	833280.0
443.10	166208.0
444.10	15530.0
445.10	917.0

Instrument :
BNA_M
ClientSampleId :
DFTPP055

m/z	Abundance
35.00	252.0
37.00	1832.0
38.00	4089.0
39.00	27504.0
40.00	2977.0
41.00	1553.0
42.00	554.0
43.00	2542.0
44.00	3734.0
45.00	2986.0
46.00	244.0
48.10	204.0
49.00	3595.0
50.00	98032.0
51.00	386432.0
52.00	20112.0
53.00	1059.0
55.00	2378.0
56.00	12076.0
57.00	26384.0
58.00	1181.0
59.10	569.0
60.00	1404.0
61.00	4869.0
62.00	6026.0
63.00	15631.0
64.00	2049.0
65.00	8191.0
66.00	718.0
67.00	716.0
68.00	7121.0
69.00	389696.0
70.00	2234.0
71.10	267.0
72.00	227.0
73.10	4203.0
74.00	39352.0
75.00	63552.0
76.10	21360.0
77.00	436800.0
78.00	30408.0
79.00	29008.0
80.00	22168.0
81.00	32256.0
82.00	7928.0
83.00	6817.0
84.00	1282.0
85.00	6395.0
86.00	9774.0
87.00	4188.0
88.00	1461.0
89.00	676.0
89.90	219.0
91.00	7297.0
92.00	8809.0
93.00	50064.0
94.00	3420.0
95.00	980.0
96.00	2546.0
97.00	1056.0
98.00	38280.0
99.00	30936.0
100.00	2786.0
101.00	18664.0
102.00	1007.0
103.00	6493.0
104.00	12016.0
105.00	11509.0
106.00	3807.0
107.00	140608.0
108.00	22080.0
109.00	3587.0
110.00	268608.0
111.00	39424.0
112.00	4754.0
113.00	1462.0
114.00	334.0
115.10	674.0
116.00	8408.0
117.00	109176.0
118.00	8050.0
119.00	1098.0
120.00	1809.0
121.00	979.0
122.00	9807.0
123.00	15188.0
124.00	6764.0
125.00	6235.0
126.10	3287.0
127.00	526912.0
128.00	39832.0
129.00	199872.0

Instrument :
BNA_M
ClientSampleId :
DFTPP055

130.00	16488.0
131.00	3441.0
132.10	1733.0
133.10	864.0
134.00	5741.0
135.00	17368.0
136.00	6852.0
137.00	7798.0
137.90	1869.0
139.10	781.0
140.00	2600.0
141.00	26256.0
142.00	9089.0
143.00	5794.0
144.00	1653.0
145.00	1123.0
146.00	5033.0
147.00	14290.0
148.00	28416.0
149.00	6469.0
150.00	1913.0
151.00	3586.0
151.80	2194.0
153.00	8367.0
154.00	6017.0
155.00	14520.0
156.00	21744.0
157.00	4525.0
158.00	5457.0
159.00	3406.0
160.00	8420.0
161.00	11041.0
162.00	3899.0
163.00	1129.0
164.10	1693.0
165.00	9093.0
166.00	7944.0
167.00	50592.0
168.00	24984.0
169.00	3632.0
170.10	1901.0
171.00	2163.0
171.90	4878.0
173.00	5982.0
174.00	11224.0
175.00	19112.0
176.10	6204.0
177.00	10112.0
178.00	3301.0
179.00	38440.0
180.00	26352.0
181.00	12640.0
182.00	1964.0
183.00	971.0
184.00	3115.0
185.00	18624.0
186.00	151040.0
187.00	43704.0
188.10	4563.0
189.00	9729.0
190.00	1458.0
191.00	4665.0
192.00	12699.0
193.00	14337.0
194.00	3107.0
195.00	1481.0
196.00	36304.0
197.00	6436.0
198.00	1134592.0
199.00	76656.0
200.00	5788.0
201.50	4970.0
203.00	7583.0
204.10	38168.0
205.00	66160.0
206.10	287424.0
207.00	40152.0
208.00	9543.0
209.00	3136.0
211.00	10466.0
213.00	916.0
214.00	382.0
215.00	3144.0
216.00	5790.0
217.00	72736.0
218.00	10349.0
219.00	736.0
221.00	49728.0
221.80	10747.0
223.00	16448.0
224.00	156544.0
225.00	42296.0
226.10	4468.0
227.00	62536.0
228.00	8671.0
229.00	14039.0

Instrument :
BNA_M
ClientSampleId :
DFTPP055

230.00	1738.0
231.00	6076.0
232.00	902.0
232.90	983.0
234.00	4048.0
235.00	4247.0
235.90	3013.0
237.00	5453.0
237.80	858.0
239.00	2242.0
240.00	1978.0
241.00	3376.0
242.00	8478.0
243.10	8844.0
244.10	126328.0
245.10	16792.0
246.00	25264.0
247.00	4796.0
248.00	1227.0
249.00	5015.0
250.00	914.0
251.00	1230.0
252.00	1539.0
253.00	2770.0
254.10	5321.0
255.00	662144.0
256.00	94616.0
257.00	6747.0
258.00	35768.0
259.00	5275.0
260.00	1130.0
261.00	1173.0
261.90	164.0
263.00	394.0
264.10	1084.0
265.00	13390.0
266.00	2166.0
267.00	329.0
268.00	196.0
270.00	804.0
271.00	1095.0
272.10	1710.0
273.00	20064.0
274.00	53656.0
275.00	310144.0
276.00	40328.0
277.00	23704.0
278.00	3751.0
279.00	1111.0
281.00	652.0
282.00	563.0
283.10	2753.0
284.00	2029.0
285.00	4093.0
286.10	845.0
288.00	212.0
289.00	1230.0
289.90	713.0
291.00	689.0
292.10	1221.0
293.00	5334.0
294.00	1276.0
295.00	2075.0
296.00	76296.0
297.00	11777.0
298.00	605.0
298.90	153.0
300.90	1021.0
302.00	1412.0
303.10	9739.0
304.10	2903.0
305.00	370.0
307.00	234.0
308.00	1534.0
309.00	817.0
310.00	1424.0
310.90	361.0
311.90	224.0
313.10	850.0
314.00	4006.0
315.00	8393.0
316.00	5394.0
317.00	795.0
319.00	154.0
320.00	520.0
321.00	2438.0
322.00	1373.0
323.10	27616.0
324.10	5099.0
325.10	603.0
326.00	659.0
327.00	5368.0
328.00	3149.0
329.00	474.0
331.00	197.0
332.10	2116.0

Instrument :
BNA_M
ClientSampleId :
DFTPP055

333.10	2862.0
334.10	17592.0
335.10	4976.0
336.00	602.0
339.10	458.0
339.90	512.0
341.00	3226.0
342.00	1165.0
342.90	231.0
345.20	228.0
346.00	6052.0
347.00	1007.0
348.00	278.0
350.00	416.0
351.00	649.0
352.10	8031.0
353.00	6201.0
354.10	10210.0
355.10	1940.0
356.10	160.0
359.00	540.0
363.60	153.0
364.10	377.0
365.00	38240.0
366.00	5703.0
367.10	387.0
370.00	746.0
371.00	2543.0
372.10	15990.0
373.10	3559.0
374.00	411.0
376.90	266.0
378.00	235.0
383.00	3916.0
384.00	1220.0
385.10	379.0
390.00	2068.0
391.00	1556.0
392.10	952.0
392.90	171.0
395.00	173.0
401.00	1226.0
402.00	6368.0
403.00	8659.0
404.10	3211.0
405.10	623.0
409.90	246.0
414.90	518.0
419.90	251.0
421.10	8500.0
422.00	8483.0
423.10	57192.0
424.10	13221.0
425.00	1316.0
438.40	301.0
438.90	301.0
441.10	164672.0
442.10	1076224.0
443.10	208128.0
444.10	19536.0
445.10	1012.0

Instrument :
BNA_M
ClientSampleId :
DFTPP055

m/z	Abundance
36.00	268.0
37.00	1038.0
38.00	2346.0
39.00	14609.0
40.00	2295.0
41.00	1403.0
42.00	571.0
43.00	2890.0
44.00	4644.0
45.00	3161.0
47.00	244.0
47.80	170.0
49.10	1826.0
50.00	49224.0
51.00	192704.0
52.00	9648.0
53.00	441.0
55.00	1372.0
56.00	6443.0
57.00	13903.0
58.00	639.0
58.90	380.0
60.00	1476.0
61.00	2564.0
62.00	2803.0
63.00	7723.0
64.00	913.0
65.00	3933.0
66.00	267.0
67.00	321.0
68.00	3597.0
69.00	195840.0
69.90	1076.0
71.10	305.0
73.00	2849.0
74.00	19784.0
75.00	30200.0
76.10	10113.0
77.00	221760.0
78.00	14136.0
79.00	14360.0
80.00	11117.0
81.00	16392.0
82.00	3980.0
83.00	4032.0
84.00	776.0
85.00	3147.0
86.00	4946.0
87.00	1972.0
88.00	773.0
89.00	410.0
91.00	3825.0
92.00	4269.0
93.00	26368.0
94.00	1841.0
95.10	537.0
96.00	1540.0
98.00	19720.0
99.00	15717.0
100.00	1422.0
101.00	10018.0
102.00	615.0
103.00	2834.0
104.00	6022.0
105.00	6080.0
106.00	1813.0
107.00	70768.0
108.00	10905.0
109.00	2377.0
110.00	134848.0
111.00	20296.0
112.00	2763.0
113.00	979.0
113.70	167.0
114.80	364.0
116.00	4122.0
117.00	55888.0
117.90	3495.0
119.00	912.0
120.00	1111.0
121.10	533.0
122.00	5190.0
123.00	8460.0
124.00	3447.0
125.00	3329.0
126.10	1664.0
127.00	283136.0
128.00	20784.0
129.00	107368.0
130.00	8417.0
131.00	1844.0
132.00	898.0

Instrument :
BNA_M
ClientSampleId :
DFTPP055

133.00	768.0
134.00	3499.0
135.00	9174.0
136.00	2916.0
137.00	4395.0
138.00	1008.0
139.00	582.0
140.00	1513.0
141.00	14214.0
142.00	4971.0
143.00	3084.0
144.10	786.0
145.00	923.0
146.00	2359.0
147.00	7503.0
148.00	16656.0
149.00	3261.0
150.00	1206.0
151.10	1747.0
151.80	1478.0
153.00	4702.0
154.00	3362.0
155.00	8545.0
156.00	11860.0
157.00	2452.0
157.90	2776.0
159.00	2189.0
160.00	4219.0
161.00	6538.0
162.00	1775.0
163.00	726.0
164.10	819.0
164.90	5775.0
166.00	4780.0
167.00	28456.0
168.00	14906.0
169.00	2389.0
170.00	1037.0
171.00	1246.0
172.00	2922.0
173.00	3458.0
174.00	5833.0
175.00	11576.0
176.00	2987.0
177.00	5862.0
177.90	2064.0
179.00	22480.0
180.00	14797.0
181.00	6661.0
182.10	1002.0
183.00	596.0
184.00	1835.0
185.00	10472.0
186.00	83624.0
187.00	23144.0
188.00	2387.0
189.00	6035.0
189.90	806.0
191.00	2963.0
192.00	6765.0
193.00	8110.0
194.00	1909.0
195.00	1057.0
196.00	20472.0
196.70	3996.0
198.00	642496.0
199.00	41784.0
200.00	3002.0
201.60	2690.0
203.00	4519.0
204.00	22344.0
205.10	36256.0
206.00	167936.0
207.10	22832.0
208.00	5920.0
209.00	2122.0
210.10	2879.0
211.10	6387.0
212.80	437.0
214.00	282.0
215.00	2239.0
216.00	3241.0
217.00	44080.0
218.00	5420.0
218.90	566.0
221.00	32176.0
223.00	9502.0
224.00	93784.0
225.00	23744.0
226.10	2469.0
227.00	38224.0
228.00	4946.0
229.00	8050.0
230.00	883.0
231.10	3350.0
232.10	713.0

Instrument :
BNA_M
ClientSampleId :
DFTPP055

233.00	771.0
234.00	2150.0
235.00	3129.0
236.00	1695.0
237.00	2966.0
237.90	542.0
239.00	1725.0
239.90	1354.0
241.00	2003.0
242.00	4602.0
243.10	4784.0
244.10	75624.0
245.10	9629.0
246.00	15598.0
247.00	3043.0
247.90	641.0
249.00	2748.0
250.00	466.0
251.00	664.0
252.00	626.0
253.00	1649.0
254.10	2494.0
255.00	402240.0
256.00	54360.0
257.00	4139.0
258.00	23112.0
259.10	3120.0
260.00	614.0
261.00	722.0
262.90	250.0
264.00	682.0
265.00	8841.0
266.00	1230.0
267.00	388.0
269.90	399.0
271.10	665.0
272.00	1437.0
273.00	12362.0
274.00	32272.0
275.00	194304.0
276.00	24320.0
277.00	15050.0
278.00	2561.0
278.90	427.0
281.10	1204.0
282.00	543.0
283.00	1585.0
284.00	1197.0
285.00	2847.0
286.00	392.0
287.90	212.0
289.10	967.0
290.00	731.0
291.10	407.0
292.10	789.0
293.00	3363.0
294.00	912.0
295.00	1044.0
296.00	51000.0
297.00	7603.0
297.90	476.0
301.00	686.0
302.00	1371.0
303.00	6537.0
304.00	1695.0
305.00	227.0
308.00	871.0
309.00	444.0
310.00	508.0
312.10	187.0
313.00	586.0
314.00	2229.0
315.00	6525.0
316.00	3523.0
317.00	684.0
320.00	204.0
321.00	1839.0
322.10	1161.0
323.10	16944.0
324.00	3685.0
325.10	245.0
326.00	538.0
327.00	3474.0
328.00	1652.0
329.10	317.0
332.00	1045.0
333.00	1689.0
334.00	11838.0
335.10	2861.0
336.00	509.0
339.10	277.0
340.00	353.0
341.10	2085.0
342.10	774.0
346.00	4601.0
347.00	750.0

Instrument :
BNA_M
ClientSampleId :
DFTPP055

350.00	232.0
350.90	297.0
352.00	5912.0
353.00	3856.0
354.10	7166.0
355.10	1142.0
358.00	162.0
359.00	537.0
364.10	353.0
365.00	27000.0
366.00	3808.0
366.80	273.0
370.00	580.0
371.10	1303.0
372.10	10785.0
373.10	2726.0
374.00	405.0
377.00	351.0
383.00	3068.0
384.00	905.0
385.00	269.0
390.10	1587.0
391.00	1117.0
392.00	882.0
394.90	164.0
397.00	178.0
400.90	730.0
402.00	4333.0
403.00	6344.0
404.00	2549.0
405.00	398.0
410.00	151.0
414.90	315.0
420.10	150.0
421.00	5694.0
422.10	5571.0
423.10	41792.0
424.10	9697.0
425.00	1028.0
439.20	202.0
440.10	1093.0
441.10	119480.0
442.10	784320.0
443.10	147200.0
444.10	13482.0
445.00	779.0

Instrument :
BNA_M
ClientSampleId :
DFTPP055