


Data Path : Z:\svoasrv\HPCHEM1\BNA\_M\Data\BM061719\  
 Data File : BM020969.D  
 Acq On : 17 Jun 2019 09:05  
 Operator : HP/JU  
 Sample : DFTPP10  
 Misc :  
 ALS Vial : 1 Sample Multiplier: 1

Instrument :  
 BNA\_M  
 ClientSampleId :  
 DFTPP10

Integration File: LSCINT.P

Method : Z:\SVOASRV\HPCHEM1\BNA\_M\METHODS\SOM-EPA-BM061419MA.M  
 Title : SVOA CALIBRATION  
 Last Update : Tue Jun 18 05:54:28 2019



AutoFind: Scans 2485, 2486, 2487; Background Corrected with Scan 2479

| Target Mass | Rel. to Mass | Lower Limit% | Upper Limit% | Rel. Abn% | Raw Abn | Result Pass/Fail |
|-------------|--------------|--------------|--------------|-----------|---------|------------------|
| 51          | 198          | 10           | 80           | 28.1      | 163640  | PASS             |
| 68          | 69           | 0.00         | 2            | 1.8       | 3509    | PASS             |
| 69          | 198          | 0.00         | 100          | 32.8      | 190848  | PASS             |
| 70          | 69           | 0.00         | 2            | 0.5       | 937     | PASS             |
| 127         | 198          | 10           | 80           | 42.3      | 246400  | PASS             |
| 197         | 198          | 0.00         | 2            | 0.7       | 4160    | PASS             |
| 198         | 198          | 100          | 100          | 100.0     | 582400  | PASS             |
| 199         | 198          | 5            | 9            | 6.7       | 39272   | PASS             |
| 275         | 198          | 10           | 60           | 31.3      | 182016  | PASS             |
| 365         | 198          | 1            | 100          | 4.1       | 23888   | PASS             |
| 441         | 198          | 0.01         | 100          | 14.6      | 85226   | PASS             |
| 442         | 198          | 50           | 100          | 94.1      | 548288  | PASS             |
| 443         | 442          | 15           | 24           | 18.9      | 103749  | PASS             |

| m/z    | Abundance |
|--------|-----------|
| 40.00  | 3298.0    |
| 44.00  | 8804.0    |
| 48.90  | 564.0     |
| 73.10  | 499.0     |
| 83.90  | 395.0     |
| 190.90 | 446.0     |
| 207.00 | 2671.0    |
| 207.90 | 590.0     |
| 209.00 | 425.0     |
| 281.00 | 659.0     |

Instrument :

BNA\_M

ClientSampleId :

DFTPP10

| m/z    | Abundance |
|--------|-----------|
| 37.00  | 894.0     |
| 38.10  | 2720.0    |
| 39.10  | 13936.0   |
| 40.00  | 3606.0    |
| 41.10  | 487.0     |
| 44.10  | 8429.0    |
| 45.00  | 679.0     |
| 49.10  | 2142.0    |
| 50.10  | 50352.0   |
| 51.10  | 193344.0  |
| 52.10  | 10191.0   |
| 53.10  | 372.0     |
| 55.00  | 1047.0    |
| 56.00  | 5673.0    |
| 57.00  | 12087.0   |
| 58.10  | 620.0     |
| 59.00  | 312.0     |
| 61.00  | 2796.0    |
| 62.00  | 3140.0    |
| 63.10  | 8561.0    |
| 64.00  | 1024.0    |
| 65.10  | 4504.0    |
| 67.00  | 491.0     |
| 68.10  | 4453.0    |
| 69.00  | 223360.0  |
| 70.10  | 1001.0    |
| 73.10  | 2544.0    |
| 74.10  | 22480.0   |
| 75.00  | 36872.0   |
| 76.10  | 12038.0   |
| 77.10  | 244032.0  |
| 78.10  | 17288.0   |
| 79.00  | 16800.0   |
| 80.00  | 12884.0   |
| 81.00  | 16824.0   |
| 82.10  | 3698.0    |
| 83.00  | 4154.0    |
| 84.00  | 789.0     |
| 85.00  | 2918.0    |
| 86.00  | 5685.0    |
| 87.00  | 2537.0    |
| 88.00  | 982.0     |
| 89.10  | 420.0     |
| 91.00  | 4269.0    |
| 92.00  | 4758.0    |
| 93.00  | 29040.0   |
| 94.00  | 2025.0    |
| 95.10  | 455.0     |
| 96.00  | 1548.0    |
| 97.10  | 662.0     |
| 98.00  | 22064.0   |
| 99.00  | 18272.0   |
| 99.90  | 1762.0    |
| 101.00 | 9998.0    |
| 101.90 | 665.0     |
| 103.00 | 3712.0    |
| 104.00 | 5973.0    |
| 105.00 | 5438.0    |
| 106.10 | 1879.0    |
| 107.00 | 75144.0   |
| 108.00 | 11662.0   |
| 109.00 | 2093.0    |
| 110.00 | 138240.0  |
| 111.00 | 21064.0   |
| 112.10 | 2460.0    |
| 113.10 | 688.0     |
| 115.10 | 381.0     |
| 116.00 | 4428.0    |
| 117.00 | 60856.0   |
| 118.00 | 5437.0    |
| 118.90 | 630.0     |
| 119.90 | 901.0     |
| 121.10 | 516.0     |
| 122.00 | 5112.0    |
| 123.00 | 8013.0    |
| 124.00 | 3976.0    |
| 125.00 | 3665.0    |
| 126.20 | 1284.0    |
| 127.00 | 272256.0  |
| 128.00 | 20432.0   |
| 129.00 | 105928.0  |
| 130.00 | 8690.0    |
| 130.90 | 1838.0    |
| 132.00 | 815.0     |
| 132.90 | 526.0     |
| 134.10 | 3194.0    |
| 135.00 | 8582.0    |
| 136.00 | 3754.0    |
| 137.10 | 5201.0    |
| 138.00 | 1210.0    |
| 139.10 | 644.0     |

Instrument :  
BNA\_M  
ClientSampleId :  
DFTPP10

|        |          |
|--------|----------|
| 140.00 | 1796.0   |
| 141.00 | 14529.0  |
| 142.00 | 4535.0   |
| 143.00 | 3153.0   |
| 144.00 | 929.0    |
| 144.90 | 660.0    |
| 146.00 | 2321.0   |
| 147.00 | 7352.0   |
| 148.00 | 16800.0  |
| 149.00 | 4583.0   |
| 150.10 | 881.0    |
| 151.00 | 1758.0   |
| 151.70 | 1101.0   |
| 153.00 | 5010.0   |
| 154.00 | 3426.0   |
| 155.00 | 7895.0   |
| 156.00 | 12203.0  |
| 157.00 | 2558.0   |
| 157.90 | 2624.0   |
| 159.00 | 1740.0   |
| 160.00 | 4261.0   |
| 161.00 | 7306.0   |
| 162.00 | 1691.0   |
| 163.00 | 897.0    |
| 164.00 | 1003.0   |
| 165.00 | 5402.0   |
| 166.00 | 4960.0   |
| 167.00 | 30520.0  |
| 168.00 | 18896.0  |
| 169.10 | 2793.0   |
| 170.10 | 1065.0   |
| 170.90 | 1341.0   |
| 172.00 | 2770.0   |
| 173.00 | 3548.0   |
| 174.00 | 6253.0   |
| 175.00 | 11728.0  |
| 176.00 | 3752.0   |
| 176.90 | 5296.0   |
| 178.00 | 1748.0   |
| 179.00 | 22536.0  |
| 180.00 | 17000.0  |
| 181.00 | 6948.0   |
| 182.10 | 1287.0   |
| 183.00 | 753.0    |
| 184.00 | 1933.0   |
| 185.00 | 10470.0  |
| 186.00 | 86360.0  |
| 187.00 | 24408.0  |
| 188.00 | 2499.0   |
| 189.00 | 5771.0   |
| 190.00 | 889.0    |
| 191.00 | 2981.0   |
| 192.00 | 7521.0   |
| 193.00 | 7256.0   |
| 194.00 | 1807.0   |
| 195.20 | 868.0    |
| 196.00 | 19376.0  |
| 197.10 | 4726.0   |
| 198.00 | 606912.0 |
| 199.00 | 41480.0  |
| 200.00 | 3269.0   |
| 201.40 | 3400.0   |
| 202.90 | 4111.0   |
| 204.00 | 21256.0  |
| 205.00 | 36840.0  |
| 206.00 | 160064.0 |
| 207.00 | 24896.0  |
| 208.00 | 6397.0   |
| 209.00 | 2178.0   |
| 210.10 | 2850.0   |
| 211.10 | 5966.0   |
| 213.90 | 336.0    |
| 215.00 | 1922.0   |
| 216.10 | 3255.0   |
| 217.00 | 42296.0  |
| 218.00 | 5171.0   |
| 219.00 | 471.0    |
| 221.00 | 27248.0  |
| 223.00 | 9669.0   |
| 224.00 | 87496.0  |
| 225.00 | 22984.0  |
| 226.10 | 2718.0   |
| 227.00 | 35936.0  |
| 228.00 | 5252.0   |
| 229.00 | 7487.0   |
| 229.90 | 873.0    |
| 231.00 | 3252.0   |
| 231.90 | 579.0    |
| 233.00 | 804.0    |
| 234.00 | 2333.0   |
| 235.00 | 2296.0   |
| 236.00 | 1759.0   |
| 237.00 | 3730.0   |
| 238.10 | 580.0    |
| 239.10 | 1477.0   |

Instrument :  
BNA\_M  
ClientSampleId :  
DFTPP10

|        |          |
|--------|----------|
| 240.00 | 1160.0   |
| 240.90 | 2085.0   |
| 242.00 | 4037.0   |
| 243.10 | 5130.0   |
| 244.00 | 72632.0  |
| 245.10 | 9512.0   |
| 246.00 | 13492.0  |
| 247.00 | 2883.0   |
| 247.90 | 837.0    |
| 249.00 | 3005.0   |
| 250.00 | 589.0    |
| 251.00 | 669.0    |
| 251.90 | 625.0    |
| 253.00 | 1444.0   |
| 254.20 | 3463.0   |
| 255.00 | 382208.0 |
| 256.00 | 57104.0  |
| 257.00 | 4704.0   |
| 258.00 | 21848.0  |
| 259.00 | 3534.0   |
| 260.00 | 754.0    |
| 261.10 | 532.0    |
| 264.10 | 555.0    |
| 265.00 | 8281.0   |
| 266.00 | 1276.0   |
| 271.10 | 1123.0   |
| 272.00 | 920.0    |
| 273.00 | 11175.0  |
| 274.00 | 30216.0  |
| 275.00 | 169536.0 |
| 276.00 | 22048.0  |
| 277.00 | 14286.0  |
| 278.00 | 2438.0   |
| 279.00 | 629.0    |
| 281.00 | 647.0    |
| 282.10 | 425.0    |
| 283.10 | 1764.0   |
| 284.00 | 1410.0   |
| 285.00 | 2543.0   |
| 286.00 | 412.0    |
| 289.10 | 485.0    |
| 289.90 | 724.0    |
| 291.00 | 437.0    |
| 292.00 | 683.0    |
| 293.00 | 3110.0   |
| 294.00 | 932.0    |
| 295.00 | 1399.0   |
| 296.00 | 47624.0  |
| 297.00 | 6739.0   |
| 298.00 | 594.0    |
| 301.00 | 465.0    |
| 302.00 | 593.0    |
| 303.00 | 6281.0   |
| 304.00 | 1765.0   |
| 307.90 | 858.0    |
| 309.00 | 580.0    |
| 310.00 | 897.0    |
| 312.90 | 466.0    |
| 314.10 | 2104.0   |
| 315.00 | 5858.0   |
| 316.10 | 2891.0   |
| 316.90 | 467.0    |
| 321.10 | 1576.0   |
| 322.10 | 986.0    |
| 323.00 | 15387.0  |
| 324.00 | 2740.0   |
| 324.90 | 335.0    |
| 326.00 | 300.0    |
| 327.00 | 3104.0   |
| 328.00 | 1497.0   |
| 329.10 | 330.0    |
| 332.00 | 1223.0   |
| 333.00 | 1745.0   |
| 334.00 | 9701.0   |
| 335.10 | 2783.0   |
| 336.00 | 469.0    |
| 341.00 | 1816.0   |
| 342.10 | 543.0    |
| 346.00 | 3263.0   |
| 347.00 | 792.0    |
| 352.00 | 4620.0   |
| 353.10 | 2790.0   |
| 354.10 | 4940.0   |
| 355.00 | 1124.0   |
| 356.00 | 314.0    |
| 359.00 | 425.0    |
| 364.00 | 305.0    |
| 365.00 | 21888.0  |
| 366.00 | 3235.0   |
| 370.10 | 503.0    |
| 371.00 | 1204.0   |
| 372.10 | 7993.0   |
| 373.10 | 2048.0   |
| 383.00 | 2183.0   |
| 384.10 | 654.0    |

Instrument :  
BNA\_M  
ClientSampleId :  
DFTPP10

|        |          |
|--------|----------|
| 390.00 | 1002.0   |
| 390.90 | 639.0    |
| 391.90 | 611.0    |
| 401.10 | 361.0    |
| 402.10 | 3252.0   |
| 403.00 | 4959.0   |
| 404.00 | 1566.0   |
| 405.00 | 509.0    |
| 421.00 | 3500.0   |
| 422.10 | 3788.0   |
| 423.00 | 23968.0  |
| 424.10 | 7176.0   |
| 425.10 | 1112.0   |
| 441.10 | 66176.0  |
| 442.10 | 430848.0 |
| 443.10 | 82776.0  |
| 444.10 | 7708.0   |
| 445.10 | 425.0    |

Instrument :  
BNA\_M  
ClientSampleId :  
DFTPP10

| m/z    | Abundance |
|--------|-----------|
| 37.00  | 1103.0    |
| 38.10  | 2425.0    |
| 39.10  | 14834.0   |
| 40.00  | 3383.0    |
| 41.10  | 458.0     |
| 44.00  | 8622.0    |
| 45.00  | 412.0     |
| 49.10  | 1649.0    |
| 50.10  | 50016.0   |
| 51.10  | 192512.0  |
| 52.10  | 9220.0    |
| 53.00  | 370.0     |
| 55.00  | 1159.0    |
| 56.00  | 6542.0    |
| 57.00  | 13695.0   |
| 58.00  | 664.0     |
| 61.10  | 2876.0    |
| 62.00  | 3451.0    |
| 63.10  | 7449.0    |
| 64.00  | 1203.0    |
| 65.10  | 3877.0    |
| 67.00  | 346.0     |
| 68.10  | 4005.0    |
| 69.00  | 225216.0  |
| 70.10  | 1076.0    |
| 73.00  | 2738.0    |
| 74.10  | 23584.0   |
| 75.10  | 36272.0   |
| 76.10  | 12446.0   |
| 77.10  | 252416.0  |
| 78.10  | 16432.0   |
| 79.00  | 16792.0   |
| 80.00  | 13353.0   |
| 81.00  | 18792.0   |
| 82.00  | 4195.0    |
| 83.10  | 4137.0    |
| 84.00  | 744.0     |
| 85.00  | 3111.0    |
| 86.00  | 5187.0    |
| 87.10  | 2224.0    |
| 87.90  | 737.0     |
| 89.00  | 449.0     |
| 91.00  | 3928.0    |
| 92.00  | 4513.0    |
| 93.00  | 28688.0   |
| 94.00  | 2177.0    |
| 95.00  | 535.0     |
| 96.00  | 1176.0    |
| 97.20  | 451.0     |
| 98.00  | 22160.0   |
| 99.00  | 18912.0   |
| 100.00 | 1593.0    |
| 101.00 | 10184.0   |
| 102.00 | 456.0     |
| 103.00 | 3516.0    |
| 104.00 | 6723.0    |
| 105.00 | 6292.0    |
| 106.10 | 2073.0    |
| 107.00 | 81616.0   |
| 108.00 | 12546.0   |
| 109.10 | 2124.0    |
| 110.00 | 148160.0  |
| 111.00 | 22144.0   |
| 112.00 | 3243.0    |
| 113.00 | 928.0     |
| 114.90 | 322.0     |
| 116.00 | 4824.0    |
| 117.00 | 64784.0   |
| 118.00 | 5120.0    |
| 119.00 | 938.0     |
| 120.00 | 1070.0    |
| 120.90 | 430.0     |
| 122.00 | 5574.0    |
| 123.00 | 8354.0    |
| 124.00 | 3623.0    |
| 125.00 | 3125.0    |
| 126.20 | 2174.0    |
| 127.00 | 295168.0  |
| 128.00 | 20872.0   |
| 129.00 | 115960.0  |
| 130.00 | 9592.0    |
| 130.90 | 1992.0    |
| 132.10 | 1146.0    |
| 133.00 | 565.0     |
| 134.00 | 2981.0    |
| 135.00 | 9858.0    |
| 136.00 | 3576.0    |
| 137.00 | 5472.0    |
| 137.90 | 1533.0    |
| 139.00 | 705.0     |
| 140.00 | 1779.0    |

Instrument :  
BNA\_M  
ClientSampleId :  
DFTPP10

|        |          |
|--------|----------|
| 141.00 | 15641.0  |
| 142.00 | 5331.0   |
| 143.00 | 3045.0   |
| 144.00 | 972.0    |
| 145.00 | 789.0    |
| 146.00 | 3126.0   |
| 147.00 | 8537.0   |
| 148.00 | 17800.0  |
| 149.00 | 4514.0   |
| 150.00 | 1528.0   |
| 151.20 | 2132.0   |
| 153.00 | 4673.0   |
| 154.00 | 4036.0   |
| 155.00 | 8914.0   |
| 156.00 | 12545.0  |
| 157.00 | 2282.0   |
| 157.90 | 2924.0   |
| 158.90 | 2205.0   |
| 160.00 | 5224.0   |
| 161.00 | 8169.0   |
| 162.00 | 2227.0   |
| 163.00 | 527.0    |
| 164.10 | 882.0    |
| 165.00 | 6637.0   |
| 166.00 | 4910.0   |
| 167.00 | 33560.0  |
| 168.00 | 19472.0  |
| 169.00 | 3023.0   |
| 169.90 | 1341.0   |
| 171.00 | 1830.0   |
| 172.00 | 3096.0   |
| 173.00 | 4159.0   |
| 174.00 | 6923.0   |
| 175.00 | 13753.0  |
| 176.00 | 3444.0   |
| 177.00 | 6272.0   |
| 178.00 | 1968.0   |
| 179.00 | 25032.0  |
| 180.00 | 19144.0  |
| 181.00 | 8868.0   |
| 182.00 | 1227.0   |
| 182.80 | 765.0    |
| 184.00 | 2183.0   |
| 185.00 | 11811.0  |
| 186.00 | 98104.0  |
| 187.00 | 27232.0  |
| 188.00 | 3187.0   |
| 189.00 | 6303.0   |
| 190.00 | 914.0    |
| 191.00 | 3241.0   |
| 192.00 | 8061.0   |
| 193.00 | 9170.0   |
| 194.00 | 1830.0   |
| 195.00 | 1254.0   |
| 196.00 | 21528.0  |
| 197.10 | 4720.0   |
| 198.00 | 698880.0 |
| 199.00 | 47096.0  |
| 200.00 | 3500.0   |
| 201.40 | 3598.0   |
| 203.00 | 5124.0   |
| 204.00 | 23824.0  |
| 205.00 | 44944.0  |
| 206.00 | 192192.0 |
| 207.00 | 27136.0  |
| 208.00 | 7289.0   |
| 209.10 | 2248.0   |
| 210.20 | 2937.0   |
| 211.10 | 7490.0   |
| 211.90 | 1095.0   |
| 213.00 | 540.0    |
| 213.90 | 333.0    |
| 215.00 | 2102.0   |
| 216.00 | 4486.0   |
| 217.00 | 51464.0  |
| 218.00 | 5687.0   |
| 219.10 | 673.0    |
| 221.00 | 31560.0  |
| 221.90 | 7078.0   |
| 223.00 | 10994.0  |
| 224.00 | 105888.0 |
| 225.00 | 27664.0  |
| 226.00 | 2777.0   |
| 227.00 | 44032.0  |
| 228.00 | 5947.0   |
| 229.00 | 8973.0   |
| 230.00 | 1314.0   |
| 231.00 | 3464.0   |
| 231.90 | 878.0    |
| 233.00 | 902.0    |
| 234.00 | 2742.0   |
| 235.00 | 3079.0   |
| 236.00 | 1870.0   |
| 237.00 | 4220.0   |
| 238.00 | 768.0    |

Instrument :  
BNA\_M  
ClientSampleId :  
DFTPP10

|        |          |
|--------|----------|
| 239.00 | 1864.0   |
| 240.00 | 1552.0   |
| 240.90 | 2561.0   |
| 242.00 | 5344.0   |
| 243.00 | 6113.0   |
| 244.00 | 83376.0  |
| 245.10 | 11739.0  |
| 246.00 | 15952.0  |
| 247.00 | 3112.0   |
| 247.90 | 981.0    |
| 249.00 | 3569.0   |
| 250.00 | 762.0    |
| 251.00 | 659.0    |
| 252.10 | 786.0    |
| 253.00 | 1617.0   |
| 254.20 | 4161.0   |
| 255.00 | 479616.0 |
| 256.00 | 70640.0  |
| 257.10 | 5513.0   |
| 258.00 | 26088.0  |
| 259.00 | 4597.0   |
| 260.00 | 617.0    |
| 261.00 | 601.0    |
| 263.00 | 382.0    |
| 264.00 | 632.0    |
| 265.00 | 10902.0  |
| 266.00 | 1638.0   |
| 266.90 | 413.0    |
| 270.10 | 492.0    |
| 271.00 | 1377.0   |
| 272.00 | 1393.0   |
| 273.00 | 14873.0  |
| 274.00 | 38104.0  |
| 275.00 | 224320.0 |
| 276.00 | 28864.0  |
| 277.00 | 17544.0  |
| 278.00 | 3604.0   |
| 278.90 | 502.0    |
| 281.00 | 989.0    |
| 282.00 | 653.0    |
| 283.00 | 2052.0   |
| 284.00 | 1611.0   |
| 285.00 | 3764.0   |
| 286.00 | 832.0    |
| 288.90 | 847.0    |
| 290.00 | 576.0    |
| 291.00 | 676.0    |
| 292.00 | 814.0    |
| 293.00 | 4599.0   |
| 294.00 | 1022.0   |
| 295.10 | 1499.0   |
| 296.00 | 61664.0  |
| 297.00 | 9083.0   |
| 298.00 | 500.0    |
| 301.00 | 846.0    |
| 302.00 | 1400.0   |
| 303.00 | 6977.0   |
| 304.10 | 1929.0   |
| 307.90 | 801.0    |
| 309.00 | 599.0    |
| 309.90 | 742.0    |
| 312.90 | 606.0    |
| 314.00 | 3174.0   |
| 315.00 | 7151.0   |
| 316.00 | 3944.0   |
| 317.00 | 633.0    |
| 321.00 | 1985.0   |
| 322.00 | 1040.0   |
| 323.00 | 20648.0  |
| 324.00 | 3509.0   |
| 325.00 | 318.0    |
| 325.90 | 546.0    |
| 327.00 | 4055.0   |
| 328.00 | 1813.0   |
| 329.10 | 300.0    |
| 331.90 | 1499.0   |
| 333.00 | 1592.0   |
| 334.10 | 13096.0  |
| 335.00 | 3327.0   |
| 335.90 | 395.0    |
| 339.00 | 393.0    |
| 341.00 | 2568.0   |
| 342.00 | 409.0    |
| 346.00 | 5504.0   |
| 346.90 | 739.0    |
| 351.10 | 499.0    |
| 352.10 | 6765.0   |
| 353.00 | 3949.0   |
| 354.10 | 6780.0   |
| 355.10 | 1429.0   |
| 359.00 | 456.0    |
| 365.00 | 27496.0  |
| 366.00 | 3986.0   |
| 370.00 | 568.0    |
| 371.00 | 2116.0   |

Instrument :  
BNA\_M  
ClientSampleId :  
DFTPP10

|        |          |
|--------|----------|
| 372.10 | 10775.0  |
| 373.10 | 2737.0   |
| 374.00 | 366.0    |
| 377.10 | 334.0    |
| 383.00 | 3348.0   |
| 384.00 | 857.0    |
| 390.10 | 1742.0   |
| 391.00 | 1037.0   |
| 392.00 | 718.0    |
| 401.00 | 766.0    |
| 402.00 | 4199.0   |
| 403.00 | 6404.0   |
| 404.00 | 1906.0   |
| 405.00 | 374.0    |
| 421.00 | 6195.0   |
| 422.10 | 5419.0   |
| 423.10 | 39648.0  |
| 424.10 | 9367.0   |
| 425.00 | 1388.0   |
| 441.10 | 103456.0 |
| 442.10 | 662912.0 |
| 443.10 | 126368.0 |
| 444.10 | 11898.0  |
| 445.10 | 770.0    |

Instrument :  
BNA\_M  
ClientSampleId :  
DFTPP10

| m/z    | Abundance |
|--------|-----------|
| 37.00  | 565.0     |
| 38.00  | 1556.0    |
| 39.10  | 8181.0    |
| 40.00  | 3246.0    |
| 41.00  | 364.0     |
| 44.00  | 8802.0    |
| 45.00  | 460.0     |
| 49.00  | 1210.0    |
| 50.10  | 25912.0   |
| 51.10  | 105064.0  |
| 52.00  | 5328.0    |
| 52.90  | 393.0     |
| 55.00  | 634.0     |
| 56.00  | 3187.0    |
| 57.00  | 7734.0    |
| 58.00  | 327.0     |
| 61.00  | 1357.0    |
| 62.00  | 1482.0    |
| 63.10  | 4350.0    |
| 64.00  | 558.0     |
| 65.10  | 2867.0    |
| 68.00  | 2069.0    |
| 69.00  | 123968.0  |
| 69.90  | 735.0     |
| 73.00  | 2157.0    |
| 74.10  | 12390.0   |
| 75.00  | 20040.0   |
| 76.00  | 7179.0    |
| 77.10  | 141056.0  |
| 78.10  | 9340.0    |
| 79.10  | 9378.0    |
| 80.00  | 7381.0    |
| 81.00  | 10008.0   |
| 82.00  | 2288.0    |
| 83.00  | 2422.0    |
| 84.00  | 531.0     |
| 85.00  | 1688.0    |
| 86.00  | 2788.0    |
| 87.10  | 1214.0    |
| 88.00  | 721.0     |
| 91.00  | 2484.0    |
| 92.00  | 2577.0    |
| 93.00  | 15544.0   |
| 93.90  | 1139.0    |
| 96.00  | 934.0     |
| 97.00  | 478.0     |
| 98.00  | 13306.0   |
| 99.00  | 11066.0   |
| 100.00 | 999.0     |
| 101.00 | 5981.0    |
| 102.00 | 359.0     |
| 103.00 | 2057.0    |
| 104.00 | 3858.0    |
| 105.00 | 3698.0    |
| 106.10 | 1401.0    |
| 107.00 | 45536.0   |
| 108.00 | 7203.0    |
| 109.00 | 1198.0    |
| 110.00 | 82184.0   |
| 111.00 | 13660.0   |
| 111.90 | 1218.0    |
| 113.00 | 598.0     |
| 115.00 | 323.0     |
| 116.10 | 2885.0    |
| 117.00 | 39608.0   |
| 118.10 | 2955.0    |
| 120.10 | 475.0     |
| 122.00 | 3467.0    |
| 123.00 | 4915.0    |
| 124.00 | 1996.0    |
| 125.10 | 2147.0    |
| 126.20 | 591.0     |
| 127.00 | 171776.0  |
| 128.00 | 11971.0   |
| 129.00 | 68088.0   |
| 130.00 | 5431.0    |
| 131.00 | 1166.0    |
| 132.10 | 694.0     |
| 132.90 | 618.0     |
| 134.00 | 2192.0    |
| 135.00 | 5835.0    |
| 136.00 | 2398.0    |
| 137.00 | 3528.0    |
| 138.00 | 806.0     |
| 139.00 | 507.0     |
| 140.00 | 950.0     |
| 141.00 | 9448.0    |
| 142.00 | 2887.0    |
| 143.00 | 2065.0    |
| 144.10 | 718.0     |
| 145.00 | 537.0     |

Instrument :  
BNA\_M  
ClientSampleId :  
DFTPP10

|        |          |
|--------|----------|
| 146.10 | 1779.0   |
| 147.00 | 5444.0   |
| 148.00 | 10340.0  |
| 149.00 | 2803.0   |
| 150.00 | 660.0    |
| 151.20 | 1393.0   |
| 153.00 | 3275.0   |
| 154.00 | 2218.0   |
| 155.00 | 5281.0   |
| 156.00 | 8057.0   |
| 157.10 | 1531.0   |
| 158.00 | 2135.0   |
| 159.00 | 1475.0   |
| 160.00 | 3320.0   |
| 161.00 | 4418.0   |
| 161.90 | 1482.0   |
| 163.90 | 514.0    |
| 165.00 | 4337.0   |
| 166.00 | 3013.0   |
| 167.00 | 20504.0  |
| 168.00 | 11512.0  |
| 169.00 | 1847.0   |
| 170.00 | 718.0    |
| 170.90 | 928.0    |
| 171.90 | 1882.0   |
| 173.00 | 2132.0   |
| 174.00 | 4660.0   |
| 175.00 | 8295.0   |
| 176.00 | 2490.0   |
| 177.00 | 3485.0   |
| 177.90 | 1154.0   |
| 179.00 | 16928.0  |
| 180.00 | 10892.0  |
| 181.00 | 4786.0   |
| 182.00 | 831.0    |
| 183.00 | 441.0    |
| 184.00 | 1073.0   |
| 185.00 | 8499.0   |
| 186.00 | 61336.0  |
| 187.00 | 17472.0  |
| 188.10 | 1699.0   |
| 189.10 | 3825.0   |
| 190.00 | 732.0    |
| 191.00 | 1719.0   |
| 192.00 | 5330.0   |
| 193.00 | 6227.0   |
| 194.00 | 1150.0   |
| 195.20 | 444.0    |
| 196.00 | 12500.0  |
| 197.10 | 3034.0   |
| 197.90 | 441408.0 |
| 199.00 | 29240.0  |
| 200.00 | 2389.0   |
| 201.40 | 1805.0   |
| 203.00 | 2934.0   |
| 204.00 | 16180.0  |
| 205.00 | 28472.0  |
| 206.00 | 121488.0 |
| 207.00 | 18944.0  |
| 208.10 | 4392.0   |
| 209.00 | 1568.0   |
| 210.20 | 1880.0   |
| 211.00 | 4962.0   |
| 213.00 | 337.0    |
| 215.00 | 1220.0   |
| 216.00 | 2175.0   |
| 217.00 | 31640.0  |
| 218.00 | 4309.0   |
| 219.00 | 380.0    |
| 221.00 | 19752.0  |
| 221.90 | 4538.0   |
| 223.00 | 7548.0   |
| 224.00 | 69104.0  |
| 225.00 | 17400.0  |
| 226.10 | 1993.0   |
| 227.00 | 28808.0  |
| 228.00 | 3570.0   |
| 229.00 | 6070.0   |
| 230.00 | 825.0    |
| 231.00 | 2044.0   |
| 232.00 | 425.0    |
| 232.90 | 675.0    |
| 234.00 | 2092.0   |
| 235.00 | 2124.0   |
| 236.00 | 1583.0   |
| 237.00 | 2456.0   |
| 237.90 | 404.0    |
| 239.00 | 1214.0   |
| 240.00 | 869.0    |
| 240.90 | 1770.0   |
| 242.00 | 3817.0   |
| 243.10 | 4668.0   |
| 244.00 | 57840.0  |
| 245.00 | 7612.0   |
| 246.00 | 11516.0  |

Instrument :  
BNA\_M  
ClientSampleId :  
DFTPP10

|        |          |
|--------|----------|
| 247.00 | 1957.0   |
| 247.90 | 663.0    |
| 249.00 | 2573.0   |
| 249.90 | 343.0    |
| 251.00 | 685.0    |
| 251.90 | 797.0    |
| 253.00 | 1479.0   |
| 254.20 | 2047.0   |
| 255.00 | 326848.0 |
| 256.00 | 46344.0  |
| 257.00 | 3579.0   |
| 258.00 | 17688.0  |
| 259.00 | 2794.0   |
| 260.10 | 419.0    |
| 261.00 | 613.0    |
| 264.10 | 485.0    |
| 265.00 | 7935.0   |
| 265.90 | 1004.0   |
| 269.90 | 376.0    |
| 271.00 | 995.0    |
| 272.10 | 996.0    |
| 273.00 | 9779.0   |
| 274.00 | 25848.0  |
| 275.00 | 152192.0 |
| 276.00 | 19368.0  |
| 277.00 | 13032.0  |
| 278.00 | 1866.0   |
| 278.90 | 378.0    |
| 280.90 | 810.0    |
| 282.00 | 604.0    |
| 283.00 | 1450.0   |
| 284.10 | 889.0    |
| 285.00 | 2413.0   |
| 286.00 | 521.0    |
| 289.00 | 544.0    |
| 290.00 | 506.0    |
| 291.00 | 336.0    |
| 291.90 | 745.0    |
| 293.00 | 2734.0   |
| 294.00 | 699.0    |
| 295.10 | 899.0    |
| 296.00 | 42272.0  |
| 297.00 | 5714.0   |
| 298.10 | 555.0    |
| 300.90 | 555.0    |
| 302.00 | 881.0    |
| 303.00 | 4951.0   |
| 304.00 | 1267.0   |
| 307.90 | 572.0    |
| 309.00 | 472.0    |
| 310.00 | 706.0    |
| 313.00 | 395.0    |
| 314.00 | 2305.0   |
| 315.00 | 5675.0   |
| 316.00 | 2788.0   |
| 317.00 | 406.0    |
| 321.00 | 1876.0   |
| 322.00 | 620.0    |
| 323.10 | 15983.0  |
| 324.00 | 2389.0   |
| 325.90 | 498.0    |
| 327.00 | 3198.0   |
| 328.00 | 1649.0   |
| 329.20 | 325.0    |
| 332.00 | 1247.0   |
| 333.00 | 1187.0   |
| 334.00 | 10096.0  |
| 335.00 | 2507.0   |
| 336.10 | 331.0    |
| 341.10 | 1740.0   |
| 342.00 | 543.0    |
| 346.00 | 3409.0   |
| 347.00 | 567.0    |
| 351.00 | 302.0    |
| 352.00 | 4604.0   |
| 353.00 | 3453.0   |
| 354.10 | 5087.0   |
| 355.10 | 847.0    |
| 359.00 | 313.0    |
| 365.00 | 22280.0  |
| 366.00 | 3530.0   |
| 370.10 | 452.0    |
| 371.10 | 1271.0   |
| 372.10 | 8219.0   |
| 373.00 | 2198.0   |
| 374.00 | 362.0    |
| 383.00 | 2406.0   |
| 384.00 | 911.0    |
| 390.00 | 1119.0   |
| 391.00 | 706.0    |
| 392.10 | 448.0    |
| 401.10 | 440.0    |
| 402.10 | 3524.0   |
| 403.00 | 4663.0   |
| 404.00 | 2007.0   |

Instrument :  
BNA\_M  
ClientSampleId :  
DFTPP10

|        |          |
|--------|----------|
| 405.00 | 353.0    |
| 421.00 | 4985.0   |
| 422.10 | 4389.0   |
| 423.10 | 30144.0  |
| 424.10 | 8155.0   |
| 425.00 | 726.0    |
| 440.20 | 388.0    |
| 441.10 | 86048.0  |
| 442.10 | 551104.0 |
| 443.10 | 102104.0 |
| 444.10 | 8912.0   |
| 445.20 | 455.0    |

|                  |
|------------------|
| Instrument :     |
| BNA_M            |
| ClientSampleId : |
| DFTPP10          |