


Data Path : Z:\svoasrv\HPCHEM1\BNA M\Data\BM061720\
 Data File : BM026425.D
 Acq On : 17 Jun 2020 13:12
 Operator : JU/CG
 Sample : DFTPP02
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 DFTPP02

Integration File: LSCINT.P

Method : Z:\SVOASRV\HPCHEM1\BNA_M\METHODS\SOM-EPA-BM061220MA.M
 Title : SVOA CALIBRATION
 Last Update : Wed Jun 17 16:26:10 2020



AutoFind: Scans 2447, 2448, 2449; Background Corrected with Scan 2442

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	46.9	98787	PASS
68	69	0.00	2	1.7	1609	PASS
69	198	0.00	100	44.5	93703	PASS
70	69	0.00	2	0.6	537	PASS
127	198	10	80	46.3	97498	PASS
197	198	0.00	2	0.5	1073	PASS
198	198	100	100	100.0	210474	PASS
199	198	5	9	6.8	14394	PASS
275	198	10	60	27.6	57997	PASS
365	198	1	100	3.7	7781	PASS
441	198	0.01	100	11.2	23590	PASS
442	198	50	100	76.9	161762	PASS
443	442	15	24	20.0	32285	PASS

m/z	Abundance
39.00	444.0
39.90	843.0
41.10	1112.0
42.10	208.0
43.10	385.0
44.00	645.0
51.10	154.0
55.10	461.0
56.10	250.0
57.10	1097.0
60.00	258.0
69.10	171.0
71.00	227.0
73.10	927.0
77.00	316.0
78.00	236.0
78.90	287.0
83.90	247.0
105.00	515.0
123.00	1240.0
123.90	275.0
132.90	154.0
146.90	178.0
179.00	271.0
193.10	215.0
207.00	1249.0
208.10	262.0
209.10	167.0
281.20	298.0

Instrument :
BNA_M
ClientSampleId :
DFTPP02

m/z	Abundance
37.00	275.0
38.10	1077.0
39.10	6762.0
40.10	920.0
41.10	764.0
43.10	571.0
44.00	671.0
44.90	430.0
49.10	695.0
50.10	20416.0
51.10	78424.0
52.10	4015.0
53.10	292.0
55.10	734.0
56.00	2324.0
57.00	5454.0
57.90	265.0
59.90	194.0
61.00	1134.0
62.10	1026.0
63.00	2853.0
64.10	446.0
65.00	1293.0
66.20	163.0
67.00	242.0
68.10	1192.0
69.00	74872.0
69.90	432.0
71.10	181.0
73.10	1389.0
74.00	7362.0
75.10	11642.0
76.10	4212.0
77.10	82232.0
78.10	5500.0
79.00	5670.0
80.00	4366.0
81.00	6054.0
82.00	1390.0
83.00	1554.0
84.10	213.0
85.00	842.0
86.10	1837.0
87.00	963.0
87.90	345.0
88.90	150.0
91.00	1177.0
92.00	1373.0
93.00	9731.0
94.00	717.0
95.10	164.0
96.00	727.0
97.00	229.0
98.10	7167.0
99.00	5846.0
100.00	527.0
101.10	3097.0
101.90	199.0
103.00	902.0
104.00	2097.0
105.00	2080.0
105.90	453.0
106.10	458.0
107.00	23424.0
108.00	3593.0
109.00	577.0
110.00	39760.0
111.00	6298.0
112.10	816.0
113.00	321.0
116.00	1202.0
117.00	20376.0
118.00	1707.0
118.80	161.0
120.00	358.0
122.00	1306.0
123.00	1935.0
124.00	1039.0
125.00	845.0
126.20	447.0
127.00	72464.0
128.00	5703.0
129.00	28848.0
130.00	2770.0
131.00	497.0
132.00	255.0
134.00	744.0
135.00	2444.0
136.00	963.0
137.00	1312.0
138.00	298.0
138.90	291.0

Instrument :
BNA_M
ClientSampleId :
DFTPP02

140.00	446.0
141.00	4076.0
142.00	1468.0
143.00	981.0
143.90	326.0
144.90	237.0
146.00	830.0
146.90	2454.0
148.00	4619.0
149.00	1139.0
150.00	175.0
151.10	525.0
151.80	338.0
153.10	1053.0
153.90	884.0
155.00	2223.0
156.00	2787.0
157.10	783.0
158.10	655.0
158.90	487.0
159.90	1219.0
161.10	1734.0
161.90	528.0
164.00	306.0
164.90	1481.0
166.00	957.0
167.00	7806.0
168.00	4943.0
169.00	651.0
170.00	324.0
171.00	339.0
172.00	681.0
173.00	761.0
174.00	1599.0
175.00	2914.0
176.00	935.0
177.00	1514.0
178.10	521.0
179.00	5744.0
180.00	3561.0
181.00	1816.0
182.00	300.0
182.90	217.0
183.90	731.0
185.00	2702.0
186.00	18640.0
187.00	5615.0
188.10	815.0
189.00	1391.0
189.90	219.0
191.10	856.0
192.00	1829.0
193.00	1891.0
194.00	554.0
194.90	182.0
195.90	3510.0
197.10	860.0
198.00	138560.0
199.00	9683.0
200.00	921.0
201.50	783.0
203.00	1073.0
204.00	5111.0
205.00	8847.0
206.10	37792.0
207.00	5940.0
208.00	1449.0
209.00	528.0
210.20	508.0
211.00	1593.0
214.90	483.0
216.00	779.0
217.00	10023.0
218.00	1421.0
220.00	263.0
221.10	5140.0
221.90	1442.0
223.00	2092.0
224.00	19624.0
225.00	4930.0
226.10	614.0
227.00	9107.0
228.00	1276.0
229.00	2045.0
230.00	188.0
231.00	787.0
233.00	239.0
234.00	457.0
234.90	604.0
235.90	396.0
236.90	850.0
238.90	395.0
240.90	387.0
242.00	1301.0
243.10	891.0
244.10	15444.0

Instrument :
BNA_M
ClientSampleId :
DFTPP02

245.10	1748.0
246.00	3102.0
247.00	795.0
248.00	186.0
249.00	504.0
250.90	170.0
251.90	177.0
253.00	362.0
254.10	641.0
255.00	78216.0
256.00	12508.0
257.10	978.0
258.00	4923.0
259.00	649.0
260.90	179.0
265.00	1925.0
266.10	240.0
271.00	210.0
272.00	197.0
273.00	2125.0
274.00	6081.0
275.00	35440.0
276.00	4958.0
277.00	3165.0
277.90	540.0
279.10	159.0
281.10	537.0
282.00	266.0
283.00	296.0
284.00	396.0
285.00	648.0
293.10	553.0
293.90	251.0
294.90	217.0
296.00	9708.0
297.00	1616.0
302.00	212.0
302.90	1197.0
304.00	287.0
307.90	166.0
310.00	173.0
314.00	422.0
314.90	1119.0
315.90	491.0
317.00	177.0
320.90	257.0
323.10	2366.0
324.00	539.0
326.90	537.0
328.10	186.0
331.90	246.0
333.00	238.0
334.00	1747.0
334.90	371.0
341.10	372.0
342.00	156.0
346.00	600.0
346.90	225.0
352.00	839.0
353.00	565.0
354.00	708.0
355.10	304.0
365.00	4160.0
365.80	545.0
371.10	234.0
372.10	1365.0
372.90	285.0
383.00	340.0
383.90	161.0
401.90	737.0
403.00	734.0
404.00	309.0
420.90	661.0
422.00	625.0
423.10	4380.0
424.10	1267.0
425.00	261.0
441.00	10980.0
442.00	74664.0
443.10	14464.0
444.10	1601.0

Instrument :
BNA_M
ClientSampleId :
DFTPP02

m/z	Abundance
35.90	225.0
37.00	628.0
38.10	1405.0
39.10	8847.0
40.00	1153.0
41.10	725.0
42.10	325.0
43.10	488.0
44.10	708.0
45.00	471.0
49.00	1090.0
50.10	30328.0
51.10	121048.0
52.10	6108.0
53.00	324.0
55.00	842.0
56.10	3665.0
57.00	8476.0
58.00	510.0
59.10	164.0
60.00	266.0
61.00	1599.0
62.10	1718.0
63.10	4627.0
64.10	787.0
65.00	2250.0
66.10	200.0
67.00	305.0
68.00	2143.0
69.00	113992.0
70.00	660.0
71.00	263.0
73.10	1890.0
74.10	12359.0
75.00	18040.0
76.10	6703.0
77.10	127256.0
78.10	8626.0
79.00	9269.0
80.00	6392.0
81.00	8123.0
82.00	2146.0
83.00	2128.0
83.90	278.0
85.00	1817.0
86.00	2880.0
87.10	1264.0
88.00	414.0
89.00	357.0
91.10	2124.0
92.00	2526.0
93.00	15229.0
94.00	1139.0
95.10	306.0
96.00	644.0
97.00	347.0
98.00	11388.0
99.00	8289.0
100.00	609.0
101.00	4867.0
102.10	265.0
103.00	1751.0
104.00	2730.0
105.00	2987.0
106.00	870.0
107.00	36632.0
108.10	5446.0
109.10	921.0
110.00	65168.0
111.00	9372.0
112.00	1152.0
113.10	406.0
114.80	200.0
116.00	2047.0
117.00	33056.0
118.00	2382.0
119.10	342.0
120.00	359.0
121.00	235.0
122.00	2310.0
123.00	3819.0
124.00	1800.0
125.00	1831.0
126.20	699.0
127.00	116360.0
128.00	8803.0
129.00	46264.0
130.10	3915.0
131.00	640.0
131.90	580.0
133.00	196.0
133.90	1302.0

Instrument :
BNA_M
ClientSampleId :
DFTPP02

135.10	3836.0
136.00	1745.0
137.10	2027.0
138.00	622.0
139.00	377.0
140.00	775.0
141.00	5914.0
141.90	2102.0
143.00	1672.0
144.00	392.0
145.00	339.0
145.90	1108.0
147.00	3491.0
148.00	8440.0
148.90	1612.0
150.10	381.0
151.10	747.0
152.00	567.0
153.00	1936.0
154.00	1640.0
155.00	3948.0
156.10	4979.0
157.10	1263.0
158.00	1143.0
159.00	862.0
160.00	2193.0
161.00	3185.0
162.00	671.0
163.00	233.0
164.00	396.0
165.00	2160.0
166.00	1812.0
167.00	14124.0
168.00	6489.0
169.10	1095.0
169.90	663.0
171.00	547.0
172.00	957.0
173.00	1405.0
174.00	2939.0
175.00	5291.0
176.10	1564.0
177.00	2087.0
178.00	838.0
179.00	9963.0
180.00	6699.0
181.00	3241.0
181.90	519.0
182.80	308.0
184.00	906.0
185.00	4753.0
186.00	35944.0
187.00	11207.0
188.00	1125.0
189.00	2257.0
190.00	297.0
191.00	1318.0
192.00	3557.0
193.10	3604.0
194.10	464.0
195.00	468.0
196.00	6893.0
197.00	1003.0
198.00	250496.0
199.00	17272.0
199.90	1143.0
201.50	988.0
202.20	251.0
203.00	2460.0
204.00	10012.0
205.00	15643.0
206.10	69496.0
207.10	9554.0
208.00	2546.0
209.00	997.0
210.30	1266.0
211.00	2918.0
213.00	256.0
214.90	914.0
216.00	1400.0
217.00	19080.0
218.00	2161.0
219.10	239.0
219.80	237.0
221.10	9714.0
221.80	2636.0
223.00	4054.0
224.00	36880.0
225.00	9509.0
226.00	1260.0
227.00	16416.0
228.10	2641.0
229.00	3310.0
230.10	432.0
231.00	1372.0
232.20	188.0

Instrument :
BNA_M
ClientSampleId :
DFTPP02

233.00	214.0
233.90	1123.0
234.90	1083.0
235.90	1009.0
237.00	1697.0
238.90	642.0
240.00	605.0
241.00	925.0
242.00	2101.0
243.00	2164.0
244.10	28608.0
245.00	3337.0
246.00	6306.0
247.00	1159.0
248.00	190.0
248.90	1047.0
250.10	234.0
251.00	243.0
252.00	283.0
253.00	744.0
254.10	984.0
255.00	153216.0
256.00	22112.0
257.00	1902.0
258.00	9960.0
259.00	1207.0
259.80	223.0
260.10	233.0
261.00	227.0
262.90	164.0
264.10	200.0
265.00	4008.0
265.90	621.0
269.90	196.0
271.10	300.0
272.00	496.0
273.00	4497.0
274.00	11262.0
275.00	66624.0
276.00	9156.0
277.00	6038.0
278.00	1125.0
278.90	182.0
281.10	336.0
282.00	231.0
283.00	761.0
284.00	368.0
285.10	1273.0
289.00	274.0
292.00	276.0
292.90	1140.0
294.00	409.0
295.00	466.0
296.00	19248.0
297.00	2891.0
298.00	218.0
301.00	255.0
302.00	422.0
303.00	2052.0
304.10	579.0
308.00	242.0
308.80	169.0
310.00	272.0
312.90	216.0
314.00	872.0
315.00	2455.0
316.00	1221.0
317.00	208.0
320.90	583.0
322.10	256.0
323.10	5470.0
324.00	1083.0
325.00	163.0
325.80	177.0
327.00	1044.0
328.10	568.0
332.00	437.0
333.00	566.0
334.00	3711.0
334.90	1336.0
335.90	154.0
341.00	785.0
341.90	189.0
346.00	1118.0
347.00	157.0
352.00	1804.0
353.10	1441.0
354.00	1975.0
355.10	475.0
359.00	172.0
365.00	8563.0
366.00	1305.0
370.00	218.0
370.90	400.0
372.00	3168.0
373.00	1076.0

Instrument :
BNA_M
ClientSampleId :
DFTPP02

382.90	742.0
384.00	290.0
390.10	342.0
391.10	331.0
392.00	200.0
400.90	210.0
402.00	1274.0
403.00	1973.0
404.00	452.0
405.00	173.0
421.10	1415.0
422.10	1825.0
423.00	10020.0
424.00	2405.0
425.00	409.0
441.10	25920.0
442.00	178560.0
443.00	35872.0
444.00	3083.0

Instrument :
BNA_M
ClientSampleId :
DFTPP02

m/z	Abundance
36.00	217.0
37.00	496.0
38.10	1159.0
39.10	7176.0
40.00	1237.0
41.10	612.0
42.00	151.0
43.00	636.0
44.10	703.0
45.00	414.0
49.10	775.0
50.10	24808.0
51.10	97352.0
52.10	4382.0
53.00	230.0
55.10	878.0
56.10	2815.0
57.00	6585.0
58.00	294.0
59.00	169.0
60.00	242.0
61.00	1188.0
62.10	1553.0
63.10	3553.0
64.00	507.0
65.00	1783.0
66.10	210.0
66.90	194.0
68.10	1494.0
69.00	92760.0
70.00	519.0
71.00	233.0
73.10	1812.0
74.00	9263.0
75.00	14967.0
76.10	5295.0
77.10	104784.0
78.10	7284.0
79.00	7078.0
80.00	5257.0
81.00	7574.0
82.00	1882.0
83.00	2288.0
84.00	298.0
85.10	1295.0
86.10	2116.0
87.00	905.0
88.00	279.0
89.00	277.0
91.00	2153.0
92.00	1787.0
93.00	13662.0
93.90	836.0
95.10	274.0
96.10	398.0
97.10	252.0
98.00	10285.0
99.00	6959.0
100.00	589.0
101.00	3979.0
101.90	326.0
103.00	1538.0
104.00	2693.0
105.00	2284.0
106.10	895.0
107.00	32192.0
108.00	4209.0
109.00	834.0
110.00	54736.0
111.00	8731.0
112.00	917.0
113.00	380.0
116.00	1602.0
117.00	29040.0
118.00	2126.0
119.00	435.0
120.00	476.0
120.80	172.0
122.00	2691.0
123.00	3145.0
124.00	1301.0
125.00	1249.0
127.00	103672.0
128.00	7754.0
129.00	43552.0
130.10	3475.0
130.90	751.0
131.90	346.0
132.90	262.0
134.00	1210.0
135.00	3607.0
136.10	1378.0

Instrument :
BNA_M
ClientSampleId :
DFTPP02

137.00	2020.0
138.00	474.0
138.80	163.0
139.90	586.0
141.00	6098.0
142.00	2056.0
143.00	1486.0
144.10	205.0
144.90	369.0
145.90	836.0
147.00	3521.0
148.00	7157.0
149.00	1782.0
150.00	361.0
151.00	583.0
151.80	427.0
153.00	1822.0
154.10	1495.0
155.00	3190.0
156.10	4630.0
157.10	774.0
158.00	1024.0
159.00	1010.0
160.00	1917.0
160.90	2469.0
162.00	689.0
163.00	343.0
163.90	320.0
165.00	2125.0
166.00	1767.0
167.00	12453.0
168.00	6037.0
169.00	1026.0
170.10	474.0
171.00	555.0
171.90	1152.0
173.00	1076.0
173.90	2690.0
175.00	4193.0
176.00	1480.0
177.00	2213.0
178.00	791.0
179.00	9194.0
180.00	6230.0
181.00	2830.0
182.00	339.0
182.80	233.0
183.90	595.0
185.00	4139.0
186.00	34224.0
187.00	9571.0
188.00	1068.0
189.00	2419.0
189.90	420.0
190.90	1386.0
192.00	2596.0
193.00	3510.0
193.90	769.0
194.90	297.0
196.00	6861.0
196.80	1454.0
197.10	1357.0
198.00	242368.0
199.00	16227.0
200.00	1568.0
201.30	999.0
203.00	2131.0
204.00	9307.0
205.10	14557.0
206.10	67544.0
207.00	9222.0
208.00	2600.0
209.00	1083.0
210.10	1091.0
211.10	2454.0
212.90	175.0
215.10	923.0
216.00	1589.0
217.00	16784.0
218.00	2403.0
219.00	178.0
220.20	507.0
221.10	9634.0
221.80	2660.0
223.00	3662.0
224.10	37160.0
225.00	9679.0
226.10	956.0
227.00	17320.0
228.00	2227.0
229.00	3069.0
230.00	485.0
231.00	1651.0
232.00	213.0
233.00	246.0
233.90	1061.0

Instrument :
BNA_M
ClientSampleId :
DFTPP02

235.00	1428.0
236.00	878.0
237.00	1401.0
239.00	719.0
240.00	612.0
241.00	938.0
242.00	2048.0
243.10	2059.0
244.00	29568.0
245.00	3840.0
246.00	6562.0
246.90	1437.0
247.90	408.0
249.10	1005.0
249.90	306.0
251.00	300.0
251.90	226.0
252.90	753.0
254.10	997.0
255.00	159360.0
256.00	22016.0
257.10	1671.0
258.00	9153.0
259.00	1484.0
260.10	457.0
261.00	295.0
263.80	228.0
265.00	3905.0
266.00	515.0
266.90	170.0
270.90	284.0
272.20	471.0
273.00	4470.0
274.00	12661.0
275.00	71928.0
276.00	9762.0
277.00	5927.0
278.00	1209.0
281.10	430.0
282.10	229.0
283.10	751.0
284.00	569.0
285.00	1108.0
289.00	229.0
291.10	189.0
292.20	247.0
292.90	1490.0
294.00	322.0
295.10	316.0
296.00	21240.0
297.10	2826.0
297.90	210.0
301.00	210.0
302.10	391.0
303.00	2617.0
304.00	725.0
308.10	314.0
308.80	165.0
309.90	239.0
313.00	217.0
314.10	1201.0
315.00	2357.0
316.10	1050.0
316.90	185.0
321.00	616.0
322.00	346.0
323.10	7155.0
324.10	1227.0
327.00	945.0
328.10	773.0
332.00	516.0
333.00	488.0
334.00	4008.0
334.90	1070.0
339.00	207.0
341.00	860.0
342.00	210.0
346.00	1559.0
347.00	363.0
352.00	2277.0
353.00	1406.0
354.00	2110.0
355.00	477.0
358.80	182.0
365.00	10622.0
366.00	1389.0
370.90	693.0
372.00	3711.0
373.00	1240.0
383.00	1094.0
383.90	197.0
390.10	629.0
390.90	345.0
392.00	286.0
401.10	317.0
402.00	1945.0

Instrument :
BNA_M
ClientSampleId :
DFTPP02

403.00	2190.0
404.10	814.0
421.00	2186.0
422.00	1584.0
423.10	12794.0
424.00	3030.0
425.10	310.0
441.00	33872.0
442.10	232064.0
443.10	46520.0
444.10	4268.0
445.00	208.0

Instrument :
BNA_M
ClientSampleId :
DFTPP02