




Data File : BM021047.D
Acq On : 19 Jun 2019 22:17
Operator : HP/JU
Sample : K3213-11
Misc :
ALS Vial : 15 Sample Multiplier: 1

Quant Time: Jun 20 03:48:57 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_M\METHODS\SOM-EPA-BM061419MA.M
Quant Title : SVOA CALIBRATION
QLast Update : Thu Jun 20 03:45:02 2019
Response via : Initial Calibration



TIC: BM021047.D

(51) 4-Nitrophenol-d4 (S)
14.604min (-0.029) 0.09ng/ul
response 890

Ion	Exp%	Act%
143.00	100	100
113.00	77.40	128.07#
41.00	36.20	1414.58#
42.00	24.30	231.33#