

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM061923\
 Data File : BM040323.D
 Acq On : 19 Jun 2023 20:28
 Operator : MA/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 SSTD0.4125

Quant Time: Jun 20 01:22:57 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\SFAM-EPA-SIM-BM061423.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Jun 15 00:31:16 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.970	152	7303	0.400	ng/ul	0.00
4) Naphthalene-d8	10.778	136	28429	0.400	ng/ul	0.00
9) Acenaphthene-d10	14.605	164	14602	0.400	ng/ul	0.00
13) Phenanthrene-d10	17.349	188	28791	0.400	ng/ul	0.00
17) Chrysene-d12	21.526	240	24386	0.400	ng/ul	# 0.00
23) Perylene-d12	23.950	264	26802	0.400	ng/ul	-0.01
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.346	96	3618	0.401	ng/ul	0.00
6) 2-Methylnaphthalene-d10	12.362	152	15181	0.421	ng/ul	-0.01
18) Fluoranthene-d10	19.375	212	27566	0.453	ng/ul	0.00
Target Compounds						
						Qvalue
2) 1,4-Dioxane	3.380	88	3897	0.411	ng/ul#	53
5) Naphthalene	10.828	128	27840	0.413	ng/ul	99
7) 2-Methylnaphthalene	12.439	142	17574	0.420	ng/ul	98
8) 1-Methylnaphthalene	12.653	142	17903	0.408	ng/ul	99
10) Acenaphthylene	14.327	152	26194	0.513	ng/ul	99
11) Acenaphthene	14.670	153	19071	0.436	ng/ul	99
12) Fluorene	15.650	166	20619	0.423	ng/ul	98
14) Pentachlorophenol	17.007	266	2779	0.473	ng/ul	99
15) Phenanthrene	17.391	178	31403	0.415	ng/ul	100
16) Anthracene	17.480	178	30098	0.469	ng/ul	99
19) Fluoranthene	19.403	202	37126	0.447	ng/ul	99
20) Pyrene	19.765	202	39631	0.459	ng/ul	99
21) Benzo(a)anthracene	21.509	228	34931	0.485	ng/ul	97
22) Chrysene	21.561	228	34208	0.450	ng/ul	97
24) Benzo(b)fluoranthene	23.210	252	36808	0.373	ng/ul	89
25) Benzo(k)fluoranthene	23.257	252	36680	0.394	ng/ul#	91
26) Benzo(a)pyrene	23.841	252	32772	0.404	ng/ul#	89
27) Indeno(1,2,3-cd)pyrene	26.467	276	41631	0.406	ng/ul#	97
28) Dibenzo(a,h)anthracene	26.484	278	33196	0.403	ng/ul	92
29) Benzo(g,h,i)perylene	27.245	276	34585	0.403	ng/ul	93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM061923\
Data File : BM040323.D
Acq On : 19 Jun 2023 20:28
Operator : MA/JU
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
SSTD0.4125

Quant Time: Jun 20 01:22:57 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\SFAM-EPA-SIM-BM061423.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Jun 15 00:31:16 2023
Response via : Initial Calibration

