



Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM061923\
 Data File : BM040306.D
 Acq On : 19 Jun 2023 08:58
 Operator : MA/JU
 Sample : DFTPP130
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 DFTPP130

Integration File: rteint.p

Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\SFAM-EPA-SIM-BM061423.M
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 Last Update : Thu Jun 15 00:31:16 2023



AutoFind: Scans 2464, 2465, 2466; Background Corrected with Scan 2455

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	34.3	425976	PASS
68	69	0.00	2	1.5	6962	PASS
69	198	0.00	100	38.5	478536	PASS
70	69	0.00	2	0.6	2669	PASS
127	198	10	80	47.0	584227	PASS
197	198	0.00	2	0.4	4580	PASS
198	198	100	100	100.0	1242539	PASS
199	198	5	9	6.8	84637	PASS
275	198	10	60	25.6	318635	PASS
365	198	1	100	3.3	41624	PASS
441	198	0.01	100	14.0	173547	PASS
442	198	50	100	87.3	1084203	PASS
443	442	15	24	19.7	213397	PASS

m/z	Abundance
35.90	400.0
37.90	229.0
39.00	770.0
39.90	1866.0
41.00	1312.0
42.00	474.0
43.00	1238.0
44.00	10156.0
45.00	996.0
47.00	212.0
47.90	168.0
48.90	980.0
50.00	426.0
51.00	1267.0
51.90	251.0
53.00	317.0
53.90	200.0
55.00	997.0
56.00	396.0
57.10	683.0
59.00	743.0
60.90	182.0
63.00	282.0
65.00	539.0
66.00	181.0
67.00	678.0
67.90	190.0
69.00	803.0
70.00	423.0
70.90	475.0
73.00	19400.0
74.00	1842.0
75.00	1691.0
75.90	229.0
76.90	1606.0
78.00	387.0
79.00	506.0
81.00	649.0
82.00	398.0
83.00	739.0
83.90	761.0
85.10	520.0
85.90	434.0
86.90	362.0
87.80	229.0
88.90	356.0
91.00	1030.0
92.00	278.0
93.00	617.0
93.90	228.0
94.90	625.0
96.00	1223.0
96.90	833.0
101.80	155.0
102.80	492.0
103.90	340.0
104.90	674.0
107.00	546.0
107.90	171.0
109.00	391.0
110.90	374.0
115.00	671.0
116.90	536.0
119.00	782.0
119.90	229.0
120.90	381.0
121.90	161.0
123.00	202.0
124.80	330.0
125.90	178.0
126.90	200.0
128.00	297.0
128.90	350.0
129.90	208.0
130.90	420.0
131.90	177.0
132.90	2088.0
134.00	304.0
134.90	400.0
137.00	162.0
141.00	310.0
142.10	184.0
143.00	169.0
145.00	340.0
147.00	7510.0
148.00	1421.0
148.90	1040.0
150.00	269.0
150.90	202.0
151.90	240.0
152.90	385.0
154.10	194.0

Instrument :
BNA_M
ClientSampleId :
DFTPP130

157.00	202.0
158.80	268.0
161.00	368.0
163.00	447.0
165.00	368.0
168.90	187.0
171.10	175.0
174.80	205.0
175.90	201.0
176.90	400.0
177.90	304.0
178.90	275.0
183.90	204.0
184.90	412.0
185.90	693.0
187.90	156.0
188.90	325.0
190.90	1739.0
191.90	268.0
192.90	1451.0
195.00	203.0
196.90	195.0
201.90	215.0
202.90	160.0
205.00	502.0
205.90	236.0
207.00	12604.0
208.00	3041.0
209.00	1639.0
209.90	271.0
221.00	3792.0
222.00	1221.0
223.00	1129.0
223.90	201.0
224.90	197.0
248.90	375.0
249.80	157.0
251.00	251.0
252.00	159.0
252.90	190.0
264.90	374.0
265.90	180.0
266.90	1587.0
267.90	354.0
268.80	377.0
281.00	5403.0
281.90	1703.0
283.00	1009.0
283.90	208.0
294.90	366.0
325.00	398.0
326.80	600.0
340.90	1884.0
341.90	657.0
342.90	263.0
355.00	2609.0
356.00	796.0
357.00	556.0
358.00	284.0
368.90	171.0
384.90	177.0
400.90	702.0
401.80	317.0
402.90	161.0
415.00	185.0
429.10	520.0
429.90	273.0
430.90	294.0

Instrument :
BNA_M
ClientSampleId :
DFTPP130

m/z	Abundance
36.00	682.0
37.00	2057.0
38.00	5672.0
39.00	36184.0
40.00	2827.0
41.00	2756.0
42.00	567.0
43.00	1962.0
44.00	10677.0
45.00	1280.0
47.00	429.0
48.00	436.0
49.00	4216.0
50.00	132928.0
51.00	505664.0
52.00	25696.0
53.00	1482.0
53.90	298.0
55.00	3817.0
56.00	14413.0
57.00	35536.0
58.00	2077.0
59.10	898.0
59.90	522.0
61.00	6274.0
62.00	7207.0
63.00	22184.0
64.00	3062.0
65.00	11601.0
66.00	1203.0
67.00	1099.0
68.00	7967.0
69.00	566592.0
70.00	3359.0
71.00	1368.0
72.00	292.0
73.00	8154.0
74.00	58104.0
75.00	92168.0
76.00	32136.0
77.00	691392.0
78.00	52144.0
79.00	41656.0
80.00	31544.0
81.00	45856.0
82.00	11179.0
83.00	11262.0
83.90	2230.0
85.00	9094.0
86.00	12749.0
87.00	6242.0
88.00	2713.0
89.00	1145.0
90.00	498.0
91.00	9334.0
92.00	11200.0
93.00	74864.0
94.00	5472.0
95.00	1821.0
96.00	4840.0
97.00	2082.0
98.00	57920.0
99.00	48920.0
100.00	4023.0
101.00	25392.0
101.90	1395.0
103.00	7553.0
104.00	14935.0
105.00	15099.0
106.00	4245.0
107.00	201856.0
108.00	29736.0
109.00	5166.0
110.00	351616.0
111.00	55520.0
112.00	7324.0
113.00	2743.0
114.10	452.0
115.00	1283.0
116.00	10066.0
117.00	183552.0
118.00	12732.0
119.00	2479.0
120.00	2335.0
121.00	1193.0
122.00	11611.0
122.90	19008.0
124.00	8744.0
125.00	8076.0
126.00	1977.0
127.00	675136.0
128.00	50648.0

Instrument :
BNA_M
ClientSampleId :
DFTPP130

129.00	253056.0
130.00	22880.0
131.00	4713.0
132.00	2298.0
132.90	2238.0
134.00	7206.0
135.00	20944.0
136.00	9181.0
137.00	12302.0
137.90	2286.0
138.90	1192.0
140.00	2590.0
141.00	30872.0
142.00	11518.0
143.00	8091.0
144.00	2474.0
145.00	2176.0
145.90	5787.0
147.00	18384.0
148.00	44400.0
149.00	9550.0
150.00	2500.0
151.00	4401.0
151.90	2838.0
152.90	10320.0
154.00	8359.0
155.00	19832.0
156.00	28064.0
157.00	5268.0
157.90	5882.0
159.00	5041.0
160.00	10274.0
161.00	15452.0
161.90	4541.0
163.00	1714.0
164.00	1474.0
165.00	11562.0
165.90	10353.0
167.00	76584.0
168.00	49008.0
169.00	6323.0
170.00	2181.0
170.90	2427.0
171.90	5830.0
172.90	7791.0
174.00	15004.0
175.00	25880.0
176.00	8017.0
177.00	12573.0
177.90	4143.0
178.90	50992.0
180.00	35864.0
181.00	17304.0
182.00	2170.0
182.90	1619.0
183.90	3905.0
185.00	24696.0
186.00	189312.0
187.00	55160.0
188.00	6209.0
188.90	12104.0
190.00	1831.0
191.00	6873.0
192.00	17344.0
193.00	20384.0
194.00	4228.0
194.90	2202.0
196.00	37944.0
196.80	4962.0
197.90	1363968.0
198.90	96456.0
200.00	7641.0
201.40	5667.0
203.00	9690.0
204.00	51640.0
205.00	86064.0
206.00	360384.0
207.00	63984.0
208.00	14473.0
209.00	6062.0
210.00	5413.0
211.00	13829.0
213.00	1045.0
214.00	394.0
214.90	3718.0
215.90	8432.0
216.90	92784.0
218.00	11981.0
218.90	1460.0
221.00	56808.0
221.80	13138.0
223.00	20320.0
224.00	201600.0
225.00	49600.0
226.00	6131.0
227.00	81080.0

Instrument :
BNA_M
ClientSampleId :
DFTPP130

228.00	12011.0
229.00	17344.0
229.90	2259.0
231.00	7177.0
232.00	1215.0
232.90	1410.0
234.00	4615.0
235.00	5873.0
235.90	4351.0
237.00	8272.0
238.00	1405.0
238.90	3259.0
240.00	2765.0
241.00	4139.0
242.00	11062.0
243.00	10623.0
244.00	160320.0
245.00	22112.0
245.90	28224.0
247.00	5889.0
248.00	1575.0
249.00	5630.0
249.90	1015.0
250.90	1385.0
252.00	1390.0
253.00	2832.0
254.00	3305.0
255.00	755456.0
256.00	111768.0
257.00	9518.0
258.00	44184.0
259.00	7065.0
260.00	1401.0
260.90	1540.0
261.80	246.0
262.90	276.0
264.00	980.0
264.90	17520.0
266.00	2561.0
266.90	979.0
267.90	435.0
268.80	160.0
269.90	920.0
271.00	1501.0
272.00	1822.0
273.00	23360.0
274.00	60192.0
275.00	317632.0
276.00	45120.0
276.90	28144.0
278.00	5501.0
279.00	1201.0
280.00	208.0
281.00	3085.0
282.00	1405.0
283.00	3355.0
283.90	1910.0
285.00	4295.0
286.00	1034.0
287.90	310.0
288.90	1252.0
289.90	860.0
290.90	649.0
292.00	1536.0
293.00	6177.0
293.90	2140.0
294.90	1717.0
296.00	96936.0
297.00	11949.0
298.00	962.0
298.90	265.0
299.80	164.0
301.00	1330.0
302.00	1745.0
303.00	10699.0
304.00	2987.0
304.90	422.0
306.10	187.0
307.00	218.0
307.90	1129.0
308.90	1212.0
310.00	1328.0
310.90	338.0
311.80	310.0
313.00	816.0
314.00	4124.0
314.90	10069.0
316.00	5740.0
317.00	1080.0
318.00	160.0
319.00	180.0
320.00	330.0
321.00	2758.0
321.90	1558.0
323.00	28072.0
324.00	5690.0

Instrument :
BNA_M
ClientSampleId :
DFTPP130

324.90	597.0
326.00	642.0
326.90	6135.0
328.00	2896.0
328.90	556.0
330.00	190.0
332.00	1810.0
333.00	2364.0
334.00	17680.0
335.00	4855.0
336.00	743.0
338.90	576.0
340.00	434.0
341.00	3911.0
342.00	865.0
342.90	262.0
346.00	7264.0
346.90	1380.0
347.80	279.0
350.00	150.0
351.00	488.0
352.00	8328.0
353.00	5382.0
354.00	9181.0
355.00	2274.0
355.90	370.0
357.00	292.0
359.00	569.0
360.70	210.0
363.90	160.0
365.00	36616.0
365.90	6443.0
366.80	409.0
370.00	697.0
371.00	1843.0
372.00	13175.0
373.00	3554.0
373.90	446.0
376.90	355.0
383.00	3713.0
384.00	1097.0
384.90	545.0
390.00	1677.0
391.00	978.0
392.00	910.0
397.00	186.0
400.90	766.0
402.00	5304.0
403.00	7270.0
404.00	2975.0
405.00	660.0
406.00	151.0
410.00	245.0
415.00	338.0
421.00	6996.0
422.00	6502.0
423.00	44776.0
424.00	13917.0
425.00	1855.0
426.00	178.0
439.00	234.0
439.90	201.0
441.00	143232.0
442.00	885184.0
443.00	180480.0
444.00	17688.0
445.00	1090.0

Instrument :
BNA_M
ClientSampleId :
DFTPP130

m/z	Abundance
36.00	789.0
37.00	2434.0
38.10	5649.0
39.00	35128.0
40.00	3550.0
41.00	2972.0
42.00	529.0
43.10	2323.0
44.00	11516.0
45.00	1410.0
45.90	182.0
47.00	375.0
47.90	418.0
49.00	4194.0
50.00	122616.0
51.00	487424.0
52.00	24344.0
53.00	1410.0
54.00	203.0
55.00	4100.0
56.00	14532.0
57.00	33640.0
58.00	1771.0
59.00	1099.0
59.90	569.0
61.00	6509.0
62.00	7328.0
63.00	20328.0
64.00	2940.0
65.00	10713.0
66.00	906.0
67.10	1399.0
68.00	8665.0
69.00	553408.0
70.00	3558.0
71.00	1322.0
71.90	511.0
73.00	8399.0
74.00	60136.0
75.00	90328.0
76.00	31176.0
77.00	681344.0
78.00	47248.0
79.00	40112.0
80.00	31608.0
81.00	45696.0
82.00	11679.0
83.00	11410.0
84.00	2494.0
85.00	9410.0
86.00	12948.0
87.00	5362.0
87.90	2716.0
89.00	1465.0
90.00	362.0
91.00	10972.0
92.00	10889.0
93.00	74680.0
94.00	5721.0
95.00	1857.0
96.00	4935.0
97.00	1748.0
98.00	57120.0
99.00	47512.0
100.00	4102.0
101.00	25016.0
101.90	1683.0
103.00	8088.0
104.00	16042.0
105.00	15065.0
106.00	4872.0
107.00	202752.0
108.00	30896.0
109.00	5188.0
110.00	353920.0
111.00	53584.0
112.00	6581.0
113.00	2290.0
114.00	516.0
114.90	1492.0
116.00	9968.0
117.00	184768.0
118.00	13209.0
119.00	2314.0
120.00	2288.0
120.90	1355.0
122.00	11486.0
123.00	18696.0
124.00	8814.0
125.00	7764.0
126.00	1789.0
127.00	681664.0

Instrument :
BNA_M
ClientSampleId :
DFTPP130

128.00	49904.0
129.00	255168.0
130.00	22296.0
131.00	4607.0
132.00	2259.0
133.00	2502.0
133.90	7364.0
135.00	20456.0
136.00	9064.0
137.00	12872.0
138.00	2645.0
139.00	1132.0
140.00	3122.0
141.00	33464.0
142.00	10881.0
143.00	7904.0
144.00	2146.0
144.90	1972.0
145.90	5707.0
147.00	18240.0
148.00	44600.0
149.00	9335.0
149.90	2937.0
151.10	4160.0
151.80	2562.0
153.00	9774.0
154.00	8213.0
155.00	19864.0
156.00	28424.0
157.00	5942.0
157.90	6237.0
158.90	4538.0
160.00	10196.0
161.00	16305.0
162.00	5006.0
163.00	1696.0
163.90	1719.0
164.90	13284.0
166.00	10721.0
167.00	80456.0
168.00	50824.0
169.00	7143.0
170.00	1910.0
170.90	3159.0
171.90	6167.0
173.00	8698.0
174.00	15579.0
175.00	28912.0
176.00	7010.0
177.00	12363.0
177.90	4610.0
178.90	53256.0
180.00	38784.0
181.00	17208.0
181.90	2471.0
182.90	1599.0
183.90	4748.0
185.00	27264.0
186.00	204672.0
187.00	59080.0
188.00	6949.0
189.00	13574.0
189.90	1993.0
191.00	7744.0
192.00	18568.0
193.00	20776.0
194.00	4436.0
195.00	1963.0
196.00	40064.0
196.80	5967.0
197.90	1499648.0
198.90	101376.0
200.00	7879.0
201.50	6618.0
203.00	10915.0
204.00	57352.0
205.00	96808.0
206.00	394816.0
207.00	65032.0
208.00	15681.0
209.00	6345.0
210.00	6652.0
211.00	15281.0
211.80	3254.0
213.00	902.0
213.90	470.0
214.90	4621.0
216.00	9360.0
216.90	107464.0
218.00	13391.0
219.00	1481.0
221.00	62928.0
221.80	14031.0
223.00	24232.0
224.00	229440.0
225.00	56912.0

Instrument :
BNA_M
ClientSampleId :
DFTPP130

226.00	6114.0
227.00	98688.0
228.00	13766.0
228.90	21040.0
230.00	2429.0
231.00	8041.0
231.90	1313.0
232.90	1855.0
234.00	5653.0
235.00	6628.0
236.00	5093.0
237.00	8748.0
238.00	1273.0
239.00	3725.0
240.00	3514.0
241.00	5464.0
242.00	12900.0
243.00	13603.0
244.00	193792.0
245.00	23816.0
246.00	34184.0
247.00	7295.0
247.90	1853.0
249.00	5891.0
250.00	1387.0
250.90	1794.0
251.90	1739.0
252.90	3513.0
254.00	4451.0
255.00	919616.0
256.00	133184.0
257.00	10388.0
258.00	53296.0
259.00	8082.0
260.00	1452.0
261.00	1787.0
261.90	186.0
263.00	603.0
264.00	1149.0
265.00	22240.0
266.00	2993.0
267.00	1416.0
267.90	296.0
268.90	259.0
270.00	1070.0
271.00	1949.0
272.00	2326.0
273.00	29496.0
274.00	71592.0
275.00	400000.0
276.00	54552.0
277.00	33688.0
278.00	5968.0
279.00	1070.0
280.00	160.0
281.00	3553.0
282.00	1640.0
283.00	4768.0
284.00	2984.0
285.00	5994.0
286.00	1062.0
288.00	344.0
288.90	1400.0
289.90	1126.0
291.00	777.0
292.00	1430.0
293.00	8063.0
293.90	2121.0
295.00	1945.0
296.00	120832.0
297.00	17392.0
297.90	1402.0
299.10	216.0
299.80	197.0
300.90	1370.0
302.00	2122.0
303.00	13126.0
304.00	3565.0
305.00	463.0
305.90	216.0
306.80	208.0
307.90	1722.0
309.00	1335.0
310.00	1539.0
311.00	329.0
311.90	461.0
313.00	935.0
314.00	5269.0
315.00	12808.0
316.00	7108.0
317.00	1454.0
318.00	175.0
318.90	250.0
320.00	436.0
321.00	4003.0
322.00	1431.0

Instrument :
BNA_M
ClientSampleId :
DFTPP130

323.00	36576.0
324.00	6708.0
324.90	644.0
325.90	747.0
327.00	8265.0
328.00	3598.0
328.90	671.0
329.90	210.0
331.00	193.0
332.00	2577.0
333.00	3078.0
334.00	24200.0
335.00	5732.0
336.10	750.0
338.90	648.0
339.90	531.0
341.00	4872.0
342.00	1088.0
343.10	350.0
344.90	164.0
346.00	10158.0
347.00	1870.0
347.80	184.0
349.80	281.0
351.00	842.0
352.00	11535.0
353.00	7716.0
354.00	11126.0
355.00	2540.0
356.00	294.0
356.90	297.0
358.00	334.0
359.00	907.0
360.00	278.0
361.10	162.0
363.90	355.0
365.00	52072.0
366.00	8051.0
367.00	703.0
370.00	897.0
371.00	3062.0
372.00	19064.0
373.00	4714.0
374.10	582.0
376.90	426.0
383.00	5712.0
384.00	1483.0
384.90	565.0
385.90	204.0
390.00	2733.0
391.00	1842.0
392.00	1200.0
392.90	378.0
395.00	160.0
395.90	241.0
400.90	1257.0
402.00	8052.0
403.00	11315.0
404.00	4084.0
404.90	822.0
410.00	272.0
415.00	589.0
421.00	9603.0
422.00	9674.0
423.00	65840.0
424.00	16952.0
425.10	2079.0
425.90	157.0
438.10	150.0
439.10	183.0
439.90	193.0
441.00	216704.0
442.00	1358336.0
443.00	264576.0
444.00	26024.0
445.00	1148.0

Instrument :
BNA_M
ClientSampleId :
DFTPP130

m/z	Abundance
36.00	694.0
37.00	1342.0
38.00	3526.0
39.00	22000.0
40.00	2659.0
41.00	2592.0
42.00	565.0
43.00	2292.0
44.00	12598.0
44.90	1136.0
46.90	482.0
48.00	292.0
49.00	3122.0
50.00	73664.0
51.00	288640.0
52.00	14826.0
53.00	1206.0
54.10	309.0
55.00	2642.0
56.00	8592.0
57.00	20968.0
58.00	1326.0
59.00	967.0
60.00	268.0
61.00	3910.0
62.00	4327.0
63.00	12298.0
64.00	1677.0
65.00	6746.0
65.90	650.0
67.00	1299.0
68.00	4823.0
69.00	318016.0
70.00	2360.0
71.00	1281.0
71.90	387.0
73.00	6947.0
74.00	34920.0
75.00	52904.0
76.00	17632.0
77.00	385792.0
78.00	26288.0
79.00	22760.0
80.00	17528.0
81.00	27304.0
82.00	7066.0
83.00	6844.0
83.90	1628.0
85.00	5738.0
86.00	7818.0
87.00	3319.0
87.90	1701.0
88.90	1210.0
89.90	156.0
91.00	7389.0
92.00	6908.0
93.00	43096.0
94.00	3305.0
95.00	1551.0
96.00	3514.0
97.00	1312.0
98.00	31664.0
99.00	27664.0
99.90	2300.0
101.00	14348.0
101.90	1096.0
103.00	4347.0
104.00	8370.0
105.00	9488.0
106.00	2822.0
107.00	116784.0
108.00	17408.0
109.00	3339.0
110.00	206272.0
111.00	30904.0
112.00	3886.0
113.00	1248.0
114.10	424.0
115.00	1322.0
116.00	5533.0
117.00	103680.0
118.00	7976.0
119.00	1615.0
119.90	1883.0
121.00	1096.0
121.90	7241.0
122.90	12026.0
123.90	6322.0
125.00	4808.0
126.00	1177.0
127.00	396480.0
128.00	29464.0

Instrument :
BNA_M
ClientSampleId :
DFTPP130

128.90	150016.0
129.90	11633.0
131.00	3055.0
132.00	1601.0
133.00	1960.0
134.00	4786.0
135.00	12050.0
136.00	4905.0
137.00	7370.0
138.00	1359.0
138.90	943.0
139.90	1761.0
140.90	19472.0
142.00	6685.0
143.00	5356.0
144.00	1164.0
145.00	1203.0
145.90	3139.0
146.90	11209.0
147.90	26688.0
149.00	5860.0
149.90	2153.0
150.90	2788.0
151.90	1822.0
152.90	6104.0
154.00	4689.0
155.00	11391.0
156.00	17280.0
157.00	3442.0
157.90	4098.0
158.90	2946.0
159.90	6277.0
161.00	10043.0
161.90	2631.0
162.90	1302.0
163.90	1062.0
164.90	8108.0
165.90	6482.0
167.00	45056.0
168.00	29816.0
169.00	4253.0
170.00	1465.0
171.00	1756.0
171.90	3868.0
173.00	4955.0
173.90	8841.0
175.00	17504.0
176.00	4153.0
176.90	8378.0
178.00	2654.0
178.90	31704.0
180.00	22176.0
181.00	11023.0
181.90	1614.0
182.80	784.0
183.90	2924.0
185.00	16392.0
186.00	119256.0
187.00	33376.0
188.00	3128.0
188.90	7223.0
189.90	1460.0
190.90	4992.0
192.00	11001.0
193.00	13369.0
194.00	3268.0
195.00	1532.0
195.90	23672.0
196.90	3395.0
197.90	864000.0
198.90	56080.0
199.90	4527.0
201.50	3558.0
202.90	6337.0
204.00	32504.0
205.00	53400.0
206.00	225728.0
207.00	41360.0
208.00	11084.0
209.00	3993.0
210.00	3836.0
211.00	9139.0
211.90	1475.0
213.00	767.0
214.00	230.0
214.90	2538.0
215.90	5717.0
216.90	62208.0
218.00	8595.0
219.00	879.0
220.10	1033.0
221.00	34088.0
221.90	7886.0
223.00	14207.0
224.00	133248.0
225.00	32144.0

Instrument :
BNA_M
ClientSampleId :
DFTPP130

226.00	3808.0
227.00	55744.0
227.90	6653.0
228.90	11026.0
229.90	1235.0
231.00	4996.0
231.90	677.0
232.90	797.0
234.00	3283.0
234.90	4322.0
235.90	2938.0
237.00	5450.0
237.90	767.0
238.90	2527.0
240.00	1928.0
240.90	3340.0
242.00	7896.0
243.00	7859.0
244.00	108920.0
245.00	14690.0
246.00	20448.0
246.90	4365.0
247.90	963.0
248.90	3972.0
249.90	893.0
251.00	1046.0
251.90	1003.0
253.00	2201.0
253.90	3130.0
255.00	544576.0
256.00	79624.0
257.00	5577.0
257.90	32688.0
258.90	4784.0
259.90	816.0
261.00	853.0
262.90	451.0
264.00	958.0
264.90	13407.0
265.90	2057.0
267.00	726.0
268.00	341.0
268.90	283.0
269.90	724.0
271.00	1257.0
272.00	1814.0
272.90	16872.0
274.00	45840.0
275.00	238272.0
276.00	32376.0
276.90	20744.0
277.90	4267.0
278.90	818.0
280.00	162.0
281.00	2850.0
281.90	1425.0
283.00	2905.0
284.00	1591.0
285.00	3689.0
286.00	704.0
287.90	279.0
288.90	888.0
290.00	725.0
290.80	374.0
291.90	934.0
292.90	5313.0
294.00	1367.0
295.00	1533.0
295.90	77280.0
296.90	9683.0
298.00	649.0
298.90	344.0
300.90	1000.0
301.90	1434.0
303.00	9104.0
304.00	2179.0
304.90	291.0
306.80	172.0
307.90	1005.0
308.90	821.0
309.90	839.0
311.00	313.0
312.90	770.0
314.00	3569.0
315.00	8065.0
316.00	4792.0
316.90	563.0
320.00	201.0
321.00	2263.0
321.90	1115.0
323.00	23592.0
324.00	4519.0
324.90	593.0
325.90	672.0
326.90	6027.0
327.90	2601.0

Instrument :
BNA_M
ClientSampleId :
DFTPP130

328.90	374.0
332.00	1544.0
333.00	2313.0
334.00	16808.0
335.00	4064.0
336.10	426.0
339.00	378.0
339.90	340.0
341.00	2778.0
342.00	841.0
346.00	6752.0
346.90	1484.0
351.00	440.0
352.00	7005.0
353.00	4986.0
354.00	7738.0
355.00	2077.0
356.00	455.0
356.90	312.0
357.90	181.0
358.90	571.0
361.00	173.0
363.90	258.0
365.00	36184.0
366.00	5707.0
367.00	449.0
370.00	615.0
371.00	2077.0
372.00	12998.0
373.00	3339.0
374.00	420.0
377.00	364.0
382.90	3667.0
384.00	1101.0
384.80	366.0
390.00	1340.0
391.00	1175.0
391.90	754.0
401.00	887.0
402.00	5778.0
403.00	7476.0
403.90	2555.0
405.00	493.0
409.90	281.0
414.90	431.0
421.00	6762.0
422.00	6880.0
423.00	46024.0
424.00	11323.0
425.00	1359.0
439.90	313.0
441.00	160704.0
442.00	1009088.0
443.00	195136.0
444.00	17528.0
445.00	1118.0

Instrument :
BNA_M
ClientSampleId :
DFTPP130