

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM062024\
 Data File : BM046150.D
 Acq On : 21 Jun 2024 09:03
 Operator : MA/JU
 Sample : PB161410BL
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 SBLK410

Manual Integrations
 APPROVED

Reviewed By :Jagrut Upadhyay 06/21/2024
 Supervised By :mohammad ahmed 06/24/2024

Quant Time: Jun 21 10:50:11 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\SFAM-EPA-SIM-BM061824.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Jun 18 13:58:29 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.417	152	1808	0.400	ng/u1	-0.05
4) Naphthalene-d8	10.198	136	4981	0.400	ng/u1	-0.05
9) Acenaphthene-d10	14.071	164	2684	0.400	ng/u1	#-0.04
13) Phenanthrene-d10	16.815	188	5736m	0.400	ng/u1	-0.05
17) Chrysene-d12	21.041	240	3957	0.400	ng/u1	0.00
23) Perylene-d12	23.116	264	3682m	0.400	ng/u1	-0.03
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.029	96	12462	5.024	ng/u1	-0.02
6) 2-Methylnaphthalene-d10	11.843	152	1901	0.284	ng/u1	-0.02
18) Fluoranthene-d10	18.857	212	4006	0.338	ng/u1	-0.03

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM062024\
Data File : BM046150.D
Acq On : 21 Jun 2024 09:03
Operator : MA/JU
Sample : PB161410BL
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
SBLK410

Manual Integrations
APPROVED

Reviewed By :Jagrut Upadhyay 06/21/2024
Supervised By :mohammad ahmed 06/24/2024

Quant Time: Jun 21 10:50:11 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\SFAM-EPA-SIM-BM061824.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Jun 18 13:58:29 2024
Response via : Initial Calibration

